

Class II Recall Analysis

OMDRHO Virtual Conference

June 23, 2021

Meredith Andress

Recall Coordinator

Office of Medical Devices and Radiological Health


Office of Regulatory Affairs

U.S. Food and Drug Administration


Disclaimer

These PowerPoint slides are the intellectual property of the U.S. Food and Drug Administration and the individual presenter and are protected under copyright Laws of the United States of America and other countries. All rights reserved.

Annual Recalls & Products Processed



FY20 Recalls Per Commodity



Learning Objectives

- Recall Review
- Previous Class I Recall Analysis Summary
- Class II Recall Analysis
- Discuss Key Takeaways from Analysis
- Identify some Best Recall Practices
- Discuss what to expect during follow-up inspections

Recall Review

Recalls 101

Recall Classifications

- **Class I:** Reasonable probability that the use of, or exposure to, a violative product will cause serious adverse health consequences or death.
- **Class II:** Use of, or exposure to, a violative product may cause temporary or medically reversible adverse health consequences or where the probability of serious adverse health consequences is remote.
- **Class III:** Use of, or exposure to, a violative product is not likely to cause adverse health consequences.


Device Classifications

- **Class III:** General Controls and Premarket Approval
- **Class II:** General Controls and Special Controls
- **Class I:** General Controls
- ❖ **Class I and II may require 510k submission**


Previous Class I Recall Analysis Summary

Background

- **118 Class 1 Recalls:**
Jan 2016 - Dec 2018
- All Office of Medical Devices and Radiological Health (OMDRHO) Divisions included in analysis – US & international





Class I Recalls by Division



Quality Issues Related to Class I Recall

44 of 74 Inspections Resulted in FDA 483;
52 of 74 Inspections Covered Class I Recalls





Class II Recall Analysis

Class II Recall Analysis


Background

- Review of 437 Class II Recalls: October 2019 - September 2020
- All OMDRHO Divisions included in analysis


Goal


- In-depth root cause analysis
- Identify trends and obtain information to
 1. Provide industry with feedback on their corrective actions
 2. Identify internal (OMDRHO) and external (Industry) best practices

Class II Recalls by OMDRHO Division




Recall Totals by Industry Type







Top Root Causes for Orthopedic Devices




Top Root Causes for Cardiovascular Devices




Top Root Causes for Radiological Devices




Top Root Causes for Chemistry Devices




Top Root Causes for Surgical Devices




Firms with Multiple FY'20 Class II Recalls



Next Inspection Identified Quality Issues



Quality Issues Related to Class II Recall



Key Takeaways

Key Takeaways

1. In our analysis of Class 1 and Class 2 recalls, top root causes were attributed to process control or device design issues
2. The root cause investigation should be conducted in a **systemic manner** to ensure the root cause(s) are accurately identified for the recall
3. Firms with multiple recalls - systemic issues

Best Practices

Production & Process Controls

Common Root Causes observed:

- Packaging/Labeling issues
- Inadequate equipment and/or process validation
- Inadequate process procedures
- Production Process Change
- Component failures
- Supplier issues

Production & Process Controls

- Manufacturers should establish production processes to ensure medical devices conform to specifications
- These include...
 - Frequency of monitoring
 - SOPs, drawings, production method and instructions (e.g., steps, sequence, startup requirements, required checks)
 - Testing/Inspection and monitoring procedures
 - Establishing process parameters and qualification of specific production equipment (process validation)
 - Documentation requirements

Identify Root Cause

Important to establish the true root cause

- Requires thorough and broad investigation including using objective evidence
- Performed to the degree commensurate with significance and risk
- Should focus on Quality System to ultimately prevent future recalls

❖ **Necessary so that appropriate corrective actions can occur**

Corrective Actions

- Actions taken to eliminate root cause are appropriate for the magnitude of the problem and risks encountered
- Effectiveness verification/validation of production process or design change to ensure the action taken is effective and does not adversely affect the finished device
- The analysis of available quality data is a proactive activity designed to identify existing or potential problems that might otherwise go undiscovered.

Follow-up Inspections

FDA Inspections

- During FDA inspections, Class II recalls are reviewed
- Coverage of CAPA subsystem verifies problems are detected and resolved.
- In covering recalls, the inspections may focus on various key elements of the CAPA process to determine if actions taken where appropriate, followed established procedure as well as the regulatory requirements/intent.
- Firms are subject to follow-up inspections even if under MDSAP

Summary

- Class I Analysis – Top root cause was Device Design
- Class II Analysis – Top Root cause was Process Control
- Recall coverage during inspections are resulting in related 483 items
- Ensure the true root cause is identified in order to take appropriate corrective actions

Resources

- **Division Contacts**

- **Division 1:** oradevices1recalls@fda.hhs.gov (CT, DC, DE, IN, KY, MA, MD, ME, MI, NH, NJ, NY, OH, PA, RI, VA, VT, WV)
- **Division 2:** oradevices2recalls@fda.hhs.gov (AL, FL, GA, IA, IL, KS, LA, MN, MO, MS, NC, ND, NE, PR, SC, SD, TN, WI and the US Virgin Islands)
- **Division 3:** oradevices3recalls@fda.hhs.gov (AK, AR, AZ, CA, CO, HI, ID, MT, NM, NV, OK, OR, TX, UT, WA, WY)

- **Industry Guidance:** <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/industry-guidance-recalls>

Questions



