



August 7, 2020

K2M, Inc.
% Megan Callanan
Senior Regulatory Affairs Specialist
Stryker
2 Pearl Court
Allendale, New Jersey 07401

Re: K201006
Trade/Device Name: K2M Navigation Instruments
Regulation Number: 21 CFR 882.4560
Regulation Name: Stereotaxic Instrument
Regulatory Class: Class II
Product Code: OLO
Dated: April 16, 2020
Received: April 17, 2020

Dear Megan Callanan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnm.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K201006

Device Name

K2M Navigation Instruments

Indications for Use (Describe)

K2M Navigation Instruments are intended to be used in the preparation and placement of K2M screws (Denali, Mesa, Everest, and Yukon) during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. MESA screw navigation is intended for open procedures only. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy. Navigated Screw Inserters are also compatible with the Medtronic IPC POWEREASE System. The K2M Navigation Instruments are not intended for navigation of occipital screws.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

510(k) Summary: K2M Navigation Instruments	
Submitter:	K2M, Inc. 600 Hope Pkwy SE Leesburg, VA 20175
Contact Person:	Name: Megan Callanan Phone: (551)262-2429 Email: megan.callanan1@stryker.com
Date Prepared:	7/30/2020
Trade Name:	K2M Navigation Instruments
Common Name:	Navigation Instruments
Proposed Class:	Class II
Classification Name:	Orthopedic / Orthopedic Stereotaxic Instruments
Regulation Number:	21 CFR 882.4560
Product Code:	OLO
Predicate Devices:	Primary Predicate: K2M Navigation Instruments (K171321) Additional Predicates: Brainlab Compatible K2M Navigation Instruments (K181890) Stryker Navigation Enabled Instruments (K183381)
Device Description:	K2M Navigation Instruments are nonsterile, reusable surgical instruments (inserters, taps, probes, and awls) and single use drill bits intended be used when implanting previously cleared components of Mesa, Denali, Everest, and Yukon Spine Systems. These instruments are designed to interface with the Medtronic StealthStation® System when used for navigation during cervical and thoracolumbar spinal surgery.
Indications for use:	K2M Navigation Instruments are intended to be used in the preparation and placement of K2M screws (Denali, Mesa, Everest, and Yukon) during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. MESA screw navigation is intended for open procedures only. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy. Navigated Screw Inserters are also compatible with the Medtronic IPC POWEREASE System. The K2M Navigation Instruments are not intended for navigation of occipital screws.

510(k) Summary: K2M Navigation Instruments

Summary of the Technological Characteristics	<p>K2M Navigation Instruments have the same technological characteristics as the predicate devices including design, material composition, intended use, function, and range of sizes. Additional instruments for use with Yukon System are identical in function and fundamental technology.</p> <p>Biocompatibility of patient-contacting materials was demonstrated by using materials that meet applicable standards or are used in 510(k) cleared devices.</p>
Summary of the Performance Data	<p>Design validation testing was conducted to ensure the K2M Navigation Instruments are acceptable for their intended use, to ensure functionality and compatibility with the Medtronic StealthStation® using the NavLock tracker, and to demonstrate substantial equivalence to the predicate instruments.</p>
Conclusion	<p>K2M Navigation Instruments are substantially equivalent to the noted predicate devices as they have the same intended use, technological characteristics, and performance specifications. Furthermore, these instruments do not introduce any new concerns related to the safety and efficacy of the associated systems. The content of this submission supports the determination of substantial equivalence.</p>