



January 19, 2021

Globus Medical Inc.
Jennifer Antonacci
Group Manager, Regulatory Affairs
2560 General Armstead Ave.
Audubon, Pennsylvania 19403

Re: K203153

Trade/Device Name: CREO Stabilization System, Navigation Instruments, ExcelsiusGPS
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral Pedicle Screw System
Regulatory Class: Class II
Product Code: NKB, KWP, OLO
Dated: October 21, 2020
Received: October 22, 2020

Dear Jennifer Antonacci:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K203153

Device Name

CREO® Stabilization System

Indications for Use (Describe)

The CREO® Stabilization System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients (including small stature) and for pediatric patients. These devices are indicated as an adjunct to fusion for the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, and failed previous fusion (pseudoarthrosis). When used as an adjunct to fusion, the CREO® Stabilization System is intended to be used with autograft and/or allograft.

In addition, the CREO® Stabilization System is intended for treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and/or ilium with removal of the implants after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/ilium.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the CREO® Stabilization System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The CREO® Stabilization System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

In order to achieve additional levels of fixation, the CREO® Stabilization System rods may be connected to the REVERE Stabilization System (4.5mm, 5.5mm, or 6.35mm rod) or ELLIPSE® Occipito-Cervico-Thoracic Spinal System (3.5mm rod) using corresponding connectors. Refer to the REVERE® or ELLIPSE® system package insert for instructions and indications of use.

In-Line Connector Growing Rods are indicated in patients under 10 years of age with potential for additional spine growth who require surgical treatment to obtain and maintain correction of severe, progressive, life-threatening, early onset spinal deformities associated with thoracic insufficiency, including early onset scoliosis, as part of a growing rod construct.

Globus Navigation Instruments are intended to be used during the preparation and placement of CREO® screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

When used for posterior fixation in conjunction with FORTRESS™ or FORTRESS-Plus™ bone cement, the CREO® Fenestrated Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion. CREO® Fenestrated screws augmented with FORTRESS™ and FORTRESS-Plus™ bone cements are for use at spinal levels where the structural integrity of the spine is not severely compromised.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

Indications for Use

510(k) Number (if known)
K203153

Device Name
Navigation Instruments

Indications for Use (Describe)

Globus Navigation Instruments are intended to be used during the preparation and placement of Globus screws (QUARTEX®, CREO®, REVERE®, REVOLVE®, ELLIPSE®, PROTEX® CT, and SI-LOK®) during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MRI based model, fluoroscopy images, or digitized landmarks of the anatomy.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number (if known)

K203153

Device Name

ExcelsiusGPS®

Indications for Use (Describe)

The ExcelsiusGPS® is intended for use as an aid for precisely locating anatomical structures and for spatial positioning and orientation of an instrument holder or guide tube to be used by surgeons for navigating and/or guiding compatible surgical instruments in open or percutaneous procedures provided that the required fiducial markers and rigid patient anatomy can be identified on CT scans or fluoroscopy. The system is indicated for the placement of spinal and orthopedic bone screws and interbody spacers, and intracranial devices such as biopsy needles, electrodes, and tubes.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary: CREO® ONE Robotic Screws

Company: Globus Medical Inc.
2560 General Armistead Ave.
Audubon, PA 19403
610-930-1800

Contact: Jennifer Antonacci, Ph.D.
Group Manager, Regulatory Affairs

Date Prepared: October 21, 2020

Device Name: CREO® Stabilization System
Navigation Instruments
ExcelsiusGPS®

Classification: CREO® Additional Implants
Per 21 CFR as follows:

§888.3070 Thoracolumbosacral pedicle screw system
§888.3050 Spinal Interlaminar Fixation Orthosis
§888.3060 Spinal Intervertebral Body Fixation Orthosis

Product Codes NKB, KWP
Regulatory Class: II, Panel Code: 87

Navigation Instruments
Per 21 CFR as follows:
§882.4560 Stereotaxic Instrument
Product Code OLO
Regulatory Class: II, Panel Code: 87

ExcelsiusGPS®
Per 21 CFR as follows:
§882.4560 Stereotaxic Instruments
Product Codes OLO
Regulatory Class: II, Panel Code: 87

Primary Predicate: CREO® Stabilization System (K124058)

Additional Predicates: CREO® Stabilization System (K143633, K191835)
Navigation Instruments (K153203, K180690)
ExcelsiusGPS® (K171651, K200047)

Purpose:

The purpose of this submission is to request clearance for CREO® ONE Robotic Screws.

Device Description:

CREO® ONE Robotic Screws are polyaxial screws with a tapered distal awl tip to allow penetration of the cortex of the pedicle, for ease of use with the ExcelsiusGPS® robot and Globus Navigated Instruments. Implants are available in a variety of sizes to accommodate individual patient anatomy. CREO® ONE screws may be used with ExcelsiusGPS® instruments and CREO® Navigation instruments.

CREO® ONE Robotic Screws are composed of titanium alloy with optional hydroxyapatite (HA) coating.

Indications for Use:

CREO® Stabilization System

The CREO® Stabilization System implants are non-cervical spinal fixation devices intended for posterior pedicle screw fixation (T1-S2/ilium), posterior hook fixation (T1-L5), or anterolateral fixation (T8-L5). Pedicle screw fixation is indicated for skeletally mature patients (including small stature) and for pediatric patients. These devices are indicated as an adjunct to fusion for the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e., fracture or dislocation), deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis, Scheuermann's Disease), tumor, stenosis, and failed previous fusion (pseudoarthrosis). When used as an adjunct to fusion, the CREO® Stabilization System is intended to be used with autograft and/or allograft.

In addition, the CREO® Stabilization System is intended for treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebra in skeletally mature patients receiving fusion by autogenous bone graft, having implants attached to the lumbosacral spine and/or ilium with removal of the implants after attainment of a solid fusion. Levels of pedicle screw fixation for these patients are L3-sacrum/ilium.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the CREO® Stabilization System implants are indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis. The CREO® Stabilization System is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

In order to achieve additional levels of fixation, the CREO® Stabilization System rods may be connected to the REVERE® Stabilization System (4.5mm, 5.5mm, or 6.35mm rod) or ELLIPSE® Occipito-Cervico-Thoracic Spinal System (3.5mm rod)

using corresponding connectors. Refer to the REVERE® or ELLIPSE® system package insert for instructions and indications of use.

In-Line Connector Growing Rods are indicated in patients under 10 years of age with potential for additional spine growth who require surgical treatment to obtain and maintain correction of severe, progressive, life-threatening, early onset spinal deformities associated with thoracic insufficiency, including early onset scoliosis, as part of a growing rod construct.

Globus Navigation Instruments are intended to be used during the preparation and placement of CREO® screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

When used for posterior fixation in conjunction with FORTRESS™ or FORTRESS-Plus™ bone cement, the CREO® Fenestrated Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion. CREO® Fenestrated screws augmented with FORTRESS™ and FORTRESS-Plus™ bone cements are for use at spinal levels where the structural integrity of the spine is not severely compromised.

Navigation Instruments

Globus Navigation Instruments are intended to be used during the preparation and placement of Globus screws (QUARTEX®, CREO®, REVERE®, REVOLVE®, ELLIPSE®, PROTEX® CT, and SI-LOK®) during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MRI based model, fluoroscopy images, or digitized landmarks of the anatomy.

ExcelsiusGPS®

The ExcelsiusGPS® is intended for use as an aid for precisely locating anatomical structures and for spatial positioning and orientation of an instrument holder or guide tube to be used by surgeons for navigating and/or guiding compatible surgical instruments in open or percutaneous procedures provided that the required fiducial markers and rigid patient anatomy can be identified on CT scans

or fluoroscopy. The system is indicated for the placement of spinal and orthopedic bone screws and interbody spacers, and intracranial devices such as biopsy needles, electrodes, and tubes.

Performance Data:

Verification and validation cadaveric testing was conducted, including planning and placement of CREO ONE screws, using the ExcelsiusGPS System and associated instruments to demonstrate that CREO ONE Robotic Screws meet performance and navigation accuracy requirements.

Technological Characteristics:

Subject CREO® ONE Robotic Screws have the same technological characteristics as the predicate devices including design, intended use, material composition, function, and range of sizes.

Basis of Substantial Equivalence:

Subject CREO® ONE Robotic Screws are similar to the predicate device with respect to technological characteristics, performance, design, and intended use. The information provided within this premarket notification supports substantial equivalence to the predicate devices.