



August 5, 2021

4WEB Medical, Inc.
% Richard Jansen, Pharm. D.
President
Silver Pine Consulting, LLC
3851 Mossy Oak Drive
Fort Myers, Florida 33905

Re: K211388

Trade/Device Name: Lateral Spine Truss System (LSTS) Interbody Fusion Device
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: MAX, OVD
Dated: July 7, 2021
Received: July 9, 2021

Dear Dr. Jansen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent L. Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211388

Device Name

Lateral Spine Truss System (LSTS) Interbody Fusion Device

Indications for Use (Describe)

The Lateral Spine Truss System (LSTS) Interbody Fusion Device is indicated for use in skeletally mature patients with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. Patients should have received 6 months of non-operative treatment prior to treatment with the devices. All LSTS Interbody Fusion Devices must be used with supplemental fixation and must be used with autograft and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft. Interbodies with 18° lordosis or greater must be used with the 4WEB Lumbar Spine Truss System Plating Solution (LSTS-PS) with integrated fixation. If using the 1-hole 4WEB LSTS-PS with integrated fixation, additional supplemental fixation is required (e.g. posterior fixation). These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared: July 6, 2021
Contact: Jesse Hunt, President
4WEB, Inc.
2801 Network Blvd., Suite 620
Frisco, TX 75034
Phone: (800) 285-7090
Fax: 972-488-1816

Regulatory Contact: Rich Jansen, Pharm. D.
Silver Pine Consulting, LLC
richj@s-pineconsulting.com

Trade Name: Lateral Spine Truss System (LSTS) Interbody Fusion Device
Product Class: Class II
Classification: 21 CFR §888.3080
Common Name: Intervertebral Body Fusion Device
Product Codes: MAX, OVD
Panel Code: 87

Purpose:

The purpose of this submission is to expand the Lateral Spine Truss System (LSTS) Interbody Fusion Device lordotic offering to include 18°, 24°, and 30° of lordosis. Additionally, the Indications for Use are being updated to include use of the 4WEB Lumbar Spine Truss System Plating Solution (LSTS-PS).

Indications for Use:

The Lateral Spine Truss System (LSTS) Interbody Fusion Device is indicated for use in skeletally mature patients with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. Patients should have received 6 months of non-operative treatment prior to treatment with the devices. All LSTS Interbody Fusion Devices must be used with supplemental fixation and must be used with autograft and/or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft. Interbodies with 18° lordosis or greater must be used with the 4WEB Lumbar Spine Truss System Plating Solution (LSTS-PS) with integrated fixation. If using the 1-hole 4WEB LSTS-PS with integrated fixation, additional supplemental fixation is required (e.g. posterior fixation). These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s).

Device Description:

The device is an open architecture truss design mathematically formulated to provide structural support with open space throughout the implant for bone growth and fusion. The 4WEB additive manufacturing process provides a hierarchical surface roughness. The implant is made from Ti6Al4V alloy.

The device is available in a variety of sizes and lordotic angles to accommodate the patient's anatomy.

Predicate Device(s):

The primary predicate device is the 4WEB Lateral Spine Truss System (LSTS) Interbody Fusion Device (K172392). Reference predicates include the 4WEB Lumbar Spine Truss System Plating Solution (LSTS-PS) (K203065) and the Zimmer-Biomet (Lanx) Timberline MPF (K131547).

Performance Standards:

Validated finite element analysis demonstrated that the product line extension for the Lateral Spine Truss System (LSTS) Interbody Fusion Device does not introduce a new worst-case compared to the previously cleared 4WEB Lumbar Interbody Fusion Devices for mechanical properties of the device.

Performance testing has been completed per the following standards for the combined Lateral Spine Truss System (LSTS) Interbody Fusion Device and the Lumbar Spine Truss System Plating Solution (LSTS-PS):

- Axial screw pushout per ASTM F543
- Static axial compression per ASTM F2077
- Static compression shear per ASTM F2077
- Dynamic axial compression fatigue per ASTM F2077
- Dynamic compression shear fatigue per ASTM F2077
- Expulsion testing
- MR Conditional testing

The results of this non-clinical testing show that the strength of the LSTS Interbody Fusion Device and LSTS Plating Solution is sufficient for its intended use and is substantially equivalent to legally marketed predicate devices.

Technological Characteristics:

4WEB, Inc. has compared these devices to the previously cleared predicate devices in regard to indications for use, materials, function, sizes and simulated testing. These comparisons demonstrate substantial equivalence to the predicate devices.

Conclusion:

4WEB, Inc. concludes that the LSTS Interbody Fusion Device is substantially equivalent to the predicate devices and raise no new questions of safety or effectiveness.