



July 9, 2021

bioMérieux, Inc.  
Debra Broyles  
Sr. Regulatory Affairs Specialist  
595 Anglum Road  
Hazelwood, Missouri 63042

Re: K201675

Trade/Device Name: VITEK 2 AST-Gram Negative Meropenem ( $\leq 0.25 - \geq 16 \mu\text{g/mL}$ )  
Regulation Number: 21 CFR 866.1645  
Regulation Name: Fully Automated Short-Term Incubation Cycle Antimicrobial Susceptibility System  
Regulatory Class: Class II  
Product Code: LON, LTW, LTT  
Dated: June 18, 2020  
Received: June 19, 2020

Dear Debra Broyles:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR

803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ribhi Shawar, Ph.D. (ABMM)  
Chief  
General Bacteriology and Antimicrobial Susceptibility  
Branch  
Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics  
and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

Device Name

Indications for Use (Describe)

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**VITEK<sup>®</sup> 2 AST-GN Meropenem  
Traditional 510(k) Submission**

---

**510(k) SUMMARY**

**VITEK<sup>®</sup> 2 AST-Gram Negative Meropenem ( $\leq 0.25$  -  $\geq 16$   $\mu\text{g/mL}$ )**

**510(k) Submission Information:**

Submitter's Name:	bioMérieux, Inc.
Address:	595 Anglum Road Hazelwood, MO 63042
Contact Person:	Debra Broyles Senior Regulatory Affairs Specialist
Phone Number:	314 -731-8805
Fax Number:	314-731-8689
Date of Preparation:	December 8, 2020

**B. Device Name:**

Formal/Trade Name:	VITEK <sup>®</sup> 2 AST-Gram Negative Meropenem ( $\leq 0.25$ - $\geq 16$ $\mu\text{g/mL}$ )
Classification Name:	21 CFR 866.1645 Fully Automated Short-Term Incubation Cycle Antimicrobial Susceptibility System Product Code(s): LON, LTW, LTT

Common Name:	VITEK <sup>®</sup> 2 AST-GN Meropenem ( $\leq 0.25$ - $\geq 16$ $\mu\text{g/mL}$ )
--------------	--

**C. Predicate Device:**

VITEK<sup>®</sup> 2 AST-GN Eravacycline ( $\leq 0.12$  -  $\geq 4$   $\mu\text{g/mL}$ ) (K191766)

**D. Device Description:**

The principle of the VITEK<sup>®</sup> 2 AST cards is based on the microdilution minimum inhibitory concentration (MIC) technique reported by MacLowry and Marsh<sup>(1)</sup> and Gerlach<sup>(2)</sup>. The VITEK<sup>®</sup> 2 AST card is essentially a miniaturized, abbreviated and automated version of the doubling dilution technique<sup>(3)</sup>.



**VITEK® 2 AST-GN Meropenem  
Traditional 510(k) Submission**

Each VITEK® 2 AST card contains 64 wells. A control well which only contains microbiological culture media is resident on all cards. The remaining wells contain premeasured portions of a specific antibiotic combined with culture media. The bacterial or yeast isolate to be tested is diluted to a standardized concentration with 0.45 – 0.5% saline before being used to rehydrate the antimicrobial medium within the card. The VITEK® 2 System automatically fills, seals and places the card into the incubator/reader. The VITEK® 2 Compact has a manual filling, sealing and loading operation. The VITEK® 2 Systems monitor the growth of each well in the card over a defined period of time. At the completion of the incubation cycle, a report is generated that contains the MIC value along with the interpretive category result for each antibiotic contained on the card.

VITEK® 2 AST-GN Meropenem ( $\leq 0.25$  -  $\geq 16$   $\mu\text{g/mL}$ ) has the following concentrations in the card: 0.5, 1, 2, 4 and 8  $\mu\text{g/mL}$  (equivalent standard method concentration by efficacy in  $\mu\text{g/mL}$ ).

**E. Substantial Equivalence Information:**

The similarities and differences of the VITEK® 2 AST-GN Meropenem ( $\leq 0.25$  -  $\geq 16$   $\mu\text{g/mL}$ ) when compared to the predicate device, VITEK® 2 AST-GN Eravacycline ( $\leq 0.12$  -  $\geq 4$   $\mu\text{g/mL}$ ), are described in the [Table 1](#) below.

**Table 1: Substantial Equivalence**

Item	Device: VITEK® 2 AST-Gram Negative Meropenem ( $\leq 0.25$ - $\geq 16$ $\mu\text{g/mL}$ )	Predicate: VITEK® 2 AST-GN Eravacycline ( $\leq 0.12$ - $\geq 4$ $\mu\text{g/mL}$ ) (K191766)
<b>Similarities</b>		
<b>Intended Use</b>	VITEK® 2 AST-Gram Negative Meropenem is designed for antimicrobial susceptibility testing of Gram negative bacilli and is intended for use with the VITEK® 2 and VITEK® 2 Compact Systems as a laboratory aid in the determination of <i>in vitro</i> susceptibility to antimicrobial agents. VITEK® 2 AST-Gram Negative Meropenem is a quantitative test. Meropenem has been shown to be active against most strains of the microorganisms listed below, according to the FDA label for this antimicrobial.	VITEK® 2 AST-Gram Negative Eravacycline is designed for antimicrobial susceptibility testing of Gram negative bacilli and is intended for use with the VITEK® 2 and VITEK® 2 Compact Systems as a laboratory aid in the determination of <i>in vitro</i> susceptibility to antimicrobial agents. VITEK® 2 AST-Gram Negative Eravacycline is a quantitative test. Eravacycline has been shown to be active against most strains of the microorganisms listed below, according to the FDA label for this antimicrobial.



VITEK® 2 AST-GN Meropenem  
Traditional 510(k) Submission

Item	Device: VITEK® 2 AST-Gram Negative Meropenem (≤0.25 - ≥16 µg/mL)	Predicate: VITEK® 2 AST-GN Eravacycline ( ≤ 0.12 – ≥ 4 µg/mL) (K191766)
<b>Similarities</b>		
	<p><b><u>Active in vitro and in clinical infections:</u></b></p> <p><i>Escherichia coli</i> <i>Klebsiella pneumoniae</i> <i>Proteus mirabilis</i> <i>Pseudomonas aeruginosa</i></p> <p><b><u>In vitro data are available, but clinical significance is unknown:</u></b></p> <p><i>Citrobacter freundii</i> <i>Citrobacter koseri</i> <i>Enterobacter cloacae</i> <i>Hafnia alvei</i> <i>Klebsiella oxytoca</i> <i>Morganella morganii</i> <i>Serratia marcescens</i></p> <p>The VITEK® 2 Gram Negative Susceptibility Card Meropenem also reports susceptibility for the following additional organisms as listed on the FDA Susceptibility Test Interpretive Criteria website: <i>Acinetobacter spp</i></p> <p>The VITEK® 2 Gram-Negative Susceptibility Card is intended for use with the VITEK® 2 Systems in clinical laboratories as an <i>in vitro</i> test to determine the susceptibility of clinically significant aerobic Gram negative bacilli to antimicrobial agents when used as instructed.</p>	<p><b><u>Active in vitro and in clinical infections:</u></b></p> <p><i>Citrobacter freundii</i> <i>Enterobacter cloacae</i> <i>Escherichia coli</i> <i>Klebsiella oxytoca</i> <i>Klebsiella pneumoniae</i></p> <p><b><u>In vitro data are available, but clinical significance is unknown:</u></b></p> <p><i>Citrobacter koseri</i> <i>Klebsiella (Enterobacter) aerogenes</i></p> <p>The VITEK® 2 Gram-Negative Susceptibility Card is intended for use with the VITEK® 2 Systems in clinical laboratories as an <i>in vitro</i> test to determine the susceptibility of clinically significant aerobic Gram negative bacilli to antimicrobial agents when used as instructed.</p>



**VITEK® 2 AST-GN Meropenem  
Traditional 510(k) Submission**

Item	Device: VITEK® 2 AST-Gram Negative Meropenem (≤0.25 - ≥16 µg/mL)	Predicate: VITEK® 2 AST-GN Eravacycline ( ≤ 0.12 – ≥ 4 µg/mL) (K191766)
<b>Similarities</b>		
<b>Test Methodology</b>	Automated quantitative antimicrobial susceptibility test for use with the VITEK® 2 and VITEK® 2 Compact Systems to determine the <i>in vitro</i> susceptibility of microorganisms	Same
<b>Inoculum</b>	Saline suspension of organism	Same
<b>Test Card</b>	Gram Negative (AST-GN) Susceptibility Card	Same
<b>Instrument</b>	VITEK® 2 and VITEK® 2 Compact Systems	Same
<b>Differences</b>		
<b>Antimicrobial Agent</b>	Meropenem	Eravacycline
<b>Concentrations</b>	0.5, 2, 6, 12	0.25, 1, 2, 4

**F. Intended Use:**

VITEK® 2 AST-Gram Negative Meropenem is designed for antimicrobial susceptibility testing of Gram negative bacilli and is intended for use with the VITEK® 2 and VITEK® 2 Compact Systems as a laboratory aid in the determination of *in vitro* susceptibility to antimicrobial agents. VITEK® 2 AST-Gram Negative Meropenem is a quantitative test. Meropenem has been shown to be active against most strains of the microorganisms listed below, according to the FDA label for this antimicrobial.

**Active *in vitro* and in clinical infections:**

- Escherichia coli*
- Klebsiella pneumoniae*
- Proteus mirabilis*
- Pseudomonas aeruginosa*

***In vitro* data are available, but clinical significance is unknown:**

- Citrobacter freundii*
- Citrobacter koseri*





**VITEK<sup>®</sup> 2 AST-GN Meropenem  
Traditional 510(k) Submission**

---

*Enterobacter cloacae*  
*Hafnia alvei*  
*Klebsiella oxytoca*  
*Morganella morganii*  
*Serratia marcescens*

The VITEK<sup>®</sup> 2 Gram Negative Susceptibility Card Meropenem also reports susceptibility for the following additional organisms as listed on the FDA Susceptibility Test Interpretive Criteria website:

*Acinetobacter spp.*

The VITEK<sup>®</sup> 2 Gram-Negative Susceptibility Card is intended for use with the VITEK<sup>®</sup> 2 Systems in clinical laboratories as an *in vitro* test to determine the susceptibility of clinically significant aerobic Gram-negative bacilli to antimicrobial agents when used as instructed.

**G. Performance Overview and Conclusion:**

VITEK<sup>®</sup> 2 AST-GN Meropenem demonstrated substantially equivalent performance when compared with the CLSI broth microdilution reference method, as defined in the FDA Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA (Issued August 28, 2009).

The Premarket Notification (510[k]) presents data in support of VITEK<sup>®</sup> 2 AST-GN Meropenem. An external evaluation was conducted with fresh and stock clinical isolates, as well as a set of challenge strains. The external evaluations were designed to confirm the acceptability of VITEK<sup>®</sup> 2 AST-GN Meropenem by comparing its performance with the CLSI broth microdilution reference method incubated at 16-20 hrs. The data is representative of performance on both the VITEK<sup>®</sup> 2 and VITEK<sup>®</sup> 2 Compact instrument platforms.

The VITEK<sup>®</sup> 2 AST-GN Meropenem demonstrated acceptable performance as presented in **Table 2** below:





**VITEK® 2 AST-GN Meropenem  
Traditional 510(k) Submission**

**Table 2: VITEK® 2 AST-GN Meropenem Performance**

	<b>%EA</b>	<b>VME</b>	<b>ME</b>	<b>mE</b>	<b>%CA</b>	<b>VME</b>	<b>ME</b>	<b>mE</b>
<b>Overall Performance</b> (with the reference method)	(1016/1070) 95.0	N/A	N/A	N/A	(1016/1070) 95.0	(0/346) 0.0	(4/680) 0.6	(50/1070) 4.7
<i>Enterobacteriales</i>	(537/569) 94.4	N/A	N/A	N/A	(552/569) 97.0	(0/65) 0.0	(4/491) 0.8	(13/569) 2.3
<i>P. aeruginosa</i>	(267/282) 94.7	N/A	N/A	N/A	(253/282) 89.7	(0/121) 0.0	(0/137) 0.0	(29/282) 10.3
<i>Acinetobacter spp.</i>	(212/219) 96.8	N/A	N/A	N/A	(211/219) 96.3	(0/160) 0.0	(0/52) 0.0	(8/219) 3.7

Reproducibility and Quality Control demonstrated acceptable results.

**H. References:**

1. MacLowry, J.D. and Marsh, H.H., Semi-automatic Microtechnique for Serial Dilution Antibiotic Sensitivity Testing in the Clinical laboratory, *Journal of Laboratory Clinical Medicine*, 72:685-687, 1968.
2. Gerlach, E.H., Microdilution 1: A Comparative Study, p. 63-76. *Current Techniques for Antibiotic Susceptibility Testing*. A. Balows (ed.), Charles C. Thomas, Springfield, IL, 1974.
3. Barry, A.L., *The Antimicrobial Susceptibility Test, Principles and Practices*, Lea and Febiger, Philadelphia, PA, 1976.