



November 19, 2021

Nvision Biomedical Technologies, Inc
Analaura Villarreal-Berain
Extremity Project Engineer
4590 Lockhill Selma
San Antonio, Texas 78249

Re: K213421

Trade/Device Name: Vector® Hammertoe Correction System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: HWC
Dated: October 11, 2021
Received: October 20, 2021

Dear Analaura Villarreal-Berain:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K213421

Device Name
VECTOR® Hammertoe Correction System

Indications for Use (Describe)

The Vector Hammertoe Correction System is indicated for the fixation of osteotomies and reconstruction of the lesser toes following correction procedures for hammertoe, claw toe, and mallet toe.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

DATE PREPARED

November 18, 2021

MANUFACTURER AND 510(k) OWNER

Nvision Biomedical Technologies, Inc.

4590 Lockhill Selma

San Antonio, TX 78249, USA

Telephone: (210) 545-3713

Fax: (866) 764-1139

Official Contact: Diana Langham, Director of Regulatory and Corporate Compliance

REPRESENTATIVE

Analaura Villarreal-Berain, Extremity Project Engineer

Nvision Biomedical Technologies

Telephone: (210) 545-3713 ext. 109

Email: analauravillarreal@nvisionbiomed.com

PROPRIETARY NAME OF SUBJECT DEVICE

VECTOR[®] Hammertoe Correction System

COMMON NAME

Bone Fixation Screw

DEVICE CLASSIFICATION

Smooth or Threaded Metallic Bone Fixation Fastener

(Classification Regulations: 21 CFR 888.3040, Product Codes: HWC, Class: II)

PREMARKET REVIEW

Orthopedic Devices

INDICATIONS FOR USE

The Vector Hammertoe Correction System is indicated for the fixation of osteotomies and reconstruction of the lesser toes following correction procedures for hammertoe, claw toe, and mallet toe.

DEVICE DESCRIPTION

The Vector Hammertoe Correction System is comprised of a sterile PEEK (polyetheretherketone) HA (hydroxyapatite) fixation device. The implants are offered in Ø3.00mm, Ø3.50mm and Ø4.00mm and in 0° angle. The system has K-wires, drill, taps, implant inserters, and sizers manufactured from medical grade stainless steel.

PREDICATE DEVICE IDENTIFICATION

The subject Vector Hammertoe Correction System is substantially equivalent to the following predicates:

<i>510(k) Number</i>	<i>Predicate Device Name / Manufacturer</i>	<i>Predicate</i>
K201850	Nvision Biomedical's Vector Hammertoe Correction System	Primary
K183055		Additional Predicate

Nvision Biomedical's nva, nvp, nvt (K193645) is also cited in this submission as a reference predicate device.

SUMMARY OF NON-CLINICAL TESTING

No FDA performance standards have been established for the Vector Hammertoe Correction System. The following was performed to demonstrate safety per methods of the previous submission:

- Engineering analysis comparing device characteristics including materials, intended use and processes (cleaning and sterilization methods)

The results of this comparison indicate that the Vector Hammertoe Correction System is substantially equivalent to the predicate devices.

EQUIVALENCE TO PREDICATE DEVICES

Nvision believes that the Vector Hammertoe Correction System modification is substantially equivalent to the predicate devices. The subject implants maintain the same features as the previously cleared devices, but steam sterilization is being added as an alternative sterilization method, which is commonly used for medical devices such as the predicates. This modification does not change the intended use or performance of the device and does not raise additional questions of substantial equivalence. These technological characteristics have undergone a comparison of characteristics to ensure the device is as safe and effective as the predicates.

CONCLUSION

Based on the comparison of device characteristics, it can be concluded that the subject device does not raise new issues of safety or efficacy compared to the predicate devices. The similar indications for use, technological characteristics, and performance characteristics for the proposed Vector Hammertoe Correction System options are assessed to be substantially equivalent to the predicate devices.