



August 13, 2020

Arthrex Inc.  
Ivette Galmez  
Senior Regulatory Affairs Specialist  
1370 Creekside Boulevard  
Naples, Florida 34108-1945

Re: K200895

Trade/Device Name: Univers Revers Modular Glenoid System, Half Augment Baseplate  
Regulation Number: 21 CFR 888.3660  
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis  
Regulatory Class: Class II  
Product Code: PHX  
Dated: June 12, 2020  
Received: July 15, 2020

Dear Ivette Galmez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Michael Owens  
Acting Assistant Director  
DHT6A: Division of Joint  
Arthroplasty Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K200895

Device Name

Univers Revers Modular Glenoid System, Half Augment Baseplate

Indications for Use (Describe)

The Univers Revers Modular Glenoid System is indicated for use in a grossly rotator cuff deficient glenohumeral joint with severe arthropathy or a previously failed joint replacement with a gross rotator cuff deficiency. The patient's joint must be anatomically and structurally suited to receive the selected implant(s), and a functional deltoid muscle is necessary to use the device.

The Univers Revers Modular Glenoid System is indicated for primary, fracture, or revision total shoulder replacement for the relief of pain and significant disability due to gross rotator cuff deficiency.

The Univers Revers Modular Glenoid System is porous coated and is intended for cementless use with the addition of screws for fixation.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

<b>Date Prepared</b>	August 7, 2020
<b>Submitter</b>	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
<b>Contact Person</b>	Ivette Galmez Senior Regulatory Affairs Specialist 1-239-643-5553, ext. 71263 Ivette.galmez@arthrex.com
<b>Name of Device</b>	Univers Revers Modular Glenoid System, Half Augment Baseplate
<b>Common Name</b>	Shoulder Prosthesis
<b>Product Code</b>	PHX
<b>Classification Name</b>	21 CFR 888.3660: Shoulder joint metal polymer semi constrained cemented prosthesis
<b>Regulatory Class</b>	II
<b>Predicate Device</b>	K193372: Univers Revers Modular Glenoid System (Augmented baseplates)
<b>Reference Device</b>	K173900: Arthrex Univers Revers Modular Glenoid System K142863: Arthrex Univers Revers Shoulder Prosthesis System
<b>Purpose of Submission</b>	This Traditional 510(k) premarket notification is submitted to obtain clearance for additional modular glenoid baseplates for use with the Arthrex Univers Revers Modular Glenoid System cleared under K193372 and K173900.
<b>Device Description</b>	The proposed devices are augmented modular glenoid baseplates made of titanium with BioSync coating. The proposed devices are half-wedge augmented modular glenoid baseplates available in two sizes (24 and 28). The baseplates are designed to be used cementless with peripheral screws and glenospheres (cleared under K193372 and K173900).
<b>Indications for Use</b>	<p>The Univers Revers Modular Glenoid System is indicated for use in a grossly rotator cuff deficient glenohumeral joint with severe arthropathy or a previously failed joint replacement with a gross rotator cuff deficiency. The patient's joint must be anatomically and structurally suited to receive the selected implant(s), and a functional deltoid muscle is necessary to use the device.</p> <p>The Univers Revers Modular Glenoid System is indicated for primary, fracture, or revision total shoulder replacement for the relief of pain and significant disability due to gross rotator cuff deficiency.</p> <p>The Univers Revers Modular Glenoid System is porous coated and is intended for cementless use with the addition of screws for fixation.</p>
<b>Summary of Technological Characteristics</b>	The proposed devices are made of the same materials as the predicate device. The proposed devices have the same intended use/indications, packaging, shelf life and sterilization as the predicate devices. The baseplates are offered in the same sizes as the predicate device. The difference with the predicate is the geometry of the baseplate augment. The augment of the proposed device is a half-wedge whereas the predicate is a full-wedge with variation in wedge angles.
<b>Performance Data</b>	<p>Mechanical testing (i.e. Rocking horse testing per ASTM F2028) was performed to demonstrate that the proposed device meets the standards requirements.</p> <p>MRI testing were conducted in accordance with FDA guidance Testing and Labeling Medical Devices for Safety in the Magnetic Resonance (MR) Environment and ASTM F2182.</p>
<b>Conclusion</b>	<p>The Univers Revers Modular Glenoid System is substantially equivalent to the predicate device in which the basic design features and intended use are the same. The mechanical testing data demonstrates that the proposed device performance is equivalent to the predicate device for the desired indications. Any differences between the proposed device and the predicate device are considered minor and do not raise questions regarding safety or effectiveness.</p> <p>Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate device.</p>