



December 20, 2021

Shenzhen Beauty Every Moment intelligent electric Co., Ltd.
% Rain Yip
Registered Engineer
Feiying Drug & Medical Consulting Technical Service Group
Contact Address

Re: K211185

Trade/Device Name: IPL Home Use Hair Removal Device

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In
Dermatology

Regulatory Class: Class II

Product Code: OHT

Dated: June 19, 2021

Received: November 22, 2021

Dear Rain Yip:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Purva Pandya
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211185

Device Name

IPL Home Use Hair Removal Device, Model(s): D-1150, D-1171, D-1153, D-1155, D-1156, D-1126, D-1178, D-1187

Indications for Use (Describe)

IPL Home Use Hair Removal Device is an over-the-counter device intended for removal of unwanted body and/or facial hair.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

"510(k) Summary" as required by 21 CFR Part 807.92.

Date: 2021-12-17

I. Submitter

Shenzhen Beauty Every Moment intelligent electric Co.,Ltd.
Rm. 501, Bldg. 4, Haoer Jiashitai Technology Park, Fengtang Avenue No. 164, Heping
Community, Fuhai Street, Bao'an District, Shenzhen, Guangdong, China
Post code: 518100
Tel.: +86 755 3325 3598

LUO HUAJUN
Certification enginner
Tel: +86 755 3325 3598
Email: 309024340@qq.com

II. Device

Name of Device: IPL Home Use Hair Removal Device
Model(s): D-1150, D-1171, D-1153, D-1155, D-1156, D-1126, D-1178, D-1187
Common or Usual Name: Light Based Over-The-Counter Hair Removal
Classification Name: Laser surgical instrument for use in general and plastic surgery and in
dermatology
Regulatory Class: II
Product Code: OHT
Regulation Number: 21 CFR 878.4810

III. Predicate Device

| <u>Manufacturer</u> | <u>Predicate Device</u> | <u>510(k) Number</u> | <u>Approval Date</u> |
|--|--|----------------------|----------------------|
| Shenzhen Bosidin Technology Co., Ltd. | IPL Home Use Hair Removal Device/D-1128 (Primary) | K192432 | Nov. 08, 2019 |
| CyDen Limited. | iPulse SmoothSkin Gold Hair Removal System | K160968 | Apr.04, 2016 |

IV. Device Description

The IPL Home Use Hair Removal Device is a personal, light-based, hair reduction device intended to be sold over-the-counter directly to the end user. The device provides hair reduction using Intense Pulsed Light (IPL) technology. It works below the skin's surface and does not involve any cutting or pulling, reducing hair growth with minimal pain. The device is only powered by the external power adapter and its IPL emission activation is by finger switch. The

device contains a Xenon lamp and a skin sensor to detect appropriate skin contact. If the device is not properly applied to the treatment area (in full contact with the skin), the device cannot be triggered a pulse emitting.

The IPL Home Use Hair Removal Device includes D-1150, D-1171, D-1153, D-1155, D-1156, D-1126, D-1178, D-1187 eight models. Their intended use, performance, structure design and operation are basically identical, with the differences contains product appearance, display contents and the number of button, lamp cartridge's quantity and appearance; in additional, due to the size of the treatment window in different lamp cartridge, the treatment area (spot size) and output density is a little different; but these differences do not affect or change the intended use of the device.

V. Indications for Use

IPL Home Use Hair Removal Device is an over-the-counter device intended for removal of unwanted body and/or facial hair.

VI. Comparison of Technological Characteristics With the Predicate Device

The IPL Home Use Hair Removal Device has the same intended use, mode of action and similar operational characteristics as the predicate devices. Any minor differences between the subject device and the listed predicate devices do no raise any issues of safety or efficacy. Performance data supports that the device is safe and as effective as the predicate device for its intended use. Therefore, the IPL Home Use Hair Removal Device may be found substantially equivalent to its predicate device.

IPL Home Use Hair Removal Device is compared with the following Predicate Devices in terms of intended use, design, material, specifications, and performance:

| <u>Comparison Elements</u> | <u>Subject Device</u> | <u>Predicate Device</u> | |
|----------------------------|--|---|--|
| | | <u>K192432 (Primary)</u> | <u>K160968</u> |
| K Number | Pending | K192432 | K160968 |
| Trade name | IPL Home Use Hair Removal Device/D-1150, D-1171, D-1153, D-1155, D-1156, D-1126, D-1178, D-1187 | IPL Home Use Hair Removal Device/D-1128 | iPulse SmoothSkin Gold Hair Removal System |
| Wavelength range | Regular window: 530-1100nm, 590-1100nm Filter window: 600-1100nm | 510-1100nm | 510-1100nm |
| Energy medium | Xenon Arc Flashlamp | Xenon Arc Flashlamp | Xenon Arc Flashlamp |
| Energy density | 2.0~4.5 J/cm ² | 2.0~4.0 J/cm ² | 3~6 J/cm ² |
| Spot size | Regular window: 3.8cm ² , 3.6cm ² , 3.0cm ² , 1.0cm ² , 2.0cm ² , 1.9cm ² , 1.8cm ² , Filter window: 3.0cm ² , 2.0cm ² | 4.5, 2.0, 2.5cm ² | 3 (3cm by 1cm) |

| <u>Comparison Elements</u> | <u>Subject Device</u> | <u>Predicate Device</u> | |
|---------------------------------|--|--|--|
| | | <u>K192432 (Primary)</u> | <u>K160968</u> |
| Pulse duration | 7.5~12ms | 7.5-14.5ms | 2-10ms |
| Pulsing control | Finger switch | Finger switch | Finger switch |
| Delivery device | Direct illumination to tissue | Direct illumination tissue | Direct illumination tissue |
| Indication for use/Intended use | IPL Home Use Hair Removal Device is an over-the-counter device intended for removal of unwanted body and/or facial hair. | IPL Home Use Hair Removal Device is an over-the-counter device intended for removal of unwanted body and/or facial hair. | The iPulse SmoothSkin Gold Hair Removal System is indicated for the removal of unwanted hair. The iPulse Smoothskin Gold is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime. |
| Location for use | OTC | OTC | OTC |

VII. Performance Data

The following performance data were provided in support of the substantial equivalence determination.

1) Biocompatibility Testing

The biocompatibility evaluation for the body-contacting components of the subject device was conducted in accordance with the "Use of International Standard ISO 10993-1, 'Biological Evaluation of Medical Devices –Part 1: Evaluation and Testing Within a Risk Management Process", as recognized by FDA. The battery of testing was performed to, and passed, including:

- ISO 10993-5 Biological Evaluation of Medical Devices –Part 5: Tests for In Vitro Cytotoxicity

- ISO 10993-10 Biological Evaluation of Medical Devices –Part 10: Tests for Irritation and Skin Sensitization

2) Electrical Safety and Eye Safety

Electrical safety and Eye safety testing was performed to, and passed, the following standards:

- IEC 60601-1-2 Medical electrical equipment –Part 1-2: General requirements for basic safety and essential performance –Collateral standard: electromagnetic compatibility
- IEC 60601-1 Medical electrical equipment –Part 1: General requirements for basic safety and essential performance
- IEC 60601-1-11 Medical Electrical Equipment –Part 1: General Requirements for Basic Safety and Essential Performance –Collateral Standard: Requirements for Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment
- IEC 60601-2-57 Medical electrical equipment –Part 2-57: Particular requirements for the basic safety and essential performance of non-laser source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use

3) Eye Safety

- IEC 62471 Photobiological safety of lamps and lamp systems

4) Software Verification and Validation

Software documentation consistent with *moderate level* of concern was submitted in this 510(k). System validation testing presented in this 510(k) demonstrated that all software requirement specifications are met and all software hazards have been mitigated to acceptable risk levels.

Summary

Based on the above performance as documented in this application, the subject device IPL Home Use Hair Removal Device was found to have a safety and effectiveness profile that is similar to the predicate device.

VIII. Conclusions

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the comparison of intended use, design, materials and performance, the subject device IPL Home Use Hair Removal Device is to be concluded substantial equivalent to its predicate devices.