



August 16, 2021

Penumbra, Inc.
Nikita Patel
Regulatory Specialist I
One Penumbra Place
Alameda, California 94502

Re: K211654

Trade/Device Name: Penumbra System (Reperfusion Catheter RED 72)
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: NRY
Dated: May 27, 2021
Received: May 28, 2021

Dear Nikita Patel:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Naira Muradyan, Ph.D.
Assistant Director
DHT5A: Division of Neurosurgical,
Neurointerventional
and Neurodiagnostic Devices
OHT5: Office of Neurological
and Physical Medicine Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K211654

Device Name
Penumbra System® (Reperfusion Catheter RED™ 72)

Indications for Use (Describe)

Penumbra Reperfusion Catheters and Separators

As part of the Penumbra System, the Reperfusion Catheters and Separators are indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments, basilar, and vertebral arteries) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.

Penumbra 3D Revascularization Device

As part of the Penumbra System, the Penumbra 3D Revascularization Device is indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.

Penumbra Aspiration Tubing

As part of the Penumbra System, the Penumbra Sterile Aspiration Tubing is indicated to connect the Penumbra Reperfusion Catheters to the Penumbra Aspiration Pump.

Penumbra Aspiration Pump

The Penumbra Aspiration Pump is indicated as a vacuum source for Penumbra Aspiration Systems.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) Summary

(as required by 21 CFR 807.92)

Pursuant to Section 12, Part (a)(i)(3A) of the Safe Medical Devices Act of 1990, Penumbra, Inc. is providing the summary of Substantial Equivalence for the subject Penumbra System® (Reperfusion Catheter RED™ 72).

1.1 Submitter

Penumbra, Inc.
 One Penumbra Place
 Alameda, CA 94502 USA

Contact Person:
 Nikita Patel
 Regulatory Specialist
 Tel: (408) 823-1820
 Fax: (510) 217-6414
 E-mail: npatel@penumbrainc.com

Date of Preparation:
 August 12, 2021

1.2 Subject Device

Penumbra System® (Reperfusion Catheter RED™ 72)

Regulatory Class: II
 Classification Panel: Neurology
 Classification Name: Percutaneous Catheter
 Regulation Number: 21 CFR 870.1250
 Product Code: NRY

1.3 Predicate Device & Reference Devices

510(k) Number	Name of Device	Name of Manufacturer
<i>Predicate Device</i>		
K173761	Penumbra System (Reperfusion Catheter JET 7)	Penumbra, Inc.
<i>Reference Device</i>		
K190010	Penumbra System® (Reperfusion Catheter JET™ 7)	Penumbra, Inc.

1.4 Device Description

The Penumbra System[®] is comprised of the following devices:

- Penumbra Reperfusion Catheter
- Penumbra Aspiration Pump
- Penumbra Aspiration Pump/Canister Tubing
- Penumbra Aspiration Tubing
- Penumbra Separator
- Penumbra 3D Revascularization Device

The Penumbra System is designed to remove thrombus from the vasculature using continuous aspiration. The Reperfusion Catheter targets aspiration from the pump directly to the thrombus. The 3D Revascularization Device is used with Reperfusion Catheters to facilitate aspiration and removal of the thrombus when needed. The Separator may be used to clear the lumen of the Reperfusion Catheter should it become blocked with thrombus. The use of the Separator may not be necessary when using a Reperfusion Catheter with an ID of 0.054 in. or larger. The Reperfusion Catheter is introduced through a guide catheter or long femoral sheath and into the intracranial vasculature and guided over a neurovascular guidewire to the site of the primary occlusion. The Penumbra Reperfusion Catheter is used with the Aspiration Pump to aspirate thrombus from an occluded vessel. As needed, a Penumbra Separator may be deployed from the Reperfusion Catheter to assist with thrombus removal. The Penumbra Separator is advanced and retracted through the Penumbra Reperfusion Catheter at the proximal margin of the primary occlusion to facilitate clearing of the thrombus from the Reperfusion Catheter tip. For the aspiration source, the Penumbra Reperfusion Catheter is used in conjunction with the Aspiration Pump, which is connected using the Penumbra Aspiration Tubing and the Penumbra Pump/Canister Tubing. The Penumbra Reperfusion Catheter is provided with a steam shaping mandrel and rotating hemostasis valve, and a peelable sheath. The Penumbra 3D Revascularization Device is provided with an introducer sheath. The Penumbra Separator is provided with an introducer and torque device. The Penumbra Reperfusion Catheters, 3D Revascularization Device and Separators are visible under fluoroscopy.

1.5 Indications For Use

Penumbra Reperfusion Catheters and Separators

As part of the Penumbra System, the Reperfusion Catheters and Separators are indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments, basilar, and vertebral arteries) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.

Penumbra 3D Revascularization Device

As part of the Penumbra System, the Penumbra 3D Revascularization Device is indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.

Penumbra Aspiration Tubing

As part of the Penumbra System, the Penumbra Sterile Aspiration Tubing is indicated to connect the Penumbra Reperfusion Catheters to the Penumbra Aspiration Pump.

Penumbra Aspiration Pump

The Penumbra Aspiration Pump is indicated as a vacuum source for Penumbra Aspiration Systems.

1.6 Comparison of Technological Characteristics with the Predicate Device

Device Attribute	Predicate Device	Subject Device
Trade Name	Penumbra System (Reperfusion Catheter JET 7)	Penumbra System (Reperfusion Catheter RED 72)
FDA Product Classification	Class II, NRY, 21 CFR 870.1250	SAME
510(k) Number	K173761	K211654
Indications for Use	<u>Penumbra Reperfusion Catheters and Separators</u> As part of the Penumbra System, the Reperfusion Catheters and Separators are indicated for use in the revascularization of patients with	SAME

Device Attribute	Predicate Device	Subject Device
	<p>acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments, basilar, and vertebral arteries) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.</p> <p><u>Penumbra 3D Revascularization Device</u> As part of the Penumbra System, the Penumbra 3D Revascularization Device is indicated for use in the revascularization of patients with acute ischemic stroke secondary to intracranial large vessel occlusive disease (within the internal carotid, middle cerebral – M1 and M2 segments) within 8 hours of symptom onset. Patients who are ineligible for intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment.</p> <p><u>Penumbra Aspiration Tubing</u> As part of the Penumbra System, the Penumbra Sterile Aspiration Tubing is indicated to connect the Penumbra Reperfusion Catheters to the Penumbra Aspiration Pump.</p> <p><u>Penumbra Aspiration Pump</u> The Penumbra Aspiration Pump is indicated as a vacuum source for Penumbra Aspiration Systems.</p>	
Principles of Operation	See Section 1.4	SAME
Device Materials	Stainless Steel, PTFE, Polyruethane, Polyether Block Amide, Nylon 12, Nitinol, Platinum/Iridium	SAME
ID Band Color	Polyolefin, PET black [white foil]	SAME
Coating	Hydrophilic coating (proprietary)	Equivalent

Device Attribute	Predicate Device	Subject Device
Min. ID	0.072 in. (1.83 mm)	SAME
Max. OD	0.085 in. (2.16 mm)	SAME
Distal Flex Length	30 cm	SAME
Coating Length	30 cm	SAME
Effective Lengths	115, 120, 125, 127, 132 cm	SAME
Accessories	Peelable Sheath, Shaping Mandrel, RHV	SAME
Packaging Materials	Polyester/Polyethylene/Tyvek, Polystyrene, SBS Paperboard	SAME
Condition Supplied	Sterile and Single Use	SAME
Sterilization Method	EO	SAME

1.7 Performance Data

The following performance data were provided in support of the substantial equivalence determination:

- Design Verification
- Biocompatibility
- Shelf-Life
- Sterilization – LAL Validation

The subject device met all established requirements.

1.7.1 Design Verification Testing

The following design verification tests were performed on the subject device:

Test	Test Method Summary	Conclusion
Dimensional/Visual Test	Confirms the units meet all dimensional and visual product specifications.	Acceptance Criteria Met
Friction Test	Confirms units meet product specification related to friction.	Acceptance Criteria Met
Fluoroscopy Test	Confirms the markerband is fluoroscopically visible.	Acceptance Criteria Met
Simulated Use Test	Confirms the functionality of units using clinically relevant benchtop model.	Acceptance Criteria Met
Particulate Test	Particulates generated during simulated use (including multiple deployment cycling) were evaluated.	Acceptance Criteria Met
Hub/Air Test	Confirms units have no leaks when tested.	Acceptance Criteria Met
Tensile Test	Confirms units meet product specification related to tensile strength.	Acceptance Criteria Met
Pressure Test	Confirms units meet product specification related to pressure.	Acceptance Criteria Met
Elongation Test	Confirms units meet product specification related to elongation.	Acceptance Criteria Met

Test	Test Method Summary	Conclusion
Corrosion Resistance Test	Confirms there is no visible corrosion on the units when tested.	Acceptance Criteria Met
Torque Strength Test	Confirms units have sufficient torque strength.	Acceptance Criteria Met
Burst Pressure Test	Confirms units can withstand sufficient pressure.	Acceptance Criteria Met
Shelf-Life	Confirms expiration date based on accelerated aging test studies.	Acceptance Criteria Met
Sterilization Test	Confirms the units are sterilized in accordance with ISO 11135 and ISO 10993-7.	Acceptance Criteria Met

1.7.2 Biocompatibility

Biocompatibility of the subject device is supported by the biocompatibility testing of the predicate and reference devices. Compared with the predicate device (K173761), all materials are the same, aside from the hydrophilic coating, which is identical to the reference device (K190010). Therefore, the previous biocompatibility testing on the predicate and reference devices applies to the subject device and additional testing is not required. The studies were selected in accordance with ISO 10993-1 guidelines (Biological Evaluation of Medical Devices) for a limited exposure (< 24 hours), externally communicating device with circulating blood contact. All studies were conducted pursuant to 21 CFR, Part 58, Good Laboratory Practices (GLP).

Tests	Acceptance Criteria	Conclusion
Cytotoxicity: MEM Elution (ISO 10993-5)	Sample extracts must have a cytotoxic reactivity score of grade 2 or lower.	Pass Non-cytotoxic
Sensitization: Magnusson-Kligman Method (ISO 10993-10)	Test Group shall yield Grade < 1 score on Magnusson and Kligman scale (provided Control Group yields Grade < 1).	Pass Non-sensitizing
Irritation: Intracutaneous Reactivity (ISO 10993-10)	The difference between the average scores for the extract of the test article and the control is ≤ 1.0 .	Pass Non-irritating
Systemic Toxicity: Acute Systemic Injection (ISO 10993-11)	Sample extracts must not cause significant biological reaction greater than control. That is: <ul style="list-style-type: none"> • Death in 2 or more animals. • Signs of toxicity in 2 or more animals (i.e. convulsions, prostration). • Weight loss > 10% in 3 or more animals. 	Pass Non-toxic
Systemic Toxicity: Material Mediated Pyrogen (ISO 10993-11)	Sample extracts must not cause a total rise in body temperature of $\geq 0.5^{\circ}\text{C}$.	Pass Non-pyrogenic
Hemocompatibility: Thrombogenicity (ISO 10993-4)	The test article must have similar or less thrombus formation than predicate.	Pass Non-thrombogenic
Hemocompatibility: Partial Thromboplastin Time (PTT) (ISO 10993-4)	Clotting times of test article must be similar to predicate values.	Pass Hemocompatible

Tests	Acceptance Criteria	Conclusion
Hemocompatibility: Complement Activation (ISO 10993-4)	The concentration of SC5b-9 of test article must be similar to predicate values.	Pass Hemocompatible
Hemocompatibility: Hemolysis, indirect contact (ISO 10993-4)	Sample extracts must be non-hemolytic ($\leq 2\%$ hemolytic index).	Pass Non-hemolytic
Hemocompatibility: Hemolysis, direct contact (ISO 10993-4)	Sample must be non-hemolytic ($\leq 2\%$ hemolytic index).	Pass Non-hemolytic

Biocompatibility test results demonstrate biological safety per ISO 10993 and USP requirements.

1.7.3 Performance Data – Animal, Clinical

No animal or clinical study was conducted as bench testing was determined sufficient for verification and validation purposes.

1.8 Conclusions

The subject Penumbra System[®] (Reperfusion Catheter RED[™] 72) is substantially equivalent to the predicate device Penumbra System (Reperfusion Catheter JET 7). The subject device has identical intended use as the predicate device. The device testing described in the 510(k) Summary demonstrates the subject device is substantially equivalent to the predicate device in regard to intended use, operating principle, design concept, fundamental technology and device performance.