



December 11, 2021

Iconic Medicare Sdn Bhd
% A.C. Thirumaran
Official Correspondent
Integrated Assessment Services Pvt Ltd
1495, Manasarovar, 16th Main Road, Anna Nagar West
Chennai, Tamil Nadu 600040
India

Re: K212182

Trade/Device Name: Iconic Nitrile Glove
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulatory Class: Class I, reserved
Product Code: LZA
Dated: October 26, 2021
Received: November 10, 2021

Dear A.C. Thirumaran:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray, III, PhD.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K212182

Device Name
Iconic Nitrile Glove

Indications for Use (Describe)

This non-powdered patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

**510(K) Summary
K212182**

ICONIC NITRILE GLOVE

Preparation Date: 12/10/2021

1. Submitter:

Mr. Tan Cho Chia
Managing Director
Company Name: Iconic Medicare Sdn Bhd.
Company Address: No. 1-2,jalan City, Icon City.14000, Bukit Mertajam, Pulau Pinang, Malaysia
Telephone No: 6045040588
Email: cctan@iconic.com.my

2. Name of the Device

Trade Name / Proprietary Name: Iconic Nitrile Glove
Device Common Name: Powder Free Nitrile Examination gloves.
Device Classification Name: Polymer Patient Examination gloves (21 CFR 880.6250). Device Class: Class I.
Product Code: LZA

3. Official Correspondent

Mr.A.C.Thirumaran
Integrated Assessment Services Private Limited
No.1495, Manasarovar, 16th Main road,
Anna Nagar west,
Chennai- 600040,
India.
Email: iasfda16@gmail.com

4. Identification of the Legally Marketed Device:

Predicate Device: Meditech Gloves SDN BHD
510k Number: - K210755
Device Name: Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue)
Device Classification Name: Patient Examination gloves (21 CFR 880.6250).
Device Class: Class I.
Product Code: LZA

5. Device Description

The subject device in this 510(k) Notification is Iconic Nitrile Glove - Powder Free Nitrile Examination Glove. The subject device is a patient examination glove made from Nitrile compound, Blue color, powder free and non-sterile (Per 21 CFR 880.6250, class I). The device meets the specifications in ASTM D6319-19 Standard specification for Nitrile Examination Gloves. The available sizes of the subject devices are Small, Medium, Large, X-Large.

6. Intended use of the Device

Iconic Nitrile Glove, a polymer patient examination glove is a disposable device & Non Sterile which is intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner. It is for over-the-counter use.

7. Technological characteristics Comparison for the proposed and predicate devices

Characteristics	Acceptance Criteria	Subject device: Iconic Nitrile Glove- Blue (Small, Medium, Large, X-Large)	Predicate Device Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue)	Remarks
Product Code	LZA	LZA	LZA	same
Intended use	A powder free patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner. The device is for over-the-counter use.	This powder free patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner. The device is for over-the-counter use.	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	same
Material used	Nitrile compound	Nitrile compound	Nitrile compound	same
Colour	N/A	Blue	Blue	same
Sterility	Sterile/Non-sterile	Non sterile	Non sterile	same
Single use	Single use	Single use	Single use	same
Dimensions	Overall Length (mm) Min 230mm	Meets ASTM D6319-19	Meets ASTM D6319- 19	same
	Width (±10mm) Small - 80 Medium- 95 Large-110 X-large-120	Meets ASTM D6319- 19	Meets ASTM D6319- 19	
	Thickness at Palm (mm) Min; 0.05 mm	Meets ASTM D6319- 19	0.06– 0.09mm	
	Thickness at Finger Tip (mm) Min 0.05 mm	Meets ASTM D6319- 19	0.07– 0.10mm	
Physical Properties	a.Before Aging (i) Tensile Strength=14 MPa, min. (ii) Ultimate Elongation= 500 % min	Meets ASTM D6319- 19	Meets ASTM D6319- 19	same

Iconic Medicare Sdn Bhd
510(k) Premarket submission
Iconic Nitrile Glove

	<p>b. After Accelerated Aging (i) Tensile Strength=14 MPa, min. (ii)Ultimate Elongation= 400 % min</p>			
--	---	--	--	--

Freedom from pinholes	AQL 2.5 Inspection Level G-1	Meets ASTM D5151- 19	Meets ASTM D5151- 19	same
Residual Powder	< 2.0 mg/pc	Meets ASTM D6124- 06 Result obtained: 0.24	Meets ASTM D6124- 06	same
Biocompatibility	ISO 10993-10:2010 Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization	Under the conditions of this study, the test article was a non- irritant.	Under the conditions of this study, the test article was a non- irritant.	same
	ISO 10993-10:2010 Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization	Under the conditions of this study, the test article was a non- sensitizer.	Under the conditions of this study, the test article was a non- sensitizer.	same
	ISO 10993-5:2009 Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity	Complies with the requirement of this standard	Complies with the requirement of this standard	Same
	ISO 10993-11:2017 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity	Not induce systemic toxicity	N/A	same

8. Summary of non-clinical testing results

Iconic Nitrile Glove was tested and found in conformance with the following standards:

ASTM D6319-19	Standard Specification for Nitrile Examination Gloves for Medical Application
ASTM D5151-19	Standard Test Method for detection of Holes in Medical Gloves
ASTM D6124-06	Standard Test Method for Residual Powder on Medical Gloves
ISO 10993-5	Biological evaluation of medical devices Part 5: Tests for in vitro cytotoxicity
ISO 10993-10	Biological evaluation on medical device Part 10: Test for irritation and Skin Sensitization
ISO 10993-11	Biological evaluation of medical devices Part 11: Tests for systemic toxicity
ISO 10993-23	Biological evaluation of medical devices - Part 23: Tests for irritation

Test Methodology	Purpose	Acceptance Criteria	Average Results				Final status	
			SMALL	Medium	Large	X Large		
ASTM D6319-19	Sterility	-	Non sterile	Non sterile	Non sterile	Non sterile	-	
	Freedom from hole - ASTM D5151-19	AQL 2.5	Pass	Pass	Pass	Pass	Pass	
	Dimensions – width, Length, Thickness	Overall Length (mm) Min 230mm.		234.9	246	236.5	241.0	Pass
		Width (±10mm) Small - 80 Medium- 95 Large-110 X-large-120		84.7	97.4	106.4	114.6	Pass
		Thickness at Palm & fingertip Min: 0.05 mm						
		Palm		0.10	0.06	0.08	0.09	Pass
		Fingertip		0.12	0.10	0.13	0.13	Pass
		Physical properties before aging, after accelerated aging	a. Before Aging					
	Tensile Strength=14 MPa, min.			21.2	25	18.9	18.4	Pass
	Ultimate Elongation= 500 % min			896.3	893.8	919.7	930.9	Pass
	b. After Accelerated Aging							
	Tensile Strength=14 MPa, min.			23.5	26.8	22.2	19.3	Pass
	Ultimate Elongation= 400 % min			879.2	787.9	855.3	858.6	Pass
	Powder-free Residue exceeds maximum limit - ASTM D6124-06	< 2.0 mg per glove		0.78	0.24	0.38	0.54	Pass
	ISO 10993-5	Test for Invitro cytotoxicity	Cytotoxic Characteristics	Pass				
ISO 10993-10	Test for irritation and Skin Sensitization	Non - Skin Sensitized	Pass					

Iconic Medicare Sdn Bhd
510(k) Premarket submission
Iconic Nitrile Glove

ISO 10993-11	Tests for systemic toxicity	Not induce systemic toxicity	Pass
ISO 10993-23	Tests for irritation	Non-Irritant	Pass

9. Summary of clinical Performance data

Not applicable - Clinical data was not used to assess performance of the subject device.

10. Conclusion

The Conclusion drawn from the Non-Clinical test demonstrates that the subject device- Iconic Nitrile Glove is as safe, as effective, and performs as well as or better than the legally marketed Predicate device cleared under K210755.