



July 6, 2022

ChoiceSpine, LLC  
Kim Finch  
Director of Regulatory Affairs  
400 Erin Drive  
Knoxville, Tennessee 37919

Re: K220024  
Trade/Device Name: ChoiceSpine Navigation System  
Regulation Number: 21 CFR 882.4560  
Regulation Name: Stereotaxic Instrument  
Regulatory Class: Class II  
Product Code: OLO  
Dated: May 6, 2022  
Received: May 10, 2022

Dear Kim Finch:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K220024

Device Name

ChoiceSpine Navigation System

Indications for Use (Describe)

The ChoiceSpine Navigation reusable instruments are intended to be used during preparation and placement of ChoiceSpine system implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. The ChoiceSpine Navigation reusable instruments are specifically designed for use with the Medtronic StealthStation S7 (V2.1.0) and S8 (V1.2.0) systems, which are indicated for any medical condition in which the use of stereotactic surgery may be appropriate and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks for the anatomy.

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**

Date: May 6, 2022

Sponsor: ChoiceSpine, LLC  
400 Erin Drive  
Knoxville, TN 37919

Phone: 865-243-3969

Fax: 865-246-3334

Contact Person: Kim Finch, Director of Regulatory Affairs

Proposed Proprietary Trade Name: ChoiceSpine Navigation System

Product Class: Class II

Classification Name: ChoiceSpine Navigation System

- 882.4560 Neurological Stereotaxic Instrument

Device Product Code: ChoiceSpine Navigation System

- OLO

Purpose of Submission: The purpose of this submission is to allow for ChoiceSpine Navigation system instruments to be compatible with the Medtronic® StealthStation® Navigation System.

Device Description: The ChoiceSpine Navigation instruments are non-sterile, reusable instruments designed to function with the Medtronic® StealthStation® System. Refer to the appropriate navigation system Instructions for Use and/or Surgical Technique Guide for details regarding navigation system use. The ChoiceSpine Navigation instruments are for use with any ChoiceSpine systems equipped with the appropriate and necessary features and geometry. These features and geometry have been tested for compatibility with the Medtronic® StealthStation®.

Indications for Use: The ChoiceSpine Navigation reusable instruments are intended to be used during preparation and placement of ChoiceSpine system implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. The ChoiceSpine Navigation reusable instruments are specifically designed for use with the Medtronic StealthStation S7 (V2.1.0) and S8 (V1.2.0) systems, which are indicated for any medical condition in which the use of stereotactic surgery may be appropriate and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks for the anatomy.

Materials:	The instruments are manufactured from previously cleared materials for surgical instruments per ASTM F899 & A564.
Predicate Devices:	<b>Primary Predicate:</b> Medtronic, Navigated CD Horizon Solera Screwdriver/Taps (K140454) <b>Additional Predicate:</b> SI-BONE, Inc. IFuse Implant System- IFuse Navigation (K203110) ChoiceSpine Navigation System (K203561) Medtronic Navigated VERTEX SELECT® Instruments (K143628)
Performance Data:	The ChoiceSpine Navigation system dimensional analysis and positional accuracy (per ASTM F2554) validations were performed in side-by-side testing of the Medtronic® StealthStation® system and predicate device. The test results demonstrate that the ChoiceSpine Navigation system performance is substantially equivalent to the predicate.
Technological Characteristics:	The ChoiceSpine Navigation System has the same technological characteristics as the predicate device. Therefore, the fundamental scientific technology of the ChoiceSpine Navigation System is the same as previously cleared devices.
Conclusion:	The overall technological characteristics, performance data, and intended use of the ChoiceSpine Navigation System are the same which lead to the conclusion that the Navigation System is substantially equivalent to the predicate devices.