



September 2, 2022

Phu Bao Group Company Limited
% Prithul Bom
Most Responsible Person
Regulatory Technology Services, LLC
1000 Westgate Drive,
Suite 510k
Saint Paul, Minnesota 55114

Re: K222426

Trade/Device Name: Perfetta Moderate Medical Face Mask
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical apparel
Regulatory Class: Class II
Product Code: FXX
Dated: August 9, 2022
Received: August 11, 2022

Dear Prithul Bom:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K222426

Device Name

Perfetta Moderate Medical Face Mask

Indications for Use (Describe)

Perfetta Moderate Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This surgical face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is single use, disposable device, provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

This 510(k) Summary as per 21 CFR 807.92

Date of submission: July 25, 2022

(1) Applicant information

510(k) Owner/ Applicant: PHU BAO GROUP COMPANY LIMITED

Address:

Head office:

PHU BAO GROUP COMPANY LIMITED

5 floor, CJ Building, 02Bis-4-6 Le Thanh Ton Street,

Ben Nghe Ward, District 1, Hochiminh city, Vietnam, 70000

Factory:

PHU BAO MEDICAL CO., LTD

Lot C21, D2 Road, Cau Tram Industrial Park,

Long Trach Ward, Can Duoc Dist.,

Long An province, Vietnam.

Contact: Le Pham Minh Ngoc (Ms.)

Email: info@phubaogroup.com.vn;

consultant.fda@phubaomedical.com

Phone: +84-28-3914 2980

Owner/Operator Number: 10074628

FEI Number: 3016615220

US Agent: THANH PHONG KIEM NGUYEN GOLUS LLC at address:

421 Sletten Drive, Lawrenceville, GA US 30046

Contact of US agent: THANH PHONG KIEM NGUYEN, Phone: 470 4189790,

Email: golus@gol.vn

Correspondent: Regulatory Technology Services at address: 1000

Westgate Drive, Suite #510, Saint Paul, Minnesota, 55114

Contact of Correspondent: Ms. Prithul Bom - Accredited Person, Reviewer

Phone 612-963-0379

Email: prithul.bom@rts3pro.com

(2) Subject device

Trade name: Perfetta Moderate Medical Face Mask
Model number: PT3B50
Common Name: Surgical Face Mask
Classification Name: Masks, Surgical
Review Panel: General Hospital
Regulation Medical Specialty: General & Plastic Surgery
Product Code: FXX
Device Classification: Class II per 21 CFR §878.4040

(3) Predicate device

Submitter: MEGASOFT(CHINA) CO., LTD
Device name: Surgical Face Mask
510(k) number: K213617
Product Code: FXX
Device Classification: Class II per 21 CFR §878.4040

(4) Description of device

The subject device is three-layers, flat-pleated mask constructed of nonwoven polypropylene materials, the inner and outer layers are made of spunbond polypropylene, and the middle layer is made of melt blown polypropylene.

The subject device is provided with ear loops. Ear loops is made of 80% polyester 20% spandex, not made with natural rubber latex.

A zinc wire nose piece with polypropylene coating is placed within the binding for comfort and individualized fit, allow the user to fit the facemask around their nose.

The subject device is provided in blue color. The blue colorant is made of polypropylene master batch.

This subject device is provided with size 17.5 cm ± 0.5 cm.

The subject device is single-use, disposable device, provided non-sterile.

(5) Indications for use

Perfetta Moderate Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This surgical face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is single use, disposable device, provided non-sterile.

(6) Comparison of Technological Characteristics and Performance testing with the Predicate device

Item	Subject Device	Predicate Device	Comparison
510 (k) number	-	K213617	Not applicable
Applicant	PHU BAO GROUP COMPANY LIMITED	MEGASOFT(CHINA) CO., LTD	Not applicable
Product Name	Perfetta Moderate Medical Face Mask	Surgical Face mask	Not applicable
Model number	PT3B50	MGSM-01	Not applicable
Device class	Class II Device, FXX (21 CFR 878.4040)	Class II Device, FXX (21 CFR 878.4040)	Same
Indications for Use	Perfetta Moderate Medical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This surgical face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is single use, disposable device, provided non-sterile.	The Surgical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is single use, disposable device(s), provided non-sterile.	Same
Material			
Outer layer	Spunbond polypropylene	Spunbond polypropylene	Same
Middle layer	Meltblown polypropylene	Meltblown polypropylene filter	Same

Inner layer	Spunbond polypropylene	Spunbond polypropylene	Same
Nose piece	Zinc wire with polypropylene coating	Polypropylene, iron and zinc	Different
Design Features	Ear loops: 80% polyester 20% spandex	Ear loops: elastics film	Similar
Mask Style	Flat Pleated	Flat Pleated	Same
Color	Blue	Blue	Same
Specification and Dimension	Length: 17.5cm±0.5cm Width: 9.5cm±0.5cm	Length: 175mm±5mm Width: 95mm±5mm	Same
Dimension ear loops	Length: 17.5cm±0.5cm Width: 0.40cm±0.02cm	-	Different
OTC use	Yes	Yes	Same
Sterility	Non-Sterile	Non-Sterile	Same
Use	Single Use, Disposable	Single Use, Disposable	Same
Performance Characteristics			
Fluid Resistance Performance ASTM F1862	Pass at 120 mmHg (Level 2 Fluid Resistance)	Pass at 120 mmHg (Level 2 Fluid Resistance)	Same
Particulate Filtration Efficiency ASTM F2299	≥98% at 0.1µm	≥98% at 0.1µm	Same
Bacterial Filtration Efficiency ASTM F2101-19	≥98%	≥98%	Same
Differential Pressure (Delta P)	< 6 mmH ₂ O/cm ²	< 6 mmH ₂ O/cm ²	Same
Flammability 16 CFR 1610	Class 1	Class 1	Same
Biocompatibility			

Cytotoxicity	Under the conditions of the studies, the subject device is non-cytotoxic	Under the conditions of the study, the device is non-cytotoxic	Same
Irritation	Under the conditions of the studies, the subject device is non-irritant	Under the conditions of the study, the device is non-irritating	Same
Sensitization	Under the conditions of the studies, the subject device is non-sensitizer	Under the conditions of the study, the device is non-sensitizing	Same

The subject device is same indication for use, design feature, design style, dimension of size, component of material (except nose piece), performance characteristics and biocompatibility with predicate device.

The subject device is different material of nose piece to predicate device. The subject device which was included the nose piece had been tested and the result was complied with acceptance criteria. This result demonstrates that the difference on material of nose piece do not affect the safety performance and effectiveness of subject device.

The component of ear loop of subject device is similar with predicate device but clearly claim the percent of composition.

The subject device is provided with detail dimension of ear loops but cannot compare with predicate device because no information about dimension of ear loops of predicate device is mentioned in its 510(k) summary. The subject device which was included the ear loop with its dimension had been tested and the result was complied with acceptance criteria. This result demonstrates that the difference on dimension of ear loop do not affect the safety performance and effectiveness of subject device.

(7) Summary of Non-clinical test

Performance characteristics			
Test Method	Subject Device	Acceptance Criteria	Conclusion
Fluid Resistance Performance ASTM F1862	≥29 out of 32 passed in 120 mmHg (Level 2)	≥29 out of 32 passed in 120 mmHg (Level 2)	Pass
Particulate Filtration Efficiency ASTM F2299	≥98% at 0.1µm	≥98% at 0.1µm	Pass

Section 5_510(k) Summary

Bacterial Filtration Efficiency ASTM F2101-19	≥98%	≥98%	Pass
Differential Pressure (Delta P)	< 6 mmH ₂ O/cm ²	< 6.0 mm H ₂ O/cm ²	Pass
Flammability 16 CFR 1610	Class 1(≥ 3.5 seconds)	Class 1 (≥ 3.5 seconds)	Pass

Performance testing were performed on three non-consecutive lots to support that the performance specification are maintained cross production lots and the lot-to-lot variability in performance is acceptable.

Sample size of each lot is complied with ISO 2859-1, general inspection level II as FDA recommendation and acceptance quality limit (AQL) of 4%.

The results of performance testing of subject device demonstrate that the subject device met all design specification as was same to predicate device, complied with requirements in guidance: “Surgical Masks – Premarket Notification [510(k)] Submissions”, complied with standard ASTM F2100 at level 2.

Biocompatibility			
Standard	Subject Device	Acceptance Criteria	Conclusion
Cytotoxicity ISO 10993-5	Under the conditions of the studies, the device is non-cytotoxic.	The device is non-cytotoxic	Pass
Skin Sensitization test ISO 10993-10	Under the conditions of the studies, the subject device is non-sensitizing	The device is non-sensitizing	Pass
Skin Irritation test ISO 10993-23	Under the conditions of the studies, the device is non-irritating.	The device is non-irritating	Pass

The biocompatibility evaluation of this subject device was conducted according to ISO 10993-1: nature of body contact of the subject device is belonged to category Surface device, intact skin, with contact duration A-limited (≤24h)

The subject device was evaluated in its final finished form, in blue color.

The results of biocompatibility testing of subject device demonstrate that the subject device is biocompatible and safe for its intended use.

(8) Summary of Clinical test

No clinical study is included in this submission.

(9) Conclusion

The conclusions drawn from the nonclinical tests demonstrate that the proposed subject device is as safe, as effective, and performs as well as the legally marketed predicate device.