



July 16, 2021

Soniquence, LLC
Ms. Suzanne Lucas, BA
Sr. Regulatory Affairs Specialist
2477 Grand Avenue
Baldwin, New York 11510

Re: K211946

Trade/Device Name: Soniquence Reusable Bipolar Cable
Regulation Number: 21 CFR 878.4400
Regulation Name: Electrosurgical Cutting and Coagulation Device and Accessories
Regulatory Class: Class II
Product Code: GEI
Dated: June 16, 2021
Received: June 23, 2021

Dear Ms. Lucas:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531 - 542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Long Chen, Ph.D.
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211946

Device Name

Soniquence Reusable Bipolar Cables

Indications for Use (Describe)

The Soniquence Reusable Bipolar Cable is intended to transfer power from an RF Generator to Bipolar Electrodes.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



K211946



510(k) SUMMARY

(As required by 21 CFR 807.92(a))

Date Prepared

July 15, 2021

Submitter's Information (807.92(a)(1))

Company Name and Address:

Soniquence, LLC
2477 Grand Avenue
Baldwin, NY 11510
Phone: (516) 634-1370
www.soniquence.com

Establishment Registration #: 3014982808

Contact Information:

Ms. Suzanne Lucas, BA
Sr. Regulatory Affairs Specialist
Soniquence, LLC
2477 Grand Avenue
Baldwin, NY 11510
Phone: (516) 634-1370
Fax: (516) 634-1371
Email: slucas@soniquence.com

Device Information (807.92(a)(2))

Trade Name: Soniquence Reusable Bipolar Cables

Common/Usual Name: Electrosurgical, Cutting & Coagulation Device & Accessories

Classification Name and Regulation: Electrosurgical Cutting and Coagulation Device and Accessories, 21 CFR 878.4400

Classification Panel: General and Plastic Surgery

Device Class/Product Code

FDA Classification: Class 2
FDA Product Code: GEI



K211946



Predicate Devices (807.92(a)(3))

- Soniquence Bipolar Electrodes (K190336)
- Tecno Instruments (K042608)

Device Description (807.92(a)(4))

The subject device, Soniquence Reusable Bipolar Cable is a bipolar accessory designed to transfer the electrosurgical power to the Bipolar electrode. It is connected to a High Frequency Generator via the male end and the female end termination connects to the standard Bipolar instrument. The cable is made of Silicone material and Polypropylene H1500 which is flexible and autoclavable. The length of the cable is 3 meters, and it is provided non-sterile.

Intended Use:

The Soniquence Reusable Bipolar Cable is intended to transfer power from an RF Generator to Bipolar Electrodes.

Substantial Equivalence Comparison (807.92(a)(6))

The Soniquence Reusable Bipolar Cables are substantially equivalent in intended use, technological characteristics, operating principle, and performance characteristics to the predicate device by Tecno (K042608).

Predicate Information	Soniquence Reusable Bipolar Cable (SUBJECT DEVICE)	Tecno Reusable Bipolar Cables (K042608) PREDICATE
Intended Use	The Soniquence Reusable Bipolar Cable is intended to transfer power from an RF Generator.to Bipolar Electrodes.	An electrosurgical cable is a device intended to connect Electrosurgical Generator with electrosurgical cutting and coagulation devices.
Product Code	GEI	GEI
Device Classification	Class II	Class II
Functions	Connect to High Frequency Generator to provide power to Bipolar Electrodes.	Connect to Electrosurgical Generator to provide power to Bipolar Electrodes.
Operating principle	Bipolar	Bipolar
Energy Source	RF Energy	RF Energy
Cable Length (m)	3m	3m
Activation Method	Footswtich	Footswtich
Materials	Silicon and Polypropylene H1500	Silicon and Santoprene
Testing standards used	IEC 60601-2-2:2017	IEC 60601-2-2:2017
Steam Sterilization Validation	ISO 17665-1:2006	ISO 17665-1:2006



K211946



Non-Clinical Testing (807.92(b)(1))

The Soniquence Reusable Bipolar Cables will be manufactured in accordance with the design control requirements of 21 CFR 820.30. Appropriate non-clinical verification and validation activities were planned and conducted to address identified risks and ensure the safety and effectiveness of the device. The following tests were successfully performed:

- IEC 60601-2-2: 2017 - Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories
- ISO 17665-5:2006 - Sterilization of health care products Moist heat Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices

Technological Characteristics

The devices are substantially equivalent to the predicate devices based on a comparison of physical and performance characteristics.

Clinical Testing (807.92(b)(2))

This premarket submission did not rely on the assessment of clinical performance data to demonstrate substantial equivalence.

Conclusion (807.92(b)(3))

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the similarity to the predicate devices in terms of technology, performance, and indications for use, Soniquence, LLC concludes that the subject device, Soniquence Reusable Bipolar Cables does not raise any issues of safety or effectiveness and are substantially equivalent to the predicate device as described above.