



November 24, 2020

Nutech Spine and Biologics
% Mr. Daniel Lanois
Consultant
SurgOp Support
101 Lamond Ct
Prosper, Texas 75078

Re: K202972

Trade/Device Name: Anterior Cervical Plate System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal Intervertebral Body Fixation Orthosis
Regulatory Class: Class II
Product Code: KWQ
Dated: March 17, 2020
Received: September 30, 2020

Dear Mr. Lanois:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K202972

Device Name

Anterior Cervical Plate System

Indications for Use (Describe)

The Nutech Spine and Biologics Anterior Cervical Plate System is intended for anterior screw fixation to the cervical spine (C2-T1) for the following indications: degenerative disc disease (as defined by neck pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), trauma (including fractures), tumors, deformity (kyphosis, lordosis or scoliosis), pseudarthrosis, failed previous fusions, spondylolisthesis, and spinal stenosis.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary (as required by 21 CFR 807.92)

Date Prepared	November 18, 2020		
Manufacturer	Nutech Spine and Biologics		
Address	600 Luckie Drive, Suite 424, Birmingham, Alabama 35223		
Telephone	205-329-7285		
Fax	888-475-7793		
Contact Person	Daniel Lanois Consultant		
Address	SurgOp Support LLC 101 Lamond Ct, Prosper, TX 75078		
Telephone	678-371-3605		
Fax			
Email	daniel@surgopsupport.com		
Trade Name	Anterior Cervical Plate System		
Common Name	Anterior Cervical Plate System		
Panel Code	Orthopedic/87		
Classification Name	Spinal Intervertebral Body Fixation Orthosis		
Class	Class II		
Regulation Number	21 CFR 888.3060		
Product Code	KWQ		
Name of Primary Predicate Device	510(k) #	Manufacturer	
Diamond Anterior Cervical Plate System	K100265	Amendia	
Description	The Anterior Cervical Plate System consists of single level titanium bone plates of various sizes and titanium bone screws in various diameters and lengths to allow for patient specific configurations. The triangular shape of the plate allows for adjacent level fusion without removal of prior Nutech Cervical Plates. Implants are made from Ti 6Al4V ELI, per ATSM F136. The instrumentation allows for simple preparation and placement of the implant construct.		
Indications and Intended Use	The Nutech Spine and Biologics Anterior Cervical Plate System is intended for anterior screw fixation to the cervical spine (C2-T1) for the following indications: degenerative disc disease (as defined by neck pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), trauma (including fractures), tumors, deformity (kyphosis, lordosis or scoliosis), pseudarthrosis, failed previous fusions, spondylolisthesis, and spinal stenosis.		
Technological Characteristics/ Substantial Equivalence	Documentation was provided to demonstrate that the Subject device, is substantially equivalent to the Predicate. The Subject device is substantially equivalent to the Predicate device in intended use, indications for use, materials, technological characteristics, and labeling.		
Performance Data	Static Compression Bending, Static Torsion, Static Tension Bending, and Dynamic Compression Bending testing (per ASTM F1717-18) confirmed that the Subject device performed as intended.		
Conclusion	Based on the intended use, indications for use, technological characteristics, materials, and comparison to Predicate devices, the Subject device has been shown to be substantially equivalent to legally marketed Predicate devices.		