



March 1, 2021

NuVasive, Incorporated
Jessica LeBlanc
Associate Manager, Regulatory Affairs
7475 Lusk Blvd.
San Diego, California 92121

Re: K210271

Trade/Device Name: NuVasive® Modulus® ALIF System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: OVD
Dated: January 29, 2021
Received: February 1, 2021

Dear Ms. LeBlanc:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for
Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210271

Device Name

NuVasive® Modulus® ALIF System

Indications for Use (Describe)

The NuVasive Modulus ALIF System is indicated for spinal fusion procedures in skeletally mature patients. The Modulus ALIF System 10°-20° lordotic cages may be used as a standalone system. The Modulus ALIF System 25°-30° lordotic cages must be used with supplemental internal spinal fixation systems (e.g., posterior pedicle screw and rod system) that are cleared by the FDA for use in the lumbar spine. The System is designed for use with autogenous and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone graft to facilitate fusion. The devices are to be used in patients who have had at least six months of non-operative treatment.

The NuVasive Modulus ALIF System is intended for use in interbody fusions in the lumbar spine from L2 to S1, following discectomy in the treatment of symptomatic degenerative disc disease (DDD), degenerative spondylolisthesis, and/or spinal stenosis at one or two adjacent levels. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The Modulus ALIF System implants can also be used as an adjunct to fusion in patients diagnosed with multilevel degenerative scoliosis and sagittal deformity; however, when used in these patients at multiple levels and for patients with degenerative spondylolisthesis and spinal stenosis at one or two adjacent levels, the Modulus ALIF System must be used with a supplemental internal spinal fixation system (e.g., pedicle screw system) cleared by the FDA for use in the lumbar spine in addition to the integrated screws.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

A. Submitted by:

Ms. Jessica LeBlanc
Associate Manager, Regulatory Affairs
NuVasive, Incorporated
7475 Lusk Blvd.
San Diego, California 92121
(951) 816-0973

Date Prepared: February 19, 2021

B. Device Name

Trade or Proprietary Name:	<i>NuVasive® Modulus® ALIF System</i>
Common or Usual Name:	Intervertebral Body Fusion Device
Classification Name:	Intervertebral Body Fusion Device
Device Class:	Class II
Classification:	21 CFR § 888.3080
Product Code:	OVD

C. Predicate Devices

The subject *NuVasive Modulus ALIF System* is substantially equivalent to the following devices:

Primary Predicate

- K203201 – NuVasive® Thoracolumbar Interbody Systems

Additional Predicates

- K193593 – NuVasive Modulus ALIF System

D. Device Description

The subject *NuVasive Modulus ALIF System* are interbody implants manufactured from titanium alloy (Ti-6Al-4V ELI) powder conforming to ASTM F3001. The solid and porous structures are simultaneously built using a powder bed fusion method. The hollow core, or graft aperture, allows for packing of graft to aid in the promotion of a solid fusion. Similarly, the macroporous internal lattice structure provides additional space for graft packing. The microporous, textured surface on the superior and inferior ends of the device serves to grip the adjacent vertebrae to resist migration of the device. The implants are available in a variety sizes and lordotic angles to suit the individual pathology and anatomical conditions of the patient. The *Modulus ALIF System* 10°-20° lordotic cages may be used as a standalone system. The *Modulus ALIF System* 25° - 30° lordotic cages must be used with supplemental internal spinal fixation systems (e.g. posterior pedicle screw and rod system) that are cleared by the FDA for use in the lumbar spine.

The purpose of this submission is to introduce a non-sterile option of the Modulus ALIF bolts.

E. Indications for Use

The *NuVasive Modulus ALIF System* is indicated for spinal fusion procedures in skeletally mature patients. The *Modulus ALIF System* 10°-20° lordotic cages may be used as a standalone system. The *Modulus ALIF System* 25°-30° lordotic cages must be used with supplemental internal spinal fixation systems (e.g., posterior pedicle screw and rod system) that are cleared by the FDA for use in the lumbar spine. The System is designed for use with autogenous and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone graft to facilitate fusion. The devices are to be used in patients who have had at least six months of non-operative treatment.

The *NuVasive Modulus ALIF System* is intended for use in interbody fusions in the lumbar spine from L2 to S1, following discectomy in the treatment of symptomatic degenerative disc disease (DDD), degenerative spondylolisthesis, and/or spinal stenosis at one or two adjacent levels. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The *Modulus ALIF System* implants can also be used as an adjunct to fusion in patients diagnosed with multilevel degenerative scoliosis and sagittal deformity; however, when used in these patients at multiple levels and for patients with degenerative spondylolisthesis and spinal stenosis at one or two adjacent levels, the *Modulus ALIF System* must be used with a supplemental internal spinal fixation system (e.g., pedicle screw system) cleared by the FDA for use in the lumbar spine in addition to the integrated screws.

F. Technological Characteristics

As was established in this submission, the subject *NuVasive Modulus ALIF System* is substantially equivalent to other predicate devices cleared by the FDA for commercial distribution in the United States. The subject device was shown to be substantially equivalent and have the same technological characteristics to its predicate devices through comparison in areas including design, intended use, material composition, and function. The device does not contain software or electrical equipment

G. Performance Data

The purpose of this submission is to introduce a non-sterile option of the Modulus ALIF bolt. No new implant designs are being introduced, therefore no performance testing was performed for this 510(k) submission and the worst-case devices included with the subject system were tested and cleared in predicate 510(k) submissions.

H. Conclusions

Based on the indications for use, technological characteristics, and comparison to predicate devices, the subject *NuVasive Modulus ALIF System* has been shown to be substantially equivalent to legally marketed predicate devices.
