



Route 92 Medical  
Kathy Tansey  
Senior Director of Regulatory Affairs and Quality Assurance  
155 Bovet Road, Suite 100  
San Mateo, California 94402

August 14, 2020

Re: K201518  
Trade/Device Name: Route 92 Medical 088 Access System  
Regulation Number: 21 CFR 870.1250  
Regulation Name: Percutaneous Catheter  
Regulatory Class: Class II  
Product Code: QJP  
Dated: June 5, 2020  
Received: June 8, 2020

Dear Kathy Tansey:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Naira Muradyan, Ph.D.  
Assistant Director (Acting)  
DHT5A: Division of Neurosurgical,  
Neurointerventional  
and Neurodiagnostic Devices  
OHT5: Office of Neurological  
and Physical Medicine Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K201518

Device Name

Route 92 Medical 088 Access System

Indications for Use (Describe)

The Route 92 Medical 088 Access System, 143 cm, is indicated for use with compatible guide catheters in facilitating the insertion and guidance of microcatheters into a selected blood vessel in the neurovascular system.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(K) SUMMARY

### 510(k) Summary

<b>Sponsor:</b>	Route 92 Medical 155 Bovet Road, Suite 100 San Mateo, CA 94402 Phone: 650-581-1179
<b>Contact:</b>	Kathy Tansey
<b>Date Prepared:</b>	June 05, 2020
<b>Device Name:</b>	Route 92 Medical 088 Access System
<b>Common Name:</b>	Percutaneous Catheter
<b>Classification Name:</b>	Percutaneous Catheter (Product Code QJP, 21 CFR 870.1250)
<b>Predicate Device:</b>	Route 92 Medical 088 Access System, 110 cm, K200121

### Device Description

The Route 92 Medical 088 Access System, 143 cm, is comprised of a Support Catheter and a Delivery Catheter. The distal portion of the Support Catheter is a single-lumen, variable stiffness catheter. Like the predicate device, the proximal portion is a stainless-steel control wire. The Delivery Catheter is a hubbed, single-lumen variable stiffness catheter. Both catheters are hydrophilically coated. The Route 92 Medical Access System, 143 cm, is intended for use with compatible guide catheters in facilitating the insertion and guidance of microcatheters into a selected blood vessel in the neurovascular system. The devices are provided sterile and non-pyrogenic and are intended for single use only.

### Indications for Use

The Route 92 Medical 088 Access System, 143 cm, is indicated for use with compatible guide catheters in facilitating the insertion and guidance of microcatheters into a selected blood vessel in the neurovascular system.

### Comparison to the Predicate Device

The method of action, design, and materials of the Route 92 Medical 088 Access System, 143 cm, are equivalent to the Predicate Device as shown in the following table.

<b>Attribute</b>	<b>Predicate Device</b> Route 92 Medical 088 Access System, 110 cm (K200121)	<b>Subject Device</b> Route 92 Medical 088 Access System, 143 cm
Indications for Use	The Route 92 Medical Access System, 110 cm, is indicated for use with compatible guide catheters in facilitating the insertion and guidance of microcatheters into a selected blood vessel in the neurovascular system.	The Route 92 Medical 088 Access System, 143 cm, is indicated for use with compatible guide catheters in facilitating the insertion and guidance of microcatheters into a selected blood vessel in the neurovascular system.
Device Description	Sterile, single-use, variable stiffness, coil-reinforced catheter with proximal control wire	Same
Targeted population	Patients requiring use of a microcatheter in the neurovascular system	Same
User	Physicians trained in neurovascular interventional techniques	Same
Anatomical Sites	Neurovasculature only	Same
Materials	Polymers and metals commonly used in the manufacture of medical devices	Same
<b>Support Catheter</b>		
Inner Diameter	0.088”	Same
Outer Diameter	Distal: 0.101 Proximal: 0.105”	Same
Length	110 cm	143 cm
<b>Delivery Catheter</b>		
Inner Diameter	0.019”	Same
Outer Diameter	Distal: 0.080” Proximal: 0.062”	Same
Length	151 cm	Same

## **Non-Clinical Testing**

### **Biocompatibility Testing**

The Route 92 Medical 088 Access System, 143 cm, is constructed using the same materials as the predicate device. All patient contacting components have been evaluated for biocompatibility in accordance with ISO 10993-1 Biological Evaluation of Medical Devices Part 1: Evaluation and Testing.

The Route 92 Medical Access System is classified per ISO 10993-1 as externally communicating with limited circulating blood contact (<24 hours). A summary of the biocompatibility testing is provided below.

<b>Test</b>	<b>Conclusions</b>
Cytotoxicity – ISO MEM Elution	The test article is non-cytotoxic.
Sensitization – ISO Guinea Pig Maximization Sensitization Test (Normal Saline and Sesame Oil)	The test article did not elicit a sensitization response.
Irritation – ISO Intracutaneous Reactivity (Normal Saline and Sesame Oil)	Requirements of the ISO intracutaneous reactivity test have been met for the test article.
Acute Systemic Toxicity – ISO Acute Systemic Injection (Normal Saline and Sesame Oil)	Requirements of the ISO acute systemic injection test have been met for the test article.
Pyrogen – Material Mediated Pyrogen (Normal Saline)	The test article is non-pyrogenic.
Hemocompatibility – Complement Activation (SC5b-9)	The test article would not be expected to result in adverse effects in vivo.
Hemocompatibility – Partial Thromboplastin Time	The test article is considered to be a non-activator of the intrinsic coagulation pathway.
Hemocompatibility – ASTM Hemolysis	The test article is considered non-hemolytic.
Hemocompatibility – Thromboresistance	The test articles have similar thromboresistance characteristics as the control devices.

### Performance Testing

The successful completion of the performance testing listed in the following table demonstrates that the Route 92 Medical 088 Access System, 143 cm, is suitable for its intended use.

<b>Test</b>	<b>Test Method</b>	<b>Results</b>
Dimensional Verification	Device dimensions were measured to confirm conformance to the specifications	<b>PASS</b> All samples met the pre-determined acceptance criteria
Luer Integrity	Tested per ISO 80369-7:2016	<b>PASS</b> All samples met the pre-determined acceptance criteria
RHV Sealing	RHV sealing around the catheter shafts was tested	<b>PASS</b> All samples met the pre-determined acceptance criteria

<b>Test</b>	<b>Test Method</b>	<b>Results</b>
Tensile Strength	The tensile strength of the catheter sections and bonds was tested	<b>PASS</b> All samples met the pre-determined acceptance criteria
Kink Resistance	Test specimen segments were formed into a defined bend diameter to evaluate kink resistance	<b>PASS</b> All samples met the pre-determined acceptance criteria
Torsion Resistance	The test specimens were rotated to evaluate integrity after rotation	<b>PASS</b> All samples met the pre-determined acceptance criteria
Tip Flexibility	Test specimens were tested for tip flexibility	<b>PASS</b> All samples met the pre-determined acceptance criteria
Air Leakage	Tested per ISO 10555-1:2013 Annex D.	<b>PASS</b> All samples met the pre-determined acceptance criteria
Liquid Leakage / Static Burst	Tested per ISO 10555-1:2013 Annex C.	<b>PASS</b> All samples met the pre-determined acceptance criteria
Dynamic Burst	Mechanical integrity was maintained up to the specified pressures	<b>PASS</b> All samples met the pre-determined acceptance criteria
Hydrophilic Coating Integrity	The integrity of the hydrophilic coating was evaluated after multiple insertion and withdrawal cycles.	<b>PASS</b> All samples met the pre-determined acceptance criteria
Particulate Recovery	After multiple insertion and withdrawal cycles, the effluent water rinsed and flushed from the devices and model was tested per USP <788>. Particulates were characterized for size ranges $\geq 10 \mu\text{m}$ , $\geq 25 \mu\text{m}$ , $\geq 50 \mu\text{m}$ , $\geq 100 \mu\text{m}$ , $\geq 200 \mu\text{m}$ , $\geq 500 \mu\text{m}$ and $\geq 1000 \mu\text{m}$	<b>PASS</b> All samples met the pre-determined acceptance criteria
Simulated Use Testing	Deliverability and compatibility with accessory devices was evaluated in a neurovascular model	<b>PASS</b> All samples met the pre-determined acceptance criteria

Test	Test Method	Results
Radiopacity	Radiopacity of the device was evaluated in an animal model under fluoroscopy	<p align="center"><b>PASS</b></p> <p align="center">All samples met the pre-determined acceptance criteria</p>
Contrast Delivery	Contrast delivery using the device was evaluated in an animal model under fluoroscopy	<p align="center"><b>PASS</b></p> <p align="center">All samples met the pre-determined acceptance criteria</p>
Animal Study	Safety and device performance were both evaluated under clinical conditions using an acute swine model with angiographic and histological assessments. The subject device was compared to a predicate control device to demonstrate substantial equivalence.	<p align="center"><b>PASS</b></p> <p align="center">All samples met the pre-determined acceptance criteria</p>
Packaging Integrity	ISO 11607-1 Part 1 ISO 11607-2 Part 2	<p align="center"><b>PASS</b></p> <p align="center">All samples met the pre-determined acceptance criteria</p>

### **Substantial Equivalence**

The Route 92 Medical 088 Access System, 143 cm, has the same intended use, the same technological characteristics and same method of action as the predicate device. The successful completion of biocompatibility testing and performance testing demonstrates that the Route 92 Medical 088 Access System, 143 cm, is substantially equivalent to the predicate device.