



Nvision Biomedical Technologies, Inc  
Analaura Villarreal Berain  
Extremity Project Engineer  
4590 Lockhill Selma  
San Antonio, Texas 78249

Re: K201850

Trade/Device Name: Vector Hammertoe Correction System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: HWC  
Dated: May 26, 2020  
Received: July 6, 2020

Dear Analaura Berain:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K201850

Device Name  
VECTOR Hammertoe Correction System

Indications for Use (Describe)

The Vector Hammertoe Correction System is indicated for the fixation of osteotomies and reconstruction of the lesser toes following correction procedures for hammertoe, claw toe, and mallet toe.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

---

---

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

---

---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **510(k) Summary**

### **DATE PREPARED**

May 26, 2020

### **MANUFACTURER AND 510(k) OWNER**

Nvision Biomedical Technologies, Inc.

4590 Lockhill Selma

San Antonio, TX 78249, USA

Telephone: (210) 545-3713

Fax: (866) 764-1139

Official Contact: Diana Langham, Director of Regulatory and Corporate Compliance

### **REPRESENTATIVE**

Analaura Villarreal Berain, Extremity Project Engineer

Nvision Biomedical Technologies

Telephone: (210) 545-3713 ext. 109

Email: analauravillarreal@nvisionbiomed.com

### **PROPRIETARY NAME OF SUBJECT DEVICE**

VECTOR™ Hammertoe Correction System

### **COMMON NAME**

Bone Fixation Screw System

### **DEVICE CLASSIFICATION**

Smooth or threaded metallic bone fixation fastener

(Classification Regulations: 21 CFR 888.3040, Product Codes: HWC, Class: II)

### **PREMARKET REVIEW**

Orthopedic Devices

### **INDICATIONS FOR USE**

The Vector Hammertoe Correction System is indicated for the fixation of osteotomies and reconstruction of the lesser toes following correction procedures for hammertoe, claw toe, and mallet toe.

### **DEVICE DESCRIPTION**

The VECTOR Hammertoe Correction System is comprised of a sterile PEEK (polyetheretherketone) HA (hydroxyapatite) fixation device. The implants are offered in Ø3.0mm (cannulated), Ø3.50mm (cannulated and non-cannulated) and Ø4.00mm (cannulated and non-cannulated) and in 0° angle. The system has K-wires, drill, taps, implant inserters, and sizers manufactured from medical grade stainless steel.

**PREDICATE DEVICE IDENTIFICATION**

The subject Vector Hammertoe Correction System is substantially equivalent to the following predicates:

<i>510(k) Number</i>	<i>Predicate Device Name / Manufacturer</i>	<i>Primary Predicate</i>
K183055	VECTOR™ Hammertoe Correction System	✓
K161449	Fusion Orthopedics HammerTech Fixation System	
K133636	Extremity Medical HammerFIX Device	

**SUMMARY OF NON-CLINICAL TESTING**

No FDA performance standards have been established for the Vector Hammertoe Correction System. The following was performed to demonstrate safety per methods of the previous submission:

- Engineering Analysis comparison of mechanical performance characteristics

The results of this engineering analysis indicate that the Vector Hammertoe Correction System is substantially equivalent to the predicate devices.

**EQUIVALENCE TO PREDICATE DEVICES**

Nvision believes that the Vector Hammertoe Correction System modification is substantially equivalent to the predicate devices. The subject implants maintain the same features as the previously cleared devices but add new diameter, length and cannulated options that are within the range of the predicates. This modification does not change the intended use or performance of the device and does not raise additional questions of substantial equivalence. These technological characteristics have undergone engineering analysis to ensure the device is as safe and effective as the predicates.

**CONCLUSION**

Based on the engineering analysis of device characteristics, it can be concluded that the subject device does not raise new issues of safety or efficacy compared to the predicate devices. The similar indications for use, technological characteristics, and performance characteristics for the proposed Vector Hammertoe Correction System options are assessed to be substantially equivalent to the predicate devices.