



March 3, 2021

Cardiovascular Systems Inc.
Erika Huffman
Regulatory Affairs Manager
1225 Old Hwy 8 NW
St. Paul, Minnesota 55112

Re: K210282
Trade/Device Name: WIRION Embolic Protection System
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: NTE
Dated: January 29, 2021
Received: February 1, 2021

Dear Erika Huffman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or

postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Finn Donaldson
Assistant Director
DHT2C: Division of Coronary
and Peripheral Intervention Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K210282

Device Name
WIRION® Embolic Protection System

Indications for Use (Describe)

The WIRION® is indicated for use as an embolic protection system to contain and remove embolic material (thrombus/debris) while performing atherectomy in calcified lesions of the lower extremities. The diameter of the vessel at the site of filter basket placement should be between 3.5mm to 6.0mm. WIRION® may be used with commercially available 0.014" guide wires.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Submitter:	Cardiovascular Systems, Inc. 1225 Old Highway 8 NW Saint Paul, MN 55112
Contact Person:	Erika Huffman, MS, RAC Regulatory Affairs Manager Cardiovascular Systems, Inc. 1225 Old Highway 8 NW St. Paul, MN 55112 Ph: 651-259-1608 ehuffman@csi360.com
Date Prepared:	January 29, 2021
Trade Name:	WIRION® Embolic Protection System
Common Name:	Embolic Protection System
Regulation Number:	870.1250, Percutaneous Catheter
Classification:	Class II
Product Code:	NTE
Predicate Device(s):	K200198 – WIRION™ Embolic Protection System
Device Description:	WIRION® is a temporary Embolic Protection System (EPS), filtering distal to the intervention site. The system is a rapid exchange, pre-crimped filter that can be used with any commercially available 0.014" guide wires.
Intended Use:	The WIRION® is indicated for use as an embolic protection system to contain and remove embolic material (thrombus/debris) while performing atherectomy in calcified lesions of the lower extremities. The diameter of the vessel at the site of filter basket placement should be between 3.5mm to 6.0mm. WIRION® may be used with commercially available 0.014" guide wires.
Comparison to Predicate Device:	The modified WIRION device is identical to the predicate device as follows: <ul style="list-style-type: none"> • Same regulation number, product code and classification • Same intended use and indications for use • Same vessel diameter range and anatomic location of use • Same filter design and materials • Same principles of operation

- Same guidewire rapid exchange (RX) design
- Same 0.014” guidewire compatibility
- Same filter delivery and retrieval methods
- Same filter dimensions and pore size
- Same filter function and placement within the vessel
- Same filter fluoroscopy markers for visualization
- Same sterilization method and SAL
- Same number of uses per device (single use)

The modified WIRION device is different from the predicate device as follows:

- Longer Retrieval Catheter RX guidewire lumen (RX port moved proximal)
- Longer shipping mandrel
- Retrieval Catheter tip material and color
- Larger sterile barrier pouch (Retrieval Catheter)
- Different pouch seal locations
- Branding
- New RX guidewire lumen tie-layer material
- Smaller nominal diameter of activating wire
- New syringe
- New tamper-evident seal on shelf box
- Clarified flushing step in Instructions for Use

Functional and Safety Testing:

Biocompatibility testing was performed to verify the changed materials and processes are acceptable. In addition, simulated use, bond and torque strength testing, and package testing were also performed to confirm pre-determined device specifications were met.

Conclusion:

No new questions of safety or effectiveness were identified compared to the predicate device and the device is expected to perform as intended under the specified use conditions.