



December 15, 2021

Wuhan Zonsen Medical Products Co.,Ltd  
% Ivy Wang  
Consultant  
Shanghai Sungo Management Consulting Company Limited  
14 th Floor, 1500# Central Avenue  
Shanghai, Shanghai 200122  
China

Re: K212861

Trade/Device Name: Surgical Gown, Reinforced Surgical Gown  
Regulation Number: 21 CFR 878.4040  
Regulation Name: Surgical Apparel  
Regulatory Class: Class II  
Product Code: FYA  
Dated: September 8, 2021  
Received: September 8, 2021

Dear Ivy Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray, III, Ph.D.  
Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K212861

Device Name  
Surgical Gown, Reinforced Surgical Gown

### Indications for Use (Describe)

Surgical Gown is intended to be worn by room personnel during surgical procedures or other invasive tests to protect both the surgical patient and operating room personnel from the transfer of microorganisms, body fluids and particulate material. This is single use, disposable device, provided sterile.

Per ANSI/AAMI PB70:2012 Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities, the surgical gown met the requirements for Level 2 classification, and the reinforced surgical gowns met the requirements for Level 3 classification.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **SECTION 5**

### **510(k) Summary**

## 510(K) Summary

K212861

Document prepared date: 2021/12/13

### A. Applicant:

Wuhan Zonsen Medical Products Co.,Ltd

Address: No.8 Jinchao Road Zhucheng Street Xinzhou District Wuhan City Hubei Province, China, 431400

Contact Person: Cynthia Ye

Tel : +86 02782737771

Submission Correspondent: `

Primary contact: Ms. Ivy Wang

Shanghai SUNGO Management Consulting Co., Ltd.

14th Floor, Dongfang Building, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-58817802

Email: [haiyu.wang@sungoglobal.com](mailto:haiyu.wang@sungoglobal.com)

Secondary contact: Mr. Raymond Luo

14th Floor, Dongfang Building, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-68828050

Email: [fda.sungo@gmail.com](mailto:fda.sungo@gmail.com)

### B. Device:

Trade Name: Surgical Gown, Reinforced Surgical Gown

Common Name: Surgical Gown

Model(s): ZSG1005, ZSG1006

#### Regulatory Information

Classification Name: Gown,

Surgical Classification: Class II

Product code: FYA

Regulation Number: 878.4040

Review Panel: Surgical Apparel

### C. Predicate device:

K211422

Level 2 Standard Surgical Gown, Level 3 Standard

Surgical Gown, Level 3 Reinforced Surgical Gown

Jiangsu Medplus Non-woven Manufacturer Co., Ltd.

### D. Intended use of the device:

Surgical Gown is intended to be worn by room personnel during surgical procedures or other invasive tests to protect both the surgical patient and operating room personnel from the transfer of microorganisms, body fluids and particulate material. This is single use, disposable device, provided sterile.

Per ANSI/AAMI PB70:2012 Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities, the surgical gown met the requirements for Level 2 classification, and the reinforced surgical gown met the requirements for Level 3 classification.

**E. Device Description:**

The proposed devices Surgical Gown / Reinforced Surgical Gown have two models: ZSG1005 and ZSG1006.

The proposed device Surgical Gown is model ZSG1005, its body, sleeve and belt are made of SMMS non-woven material, and cuff is made of cotton. The proposed device is available in M, L, XL sizes. This proposed device can meet the requirements for Level 2 per ANSI/AAMI PB70:2012.

The proposed device Reinforced Surgical Gown is model ZSG1006, its body, sleeve and belt are made of SMMS non-woven material, and cuff is made of cotton. The reinforced and critical zone is front chest and sleeves. This zone is reinforced with PP/PE composite breathable film. The proposed device is available in M, L, XL sizes. This proposed device can meet the requirements for Level 3 per ANSI/AAMI PB70:2012.

The proposed devices are disposable medical devices and provided in sterile.

**F. Comparison with predicate device**

**Table 1 General Comparison**

<b>Device</b>	<b>Predicate Device</b>	<b>Proposed Device</b>	<b>Remark</b>
<b>Manufacturer</b>	Jiangsu Medplus Non-woven Manufacturer Co., Ltd.	Wuhan Zonsen Medical Products Co.,Ltd	-
<b>510K number</b>	K211422	K212861	-
<b>Model Name</b>	Level 2 Standard Surgical Gown, Level 3 Standard Surgical Gown, Level 3 Reinforced Surgical Gown	Surgical Gown, Reinforced Surgical Gown,	-
<b>Classification</b>	Class II Device, FYA (21 CFR878.4040)	Class II Device, FYA (21CFR878.4040)	Same
<b>Intend use</b>	Surgical gown is intended to be worn by operating room personnel during surgical procedure to protect both the surgical patient and the operating room personnel from transfer of microorganisms, body fluids, and particulate material.	Surgical Gown is intended to be worn by room personnel during surgical procedures or other invasive tests to protect both the surgical patient and operating room personnel from the transfer of microorganisms, body fluids and particulate material. This is single use, disposable device, provided sterile.	Similar. No issues of safety or efficacy raised.
<b>Style</b>	Non-reinforced/Reinforced	Non-reinforced/Reinforced	Same
<b>Use</b>	Single Use; Disposable; Sterile	Single Use; Disposable; Sterile	Same
<b>Color</b>	Blue	Blue	Same

<b>Labeling</b>	Conform with 21CFR Part 801	Conform with 21CFR Part 801	Same
-----------------	-----------------------------	-----------------------------	------

**Table 2 Safety and Performance Comparison**

<b>Item</b>	<b>Predicate Device</b>	<b>Proposed Device</b>	<b>Remark</b>
<b>Weight per square (g)</b>	Level 2 Standard Surgical Gown: 35g/m <sup>2</sup> ; Level 3 Standard Surgical Gown: 43g/m <sup>2</sup> Level 3 Reinforced Surgical Gown: 35g/m <sup>2</sup> and 28g/m <sup>2</sup>	Surgical Gown: 45g/ m <sup>2</sup> ; Reinforced Surgical Gown: 45g/ m <sup>2</sup>	Difference resolved by performance testing
<b>Size</b>	XS, S, M, L, XL, XXL, XXXL	M, L, XL	Different. No affect on safety or efficacy
<b>Flammability</b>	Class I	Class I	Same
<b>Hydrostatic pressure</b>	Level 2 Standard Surgical Gown: >20 cm; Level 3 Standard Surgical Gown: >50 cm; Level 3 ReinforcedSurgical Gown: >50 cm	Surgical Gown: >20 cm; Reinforced Surgical Gown: >50 cm	Same
<b>Water impact</b>	≤1.0 g	≤1.0 g	Same
<b>Breaking strength</b>	>20N	>20N	Same
<b>Tearing strength</b>	>20N	>20N	Same
<b>Linting</b>	Log <sub>10</sub> (particle count) <4	Log <sub>10</sub> (particle count) <4	Same
<b>Barrier protection level</b>	Level 2 and 3 per AAMI PB 70	Level 2 and 3 per AAMI PB 70	Same
<b>Material</b>	Level 2 Standard Surgical Gown and Level 3 Standard Surgical Gown: SMS nonwoven, Polyester and Polyamide;  Level 3 Reinforced Surgical Gown: SMS nonwoven,	Surgical Gown: SMMS non-woven, Cotton, and Nylon  Reinforced Surgical Gown: SMMS non-woven, Cotton, Nylon, Polypropylene and	Similar

	Polyester, Polyamide and Hydrophilic nonwoven	Polyethylene	
<b>Sterility</b>	Sterile	Sterile	Same
<b>Cytotoxicity</b>	Under the conditions of the study, the device is non-toxic, non-irritating, and non-sensitizing.	Under the conditions of the study, the device is non-toxic, non-irritating, and non-sensitizing.	Same
<b>Irritation</b>			
<b>Sensitization</b>			

### G. Non-Clinical Test Conclusion

Non-clinical tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device met its acceptance criteria or testing endpoint safe levels using the following standards:

- ISO 10993-5: 2009 Biological Evaluation of Medical Devices -- Part 5: Tests For In Vitro Cytotoxicity
- ISO 10993-10: 2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization
- CPSC 16 CFR Part 1610-2008, Standard for the Flammability of clothing textiles;
- AATCC 127-2014, Water Resistance: Hydrostatic Pressure Test;
- AATCC 42-2013, Water Penetration Resistance: Impact Penetration Test;
- ASTM D5034-09, Standard Test Method for Breaking Strength and Elongation of Textile Fabrics (Grab Test);
- ASTM D5587-15, Standard Test Method for Tearing Strength of Fabrics by Trapezoid Procedure;
- AAMI/ANSI PB70:2012, Liquid Barrier Performance and Classification of protective Apparel and Drapes Intended For Use In Health Care Facilities.

**Table 3 Performance Testing**

<b>Name of Testing Methodology</b>	<b>Purpose</b>	<b>Acceptance Criteria</b>	<b>Results</b>
<b>Flammability</b>	The test was performed in accordance with 16 CFR Part 1610 Standard for the Flammability of Clothing Textiles to evaluate the flammability of the test sample.	Meets Class 1 requirements	<b>PASS</b>  Surgical Gown: Class 1  Reinforced Surgical Gown: Class 1
	The test was performed in accordance with AATCC 127: 2017 Water Resistance: Hydrostatic Pressure Test to	Surgical Gown (Level2): >20cm;	<b>PASS</b>  Surgical Gown:



<b>Hydrostatic pressure</b>	determine the hydrostatic pressure of the test sample.	Reinforced Surgical Gown (Level 3): >50 cm	60cm Reinforced Surgical Gown: 95cm
<b>Water impact</b>	The test was performed in accordance with AATCC 42: 2017 Water Resistance: Impact Penetration Test to evaluate the water impact of the test sample.	≤1.0 g	<b>PASS</b> Surgical Gown: 0.1g Reinforced Surgical Gown:0g
<b>Breaking strength</b>	The test was performed In accordance with ASTM D5034: 2009(2017) Standard. Test Method for Breaking Strength and Elongation of Textile Fabrics (Grab Test) to evaluate the breaking strength of the test sample.	>20N	<b>PASS</b> Surgical Gown: 124N Reinforced Surgical Gown: 123N
<b>Tearing strength</b>	The test was performed in accordance with ASTM D5587: 2015(2019) Standard Test Method for Tearing Strength of Fabrics by Trapezoid Procedure to evaluate the tearing strength of the test sample.	>20N	<b>PASS</b> Surgical Gown: 52N Reinforced Surgical Gown:61N
<b>Linting</b>	The test was performed in accordance with ISO 9073-10: 2003 Textiles-Test Methods for Nonwovens-Part 10: Lint and Other Particles Generation in the Dry State to evaluate the linting of the test sample.	Log <sub>10</sub> (particle count) < 4	<b>PASS</b> Surgical Gown: 2.11 Reinforced Surgical Gown: 1.88

**Table 4 Biocompatibility Testing**

Item	Purpose	Acceptance Criteria	Result
------	---------	---------------------	--------

<b>Cytotoxicity</b>	The purpose of the biocompatibility testing is to demonstrate the biocompatibility of the subject device.	Non-Cytotoxic	<b>PASS</b> Under the conditions of the study, the device is non-cytotoxic.
<b>Irritation</b>		Non-Irritating	<b>PASS</b> Under the conditions of the study, the device is non-irritating.
<b>Sensitization</b>		Non-Sensitizing	<b>PASS</b> Under the conditions of the study, the device is non-sensitizing

#### **H. Clinical Test Conclusion**

No clinical study is included in this submission.

#### **I. Conclusion**

Based on the comparison and analysis above, and the non-clinical tests performed, the proposed devices are determined to be as safe, as effective, and performs as well as the legally marketed predicate device under K211422.