



October 21, 2021

Bryan Medical, Inc.
% Martha Russell
Director of Regulatory Affairs
AlvaMed, Inc.
1116 Great Plain Avenue, Suite 1
Needham, Massachusetts 02492

Re: K213028

Trade/Device Name: Tenax Laser Resistant Endotracheal Tube
Regulation Number: 21 CFR 868.5730
Regulation Name: Tracheal tube
Regulatory Class: Class II
Product Code: BTR, BSK
Dated: September 17, 2021
Received: September 21, 2021

Dear Martha Russell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531 - 542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Todd Courtney
Assistant Director
DHT1C: Division of Sleep Disordered
Breathing, Respiratory and
Anesthesia Devices
OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT and Dental Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K213028

Device Name

Tenax® Laser Resistant Endotracheal Tube

Indications for Use (Describe)

The Tenax® Laser Resistant Endotracheal Tube is intended for endotracheal intubation. It is indicated for use for all types of surgical procedures involving carbon dioxide (10.60 microns), KTP (532 nm) or Blue Laser (445nm) laser use (normal pulsed or continuous beam delivery in the non-contact mode), when endotracheal intubation is required to administer anesthetic gases or to overcome emergency obstruction of an airway.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.


This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

	Special 510(k): Modification to Tenax® Laser Resistant Endotracheal Tube	Page 1
---	--	--------

1.0 SPECIAL 510(K) SUMMARY FOR TENAX® LASER RESISTANT ENDOTRACHEAL TUBE

This Special 510(k) for the Tenax® Laser Resistant Endotracheal Tube is submitted based on the FDA Guidance document “The Special 510(k) Program: Guidance for Industry and Food and Drug Administration Staff” issued September 13, 2019. This Special 510(k) addresses labeling revisions to the Tenax® Instructions for Use to add the use of a 445 nm Blue Laser, and that the device is not made with natural rubber latex. In addition, the Indications for Use statement was modified to add the 445nm Blue Laser.

1.1. Submitter

Bryan Medical, Inc.
5725 Dragon Way, Suite 300
Cincinnati, OH 45227

Phone: (617) 517-4932

Fax: (617) 249-0955

Contact Person: Martha Kamrow-Russell, AlvaMed, Inc.

Date Prepared: September 17, 2021

1.2. Device

Name of Device: Tenax® Laser Resistant Endotracheal Tube

Common or Usual Name: Endotracheal tube

Classification Name: Tracheal tube (21 CFR 868.5730)

Regulatory Class: II

Product Code: BTR, BSK

1.3. Predicate Device

Tenax® Laser Resistant Endotracheal Tube, K200761


1.4. Device Description

The Tenax® Laser Resistant Endotracheal Tube is an endotracheal tube with two inflatable cuffs. It is indicated for use in surgical procedures where intubation is required in the presence of KTP, 445nm Blue Laser, or CO2 lasers.


The device is provided sterile (EO). After use, the device is discarded and disposed of in accordance with local regulations. There are no associated device accessories.

1.5. Indication For Use

The Tenax® Laser Resistant Endotracheal Tube is intended for endotracheal intubation. It is indicated for use for all types of surgical procedures involving carbon dioxide (10.60 microns), KTP (532 nm), or Blue (445 nm) laser use (normal pulsed or continuous beam delivery in the non-contact mode), when endotracheal intubation is required to administer anesthetic gases or to overcome emergency obstruction of an airway.


	Special 510(k): Modification to Tenax® Laser Resistant Endotracheal Tube	Page 2
---	--	---------------

Intended Population:	Adults (≥22 years old) requiring endotracheal intubation during laser surgery procedures in the upper gastro-respiratory tract.
Intended Environment of Use:	Trachea/Upper GI Tract
Contraindications:	The Tenax® Laser Resistant Endotracheal Tube should not be used in patients with narrow airways, which could restrict ventilation, resulting in excessive elevation of intratracheal pressure. Do not use in any patient suffering from conditions that may preclude endotracheal intubation. Do not use this device for any purpose other than its intended use.

	Special 510(k): Modification to Tenax® Laser Resistant Endotracheal Tube	Page 3
---	--	---------------

1.6. Comparison of Technological Characteristics with the Predicate Device

	Subject Device: Tenax® Laser Resistant Endotracheal Tube	Predicate Device: Tenax® Laser Resistant Endotracheal Tube
Manufacturer	Bryan Medical, Inc.	Bryan Medical, Inc.
510(k) Number	To be assigned.	K200761
Product Code	BTR, BSK	BTR, BSK
Regulation Number	868.5730	868.5730
Regulation Description	Tracheal tube	Tracheal tube
Common Name	Endotracheal tube	Endotracheal tube
Indications for Use	The Tenax® Laser Resistant Endotracheal Tube is intended for endotracheal intubation. It is indicated for use for all types of surgical procedures involving carbon dioxide (10.60 microns), Blue (445 nm), or KTP (532 nm) laser use (normal pulsed or continuous beam delivery in the non-contact mode), when endotracheal intubation is required to administer anesthetic gases or to overcome emergency obstruction of an airway.	The Tenax® Laser Resistant Endotracheal Tube is intended for endotracheal intubation. It is indicated for use for all types of surgical procedures involving carbon dioxide (10.60 microns) or KTP (532 nm) laser use (normal pulsed or continuous beam delivery in the non-contact mode), when endotracheal intubation is required to administer anesthetic gases or to overcome emergency obstruction of an airway.
Sterility	Sterile (ethylene oxide)	Sterile (ethylene oxide)
Single Use	Yes	Yes

	Special 510(k): Modification to Tenax® Laser Resistant Endotracheal Tube	Page 4
---	--	--------

Both devices are laser resistant endotracheal tubes. They share an identical indication for use (outside of the addition of the Blue 445 nm laser for the subject device) and are intended for use in the same manner. The subject and predicate devices also share the following technological characteristics:

- Laser resistance using an inner aluminum scattering layer,
- Atraumatic outer layer for protection of patient tissue,
- Inflatable cuffs,
- Use of internal dye as a warning indicator to alert practitioners of device puncture or damage.

1.7. Performance Data

The following performance data were provided in support of the substantial equivalence determination:

- Laser resistance

1.8. Conclusion

The non-clinical data support the safety of the device and the verification and validation testing demonstrate that the Tenax® Laser Resistant Endotracheal Tube shall perform as intended in the specified use conditions. The data demonstrate that the Tenax® Laser Resistant Endotracheal Tube performs comparably to the predicate device currently marketed for the same intended use.