



December 14, 2021

Weihai Hongyu Nonwoven Fabric Products Co., Ltd.
Ray Wang
General Manager
Beijing Believe-Med Technology Service Co., Ltd.
Rm.912, Building #15, XiYueHui, No.5, YiHe North Rd.,
FangShan District
Beijing, Beijing 102401
China

Re: K213029

Trade/Device Name: Nitrile Examination Gloves (Powder free, Blue)
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulatory Class: Class I, reserved
Product Code: LZA
Dated: September 21, 2021
Received: September 21, 2021

Dear Ray Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray III, PhD
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K213029

Device Name
Nitrile Examination Gloves (Powder free, Blue)

Indications for Use (Describe)

The Nitrile Examination Gloves (Powder free, Blue) is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) Number: K213029

1. Date of Preparation: 11/23/2021
2. Sponsor

Weihai Hongyu Nonwoven Fabric Products Co., Ltd.

No.567, Gushan Road, Area of Economy and Technique,264207, Weihai City, Shandong Province, P.R. China.

Contact Person: Dong Bin

Position: General Manager

Tel: +86-15662383633

Fax: +86-631-3636942

Email: dbfeel@163.com

3. Submission Correspondent

Beijing Believe-Med Technology Service Co., Ltd.

Rm.912, Building #15, XiYueHui, No.5, YiHe North Rd., FangShan District, Beijing, China,102401

Contact Person: Ray Wang

Position: General Manager

Tel: +86-18910677558

Fax: +86-10-56335780

Email: information@believe-med.com

4. Proposed Device Identification

Trade Name: Nitrile Examination Gloves (Powder free, Blue)

Common Name: NITRILE Patient Examination Gloves (Powder Free)

Regulatory Information:

Classification: I

Product Code: LZA

Regulation Number: 21 CFR 880.6250

Review Panel: General Hospital

Indication For Use Statement:

The Nitrile Examination Gloves (Powder free, Blue) is a disposable device intended for medical purposes that is worn on the examiner’s hands to prevent contamination between patient and examiner.

5. Predicate Device Identification

510(k) Number: K150340

Product Name: POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue)

Manufacturer: HEBEI HONGSEN PLASTICS TECHNOLOGY CO., LTD

6. Device Description

The proposed device, Nitrile Examination Gloves (Powder free, Blue) are disposable devices intended for medical purposes that is worn on the examiner’s hands to prevent contamination between patient and examiner.

The proposed devices are Powder Free Nitrile Examination Gloves and includes variations of different size. The color of the proposed device is Blue.

The proposed device is provided non-sterile.

The proposed device is made of Nitrile.

Table 1 Device Size Specifications

Designation	Size					Tolerance
	XS	S	M	L	XL	
Length, mm	220	220	230	235	235	min
Width, mm	70	80	95	110	120	±10
Thickness, mm:						
Finger	0.05				min	
Palm	0.05				min	
Cuff	0.05				min	

Table 2 Performance and Physical Specifications

Before Aging		After Aging		Pinhole AQL
Tensile Strength	Ultimate Elongation	Tensile Strength	Ultimate Elongation	
14 MPa, min	500 % min	14 MPa, min	400 % min	

The above data of size, performance, and physical specifications of proposed gloves meet all the current specifications listed in the ASTM standard D6319.

7. Comparison of Technological Characteristics

Table 1 General Comparison

ITEM	Proposed Device(K213029) Nitrile Examination Gloves (Powder free, Blue)	Predicate Device (K150340) POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue)	Remark
Product Code	LZA	LZA	SAME
Regulation No.	21 CFR 880.6250	21 CFR 880.6250	SAME
Class	I	I	SAME
Intended Use	The Nitrile Examination Gloves (Powder free, Blue) is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.	The POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue) is a disposable device intended for medical purposes that is worn on the examiner's hands to prevent contamination between patient and examiner.	SAME
Powdered or Powdered free	Powdered free	Powdered free	SAME

Table 2 Device Dimensions Comparison

Proposed Device(K213029) Nitrile Examination Gloves (Powder free, Blue)	Designation	Size					Tolerance
		XS	S	M	L	XL	
	Length, mm	220	220	230	235	235	min
	Width, mm	70	80	95	110	120	±10
Thickness, mm:							
	Finger	0.05					min
	Palm	0.05					min
	Cuff	0.05					min
Predicate Device (K150340) POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue)	Designation	Size					Tolerance
		XS	S	M	L	XL	
	Length, mm	230	230	230	230	230	min
	Width, mm	70	80	95	110	120	±10
Thickness, mm:							
	Finger	0.10-0.12					±0.03
	Palm	0.08-0.10					±0.03
	Cuff	0.06-0.09					±0.03
Remark		Similar					

Different Analysis:

The proposed device has different size specification to the predicate device, but all proposed devices are meet the specifications of ASTM D 6319.

So we consider this as the proposed device is similar to the predicate device.

Table 3 Performance Comparison

ITEM			Proposed Device(K213029) Nitrile Examination Gloves (Powder free, Blue)	Predicate Device (K150340) POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue)	Remark
Colorant			Blue	White, Cobalt Blue, Black, Ice Blue	Different
Single Use			Yes	Yes	SAME
Physical Properties	Before	Tensile Strength	14 MPa, min	15 MPa, min	Different
	Aging	Ultimate Elongation	500 % min	500 % min	SAME
	After Aging	Tensile Strength	14 MPa, min	14 MPa, min	SAME

	Ultimate Elongation	400 % min	400 % min	SAME
	Comply with ASTM D6319		Comply with ASTM D6319	SAME
Freedom from Holes	Be free from holes when tested in accordance with ASTM D5151, AQL 2.5		Be free from holes when tested in accordance with ASTM D5151, AQL 1.5	Different
Powder Content	Less than 2 mg per glove when tested in accordance with ASTM D6124		Meet the requirements of ASTM 6124	SAME

Different Analysis:

The proposed device has different color to the predicate device, this different may causes potential biocompatibility risk, for this risk we conducted the biocompatibility test according to the ISO 10993-5, and ISO 10993-10 the test results showed that the proposed devices with blue colorant did not induce skin irritation and showed no significant evidence of causing skin sensitization and no cytotoxic potential.

Different Analysis:

The proposed device has different Tensile Strength before aging specification to the predicate device, but all proposed device meets the specification requirements of ASTM D 6319.

Different Analysis:

The proposed device has different AQL to the predicate device, but all proposed device meets the specification requirements of ASTM D 6319.

Table 4 Safety Comparison

ITEM		Proposed Device (K213029) Nitrile Examination Gloves (Powder free, Blue)	Predicate Device (K150340) POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue)	Remark
Material		Nitrile	Nitrile	SAME
Biocompatibility	Irritation	Under the conditions of the study, not an irritant	Under the conditions of the study, not an irritant	SAME
	Sensitization	Under conditions of the study, not a sensitizer.	Under conditions of the study, not a sensitizer.	
	Cytotoxic	Under the conditions of the study, there was no evidence of cytotoxic potential from the extract.	Not Available	Different
Label and Labeling		Meet FDA's Requirements	Meet FDA's Requirements	SAME

Different Analysis:

The proposed device has conducted the acute cytotoxic testing, and the test results shown that there was

no evidence of cytotoxic potential.

8. Summary of Non-Clinical Testing

Bench tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device complies with the following standards:

ASTM D6319-19, Standard Specification for Nitrile Examination Gloves for Medical Application.

ASTM D5151-19, Standard Test Method for Detection of Holes in Medical Gloves.

ASTM D6124-17, Standard Test Method for Residual Powder on Medical Gloves.

ISO 10993-10:2010, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization.

ISO 10993-5:2009, Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity.

Table 5 Performance Test Results Summary

Test Method	Purpose	Acceptance Criteria	Results
ASTM D5151	Testing for Freedom from holes	Freedom from holes AQL 2.5	No water leakage is inspected form 200 samples
ASTM D6124	Determine the powder residue for powder free gloves	<2.0 mg per glove	Residual Powder of Size XS: Average 0.28 mg; Residual Powder of Size S: Average 0.31 mg; Residual Powder of Size M: Average 0.34 mg; Residual Powder of Size L: Average 0.31 mg; Residual Powder of Size XL: Average 0.40mg;
ASTM D412 ASTM D573	Testing for Physical property characteristics	Before Aging Tensile Strength: 14 MPa min. Ultimate Elongation: 500% min. After Aging Tensile Strength: 14 MPa min. Ultimate Elongation: 400% min.	Before Aging Tensile Strength: ≥ 19MPa; Ultimate Elongation: ≥ 500%. After Aging Tensile Strength: ≥ 18 MPa; Ultimate Elongation: ≥ 473%.
ASTM D412 ASTM D3767	Testing For physical dimensions specification	Length: 220 mm min. for size (XS, S); Length: 230 mm min. for size (M); Length: 235 mm min. for size (L, XL) Width: 70±10 mm for XS; 80±10 mm for S; 95±10 mm for M; 110±10 mm for L; 120±10 mm for XL. Cuff Thickness: ≥0.05 mm; Finger Thickness: ≥0.05 mm; Palm Thickness: ≥0.05 mm; All acceptance criteria above meet the requirements in Table 1 Dimensions and Tolerances of ASTM D6319	Length of Size XS: ≥ 223 mm; Width of Size XS: 75±2 (73-77) mm; Cuff Thickness of Size XS: ≥0.06 mm; Palm Thickness of Size XS: ≥0.07 mm; Finger Thickness of Size XS: ≥0.10 mm. Length of Size S: ≥ 224mm; Width of Size S: 85±2 (85-87) mm; Cuff Thickness of Size S: ≥0.06 mm; Palm Thickness of Size S: ≥0.07 mm; Finger Thickness of Size S: ≥0.10 mm. Length of Size M: ≥ 233 mm;

			<p>Width of Size M: 95±3 (95-98) mm; Cuff Thickness of Size M: ≥0.06 mm; Palm Thickness of Size M: ≥0.07 mm; Finger Thickness of Size M: ≥0.10 mm.</p> <p>Length of Size L: ≥ 236mm; Width of Size L: 115±4 (115-119) mm; Cuff Thickness of Size L: ≥0.06 mm; Palm Thickness of Size L: ≥0.07 mm; Finger Thickness of Size L: ≥0.10 mm.</p> <p>Length of Size XL: ≥ 236 mm; Width of Size XL: 125±2 (115-117) mm; Cuff Thickness of Size XL: ≥0.06 mm; Palm Thickness of Size XL: ≥0.07 mm; Finger Thickness of Size XL: ≥0.10 mm.</p>
ISO 10993-5	Evaluate the endpoint of Cytotoxicity for biocompatibility	The test article showed no evidence of cytotoxic potential from the extract.	The test article showed no evidence of systemic toxicity from the extract.
ISO 10993-10	Evaluate the endpoint of irritant for biocompatibility	The test article showed no irritation on the skin.	The test article showed no irritation on the skin.
	Evaluate the endpoint of sensitization for biocompatibility	The test article showed no evidence of causing delayed dermal contact sensitization.	The test article showed no evidence of causing delayed dermal contact sensitization in the guinea pig.

9. Summary of Clinical Testing

Clinical Testing is not applicable.

10. Conclusion

Based on the nonclinical tests performed, the subject device is as safe, as effective, and performs as well as the legally marketed predicate device, POWDER FREE Nitrile GLOVES (White, Cobalt Blue, Black, Ice Blue) cleared under K150340.