



January 6, 2022

DePuy Orthopaedics Inc.  
% Paul Shin  
Project Lead, Regulatory Affairs  
DePuy (Ireland)  
Loughbeg, Ringaskiddy  
Cork, Cork P43ED82  
Ireland

Re: K213839

Trade/Device Name: DePuy Corail AMT Hip Prosthesis

Regulation Number: 21 CFR 888.3353

Regulation Name: Hip Joint Metal/Ceramic/Polymer Semi-Constrained Cemented Or Nonporous  
Uncemented Prosthesis

Regulatory Class: Class II

Product Code: LZO, MEH, KWL, KWY

Dated: December 8, 2021

Received: December 9, 2021

Dear Paul Shin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Limin Sun, Ph.D.  
Acting Assistant Director  
DHT6A: Division of Joint  
Arthroplasty Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: 06/30/2023  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)

K213839

Device Name

DePuy Corail AMT Hip Prosthesis

Indications for Use (Describe)

Total hip replacement or hip arthroplasty is indicated in the following conditions:

1. A severely painful and/or disabled joint from osteoarthritis, traumatic arthritis, rheumatoid arthritis, or congenital hip dysplasia.
2. Avascular necrosis of the femoral head.
3. Acute traumatic fracture of the femoral head or neck.
4. Failed previous hip surgery including joint reconstruction, internal fixation, arthrodesis, hemi-arthroplasty, surface replacement arthroplasty, or total hip replacement.
5. Certain cases of ankyloses.

Partial hip replacement or hip hemi-arthroplasty is indicated in the following conditions:

1. Acute fracture of the femoral head or neck that cannot be appropriately reduced and treated with internal fixation.
2. Fracture dislocation of the hip that cannot be appropriately reduced and treated with internal fixation.
3. Avascular necrosis of the femoral head.
4. Non-union of femoral neck fractures.
5. Certain high subcapital and femoral neck fractures in the elderly.
6. Degenerative arthritis involving only the femoral head in which the acetabulum does not require replacement.
7. Pathology involving only the femoral head/neck and/or proximal femur that can be adequately treated by hip hemiarthroplasty.

HA coated stems of the Corail Hip system are indicated for cementless use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

**510(K) SUMMARY**

(As required by 21 CFR 807.92)

<b>Submitter Information</b>	
Name	DePuy Orthopaedics, Inc.
Address	700 Orthopaedic Drive Warsaw, IN 46852
Phone number	(574) 404-8348
Fax number	N/A
Establishment Registration Number	1818910
Name of contact person	Paul Shin
Date prepared	December 7, 2021
<b>Name of device</b>	
<b>Trade or proprietary name</b>	DePuy Corail AMT Hip Prosthesis
<b>Common or usual name</b>	Uncemented hip prosthesis
<b>Classification name</b>	Hip joint metal/ceramic/polymer semi-constrained cemented or nonporous uncemented prosthesis Hip joint femoral (hemi-hip) metallic cemented or uncemented prosthesis Hip joint femoral (hemi-hip) metal/polymer cemented or uncemented prosthesis
<b>Class</b>	II
<b>Classification panel</b>	87 Orthopedics
<b>Regulation</b>	21 CFR 888.3353, 888.3360, 888.3390
<b>Product Code(s)</b>	LZO, MEH, KWL, KWY
<b>Legally marketed device(s) to which equivalence is claimed</b>	DePuy Corail AMT Hip Prosthesis (K042992, K093736, K190344, K192946, K203167).
<b>Reason for 510(k) submission</b>	The purpose of this submission is to implement the following: <ul style="list-style-type: none"> <li>• Addition of DePuy Ireland as a hydroxyapatite (HA) coating site.</li> <li>• Increase in the shelf life of <b>specific product codes</b> of Corail AMT Hip stems from 5 years to 10 years.</li> <li>• Addition of an alternate manufacturing process flow.</li> </ul>
<b>Device description</b>	The DePuy Corail AMT hip stem family are manufactured from forged titanium alloy (Ti6Al4V) and plasma-sprayed with a hydroxyapatite (HA) coating for

	bone fixation. The stem consists of a wide range of stem neck designs and sizes allowing an accurate anatomical match for each patient. Corail AMT stems are available with or without a collar, with various neck angles, and with various neck offsets.
<b>Intended use of the device</b>	Total hip arthroplasty and hemi-hip arthroplasty
<b>Indications for use</b>	<p><b>Total hip replacement or hip arthroplasty is indicated in the following conditions:</b></p> <ol style="list-style-type: none"> <li>1. A severely painful and/or disabled joint from osteoarthritis, traumatic arthritis, rheumatoid arthritis, or congenital hip dysplasia.</li> <li>2. Avascular necrosis of the femoral head.</li> <li>3. Acute traumatic fracture of the femoral head or neck.</li> <li>4. Failed previous hip surgery including joint reconstruction, internal fixation, arthrodesis, hemi-arthroplasty, surface replacement arthroplasty, or total hip replacement.</li> <li>5. Certain cases of ankylosis.</li> </ol> <p><b>Partial hip replacement or hip hemi-arthroplasty is indicated in the following conditions:</b></p> <ol style="list-style-type: none"> <li>1. Acute fracture of the femoral head or neck that cannot be appropriately reduced and treated with internal fixation.</li> <li>2. Fracture dislocation of the hip that cannot be appropriately reduced and treated with internal fixation.</li> <li>3. Avascular necrosis of the femoral head.</li> <li>4. Non-union of femoral neck fractures.</li> <li>5. Certain high subcapital and femoral neck fractures in the elderly.</li> <li>6. Degenerative arthritis involving only the femoral head in which the acetabulum does not require replacement.</li> <li>7. Pathology involving only the femoral head/neck and/or proximal femur that can be adequately treated by hip hemi-arthroplasty.</li> </ol> <p>HA coated stems of the Corail Hip system are indicated for cementless use only.</p>

<b>SUMMARY OF THE TECHNOLOGICAL CHARACTERISTICS OF THE DEVICE COMPARED TO THE PREDICATE DEVICE</b>				
<b>Characteristics</b>	<b>Subject Device</b>  DePuy Corail AMT Hip Prosthesis  DePuy Corail Revision Hip Prosthesis	<b>Primary Predicate</b>  DePuy Corail AMT Hip Prosthesis (K042992)	<b>Additional Predicates</b>  DePuy Corail AMT Hip Prosthesis (K203167, K190344)	<b>Additional Predicates</b>  DePuy Corail Revision Hip Prosthesis (K203167, K192946, K093736)
<b>Intended Use</b>	Total Hip Arthroplasty	Total Hip Arthroplasty	Total Hip Arthroplasty	Total Hip Arthroplasty
<b>Material</b>	Titanium Alloy (Ti6Al4V) with plasma sprayed HA coating	Titanium Alloy (Ti6Al4V) with plasma sprayed HA coating	Titanium Alloy (Ti6Al4V) with plasma sprayed HA coating	Titanium Alloy (Ti6Al4V) with plasma sprayed HA coating
<b>Fixation</b>	Uncemented	Uncemented	Uncemented	Uncemented
<b>Sterilization Method</b>	Gamma	Same	Same	Same
<b>Packaging</b>	Nylon Inner Pouch and outer PETG blister with Tyvek peel lid	Same	Same	Same
<b>Shelf life</b>	10 Years	5 Years	10 Years	10 Years
<b>PERFORMANCE DATA</b>				
<b>SUMMARY OF NON-CLINICAL TESTS CONDUCTED FOR DETERMINATION OF SUBSTANTIAL EQUIVALENCE</b>				

The following tests were performed on the subject devices to demonstrate substantial equivalence of safety and effectiveness with the predicate devices:

- HA coating characterization supporting addition of alternate HA coating site
  - Completed the full suite of HA coating characterization and testing per FDA Guidance 510(k) *Information Needed for Hydroxyapatite Coated Orthopedic Implants* 1997
- Conformational design verification testing supporting addition of alternate HA coating site
  - Body fatigue testing per ISO 7206-4
  - Neck fatigue testing per ISO 7206-6
- Packaging stability testing supporting increase in shelf life
  - Visual inspection per ASTM F1886-16: Standard test method for determining integrity of seals for flexible packaging by visual inspection
  - Dye Leak per ASTM F1929-15: Standard test method for detecting seal leaks in porous medical packaging by dye penetration
  - Seal strength per ASTM F88/F88M-15: Standard test method for seal strength of flexible barrier materials
- HA coating testing on shelf-aged product per
  - ISO-13779-2: Implants for surgery – Hydroxyapatite Part 2: Thermally Sprayed coatings of Hydroxyapatite
  - ISO-13779-3: Implants for surgery – Hydroxyapatite Part 3: Chemical analysis and characterization of crystallinity and phase purity.
  - ASTM F1854 – Standard test method for stereological evaluation of porous coatings on medical implants.
  - ASTM E2109 – Standard test methods for determining area percentage porosity in thermal sprayed coatings.

**SUMMARY OF CLINICAL TESTS CONDUCTED FOR DETERMINATION OF SUBSTANTIAL EQUIVALENCE AND/OR OF CLINICAL INFORMATION**

No clinical tests were conducted to demonstrate substantial equivalence.

**CONCLUSIONS DRAWN FROM NON-CLINICAL AND CLINICAL DATA**

The subject DePuy Corail AMT Hip Prosthesis devices are substantially equivalent to the predicate devices (K042992, K093736, K190344, K192946, K203167).