



May 25, 2022

Medtronic Sofamor Danek USA, Inc.
Elizabeth Hamilton
Regulatory Affairs Specialist
1800 Pyramid Place
Memphis, Tennessee 38132

Re: K221180

Trade/Device Name: Anteralign™ Spinal System with Titan NanoLOCK™ Surface Technology
Navigated Instruments

Regulation Number: 21 CFR 882.4560

Regulation Name: Stereotaxic Instrument

Regulatory Class: Class II

Product Code: OLO

Dated: April 21, 2022

Received: April 25, 2022

Dear Elizabeth Hamilton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K221180

Device Name

Anteralign™ LS Spinal System with Titan nanoLOCK™ Surface Technology Navigated Instruments

Indications for Use (Describe)

Medtronic Navigated Surgical Instruments are intended to be used during preparation and placement of Medtronic implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. Medtronic Navigated Reusable Instruments are compatible with various Medtronic spinal implant systems.

Navigated surgical instruments are specifically designed for use with the StealthStation™ System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, long bone, or vertebra can be identified relative to a CT or MR-based model, fluoroscopy images, or digitized landmarks of the anatomy. An instrument may incorporate a measuring function, which has uses described on the label and the instruments.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) SUMMARY
MEDTRONIC Sofamor Danek
20 April 2022

| | |
|----------------------------|--|
| Submitter: | Medtronic Sofamor Danek, USA Inc. 1800 Pyramid Place Memphis, Tennessee 38132 Telephone: (901)399-3395 Fax: (901) 346-9738 |
| Contact Person | Elizabeth Hamilton Sr. Regulatory Affairs Specialist Email : elizabeth.c.hamilton@medtronic.com |
| Alternate Contact: | Raphael McInnis Sr. Regulatory Affairs Manager Email : raphael.mcinniss@medtronic.com |
| Date Prepared: | 20 April 2022 |
| Name of Device | Anteralign™ LS Spinal System with Titan nanoLOCK™ Surface Technology Navigated Instruments |
| Common Name | Navigated Instruments |
| Classification Name | Stereotaxic Instruments (21 CFR 882.4560) |
| Regulatory Class | Class II |
| Product Code | OLO |
| Predicate Devices | Primary Predicate -Anteralign™ TL Spinal System with Titan nanoLOCK™ Surface Technology Navigated Instruments (K212524, S.E. 12/08/2021) |

Description of Device

Medtronic Navigated Reusable Instruments are spine preparation instruments made of high-grade stainless steel. These instruments are specifically designed for use in procedures where the use of stereotactic surgery may be appropriate. Placing Medtronic single-use sterile spheres on each of the tracker passive stems allows a Medtronic computer-assisted surgery system such as the StealthStation™ Image Guidance System to track the instruments in the surgical field. An instrument may incorporate a measuring function which has uses as described on the label and the instrument. Medtronic Navigated Reusable Instruments are compatible with various Medtronic spinal implant systems.

Indications for Use

Medtronic Navigated Manual Reusable Instruments for Use with StealthStation™ System

Medtronic Navigated Surgical Instruments are intended to be used during preparation and placement of Medtronic implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. Medtronic Navigated Reusable Instruments are compatible with various Medtronic spinal implant systems.

Navigated surgical instruments are specifically designed for use with the StealthStation™ System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, long bone, or vertebra can be identified relative to a CT or MR-based model, fluoroscopy images, or digitized landmarks of the anatomy. An instrument may incorporate a measuring function, which has uses described on the label and the instruments.

Comparison of Technological Characteristics with Predicate Devices

The subject Anteralign™ LS Spinal System with Titan nanoLOCK™ Surface Technology Navigated Instruments have an identical intended use, fundamental scientific technology, sterilization method, materials, and compatibilities with StealthStation™ System as the recently cleared predicate devices listed above. Please refer to the substantial equivalence section of this submission for more details.

Performance Data

Testing was rationalized or completed new to ensure the functionality and compatibility of the subject instruments when used with StealthStation™ System. The following table summarizes the performance testing that was rationalized or completed:

| Test | Description |
|--|--|
| Navigation Accuracy Analysis | Confirmed navigated instrument accuracy in both 2D and 3D space |
| Anatomical Simulated Use | Confirmed instrument functionality under expected use conditions |
| Navigated Simulated Use | Confirmed navigation system functionality under expected use conditions |
| CAD Model Evaluation | Verified that the CAD models are accurately reflected in the application software |
| Implant/ Instrument Mating Conditions | Verified that the instruments can be assembled with the appropriate devices according to their intended use |
| Spine Tools Package Functional Testing | Verified that the Spine Tools package has met the required interface needs of the spine application software |

Conclusion

Based on the information contained in this submission, Medtronic believes that the subject instruments are substantially equivalent to the predicate device.