



Zimmer, Inc.
Caleb Barylski
Specialist, Regulatory Affairs
1800 W. Center Street
Warsaw, Indiana 46580

February 28, 2020

Re: K192660

Trade/Device Name: Zimmer M/L Taper Hip Prosthesis

Regulation Number: 21 CFR 888.3353

Regulation Name: Hip Joint Metal/Ceramic/Polymer Semi-Constrained Cemented Or Nonporous
Uncemented Prosthesis

Regulatory Class: Class II

Product Code: LZ0, MEH

Dated: January 29, 2020

Received: January 30, 2020

Dear Caleb Barylski:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Vesa Vuniqi, MS
Assistant Director
DHT6A: Division of Joint Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K192660

Device Name

Zimmer M/L Taper Hip Prosthesis

Indications for Use (Describe)

Total hip replacement for the following: severe hip pain and disability due to rheumatoid arthritis, osteoarthritis, traumatic arthritis, polyarthritis, collagen disorders, avascular necrosis of the femoral head, nonunion of previous fractures of the femur; congenital hip dysplasia, protrusio acetabuli, slipped capital femoral epiphysis; disability due to previous fusion; previously failed endoprostheses, and/or total hip components in the affected extremity and acute femoral neck fractures.

Hemi-hip replacement for the following: fracture dislocation of the hip; elderly, debilitated patients when a total hip replacement is contraindicated; irreducible fractures in which adequate fixation cannot be obtained; certain high subcapital fractures and comminuted femoral neck fractures in the aged; nonunion of femoral neck fractures; secondary avascular necrosis of the femoral head; pathological fractures of the femoral neck; and osteoarthritis in which the femoral head is primarily affected.

The femoral stem is for cementless use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

In accordance with 21 CFR §807.92 and the Safe Medical Devices Act of 1990, the following information is provided for the VerSys Cemented Revision/Calcar 510(k) premarket notification. The submission was prepared in accordance with the FDA guidance document, 'Format for Traditional and Abbreviated 510(k)s,' issued on September 13, 2019.

Sponsor: Zimmer, Inc.
P.O. Box 708
Warsaw, IN 46581-0708
Establishment Registration Number: 1822565

Contact Person: Caleb Barylski
Regulatory Affairs Specialist
Telephone: (574) 371-0250

Date: 27 February 2020

Subject Device: **Trade Name:** Zimmer M/L Taper Hip Prosthesis
Common Name: Hip Prosthesis

Classification Name:

- LZO – Prosthesis, Hip, Semi-Constrained, Metal/Ceramic/Polymer, Cemented or Non-Porous, Uncemented (21 CFR 888.3353 Hip joint metal/ceramic/polymer semi-constrained cemented or nonporous uncemented prosthesis)
- MEH – Prosthesis, Hip, Semi-Constrained, Uncemented, Metal/Polymer, Non-Porous, Calcium-Phosphate (21 CFR 888.3353 Hip joint metal/ceramic/polymer semi-constrained cemented or nonporous uncemented prosthesis)

Predicate Device(s):

K032726	ZIMMER M/L TAPER HIP PROSTHESIS, 7711 SERIES	Zimmer, Inc
K042337	ZIMMER M/L TAPER HIP PROSTHESIS WITH CALCICOAT CERAMIC COATING	Zimmer, Inc
K060040	ZIMMER M/L	Zimmer, Inc

Device Description

The Zimmer M/L Taper Hip Prosthesis is a modular, titanium alloy femoral stem designed to replace the proximal femur in total hip arthroplasty. It is flat, collarless and features a proximal-to-distal taper in the mediolateral (M/L) plane. The wedge shaped prosthesis is designed for cementless use and is circumferentially coated with titanium alloy plasma spray over the proximal body region. A subset of the subject devices are coated with Calcicoat hydroxyapatite tricalcium phosphate coating (HA/TCP) over the coating.

Indications for Use:

Total hip replacement for the following: severe hip pain and disability due to rheumatoid arthritis, osteoarthritis, traumatic arthritis, polyarthritis, collagen disorders, avascular necrosis of the femoral head, nonunion of previous fractures of the femur; congenital hip dysplasia, protrusion acetabuli, slipped capital femoral epiphysis; disability due to previous fusion; previously failed endoprostheses, and/or total hip components in the affected extremity and acute femoral neck fractures.

Hemi-hip replacement for the following: fracture dislocation of the hip; elderly, debilitated patients when a total hip replacement is contraindicated; irreducible fractures in which adequate fixation cannot be obtained; certain high subcapital fractures and comminuted femoral neck fractures in the aged; nonunion of femoral neck fractures; secondary avascular necrosis of the femoral head; pathological fractures of the femoral neck; and osteoarthritis in which the femoral head is primarily affected.

The femoral stem is for cementless use only.

Summary of Technological Characteristics:

The rationale for substantial equivalence is based on consideration of the following characteristics:

- **Intended Use:** Identical
- **Indications for Use:** Identical
- **Materials:** Substantially Equivalent
- **Design Features:** Substantially Equivalent
- **Sterilization:** Identical

Summary of Performance Data (Nonclinical and/or Clinical)

Non-Clinical Tests:

- **Performance Evaluation** – Performance testing was completed on the Zimmer M/L Taper Hip Prosthesis line extensions to determine equivalence to legally marketed devices.
- **Compatibility** – Fatigue testing and range of motion testing was conducted to show compatibility and safe and effective use with acquired devices.

Clinical Tests:

- Clinical test data is not provided for the subject device.

Substantial Equivalence Conclusion

The data presented in this submission demonstrate that the changes do not affect the safety and/or effectiveness of the subject devices and that the subject devices perform in a substantially equivalent manner to the legally marketed predicate devices.