



April 2, 2020

DeRoyal Industries, Inc.  
Sarah Bennett  
Senior Regulatory Affairs Specialist  
200 DeBusk Lane  
Powell, Tennessee 37849

Re: K200319

Trade/Device Name: HYDRO-TEMP Neonatal Skin Surface Temperature Sensor, HYDRO-TEMP Neonatal Skin Surface Temperature Sensor with Interface Cable

Regulation Number: 21 CFR 880.5130

Regulation Name: Infant Radiant Warmer

Regulatory Class: Class II

Product Code: FMT

Dated: February 5, 2020

Received: February 7, 2020

Dear Sarah Bennett:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Tina Kiang, Ph.D.  
Director  
DHT3C: Division of Drug Delivery and  
General Hospital Devices,  
and Human Factors  
OHT3: Office of Gastrorenal, ObGyn,  
General Hospital and Urology Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K200319

Device Name

HYDRO-TEMP Neonatal Skin Surface Temperature Sensor

Indications for Use (Describe)

The HYDRO-TEMP Neonatal Skin Surface Temperature Sensor is a non-sterile device intended for connection to an infant radiant warmer to monitor the neonatal patient's skin surface temperature.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**Corporate**  
200 DeBusk Lane  
Powell, TN 37849 USA  
865.938.7828  
800.251.9864  
www.deroyal.com

DeRoyal Industries, Inc.  
Traditional 510(k) Submission – HYDRO-TEMP Neonatal Skin  
Surface Temperature Sensor

**510(k) Summary**

**Date prepared:** February 5, 2020

**510(k) Owner:** DeRoyal Industries, Inc.  
200 DeBusk Lane  
Powell, TN 37849  
Owner/Operator #1044833

**510(k) Contact:** Sarah Bennett  
Senior Regulatory Affairs Specialist  
P: 865-362-6112  
F: 865-362-3741  
sabennett@deroyal.com

**Contract Manufacturer:** DeRoyal Cientifica de Latinoamerica  
S.R.L.  
Global Park  
Box 180-3006, 602 Parkway  
La Aurora, Heredia Heredia, Costa  
Rica 146

DeRoyal Industries, Inc.  
185 Richardson Way  
Maynardville, TN 37807

**Trade Name:** HYDRO-TEMP Neonatal Skin  
Surface Temperature Sensor

HYDRO-TEMP Neonatal Skin  
Surface Temperature Sensor with  
Interface Cable

**Common Name:** Skin Surface Temperature Sensor

**Classification Name:** Warmer, Infant Radiant

**Device Product Code:** FMT

**Regulatory Class:** Class II

**Regulation Number:** 21 CFR 880.5130

**Classification Panel:** General Hospital



**Corporate**

200 DeBusk Lane  
Powell, TN 37849 USA  
865.938.7828  
800.251.9864  
www.deroyal.com

DeRoyal Industries, Inc.  
Traditional 510(k) Submission – HYDRO-TEMP Neonatal Skin  
Surface Temperature Sensor

**Predicate Devices:** HYDRO-TEMP NICU and Skin  
Surface Temperature Sensor  
[K925006]

**Device Description**

The HYDRO-TEMP Neonatal Skin Surface Temperature Sensor consists of a probe used to monitor a neonate’s skin surface temperature used as an accessory while the patient is on an infant radiant warmer. The probe is a wire set with a thermistor chip at the distal end that passively modifies the electrical current traveling through the device. A colored connector at the proximal end of the wire set interfaces with the warming bed or a cable that interfaces with the bed. The probe is packaged with an adhesive pad used to affix the sensor to the patient’s skin surface.

The device is non-sterile and individually packaged in a resealable plastic bag. The probe covers that are packaged with the probe are cleared under K191425. The probes are compatible with the following warming beds: Datex-Ohmeda (Ohmeda, Care Plus, and IWS brands), Draeger (Babytherm, Isolette, Babyleo, Caleo, Resuscitaire, ICS, and Globe-Trotter brands), Atom (Transcapsule, Infa Warmer, INCU 1, and Sunflower brands), and GE (Giraffe and Panda brands).

**Indications for Use**

The HYDRO-TEMP Neonatal Skin Surface Temperature Sensor is a non-sterile device intended for connection to an infant radiant warmer to monitor the neonatal patient’s skin surface temperature.

**Differences to Predicate Device**

The encapsulation method of the subject device is different that the predicate device. The predicate device, the current encapsulation method utilizes a molded plastic cap that is filled with epoxy glue, and the thermistor is inserted into this cap. In the subject device, these materials would be replaced with a UV-cured adhesive. The thermistor is dipped into the adhesive and cured under UV light. This is the only component change in the device. The thermistor cap directly contacts the patient’s intact skin.

**Summary of Technological Characteristics**

Characteristic	HYDRO-TEMP Neonatal Skin Surface Temperature Sensor	HYDRO-TEMP NICU and Skin Surface Temperature Sensor Probe (K925006)



**Corporate**

200 DeBusk Lane  
 Powell, TN 37849 USA  
 865.938.7828  
 800.251.9864  
 www.deroyal.com

DeRoyal Industries, Inc.  
 Traditional 510(k) Submission – HYDRO-TEMP Neonatal Skin  
 Surface Temperature Sensor

Prescription Only	Yes	Yes
Use Environment	NICU	NICU
Design	A wire set packaged with an adhesive pad. The wire set has a thermistor chip at the distal end and a colored connector at the proximal end.	A wire set packaged with an adhesive pad. The wire set has a thermistor chip at the distal end and a colored connector at the proximal end.
Materials	Probe Cover – Adhesive pad with polyester film backing and release liner Wire – Copper with PVC insulation Thermistor – Ceramic Connector – Plastic-molded metal connector Strain Relief – PVC Cap – UV-curing Adhesive	Probe Cover – Adhesive pad with polyester film backing and release liner Wire – Copper with PVC insulation Thermistor – Ceramic Connector – Plastic-molded metal connector Strain Relief – PVC Cap – PVC and epoxy glue
Accuracy	±0.2°C	±0.2°C
Sterility	Non-Sterile	Non-Sterile
Biocompatibility	ISO 10993-1 compliant	ISO 10993-1 compliant
Shelf life	1 year (intent to extend to 5 years once testing completed)	3 years

**Summary of Performance Tests**

The proposed device has been tested and/or evaluated according to the following standards as applicable: ISO 10993-1, IEC 60601-2-21, ISO 80601-2-56, IEC 60601-1, IEC 60601-1-2.

The following biocompatibility tests were performed on the final, finished proposed device in accordance with ISO 10993 and FDA guidance: cytotoxicity, sensitization, and irritation. These results show the proposed device is biologically safe for its intended use.

Accuracy testing was performed to IEC 60601-2-21 demonstrating the tolerance range of temperature readings with the proposed change. Additionally, a time response test in accordance with ISO 80601-2-56 also was performed. The results of this testing demonstrate proposed modification to the thermistor encapsulation does not impact the device’s performance of its intended use. Evaluation and testing according to IEC 60601-1 and IEC 60601-1-2 also was performed to ensure the change to the thermistor encapsulation did not affect the electrical safety of the device.

All testing was performed on final, finished product manufactured with the proposed modification. The test results met the requirements of the aforementioned standards and demonstrate that



**Corporate**

200 DeBusk Lane  
Powell, TN 37849 USA  
865.938.7828  
800.251.9864  
[www.deroyal.com](http://www.deroyal.com)

DeRoyal Industries, Inc.  
Traditional 510(k) Submission – HYDRO-TEMP Neonatal Skin  
Surface Temperature Sensor

the proposed modification does not impact the safety or efficacy of the device.

**Conclusion**

The results of performance testing demonstrate the HYDRO-TEMP Neonatal Skin Surface Temperature Sensor is substantially equivalent to the predicate device. The difference between the predicate device and subject device is change to the thermistor encapsulation method and all other features are the same.