



April 30, 2020

CarboFix Orthopedics Ltd.  
Ms. Hila Wachsler-Avrahami  
Regulatory Affairs  
11 Ha'hoshlim St.  
Herzeliya, 4672411  
Israel

Re: K200519

Trade/Device Name: CarboClear® II Pedicle Screw System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Thoracolumbosacral Pedicle Screw System  
Regulatory Class: Class II  
Product Code: NKB  
Dated: February 27, 2020  
Received: March 2, 2020

Dear Ms. Wachsler-Avrahami:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.  
Acting Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K200519

Device Name

CarboClear® II Pedicle Screw System

Indications for Use (Describe)

The CarboClear® II Pedicle Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(K) Summary**  
**CarboFix Orthopedics Ltd.**  
**CarboClear® II Pedicle Screw System**

**Applicant Name**

CarboFix Orthopedics Ltd.  
11 Ha'hoshlim St., Herzeliya 4672411, Israel

**Contact Person**

Hila Wachsler-Avrahami  
CarboFix Orthopedics Ltd.  
11 Ha'hoshlim St., Herzeliya 4672411, Israel  
Tel: +972 9 9511511, Fax: +972 9 9548939

**Date Prepared**

February 27, 2020

**Trade/Proprietary Name**

CarboClear® II Pedicle Screw System

**Common Name**

Pedicle Screw System

**Regulation Number and Device Class**

21 CFR §888.3070; Class II

**Product Code, Regulatory Description and Review Panel**

NKB; Thoracolumbosacral pedicle screw system; Orthopedic

## **Predicate Devices**

### Primary Predicate Device

- CarboClear Pedicle Screw System (CarboFix Orthopedics Ltd.; K173487, K182377)

### Additional Predicate Devices

- S4<sup>®</sup> Spinal System (Aesculap Spine; K130291 and more)
- CD Horizon Pedicle Screws (Medtronic; K182928 and more)
- Mesa Spinal System (K2M, Inc.; K052398 and more)

### Reference Device

- Piccolo Composite Nailing Systems (CarboFix Orthopedics Ltd.; K102369, K123810 and more)

## **Indications for Use**

The CarboClear<sup>®</sup> II Pedicle Screw System is intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the thoracic and lumbar spine in whom life expectancy is of insufficient duration to permit achievement of fusion.

## **System Description**

The CarboClear<sup>®</sup> II Pedicle Screw System is composed of implants in various dimensions, used to build a spinal construct; and of a set of instruments, intended to assist in the insertion and placement of the implants.

The CarboClear<sup>®</sup> II implants include pedicle screws, rods, locking element and transverse connectors. The implants are made of carbon fiber-reinforced polyetheretherketone (CFR-PEEK). The threaded portion of the pedicle screws is encased within a thin titanium shell and includes a small tantalum marker.

The implants are supplied sterile, and are intended for single use.

The locking mechanism of the screws is being modified to lower their profile and simplify the surgical technique. In addition, larger diameter and longer screws are being added to the system.

### **Performance Data and Substantial Equivalence**

The CarboClear<sup>®</sup> II Pedicle Screw System intended use, material, design, dimensions, technological characteristics and principles of operation are substantially equivalent to those of the primary predicate device. Aside for larger screws, the only difference between the systems is a modification of the locking mechanism.

Performance testing was conducted according to ASTM F1717 (static and dynamic compression bending, and static torsion) and ASTM F1798 (axial and torsional grip, and static flexion-extension). The test results are substantial equivalent to those of the predicate devices, demonstrating substantially equivalent mechanical performance of the subject device.

### **Conclusion**

Based on the information provided in this Premarket Notification, the subject CarboClear<sup>®</sup> II Pedicle screw System is substantially equivalent to its predicate devices.