



GE Medical Systems Ultrasound and Primary Care Diagnostics, LLC
% Tracey Ortiz
Regulatory Affairs Director
9900 W. Innovation Drive
WAUWATOSA WI 53226

July 1, 2020

Re: K200138

Trade/Device Name: Versana Premier
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: Class II
Product Code: IYN, IYO, ITX
Dated: June 1, 2020
Received: June 2, 2020

Dear Tracey Ortiz:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see

<https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.
Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K200138

Device Name

Versana Premier

Indications for Use (Describe)

The Versana Premier is a general purpose diagnostic ultrasound system intended for use by qualified and trained healthcare professionals for ultrasound imaging, measurement, display and analysis of the human body and fluid. Versana Premier clinical applications include : Fetal/Obstetrics, Abdominal (includes GYN, Urology), Pediatric, Small Organ (includes breast, testes, thyroid), Cardiac (includes Adult and Pediatric), Vascular/Peripheral Vascular, Musculoskeletal Conventional, Musculoskeletal Superficial, Transcranial, Transrectal, Transvaginal, Interventional guidance (includes tissue biopsy, fluid drainage, vascular and non-vascular access). Modes of operation include : B, M, PW Doppler, CW Doppler, Color Doppler, Color M Doppler, Power Doppler, Harmonic Imaging, Coded Pulse, 3D/4D Imaging mode and Combined modes: B/M, B/Color, B/PWD, B/Color/PWD, B/Power/PWD. Versana Premier is intended to be used in a hospital or medical clinic.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



GE Healthcare

510(k) Premarket Notification Submission

510(k) Summary

K200138

In accordance with 21 CFR 807.92 the following summary of information is provided:

Date: May 28, 2020

Submitter: GE Medical Systems Ultrasound and Primary Care Diagnostics
9900 Innovation Drive
Wauwatosa, WI 53226

Primary Contact Person: Tracey Ortiz
Regulatory Affairs Director
GE Healthcare
T:(262)676-6120

Secondary Contact Person: Karim Amrouche
Regulatory Affairs
GE Healthcare

Device Trade Name: Versana Premier
Common/Usual Name: Diagnostic Ultrasound System
Classification Names: Class II
IYN (primary), IYO, ITX (secondary)

Product Code(s):
Ultrasonic Pulsed Doppler Imaging System. 21CFR 892.1550, 90-IYN; Ultrasonic Pulsed Echo Imaging System, 21CFR 892.1560, 90-IYO; Diagnostic Ultrasound Transducer, 21 CFR 892.1570, 90-ITX

Primary Predicate Device: Versana Premier (K182277)

Reference Device(s): LOGIQ P6 (K073297)
Versana Active (K191798)

Device Description: The Versana Premier is a general purpose, diagnostic ultrasound system for use by qualified and trained healthcare professionals. The system is a mobile console that includes an operator control panel, display monitor and transducers.

The console provides digital acquisition, processing and display capability. The system has an internal battery to allow for acquisition while the system is not plugged into a power source.

The operator control panel includes function keys, trackball, an alfa-numeric keyboard and a touch panel as input sources of the device.



GE Healthcare

510(k) Premarket Notification Submission

The variety of transducers include convex, linear, sector, dual and mechanical 4D transducers. The access types include trans- body surface, transrectal, transvaginal and transcranial.

Data can be imported or exported by DVD, USB, LAN or WiFi if the USB wireless adapter is connected to the system. An external ECG module has been verified to use as input for gating/triggering during scanning.

The system has a HDMI port, VGA connection port, Audio out port, S-Video port, and a Composite Out port connection. The system has an external AC outlet to allow connection of a printer and sit in the printer box of the console and an option for external Printer USB Isolator for other printers to connect. The system supports one way, Bluetooth communication capability from the system to a personal device to allow for sharing of the patient's data/images when the Bluetooth USB adapter is connected to the system.

Intended Use: The Versana Premier is a general purpose diagnostic ultrasound system intended for use by qualified and trained healthcare professionals for ultrasound imaging, measurement, display and analysis of the human body and fluid. Versana Premier clinical applications include : Fetal/Obstetrics, Abdominal (includes GYN, Urology), Pediatric, Small Organ (includes breast, testes, thyroid), Cardiac (includes Adult and Pediatric), Vascular/Peripheral Vascular, Musculoskeletal Conventional, Musculoskeletal Superficial, Transcranial, Transrectal, Transvaginal, Interventional guidance (includes tissue biopsy, fluid drainage, vascular and non-vascular access). Modes of operation include : B, M, PW Doppler, CW Doppler, Color Doppler, Color M Doppler, Power Doppler, Harmonic Imaging, Coded Pulse, 3D/4D Imaging mode and Combined modes: B/M, B/Color, B/PWD, B/Color/PWD, B/Power/PWD. Versana Premier is intended to be used in a hospital or medical clinic.

Technology: The Versana Premier employs the same fundamental scientific technology as its predicate device and reference devices.



GE Healthcare

510(k) Premarket Notification Submission

Determination of Substantial Equivalence: Comparison to Predicate Devices
The proposed Versana Premier system is substantially equivalent to the predicate Versana Premier device with regard to intended use, imaging capabilities, technological characteristics and safety effectiveness.

- The proposed Versana Premier and predicate Versana Premier (K182277) have similar clinical indications for use except for the following changes:
- Vascular has been added with peripheral vascular, this clarifies vessels and is not a change in imaging or product. No impact to safe or effective use.
- Applications that were listed in the notes of the probe IFU tables have been added in the written statement, so this is really not a change from the predicate.
- Modes of operation & use environment are added to the written statement as per the 2019 ultrasound guidance. There was no change in the modes or the use environment from the predicate.
- A new transducer, E7C8L-RS, is being added to the Versana Premier. The E7C8L-RS transducer has similar applications as the reference device ERB transducer on LOGIQ P6, cleared in K073297. Interventional needle guidance use is expanding to include non-vascular access but does not include brachytherapy.
- E7C8L-RS transducer materials are similar to the E8Cs-RS cleared on predicate Versana Premier (K182277).
- Adding M-mode to L6-12-RS, 12L-RS and LK760-RS already cleared on Versana Active (K191798);
- Adding Transcranial clinical application to the 8C-RS transducer.



GE Healthcare

510(k) Premarket Notification Submission

Summary of Non-Clinical Tests:

Versana Premier has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic, and mechanical safety, and has been found to comply with applicable medical device safety standards. The Versana Premier complies with voluntary standards:

- AAMI/ANSI ES60601-1, Medical Electrical Equipment – Part 1: General requirements for basic safety and essential performance - 2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012
- IEC 60601-1-2, Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests - Edition 4.0, 2014
- IEC 60601-2-37, Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment - Edition 2.1, 2015
- ISO 10993-1, Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process - Fourth edition, 2009
- IEC 62359, Ultrasonics - Field characterization - Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields - Edition 2.1, 2017
- ISO 14971, Application of risk management to medical devices, 2007
- NEMA PS 3.1 - 3.20, Digital Imaging and Communications in Medicine (DICOM) Set , 2016



GE Healthcare

510(k) Premarket Notification Submission

The following quality assurance measures are applied to the development of the system:

- Risk Analysis
- Requirements Reviews
- Design Reviews
- Testing on unit level (Module verification)
- Integration testing (System verification)
- Performance testing (Verification)
- Safety testing (Verification)

Transducer material and other patient contact materials are biocompatible.

Summary of Clinical Tests:

The subject of this premarket submission, Versana Premier, did not require clinical studies to support substantial equivalence.

Conclusion: GE Healthcare considers the Versana Premier to be as safe, as effective, and performance is substantially equivalent to the predicate devices.