



October 25, 2021

Canadian Pioneer Medical Technology Corporation
Rashid Reza Mir Sayah
Managing Director
unit 2, 210 Drumlin Circle, Concord
Vaughan, Ontario L4k 3E3
Canada

Re: K212793

Trade/Device Name: Laser Treatment System, Model: Hera, Armo

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In
Dermatology

Regulatory Class: Class II

Product Code: GEX

Dated: August 26, 2021

Received: September 1, 2021

Dear Rashid Reza Mir Sayah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmnm.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Purva Pandya
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K212793

Device Name

Laser Treatment System, Model: Hera, Armo

Indications for Use (Describe)

The Laser Treatment System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Section 5 510(k) summary

I Submitter

Canadian Pioneer Medical Technology Corporation
Unit 2-210 Drumlin Circle, Vaughan, Ontario, L4K 3E3, Canada

Contact person:

Dr. Rashid Reza Mir Sayah

Managing Director

Phone: 4377727788

Email:

Canadianpioneer@yahoo.com

Canadianpioneermedical@gmail.com

Date of preparation: October 7, 2021

II Subject Device

Trade Name of Device: Laser Treatment System, Model: Hera, Armo

Common name: Powered Laser Surgical Instrument

Classification Name: Laser Surgical Instrument For Use In General And Plastic Surgery
And In Dermatology

Regulation Number: 21 CFR 878.4810

Regulatory Class: II

Product code: GEX

Review Panel: General & Plastic Surgery

III Predicate Devices

Trade Name of Devices: Laser Treatment System, Model: NYX and Grace

Common name: Powered Laser Surgical Instrument

Regulation Number: 21 CFR 878.4810

Regulatory Class: II

Product code: GEX

Review Panel: General & Plastic Surgery

510(k) number: K210033

Manufacturer: Canadian Pioneer Medical Technology Corporation

IV Device description

The Laser Treatment System consists of the main unit and a hand piece. The system uses a diode laser as an active medium placed in an optical cavity to produce amplified beam at the

wavelength of 808 nm. A microprocessor is used to control electronics for the front panel. A self-contained water cooling system is built into the power supply unit.

The Diode Laser Therapy Devices is designed to be used in dermatological practice for stable, long term hair reduction. The principle of laser hair removal is selective photothermolysis. The wavelength of 808nm would be able to effectively penetrate deep into and absorbed by the target chromophore. The laser power is delivered to the treatment region via a delivery system.

The proposed device includes power supply system, delivery system, control system, cooling system, laser system.

The models object of this submission are Armo and Hera:

Hera has one handpiece connector.

Armo has two handpiece connectors.

The 808 nm handpieces with different spot sizes (3 spot sizes : 12×12mm,12×20mm,15×27m) are available for both models.

V Indications for use

The Laser Treatment System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.

Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.

VI Comparison of technological characteristics with the predicate devices

The indication of proposed device is covered by the predicated devices. The proposed device is only intended to use for hair removal. The device includes three models of handpieces for clearance in this submission. The differences between models are their spot sizes and Density .These are covered by the predicated devices.

Device feature	Laser Treatment System (subject device)	Laser treatment system Nyx and Grace (K210033)
Product code	GEX	GEX
Regulation number	21 CFR 878.4810	21 CFR 878.4810
Indications for use	The Diode Laser Treatment System is intended for hair	The Diode Laser Treatment

	<p>removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.</p> <p>□Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.</p>	<p>System is intended for hair removal, permanent hair reduction on all skin types (Fitzpatrick skin type I-VI), including tanned skin.□Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime.</p>
Operation principle	<p>Melanin could absorb the energy from the laser, which would result in temperature rapid increase, to destroy surrounding hair follicles, and finally remove hair.</p>	<p>Melanin could absorb the energy from the laser, which would result in temperature rapid increase, to destroy surrounding hair follicles, and finally remove hair.</p>
Laser classification	Class IV	Class IV
Light delivery system	808 nm	808nm
Spot Size	12×12mm 12×20mm	12×12mm

	15×27mm	
Fluence	1~120J/cm ² at spot size 12×12mm 1~70 J/cm ² at spot size12×20mm 1~40 J/cm ² at spot size15×27mm	1~120 J/cm ²
Frequency	1-10Hz	1-10 Hz
Pulse Duration	10-300ms	10-300 ms
Power supply	100-240V AC, 50/60Hz	100-240V AC, 50/60Hz
Patient contact material	Sapphire in handpiece and handpiece tip (stainless steel)	Sapphire in handpiece and handpiece tip (stainless steel)
Biocompatibility	Comply with ISO10993-1	Comply with ISO10993-1
Electrical Safety	Comply with IEC60601-1, IEC60601-2-22	Comply with IEC60601-1, IEC60601-2-22
EMC	Comply with IEC60601-1-2,	Comply with IEC60601-1-2,
Laser safety	Comply with IEC60825-1, IEC60601-2-22	Comply with IEC60825-1, IEC60601-2-22

VII Performance data

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility testing

Biocompatibility of the Laser Treatment System was evaluated in accordance with ISO 10993-1:2009 for the body contact category of "Surface –intact skin" with a contact duration of "Limited (< 24 hours)". The following tests were performed, as recommended: Cytotoxicity, Irritation and Sensitization. All evaluation acceptance criteria were met

Electrical safety and electromagnetic compatibility (EMC)

Electrical safety and EMC testing were conducted on the Laser Treatment System. The system has been tested to comply with the following standards:

- ANSI/AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012" Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance;
- IEC 60601-1-2:2014, Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance- Collateral standard: Electromagnetic compatibility- Requirements and tests.

Bench Testing

- IEC 60601-2-22:2007 + A1:2012, Medical Electrical Equipment - Part 2-22: Particular Requirements For Basic Safety And Essential Performance Of Surgical, Cosmetic, Therapeutic And Diagnostic Laser Equipment;
- IEC 60825-1: 2007, Safety of laser products - Part 1: Equipment classification and requirements.

Software Verification and Validation Testing

Software verification and validation testing were conducted, and documentation was provided as recommended by FDA's Guidance for Industry and FDA Staff, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices".

VIII Conclusion

The non-clinical testing demonstrates that the device is as safe, as effective and performs as well as the legally marketed device.