



March 14, 2022

DWFritz Automation, Inc.  
% John Gillespy  
President  
FDA 510k Consultants, LLC  
1100 Del Lago Circle  
Palm Beach Gardens, Florida 33410

Re: K213894  
Trade/Device Name: DWFritz ASM2000  
Regulation Number: 21 CFR 878.4040  
Regulation Name: Surgical apparel  
Regulatory Class: Class II  
Product Code: FXX  
Dated: December 10, 2021  
Received: December 14, 2021

Dear John Gillespy:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Bifeng Qian  
Acting Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K213894

Device Name  
DWFritz ASM2000

Indications for Use (Describe)

When properly worn, DWFritz ASM2000 surgical face mask is intended to protect both the patient and healthcare worker from transfer of microorganisms, body fluids and particulate matter. The device is non-sterile and for single use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary\_K213894

Prepared in accordance with the requirements of 21 CFR Part 807.92.

1. 510(k) Submitter: DWFritz Automation, Inc.  
9600 SW Boeckman Rd., Wilsonville OR 97070  
Phone: 503-885-7030  
Email: cpsmith@dwfritz.com
2. Company Contact: Curtis Smith, Senior Quality Manager
3. Date of Submission: December 10, 2021
4. 510(k) Preparer: John F. Gillespy, MBA  
FDA 510k Consulting, LLC  
Palm Beach Gardens, FL 33410  
Phone: 386-243-4332  
Email: john@fda510kconsultants.com
5. Device Classification: Trade name: DWFritz ASM2000  
Common name: Surgical Face Mask  
Device: Surgical Mask  
Class: II  
Regulation #: 878.4040  
Product Code: FXX
6. Predicate: Applicant: Mexpo International, Inc.  
Device: Avianz Surgical Face Mask  
510(k) Number: K200847



### 7. Device Description

**DWFritz ASM2000** is a 3-ply, flat-pleated style surgical face mask. The device is manufactured with three layers. The inner and outer layers are made of spunbond polypropylene, and the middle layer is made of melt blown polypropylene filter. Each mask contains ear loops made of polyester filament and spandex fiber to secure the mask over the user's face and mouth with galvanized wire nose bridge to fit firmly over the nose.

This device is not made from any natural rubber latex.

The product is sold non-sterile and intended to be disposable for single-use.

8. Indications For Use

When properly worn, **DWFritz ASM2000** surgical face mask is intended to protect both the patient and healthcare worker from transfer of microorganisms, body fluids and particulate matter. This device is non-sterile and for single use only.

The device is intended for OTC use.

9. Comparison To Predicate Device

**DWFritz ASM2000** is compared to a predicate device, Avianz Surgical Face Mask (K200847).

**Comparison Table**

Device	Proposed Device	Predicate Device	Comparison
<b>Manufacturer</b>	DWFritz Automation, Inc.	Mexpo International, Inc.	-
<b>510K Number</b>	K213894	K200847	-
<b>Product Name</b>	Surgical Face Mask	Surgical Face Mask	Same
<b>Product Code</b>	FXX	FXX	Same
<b>Classification</b>	Class II (21 CFR 878.4040)	Class II (21 CFR 878.4040)	Same
<b>Intended Use</b>	When properly worn, <b>DWFritz ASM2000</b> surgical face mask is intended to protect both the patient and healthcare worker from transfer of microorganisms, body fluids and particulate matter. This device is non-sterile and for single use only.	When properly worn, the surgical face masks are intended to protect both patient and healthcare workers from transfer of microorganisms, body fluids and airborne particles. This device is non-sterile and for single use only.	Same
<b>Model</b>	3-ply, flat-pleated style with ear loops and nose bridge	3-Ply, Flat-Pleated Style with ear loops and nose bridge	Same
<b>Materials, Dimensions, &amp; Other Features</b>			
<b>Outer Facing Layer</b>	Spunbond Polypropylene	Spunbond Polypropylene	Same
<b>Middle Layer</b>	Meltblown Polypropylene	Meltblown Polypropylene	Same
<b>Inner Facing Layer</b>	Spunbond Polypropylene	Spunbond Polypropylene	Same
<b>Nose Bridge</b>	Galvanized wire with polyester coating	Galvanized wire with polyester coating	Same

<b>Ear Loops</b>	Not made with natural rubber latex	Not made with natural rubber latex	Same
<b>Color</b>	Blue (inner facing) & white	White	Similar
<b>Dimension (W X L)</b>	9.8 X 17.2 ± 0.5cm	9.5 X 17.5 ± 0.5cm	Similar
<b>Use</b>	OTC, single use, disposable	OTC, single use, disposable	Same
<b>Sterility</b>	Non-Sterile	Non-Sterile	Same
<b>Testing</b>			
<b>Performance Level:</b>			
<b>ASTM F2100:2019</b>	Level 2	Level 2	Same
<b>Bacterial Filtration</b>	99.9%	98%	Both Pass
<b>Particle Filtration</b>	98.4%	98%	Both Pass
<b>Fluid Resistance</b>	94 of 96 pass @ 120 mmHg	29 of 32 pass @ 120 mmHg	Both Pass
<b>Differential Pressure</b>	< 4.7 mmH <sub>2</sub> O/cm <sup>2</sup>	< 5.0 mmH <sub>2</sub> O/cm <sup>2</sup>	Both Pass
<b>Flammability</b>	Class 1	Class 1	Same
<b>Biocompatibility:</b>			
<b>Cytotoxicity</b>	Not Cytotoxic	Not Cytotoxic	Same
<b>Sensitization</b>	Not Skin Sensitive	Not Skin Sensitive	Same
<b>Irritation</b>	Not Skin Irritating	Not Skin Irritating	Same

**DWFritz ASM2000** is comparable to the predicate device (K200847) in terms of materials, performance, and biocompatibility.

#### 10. Summary Of Non-Clinical Testing

The following performance data has been provided to demonstrate that **DWFritz ASM2000** meets the acceptance criteria in the standard.

#### **Non-Clinical Testing Table**

<b>Test Standard</b>	<b>Purpose</b>	<b>Acceptance Criteria</b>	<b>Results</b>
<b>ISO 10993-05</b>	Cytotoxicity	< Grade 2 cytotoxic effect	Grade 1 (slight reactivity)
<b>ISO 10993-10</b>	Skin Sensitization	No sensitization reaction	No sensitization reaction
<b>ISO 10993-10</b>	Skin Irritation	No irritation reaction	No irritation reaction
<b>ASTM F2101</b>	Bacterial Filtration Efficiency (BFE)	98%	99.9%
<b>ASTM F2299</b>	Particulate Filtration Efficiency (PFE)	98%	98.8-99.7% in 32 test articles in each of three lots (96 total test articles)

<b>ASTM F1862</b>	Synthetic Blood Fluid Resistance	29 out of 32 pass at 120 mmHg	30, 31, and 32 pass at 120 mmHg, out of 32 test articles in each of three lots (96 total test articles)
<b>MIL-M-36954C</b>	Differential Pressure (Delta P)	< 5.0 mmH <sub>2</sub> O/cm <sup>2</sup>	2.9 - 4.7 mmH <sub>2</sub> O/cm <sup>2</sup> in 32 test articles in each of three lots (96 total test articles)
<b>16 CFR 1610</b>	Flammability	IBE = Test article ignited but extinguished, burn time > 3.5 sec (Class 1)	IBE for all 14 test articles in each of 3 lots (42 total test articles)

## 11. Patient-Contacting Materials

Materials composition follows:

- Outer facing layer – spunbond polypropylene (blue)
- Middle layer – meltblown polypropylene filter (white)
- Inner facing layer – spunbond polypropylene (white)
- Nose bridge – Galvanized wire with polyester coating
- Ear loops – polyester filament & spandex fiber, not made from natural rubber latex (white)

The outer & inner layers and ear loops touch the patient's face and hands. Type and duration of contact with patient follows:

- Surface contact – intact skin
- Limited duration – less than 24 hours

## 12. Software Verification and Validation... The mask contains no software or firmware.

The conclusion drawn from the nonclinical tests demonstrates that the subject device, **DWFritz ASM2000**, is as safe, as effective, and performs as well as or better than the legally marketed predicate device cleared under K200847.