



November 25, 2020

Fusion Orthopedics, LLC  
Kolby Black  
4135 S. Power Rd., Suite 110  
Mesa, Arizona 85212

Re: K193377/S001

Trade/Device Name: TopLock Anchor System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: MBI  
Dated: October 28, 2020  
Received: October 30, 2020

Dear Mr. Black:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see

<https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laura C. Rose, Ph.D.  
Assistant Director  
DHT6C: Division of Restorative, Repair,  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K193377

Device Name  
TopLock Anchor System

### Indications for Use (Describe)

The TopLock Anchor System is intended to secure soft tissue to bone of:

The Shoulder:

Bankart Repair, SLAP Lesion Repair, Acromio-Clavicular Separation, Rotator Cuff Repair, Capsule Repair, Biceps Tenodesis, Deltoid Repair

The Elbow:

Ulnar or Radial Collateral Ligament Reconstruction, Bicep Tendon Reconstruction, Tennis Elbow Repair

The Hand and Wrist:

Scapholunate Ligament Reconstruction, Ulnar / Radial Collateral Ligament Reconstruction, Extensor Tendon Repair, Flexor Tendon Repair

The Knee:

Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Posterior Oblique Ligament Repair, Extra Capsular Reconstruction, Iliotibial Band Tenodesis, Patellar Tendon Repair

The Ankle and Foot:

Lateral Ligament Stabilization Repair, Medial Ligament Stabilization Repair, Achilles Tendon Repair / Reconstruction, Metatarsal Ligament Repair, Hallux Valgus Reconstruction, Mid and Rear Foot Reconstruction

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

### 510(k) Summary: TopLock Anchor System

In accordance with 21 CFR 807.92 of the Federal Code of Regulations

<b>Date Prepared</b>	November 24, 2020
<b>Submitted By</b>	Fusion Orthopedics, LLC 4135 S. Power Rd., Suite 110 Mesa, AZ 85212 800-403-6876
<b>Primary Contact</b>	Kolby Black 4135 S. Power Rd., Suite 110 Mesa, AZ 85212 800-403-6876 Tele e-mail: FDA@fusionorthopedics.com
<b>Trade Name</b>	TopLock Anchor System
<b>Submission Number</b>	K193377
<b>Classification Name</b>	Fastener, Fixation, Non-degradable, Soft Tissue
<b>Class</b>	II
<b>Product Code</b>	MBI
<b>CFR Section</b>	21 CFR Section 888.3040
<b>Device Panel</b>	Orthopedic
<b>Primary Predicate Device</b>	Core Essence reNOVO Suture Anchors (K071520)
<b>Secondary Predicate Device</b>	Arthrex PEEK Mini PushLock Anchors (K101679)
<b>Reference Devices</b>	Core Essence Orthopaedics, Inc., TAC-tite™ PEEK (K111716) Parcus Medical Twist & Triple Twist PEEK Screw-In Suture Anchors (K120942) Parcus Medical PEEK CF Push-In Anchors (K102162)
<b>Device Description</b>	The TopLock Anchor System consists of a collection of barbed suture anchors, manufactured from medical grade Polyetheretherketone or PEEK (ASTM F2026). The anchors are offered from diameters of 3.5mm to 5.0mm with each size offering two eyelet size options to accommodate suture and suture tape. The size range and configurations are offered for surgical treatment of ligament, tendon and soft tissue pathologies of the shoulder and other joints. The specialized instruments are made primarily of surgical grade stainless steel (ASTM F899).
<b>Indications for Use</b>	The TopLock Anchor System is intended to secure soft tissue to bone of: The Shoulder: Bankart Repair SLAP Lesion Repair Acromio-Clavicular Separation Rotator Cuff Repair

	<p>Capsule Repair Biceps Tenodesis Deltoid Repair</p> <p>The Elbow: Ulnar or Radial Collateral Ligament Reconstruction Bicep Tendon Reconstruction Tennis Elbow Repair</p> <p>The Hand and Wrist: Scapholunate Ligament Reconstruction Ulnar / Radial Collateral Ligament Reconstruction Extensor Tendon Repair Flexor Tendon Repair</p> <p>The Knee: Medial Collateral Ligament Repair Lateral Collateral Ligament Repair Posterior Oblique Ligament Repair Extra Capsular Reconstruction Iliotibial Band Tenodesis Patellar Tendon Repair</p> <p>The Ankle and Foot: Lateral Ligament Stabilization Repair Medial Ligament Stabilization Repair Achilles Tendon Repair / Reconstruction Metatarsal Ligament Repair Hallux Valgus Reconstruction Mid and Rear Foot Reconstruction</p>
<b>Materials</b>	<p>Polyetheretherketone (PEEK) (ASTM F2026) Ultra High Molecular Weight Polyethylene Suture (UHMWPE) (ASTM F2848) Stainless Steel (ASTM F899) Aluminum (ASTM B209) Nylon</p>
<b>Substantial Equivalence Claimed to Predicate Devices</b>	<p>The TopLock Anchor System is substantially equivalent to the predicate devices in terms of design, materials used, mechanical safety and performances.</p>
<b>Non-clinical Test Summary</b>	<p>Validations were performed on the cleaning, packaging and sterilization of the implants and associated surgical instruments. Mechanical Testing was performed to show design equivalence including insertion testing, axial pull-out testing, cyclic fatigue testing, and post fatigue pullout testing. Pyrogenicity was evaluated using the Limulus amoebocyte lysate (LAL) assay. The testing demonstrated that the subject device meets the recommended maximum endotoxin levels of 20 (EU)/Device.</p>
<b>Clinical Test Summary</b>	<p>No clinical studies were performed.</p>
<b>Conclusions: Non- clinical and Clinical</b>	<p>Fusion Orthopedics LLC considers the TopLock Anchor System to be equivalent to the predicate devices listed above. This conclusion is based upon the devices' similarities in principles of operation, technology, materials and indications for use.</p>