



March 3, 2021

Orthofix Inc.
Ms. Natalia Volosen
Regulatory Affairs Principal
3451 Plano Parkway
Lewisville, Texas 75056

Re: K203576
Trade/Device Name: FORZA[®] Ti Spacer System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: February 4, 2021
Received: February 5, 2021

Dear Ms. Volosen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for
Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K203576

Device Name
FORZA® Ti Spacer System

Indications for Use (Describe)

The FORZA Ti Spacer System is indicated for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels in the lumbar spine (L2-S1). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies. DDD patients may also have up to Grade 1 spondylolisthesis at the involved levels. These patients may have had a previous non-fusion surgery at the involved levels.

The FORZA Ti Spacer System is intended for use with autograft and/or allograft comprised of cancellous and/or corticocancellous bone graft and supplemental fixation system.

Patients must have undergone a regimen of at least six months of non-operative treatment prior to being treated with the FORZA Ti Spacer System.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary K203576

Device Trade Name: FORZA® Ti Spacer System

Manufacturer: Orthofix Inc.
3451 Plano Parkway
Lewisville, TX 75056, USA
Phone: (214) 937-2145
Fax: (214) 937-3322

Contact Person: Ms. Natalia Volosen
Regulatory Affairs Principal
nataliavolosen@orthofix.com

Date Prepared: February 4, 2021

Registration Number: 2183449

Product Code: MAX

Classifications: Class II – 21 CFR §888.3080, Intervertebral body fusion device

Primary Predicate: K200052 – FORZA PTC Spacer System, SE 4/9/2020

Reason for the 510(k) Submission: Addition of 3D printed interbody devices

Device Description:

FORZA Ti Spacer System is comprised of a variety of 3D Titanium printed implants (conforming to ASTM F3001) that have porous end plates and a functional gradient porous structure. The implant superior and inferior Titanium plate surface provides increased stability for the implant. The titanium porous structure have open macroscopic 3D pores with a microscopic roughened surface and nano-scale features. FORZA Ti Spacer System is implanted in the intervertebral disc space and is intended to facilitate vertebral fusion by stabilizing adjacent vertebrae, maintaining disc height, and preventing the collapsing of one vertebrae onto another.

FORZA Ti Spacer System is not intended to be used as a stand-alone device and must be used with supplemental fixation. FORZA Ti implants are provided sterile.

Indications for Use and Intended Use:

FORZA Ti Spacer System is indicated for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels in the lumbar spine (L2-S1). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies. DDD patients may also have up to Grade 1 spondylolisthesis at the involved levels. These patients may have had a previous non-fusion surgery at the involved levels. FORZA Ti Spacer System is intended for use with autograft and/or allograft comprised of cancellous and/or corticocancellous bone graft and supplemental fixation system.

Patients must have undergone a regimen of at least six months of non-operative treatment prior to being treated with the FORZA Ti Spacer System.

Performance Testing Summary:

FORZA Ti spacer was evaluated via mechanical testing per ASTM F2077 (including static and dynamic axial compression, static and dynamic compression shear), ASTM F2267 (subsidence) and characterization of particles per ASTM F1877. The results demonstrated the performance of FORZA Ti spacer is substantially equivalent to the predicate devices.

Substantial Equivalence:

The subject device is substantially equivalent to the FORZA PTC Spacer System (K200052) predicate device with respect to indications, design, materials, function, and performance.

Conclusion:

FORZA Ti Spacer System have the same intended use, indications for use, technological characteristics, materials, the same principles of operation and same design as the predicate device FORZA PTC Spacer System (K200052).