



February 3, 2023

Beckman Coulter, Inc.  
Kate Oelberg  
Senior Staff Quality and Regulatory Affairs  
1000 Lake Hazeltine Drive  
Chaska, MN 55318

Re: K223679  
Trade/Device Name: Access AMH  
Regulation Number: 21 CFR 862.1092  
Regulation Name: Anti-Mullerian Hormone Test System  
Regulatory Class: Class II  
Product Code: PQO  
Dated: December 8, 2022  
Received: December 8, 2022

Dear Kate Oelberg:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Paula Caposino -S

Digitally signed by Paula  
Caposino -S  
Date: 2023.02.03 17:08:40  
-05'00'

Paula Caposino, Ph.D.

Acting Deputy Director

Division of Chemistry

and Toxicology Devices

OHT7: Office of In Vitro Diagnostics

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K223679

Device Name  
Access AMH

### Indications for Use (Describe)

The Access AMH assay is a paramagnetic particle chemiluminescent immunoassay for the quantitative determination of anti-Müllerian hormone (AMH) levels in human serum and lithium heparin plasma using the Access Immunoassay Systems as an aid in the assessment of ovarian reserve in women presenting to fertility clinics. This system is intended to distinguish between women presenting with AFC (antral follicle count) values > 15 (high ovarian reserve) and women with AFC values ≤ 15 (normal or diminished ovarian reserve). The Access AMH is intended to be used in conjunction with other clinical and laboratory findings such as antral follicle count, before starting fertility therapy. The Access AMH is not intended to be used for monitoring of women undergoing controlled ovarian stimulation in an Assisted Reproduction Technology program.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



Access AMH  
510(k) Summary

## 510(k) Summary

510(k) Number k223679

Date Prepared: February 3, 2023

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92

The assigned 510(k) number is k223679.

**Submitted By:**

Beckman Coulter, Inc.  
1000 Lake Hazeltine Drive  
Chaska, MN 55318

**Primary Contact:**

Kate Oelberg  
Senior Staff Quality and Regulatory Affairs  
Phone: (612) 431-7315  
Email: [kmoelberg@beckman.com](mailto:kmoelberg@beckman.com)

**Alternate Contact:**

Madhuri Boppana  
Regulatory Affairs Analyst, Senior  
Email: [mboppana@beckman.com](mailto:mboppana@beckman.com)

**Device Name:**

**Common Name:** Anti-Müllerian hormone test system

**Trade Name:** Access AMH

**Classification Name:** Anti-Müllerian hormone test system

**Classification Regulation:** (21 CFR 862.1092)

**Predicate Device:**

**Device Name:** Access AMH

**510(k) Numbers:** k170524

## Device Description

The Access Anti-Mullerian Hormone Assay, Access Anti-Mullerian Hormone Calibrators, and the Access Immunoassay analyzers comprise the Dxl 9000 Access Immunoassay System for the quantitative determination of Anti-Mullerian Hormone levels in human serum and lithium heparin plasma using the Dxl 9000 Access Immunoassay system.

## Intended Use

The Access AMH assay is a paramagnetic particle chemiluminescent immunoassay for the quantitative determination of anti-Mullerian hormone (AMH) levels in human serum and lithium heparin plasma using the Access Immunoassay Systems as an aid in the assessment of ovarian reserve in women presenting to fertility clinics. This system is intended to distinguish between women presenting with AFC (antral follicle count) values  $> 15$  (high ovarian reserve) and women with AFC values  $\leq 15$  (normal or diminished ovarian reserve). The Access AMH is intended to be used in conjunction with other clinical and laboratory findings such as antral follicle count, before starting fertility therapy. The Access AMH is not intended to be used for monitoring of women undergoing controlled ovarian stimulation in an Assisted Reproduction Technology program.

## Comparison of Technological Characteristics to the Predicate

Parameter	Access AMH Assay on Access 2 Immunoassay System Predicate – k170524	Access AMH Assay on Dxl 9000 Access Immunoassay Analyzer
Intended use	The Access AMH assay is a paramagnetic particle chemiluminescent immunoassay for the quantitative determination of anti-Müllerian hormone (AMH) levels in human serum and lithium heparin plasma using the Access Immunoassay Systems as an aid in the assessment of ovarian reserve in women presenting to fertility clinics.  This system is intended to	Same

<b>Parameter</b>	<b>Access AMH Assay on Access 2 Immunoassay System Predicate – k170524</b>	<b>Access AMH Assay on Dxl 9000 Access Immunoassay Analyzer</b>
	distinguish between women presenting with AFC (antral follicle count) values >15 (high ovarian reserve) and women with AFC values ≤15 (normal or diminished ovarian reserve). The Access AMH is intended to be used in conjunction with other clinical and laboratory findings such as antral follicle count, before starting fertility therapy. The Access AMH is not intended to be used for monitoring of women undergoing controlled ovarian stimulation in an Assisted Reproduction Technology program.	
<b>Analyte Measured</b>	Anti-Müllerian hormone (AMH) levels in human serum and lithium heparin plasma	Same
<b>Technology</b>	Two-site immunoenzymatic (sandwich) assay	Same
<b>Format</b>	Chemiluminescent	Same
<b>Method</b>	Automated	Same
<b>Calibration</b>	Utilizes a stored calibration curve	Same
<b>Sample Type</b>	Serum or plasma	Same
<b>Reagent Pack formulation and packaging</b>	Access Reagent Pack formulation and packaging.	Same
<b>Measuring Range</b>	0.08-24 ng/mL (0.57-171 pmol/L).	Same
<b>Reagent Stability</b>	Unopened at 2° to 10°C until stated expiration date	Same
<b>Substrate</b>	Access Substrate	Lumi-Phos PRO Substrate
<b>Instrument</b>	Access 2 Immunoassay system	Dxl 9000 Access Immunoassay Analyzer

## Summary of Studies

**Method Comparison:** The results of the Access AMH assay on the Dxl 9000 are comparable to those of the Access AMH assay on Access 2. Further, the estimated bias at concentrations corresponding to reference limits defined on the predicate system suggest that such values have not changed appreciably on the Dxl 9000 system.

N	Concentration Range* (ng/mL)	Slope	Slope 95% CI	Intercept	Intercept 95% CI	Correlation Coefficient R
126	0.11 - 22	1.02	1.00 - 1.03	0.011	-0.013 - 0.067	1.00

\*Range is Access 2 values

**Imprecision:** The within-laboratory (total) % CV ranged from 2.2% to 5.4%, for AMH concentrations > 0.16 ng/mL. Access AMH Advanced assay meets total error and bias acceptance criteria with maximum Total Error (TE) of 16%.

ng/mL			Repeatability (Within-Run)		Between-Run		Between-Day		Within-Laboratory	
Sample	N	Mean	SD	%CV	SD	%CV	SD	%CV	SD	%CV
Sample 1	88	0.29	0.010	3.6	0.006	2.0	0.006	2.0	0.013	4.6
Sample 2	88	0.80	0.030	3.7	0.000	0.0	0.000	0.0	0.030	3.7
Sample 3	88	1.0	0.03	3.0	0.02	2.1	0.03	3.0	0.05	4.8
Sample 4	88	2.4	0.11	4.5	0.00	0.0	0.05	2.0	0.12	5.0
Sample 5	88	5.1	0.18	3.4	0.03	0.7	0.09	1.7	0.20	3.9
Sample 6	88	6.9	0.19	2.8	0.00	0.0	0.16	2.3	0.25	3.6
Sample 7	88	13	0.4	3.1	0.3	1.9	0.3	2.6	0.6	4.5
Sample 8	88	16	0.7	4.4	0.0	0.0	0.5	3.1	0.8	5.4
Sample 9	88	19	0.5	2.6	0.3	1.8	0.3	1.4	0.6	3.4

**Linearity:** The results of this study met the acceptance criterion and indicate that the Access AMH assay is linear on the Dxl 9000 Immunoassay System throughout the analytical measuring interval (0.08 - 24 ng/mL (0.57 - 171 pmol/L)).

**Limit of Blank (LoB):** The LoB for AMH was determined to be 0.001 ng/mL on the Dxl 9000 Immunoassay System. The results of this study demonstrate that the AMH assay met the acceptance criterion of  $\leq 0.01$  ng/mL.

**Limit of Detection (LoD):** The LoD estimate for the AMH assay is 0.002 ng/mL on the Dxl 9000 Immunoassay System. The results of this study demonstrate that the AMH assay met the acceptance criterion of  $\leq 0.02$  ng/mL on the Dxl 9000 Access Immunoassay System.

**Limit of Quantitation (LoQ):** The results provided demonstrate the 20% CV LoQ estimate for the Access AMH assay is 0.003 ng/mL (0.02 pmol/L). The results provided demonstrate the LoQ at 20% within-laboratory (total) CV estimate of the Access AMH assay to be  $\leq 0.08$  ng/mL (0.57 pmol/L).

**Reproducibility:** The study shows that the Access AMH assay meets design input requirements for reproducibility on Dxl 9000 with an SD  $\leq 0.042$  ng/mL for values  $\leq 0.16$  ng/mL and CV  $\leq 13.0\%$  for values  $> 0.16$  ng/mL.

**Conclusion:** The information provided in this submission demonstrates that the proposed new device is substantially equivalent to the predicate device.