



January 14, 2022

BRM Extremities Srl  
% Margeaux Rogers  
Associate Director, Regulatory Affairs  
MCRA, LLC  
1050 K Street NW, Suite 1000  
Washington, District of Columbia 20001

Re: K203773

Trade/Device Name: BRM TOOL Screws  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: HWC  
Dated: December 23, 2020  
Received: December 23, 2020

Dear Margeaux Rogers:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, MPH  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K203773

Device Name

BRM TOOL Screws

Indications for Use (Describe)

The BRM TOOL Screws are indicated for use for aligned bone fracture repair and arthrodesis, osteotomy, joint fusion and bone fragment fixation appropriate with the size of the screw.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

**Manufacturer:** BRM Extremities S.r.l.  
Registered Office: Via Lorenzo Mascheroni 29  
20145 Milano (MI) - Italy  
Headquarter: Via Papa Giovanni XXIII, 9  
23862 Civate (LC) – Italy

**Contact:** Ing. Andrea De Maglio  
Regulatory Affairs Manager  
BRM Extremities Srl  
Tel. +39 0341 1693087  
Cell. +39 389 1938034  
Email. a.demaglio@brm-extremities.com

**Prepared By:** MCRA, LLC  
703 8<sup>th</sup> Street NW  
Washington, DC 20001  
Phone: 202.552.5800

**Date Prepared:** January 13<sup>th</sup>, 2022

**Device Trade Name:** BRM TOOL Screws

**Device Common Name:** Smooth or Threaded Metallic Bone Fixation Fastener

**Classification:** 21 CFR 888.3040 – Screw, Fixation, Bone  
Class II

**Product Codes:** HWC

### Indications for Use:

The BRM TOOL Screws are indicated for use for aligned bone fracture repair and arthrodesis, osteotomy, joint fusion and bone fragment fixation appropriate with the size of the screw.

### Device Description:

The BRM TOOL Screws are interfragmentary compression osteosynthesis screws, i.e., devices used in surgical orthopedic interventions to provide support to the bone, ligament, tendon or cartilage structure, with the aim to reduce fractures in several parts of the skeleton, particularly of bone epiphyses. The application field excludes spinal, rib cage and skull bones. The BRM TOOL Screws are available in Titanium alloy.

**Predicate Devices:**

The BRM Screw Family of devices are substantially equivalent to the primary predicate OrthoSolutions System26 Cannulated Extremity Screws (K163489) and reference predicate OVERMED MINIARS Screw (K143596) with respect to intended use, indications for use, design and materials.

**Technological Comparison:**

Testing included torsional yield strength, driving torque and pullout strength testing. Results were determined to meet the pre-defined acceptance criteria per ASTM F543 and ASTM F1264. Additionally, the BRM Screw Family is in compliance with LAL testing requirements for orthopedic devices per AAMI ST-72 testing.

**Substantial Equivalence:**

The BRM Screw Family and the legally marketed predicate devices have the same intended use and indications for use, similar dimensions, geometry and materials. Additionally, information presented in the Mechanical Testing section demonstrate that the screws met the pre-defined acceptance criteria per ASTM F543 and ASTM F1264.