



November 12, 2020

Calvary Spine, LLC
% J.D. Webb
Official Correspondent
The OrthoMedix Group, Inc.
4313 W. 3800 S.
West Haven, Utah 84401

Re: K201568
Trade/Device Name: Calvary Spine Pedicle Screw System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral Pedicle Screw System
Regulatory Class: Class II
Product Code: NKB
Dated: October 12, 2020
Received: October 15, 2020

Dear J.D. Webb:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K201568

Device Name

Calvary Spine Pedicle Screw System

Indications for Use (Describe)

The Calvary Pedicle Screw System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of thoracic, lumbar, and sacral spine (T1 to S1/ilium): degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, deformities or curvatures (i.e. scoliosis, kyphosis, and/or lordosis), spinal stenosis, spinal tumor, pseudarthrosis and failed previous fusion.

When used for posterior, non-cervical (T1 to S1/ilium), pedicle screw fixation in pediatric patients, the Calvary Spine Pedicle Screw System implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e., scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the Calvary Spine Pedicle Screw System is intended to treat pediatric patients diagnosed with spondylolisthesis/spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion.

This system is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary: Calvary Spine Pedicle Screw System

In accordance with 21 CFR 807.92 of the Federal Code of Regulations

Date Prepared	September 14, 2020
Submitted By	Calvary Spine, LLC 308 North Wind Road Towson, MD 21204
Primary Contact	J.D. Webb 4313 W. 3800 S West Haven, UT 84401 512-590-5810 Tele e-mail: jdwebb@orthomedix.net
Trade Name	Calvary Spine Pedicle Screw System
Common Name	Thoracolumbosacral Pedicle Screw System
Classification Name	Thoracolumbosacral Pedicle Screw System
Class	II
Product Code	NKB
CFR Section	21 CFR section 888.3070
Device Panel	Orthopedic
Primary Predicate Device	S100 Pedicle Screw System, Renovis Surgical Technologies (K101682)
Reference Predicate Devices	EXPEDIUM®/Viper Spine System, Medos International SARL (K111136 / K131802 / K200245) Synergy VLS – open, Cross Medical (K940631 / K950099) PWB (now Synergy), Cross Medical (K920116) Moss Miami SS, DePuy Spine (K000536) Moss Miami Ti, DePuy Spine (K955348)
Device Description	Calvary Spine Pedicle Screw System includes screws and connecting components in a wide variety of sizes and shapes, which can be locked in various configurations, each assembly being tailor-made. Calvary pedicular screws must be used with Ø 5.5mm straight and curved rods.
Materials	Ti-6Al-4V ELI per ASTM F136 CP Titanium per ASTM F67 CoCrMo per ASTM F1582

Intended Use	The Calvary Spine Pedicle Screw System is a spinal fixation system intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine.
Substantial Equivalence Claimed to Predicate Devices	The Calvary Spine Pedicle Screw System is substantially equivalent to the predicate devices in terms of intended use, design, materials used, mechanical safety and performances.
Indications for Use	<p>The Calvary Spine Pedicle Screw System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of thoracic, lumbar, and sacral spine (T1 to S1/ilium): degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), degenerative spondylolisthesis with objective evidence of neurological impairment, fracture, dislocation, deformities or curvatures (i.e. scoliosis, kyphosis, and/or lordosis), spinal stenosis, spinal tumor, pseudarthrosis and failed previous fusion.</p> <p>When used for posterior, non-cervical (T1 to S1/ilium), pedicle screw fixation in pediatric patients, the Calvary Spine Pedicle Screw System implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e., scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the Calvary Spine Pedicle Screw System is intended to treat pediatric patients diagnosed with spondylolisthesis/ spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion.</p> <p>This system is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.</p>
Summary of the technological characteristics compared to predicate	<p><u>Intended Use</u></p> <p>The Calvary Spine Pedicle Screw System and the predicate device are spinal fixation systems intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine.</p> <p><u>Indications for Use</u></p> <p>All of the devices comply with the indications for use specified in 21 CFR section 888.3070 for thoracolumbosacral pedicle screw systems.</p> <p><u>Material</u></p> <p>The Calvary Spine Pedicle Screw System uses the same material as the predicate.</p> <p><u>Design</u></p> <p>The Calvary Spine Pedicle Screw System and the predicate are equivalent in terms of shape, sizes, material, and manufacturing process.</p> <p><u>Strength</u></p> <p>The Calvary Spine Pedicle Screw System has greater or equivalent strength values compared to other devices cleared for use in the thoracolumbosacral spine.</p>
Non-clinical Test Summary	<p>The following analyses were conducted:</p> <ul style="list-style-type: none"> • Static and dynamic compression per ASTM F1717 • Static torsion per ASTM F1717

	The results of these evaluations indicate that the Calvary Spine Pedicle Screw System is equivalent to predicate devices.
Clinical Test Summary	No clinical studies were performed
Conclusions: Non-clinical and Clinical	Calvary Spine, LLC considers the Calvary Spine Pedicle Screw System to be equivalent to the predicate devices listed above. This conclusion is based upon the devices' similarities in principles of operation, technology, materials and indications for use