



November 13, 2020

Paragon 28, Inc.
% Dave McGurl
Director, Regulatory Affairs
Musculoskeletal Clinical Regulatory Advisers, LLC
1050 K Street NW, Suite 1000
Washington, District of Columbia 20001

Re: K202373

Trade/Device Name: APEX 3D Total Ankle Replacement System
Regulation Number: 21 CFR 888.3110
Regulation Name: Ankle Joint Metal/Polymer Semi-Constrained Cemented Prosthesis
Regulatory Class: Class II
Product Code: HSN
Dated: September 18, 2020
Received: September 18, 2020

Dear Dave McGurl:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Ting Song, Ph.D., R.A.C.
Assistant Director
DHT6A: Division of Joint
Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K202373

Device Name

APEX 3D Total Ankle Replacement System

Indications for Use (Describe)

The APEX 3D Total Ankle Replacement System is indicated as a total ankle replacement in primary surgery for patients with ankle joints damaged by severe rheumatoid, post-traumatic, or degenerative arthritis. Revision surgery for these patients is also indicated for patients with sufficient bone stock present. Components are intended for cemented use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

510(k) Number: K202373

Manufacturer: Paragon 28, Inc.
14445 Grasslands Dr.,
Englewood, CO 80112

Contact: Ms. Haylie Hertz
Regulatory Affairs Specialist
Phone: 855-786-2828
hhertz@paragon28.com

Correspondent: Musculoskeletal Clinical Regulatory Advisers, LLC
1050 K Street, NW, Suite 1000
Washington, DC 20001
Phone: 202.552.5800

Date Prepared: August 21, 2020

Device Trade Name: APEX 3D Total Ankle Replacement System

Device Common Name: Total Ankle Prosthesis

Classification: 21 CFR 888.3110
Class II

Product Codes: HSN

Indications for Use:

The APEX 3D Total Ankle Replacement System is indicated as a total ankle replacement in primary surgery for patients with ankle joints damaged by severe rheumatoid, post-traumatic, or degenerative arthritis. Revision surgery for these patients is also indicated for patients with sufficient bone stock present. Components are intended for cemented use only.

Device Description:

The Paragon 28 Laser Alignment Guide is a line extension to the APEX 3D Total Ankle Replacement System instrumentation (K192994). The Paragon 28® APEX 3D™ Total Ankle

Replacement System Laser Alignment Guide is designed to provide a visual guide to verify alignment of the instrumentation associated with the APEX 3D™ Total Ankle Replacement System. Once inserted into the proper instrumentation, the Laser projects a line onto the patient to be used by surgeons as a visual reference for alignment. The Laser Alignment Guide is a single-use electrical device.

Predicate Devices:

The Paragon 28 Laser Alignment Guide is substantially equivalent to the APEX 3D Total Ankle Replacement System Instrumentation (K192994) and the Corin Optimized Positioning System (K152893). The OEC Elite (K192819) is considered a reference device.

Substantial Equivalence:

The subject Laser Alignment Guide is substantially equivalent to the predicate systems with respect to intended use, indications, and design. The subject Laser Alignment Guide possesses similar technological characteristics as the predicate devices, including design, material, chemical composition, and energy source. Non-clinical testing performed included a cadaveric design evaluation of the Laser line, a stability assessment, and electrical safety testing. The results of this testing demonstrate that the subject system performs as intended.

Conclusion:

The APEX 3D Total Ankle Replacement System Laser Alignment Guide possesses the same intended use and technological characteristics as the predicate devices. Therefore, the Laser Alignment Guide is substantially equivalent for its intended use.