



November 5, 2020

Spine Wave, Inc.  
Amy Noccioli  
Sr. Regulatory Affairs Specialist  
Three Enterprise Drive, Suite 210  
Shelton, Connecticut 06484

Re: K202476

Trade/Device Name: Salvo® Spine System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Thoracolumbosacral Pedicle Screw System  
Regulatory Class: Class II  
Product Code: NKB  
Dated: October 26, 2020  
Received: October 27, 2020

Dear Amy Noccioli:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.  
Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K202476

Device Name

Salvo® Spine System

Indications for Use (Describe)

The Salvo® Spine System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease; spondylolisthesis; fracture; dislocation; scoliosis; kyphosis; spinal tumor; and failed previous fusion (pseudarthrosis).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**510(k) Summary  
Salvo<sup>®</sup> Spine System**

**1. Submitter Information**

*Submitter:* Spine Wave, Inc.  
*Address:* Three Enterprise Drive  
Suite 210  
Shelton, CT 06484  
*Telephone:* 203-712-1842  
*Fax:* 203-944-9493  
  
*Contact:* Amy Nocchioli  
*Date Prepared:* August 27, 2020

**2. Device Information**

*Trade Name:* Salvo<sup>®</sup> Spine System  
*Common Name:* Thoracolumbosacral Pedicle Screw System  
*Classification:* Class II per 21 CFR 888.3070  
*Classification Name:* Thoracolumbosacral Pedicle Screw System  
*Product Code:* NKB

**3. Purpose of Submission**

The purpose of this submission is to gain clearance for additional yokes, rods, and connectors to the Salvo<sup>®</sup> Spine System.

**4. Predicate Device Information**

The Salvo<sup>®</sup> Spine System described in this submission is substantially equivalent to the following predicate:

<b>Primary Predicate Device</b>	<b>Manufacturer</b>	<b>510(k) No.</b>
Salvo <sup>®</sup> Spine System	Spine Wave, Inc.	K191045

  

<b>Reference Predicate Device</b>	<b>Manufacturer</b>	<b>510k No.</b>
CapSure <sup>®</sup> PS Spine System	Spine Wave, Inc.	K172175

## **5. Device Description**

The Salvo<sup>®</sup> Spine System is a thoracolumbosacral pedicle screw system designed to be implanted through a posterior surgical approach. The implants are manufactured from titanium alloy (Ti-6Al-4V ELI) per ASTM F136, commercially pure titanium per ASTM F67, cobalt-chromium alloy per ASTM F1537, and PEEK-OPTIMA LT1 per ASTM F2026. The Salvo<sup>®</sup> Spine System consists of a variety of screws, rods, and connectors that create a rigid construct as an adjunct to fusion for internal fixation and stabilization of the thoracic, lumbar, and sacral spine.

## **6. Indications for Use**

The Salvo<sup>®</sup> Spine System is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the thoracic, lumbar, and sacral spine: degenerative disc disease; spondylolisthesis; fracture; dislocation; scoliosis; kyphosis; spinal tumor; and failed previous fusion (pseudarthrosis).

## **7. Comparison of Technological Characteristics**

The Salvo<sup>®</sup> Spine System has technological characteristics equivalent to those of the predicate device, including intended use, performance, design, and material composition.

## **8. Performance Data**

Nonclinical testing was performed on the Salvo<sup>®</sup> Spine System to support substantial equivalence to the predicate device. The following testing was performed:

- Static and dynamic axial compression testing per ASTM F1717
- Static torsion testing per ASTM F1717
- Static pull-apart testing per ASTM F1798
- Axial and torsional grip testing per ASTM F1798

## **9. Conclusion**

The indications for use, technological characteristics, and performance testing show that the Salvo<sup>®</sup> Spine System is substantially equivalent to the predicate device and does not present any new issues of safety or effectiveness.