



May 27, 2022

In2Bones USA, LLC
Christine Scifert
VP of Quality and Regulatory
6000 Poplar Ave, Suite 115
Memphis, Tennessee 38119

Re: K213698

Trade/Device Name: CoLink® PCR Plating System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: HRS, HWC
Dated: April 26, 2022
Received: April 28, 2022

Dear Christine Scifert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K213698

Device Name
CoLink® PCR Plating System

Indications for Use (Describe)

The CoLink® PCR Plating System is indicated for stabilization and fixation of fractures, revision procedures, joint fusion, osteotomies and reconstruction of the small bones in the hand, wrist, foot and ankle.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary
CoLink® PCR Plating System
May 25, 2022

Company: In2Bones USA, LLC
6000 Poplar Ave, Suite 115
Memphis, TN 38119
901-260-7931

Primary Contact: Christine Scifert

Trade Name: CoLink® PCR Plating System

Common Name: Plate, Fixation, Bone
Screw, Fixation, Bone

Classification: II

Regulation Number: 888.3030 - Single/multiple component metallic bone fixation appliances
and accessories (Primary)
888.3040 - Smooth or threaded metallic bone fixation fastener

Panel: 87-Orthopedic

Product Code(s): HRS, HWC

Device Description: The In2Bones CoLink® PCR Plating System includes plates and screws and surgical instruments intended for treating fusions and fractures of the extremities. These subject devices are part of the overarching CoLink® Plating System and will be commonly referred to as the CoLink® PCR Plating System. The CoLink® PCR Plating System consists of the MTP Narrow NX and Universal style plates available in Carbon Fiber Reinforced PEEK-OPTIMA™ with Barium Sulfate. The plates are used with previously cleared 3.0mm and 3.5mm titanium alloy locking and non-locking screws.

Indications for Use: The CoLink® PCR Plating System is indicated for stabilization and fixation of fractures, revision procedures, joint fusion, osteotomies and reconstruction of the small bones in the hand, wrist, foot and ankle.

Substantial Equivalence: The subject components were demonstrated to be substantially equivalent to the following systems previously cleared by the FDA:

Primary Predicate

- K163293 – In2Bones CoLink® Plating System

Additional Predicate

- K170401 – CarboFix Orthopedics Piccolo Composite Plate System

The subject CoLink® PCR Plating System plates are similar in geometry to the previously cleared CoLink® Plating System and is made of Carbon Fiber Reinforced PEEK-OPTIMA™ with Barium Sulfate. The CoLink® PCR Plating System has been demonstrated to be substantially equivalent to the previously cleared devices identified above as the products are similar in indications, materials and geometry.

Performance Testing: The CoLink® PCR Plating System was tested in static and dynamic four-point bending per ASTM F382 to show the subject plate is substantially equivalent to the predicate plates. The 3.0mm and 3.5mm screws are identical to previously cleared screws and no new worst-case plate or screws were added. The CoLink® PCR Plating System was validated per ISO 10993-1 for biocompatibility, ISO 11137-2 for gamma sterilization and ISO 11607-1 for packaging. Endotoxin testing was completed per ANSI/AAMI ST72.

Conclusion

Based on the test results and the comparison to the predicate devices, the subject device is determined to be substantially equivalent to the predicate devices.