



CoV-2 IgM  
6R87

H14977R02

# AdviseDx SARS-CoV-2 IgM

FOR USE I

ARCHITECT

Read Highlighted Changes: Revised September 2021.

REF 6R87-20

REF 6R87-30

**For use under an Emergency Use Authorization (EUA) Only**  
Prescription Use only.

For *In Vitro* Diagnostic Use Only.

Instructions must be carefully followed. Reliability of assay results cannot be guaranteed if there are any deviations from these instructions.

For laboratory professional use only.

## NAME

AdviseDx SARS-CoV-2 IgM (also referred to as CoV-2 IgM or SARS-CoV-2 IgM on the reagent bottle label)

## INTENDED USE

The AdviseDx SARS-CoV-2 IgM assay is a chemiluminescent microparticle immunoassay (CMIA) intended for the qualitative detection of IgM antibodies to SARS-CoV-2 in human serum, serum separator tube, and plasma (dipotassium EDTA, tripotassium EDTA, lithium heparin, lithium heparin separator tube, sodium heparin). The AdviseDx SARS-CoV-2 IgM assay is intended for use as an aid in identifying individuals with an adaptive immune response to SARS-CoV-2, indicating recent or prior infection. At this time, it is unknown for how long antibodies persist following infection and if the presence of antibodies confers protective immunity. The AdviseDx SARS-CoV-2 IgM assay should not be used to diagnose acute SARS-CoV-2 infection.

Testing is limited to laboratories certified under the Clinical Laboratory Improvement Amendments (CLIA) of 1988, 42 U.S.C 263a, that meet requirements to perform moderate or high complexity tests.

Results are for the detection of SARS CoV-2 antibodies. IgM antibodies to SARS-CoV-2 are generally detectable in blood several days after initial infection, although the duration of time antibodies are present post-infection is not well characterized. Individuals may have detectable virus present for several weeks following seroconversion.

Laboratories within the United States and its territories are required to report all results to the appropriate public health authorities.

The sensitivity of the AdviseDx SARS-CoV-2 IgM assay early after infection is unknown. Negative results do not preclude acute SARS-CoV-2 infection. If acute infection is suspected, direct testing for SARS-CoV-2 is necessary.

False positive results for the AdviseDx SARS-CoV-2 IgM assay may occur due to cross-reactivity from pre-existing antibodies or other possible causes. Due to the risk of false positive results, confirmation of positive results should be considered using a second, different IgM assay.

Samples should only be tested from individuals with 15 days to 30 days post symptom onset. SARS-CoV-2 antibody negative samples collected 15 days or more post symptom onset should be reflexed to a test that detects and reports SARS-CoV-2 IgG.

**The AdviseDx SARS-CoV-2 IgM assay is only for use under the Food and Drug Administration's Emergency Use Authorization.**

## SUMMARY AND EXPLANATION OF THE TEST

The AdviseDx SARS-CoV-2 IgM assay is designed to detect immunoglobulin class M (IgM) antibodies to the spike protein of SARS-CoV-2 in serum and plasma from individuals who are suspected to have had coronavirus disease (COVID-19) or in serum and plasma of subjects that may have been infected by SARS-CoV-2.

COVID-19 is defined as illness caused by a novel coronavirus now called severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2, formerly called 2019-nCoV).<sup>1</sup> On March 11, 2020, the World Health Organization (WHO) declared COVID-19 a global pandemic.<sup>2</sup> The incubation period of COVID-19 ranges between 1 and 14 days, with the majority of cases manifesting within 3 to 5 days. The most common symptoms of COVID-19 are fever, tiredness, dry cough, and difficulty breathing. A severe acute respiratory distress syndrome (ARDS) may develop.<sup>3</sup> Reported case fatality rates depend on geographic location,<sup>4</sup> age, and comorbidities.

The causative agent of COVID-19 is a beta coronavirus and belongs to a family of viruses that may cause respiratory symptoms ranging from common cold to severe pneumonia. These viruses are common in animals worldwide and may eventually transfer to humans, as has likely happened with SARS-CoV-2.<sup>1</sup>

The host immune system reacts to the infection by SARS-CoV-2 by producing specific antibodies. These antibodies have been reported to appear in serum or plasma of infected individuals after the detection of viral ribonucleic acid (RNA) in swabs<sup>5</sup> in as early as a few days to 2 weeks after the onset of symptoms.<sup>6</sup> Specific IgM antibodies to SARS-CoV-2 may be detectable in COVID-19 patients during the symptomatic phase of the disease after RNA is no longer detectable.<sup>5, 6</sup> The persistence of IgM antibodies allows identification of people who have been recently infected and evaluation of disease courses including recovery from the illness.<sup>7</sup> It is unknown if antibodies to SARS-CoV-2 confer immunity to infection. SARS-CoV-2 IgM, together with IgG serological assays, will likely play an important role in research and surveillance.<sup>8</sup>

## BIOLOGICAL PRINCIPLES OF THE PROCEDURE

This assay is an automated, two-step immunoassay for the qualitative detection of IgM antibodies to SARS-CoV-2 in human serum and plasma using chemiluminescent microparticle immunoassay (CMIA) technology.

Sample, SARS-CoV-2 antigen coated paramagnetic microparticles, and assay diluent are combined and incubated. The IgM antibodies to SARS-CoV-2 present in the sample bind to the SARS-CoV-2 antigen coated microparticles. The mixture is washed. Anti-human IgM acridinium-labeled conjugate is added to create a reaction mixture and incubated. Following a wash cycle, Pre-Trigger and Trigger Solutions are added.

The resulting chemiluminescent reaction is measured as a relative light unit (RLU).

The presence or absence of IgM antibodies to SARS-CoV-2 in the sample is determined by comparing the chemiluminescent RLU in the reaction to the calibrator RLU, which is calculated by the system as an Index (S/C).

For additional information on system and assay technology, refer to the ARCHITECT System Operations Manual, Section 3.

## REAGENTS

### Kit Contents

AdviseDx SARS-CoV-2 IgM Reagent Kit 6R87

NOTE: Some kit sizes may not be available for use on all ARCHITECT i Systems. Please contact your local distributor.

Volumes (mL) listed in the following table indicate the volume per bottle.

REF	6R87-20	6R87-30
Tests per kit	100	500
Number of kits per box	1	1
Tests per box	100	500
<b>M I</b>	6.6 mL	27.0 mL
<b>CONJUGATE</b>	5.9 mL	26.3 mL
<b>ASSAY I</b>	11.4 mL	40.7 mL

**M I** Purified SARS-CoV-2 recombinant antigen coated microparticles in TRIS buffer with surfactant. Minimum concentration: 0.0675% solids. Preservatives: ProClin 950 and sodium azide.

**CONJUGATE** Anti-human IgM (mouse, monoclonal) acridinium-labeled conjugate in MES buffer with surfactants and protein (bovine) stabilizer. Minimum concentration: 20 ng/mL. Preservatives: ProClin 300 and antimicrobial agent.

**ASSAY I** TRIS buffer and detergent. Preservatives: ProClin 300, sodium azide, and antimicrobial agents.

### Warnings and Precautions

#### For Use Under An Emergency Use Authorization Only.

This assay is only for *in vitro* diagnostic use under the FDA Emergency Use Authorization.

- IVD**
- For *In Vitro* Diagnostic Use
- Rx ONLY**
- This product has not been FDA cleared or approved but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories.
- This product has been authorized only for detecting the presence of IgM antibodies to SARS-CoV-2, not for any other viruses or pathogens.
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

#### Safety Precautions

**CAUTION:** This product requires the handling of human specimens. It is recommended that all human-sourced materials and all consumables contaminated with potentially infectious materials be considered potentially infectious and handled in accordance with the OSHA Standard on Bloodborne Pathogens. Biosafety Level 2 or other appropriate regional, national, and institutional biosafety practices should be used for materials that contain, are suspected of containing, or are contaminated with infectious agents.<sup>9-12</sup>

The following warnings and precautions apply to: <b>M I</b>	
<b>WARNING</b>	Contains methylisothiazolone and sodium azide.
H317	May cause an allergic skin reaction.
EUH032	Contact with acids liberates very toxic gas.
<b>Prevention</b>	
P261	Avoid breathing mist / vapors / spray.
P272	Contaminated work clothing should not be allowed out of the workplace.
P280	Wear protective gloves / protective clothing / eye protection.
<b>Response</b>	
P302+P352	IF ON SKIN: Wash with plenty of water.
P333+P313	If skin irritation or rash occurs: Get medical advice / attention.
P362+P364	Take off contaminated clothing and wash it before reuse.
<b>Disposal</b>	
P501	Dispose of contents / container in accordance with local regulations.

The following warnings and precautions apply to: <b>CONJUGATE</b>	
	
<b>WARNING</b>	Contains methylisothiazolones.
H317	May cause an allergic skin reaction.
H402	Harmful to aquatic life.
H412	Harmful to aquatic life with long lasting effects.
<b>Prevention</b>	
P261	Avoid breathing mist / vapors / spray.
P272	Contaminated work clothing should not be allowed out of the workplace.
P273	Avoid release to the environment.
P280	Wear protective gloves / protective clothing / eye protection.
<b>Response</b>	
P302+P352	IF ON SKIN: Wash with plenty of water.
P333+P313	If skin irritation or rash occurs: Get medical advice / attention.
P362+P364	Take off contaminated clothing and wash it before reuse.
<b>Disposal</b>	
P501	Dispose of contents / container in accordance with local regulations.

The following warnings and precautions apply to: <b>ASSAY I</b>	
<b>WARNING</b>	Contains alcohols, C11-C15-secondary, ethoxylated, methylisothiazolones and sodium azide.
H319	Causes serious eye irritation.
H317	May cause an allergic skin reaction.
H402	Harmful to aquatic life.
H412	Harmful to aquatic life with long lasting effects.
EUH032	Contact with acids liberates very toxic gas.

Prevention	
P261	Avoid breathing mist / vapors / spray.
P264	Wash hands thoroughly after handling.
P272	Contaminated work clothing should not be allowed out of the workplace.
P280	Wear protective gloves / protective clothing / eye protection.
P273	Avoid release to the environment.
Response	
P305+P351+P338	IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
P337+P313	If eye irritation persists: Get medical advice / attention.
P302+P352	IF ON SKIN: Wash with plenty of water.
P333+P313	If skin irritation or rash occurs: Get medical advice / attention.
P362+P364	Take off contaminated clothing and wash it before reuse.
Disposal	
P501	Dispose of contents / container in accordance with local regulations.

Follow local chemical disposal regulations based on your location along with recommendations and content in the Safety Data Sheet to determine the safe disposal of this product.

For the most current hazard information, see the product Safety Data Sheet.

Safety Data Sheets are available at [www.corelaboratory.abbott](http://www.corelaboratory.abbott) or contact your local representative.

For a detailed discussion of safety precautions during system operation, refer to the ARCHITECT System Operations Manual, Section 8.

#### Reagent Handling

- Reagents are shipped on wet ice.
- Do not pool reagents within a kit or between kits.
- Before loading the reagent kit on the system for the first time, the microparticle bottle requires mixing to resuspend microparticles that may have settled during shipment. For microparticle mixing instructions, refer to the PROCEDURE, Assay Procedure section of this package insert.
- **Septums MUST be used to prevent reagent evaporation and contamination and to ensure reagent integrity. Reliability of assay results cannot be guaranteed if septums are not used according to the instructions in this package insert.**
  - To avoid contamination, wear clean gloves when placing a septum on an uncapped reagent bottle.
  - Once a septum has been placed on an open reagent bottle, do not invert the bottle as this will result in reagent leakage and may compromise assay results.
  - Over time, residual liquids may dry on the septum surface. These are typically dried salts and have no effect on assay efficacy.
- When handling conjugate vials, change gloves that have contacted human serum or plasma, since introduction of human IgM will result in a neutralized conjugate.

For a detailed discussion of reagent handling precautions during system operation, refer to the ARCHITECT System Operations Manual, Section 7.

#### Reagent Storage

- Do not freeze.
- Protect from light.
- Reagents should not be used beyond their expiration date.

	Storage Temperature	Maximum Storage Time	Additional Storage Instructions
<b>Unopened</b>	2 to 8°C	Until expiration date	Store in upright position.
<b>Onboard</b>	System Temperature	30 days	
<b>Opened</b>	2 to 8°C	Until expiration date	Store in upright position. If the microparticle bottle does not remain upright (with a septum installed) while in refrigerated storage off the system, the reagent kit must be discarded.

Reagents may be stored on or off the ARCHITECT i System. If reagents are removed from the system, store them at 2 to 8°C (with septums and replacement caps) in an upright position. For reagents stored off the system, it is recommended that they be stored in their original trays and boxes to ensure they remain upright.

For information on unloading reagents, refer to the ARCHITECT System Operations Manual, Section 5.

#### Indications of Reagent Deterioration

Deterioration of the reagents may be indicated when a calibration error occurs or a control value is out of the specified range.

Associated test results are invalid, and samples must be retested. Assay recalibration may be necessary.

For troubleshooting information, refer to the ARCHITECT System Operations Manual, Section 10.

#### INSTRUMENT PROCEDURE

The AdviseDx SARS-CoV-2 IgM assay file must be installed on the ARCHITECT i System prior to performing the assay.

For detailed information on assay file installation and viewing and editing assay parameters, refer to the ARCHITECT System Operations Manual, Section 2.

For information on printing assay parameters, refer to the ARCHITECT System Operations Manual, Section 5.

For a detailed description of system procedures, refer to the ARCHITECT System Operations Manual.

#### SPECIMEN COLLECTION AND PREPARATION FOR ANALYSIS

##### Specimen Types

The specimen types listed below may be used with this assay.

Specimen Types	Collection Tubes
Serum	Serum
	Serum separator
Plasma	Dipotassium EDTA
	Tripotassium EDTA
	Lithium heparin
	Lithium heparin separator
	Sodium heparin

- Each laboratory is responsible for following their own procedures to establish the use of additional tube or collection types.
- Performance has not been established for the use of cadaveric specimens or the use of bodily fluids other than human serum/plasma.
- Liquid anticoagulants may have a dilution effect resulting in lower Index (S/C) values for individual specimens.

The instrument does not provide the capability to verify specimen types. It is the responsibility of the operator to verify that the correct specimen types are used in the assay.

##### Specimen Conditions

- Do not use:
  - heat-inactivated specimens

- pooled specimens
- grossly hemolyzed specimens
- specimens with obvious microbial contamination
- specimens with fungal growth
- For accurate results, serum and plasma specimens should be free of fibrin, red blood cells, and other particulate matter. Serum specimens from patients receiving anticoagulant or thrombolytic therapy may contain fibrin due to incomplete clot formation.
- To prevent cross contamination, use of disposable pipettes or pipette tips is recommended.

### Preparation for Analysis

- Follow the tube manufacturer's processing instructions for collection tubes. Gravity separation is not sufficient for specimen preparation.
- Specimens should be free of bubbles. Remove bubbles with an applicator stick before analysis. Use a new applicator stick for each specimen to prevent cross contamination.

To ensure consistency in results, recentrifuge specimens prior to testing if

- they contain fibrin, red blood cells, or other particulate matter.

NOTE: If fibrin, red blood cells, or other particulate matter are observed, mix by low speed vortex or by inverting 10 times prior to recentrifugation.

Prepare frozen specimens as follows:

- Frozen specimens must be completely thawed before mixing.
- Mix thawed specimens thoroughly by low speed vortex or by inverting 10 times.
- Visually inspect the specimens. If layering or stratification is observed, mix until specimens are visibly homogeneous.
- If specimens are not mixed thoroughly, inconsistent results may be obtained.
- Recentrifuge specimens that contain particulate matter.

Recentrifugation of Specimens

- Transfer specimens to a centrifuge tube and centrifuge.
- Transfer clarified specimen to a sample cup or secondary tube for testing. For centrifuged specimens with a lipid layer, transfer only the clarified specimen and not the lipemic material.

### Specimen Storage

Specimen Type	Temperature	Maximum Storage Time	Special Instructions
Serum/Plasma	Room temperature (15 to 30°C)	2 days	Specimens may be stored on or off the clot, red blood cells, or separator gel.
	2 to 8°C	7 days	Specimens may be stored on or off the clot, red blood cells, or separator gel.

If testing will be delayed longer than 7 days at 2 to 8°C storage time, remove serum or plasma from the clot, red blood cells, or separator gel and store frozen (-20°C or colder).

It is the responsibility of the individual laboratory to determine specific specimen stability criteria for their laboratory per their laboratory workflow.

For additional information on sample handling and processing, refer to CLSI GP44-A4.<sup>13</sup> The storage information provided here is based on data maintained by the manufacturer.

Frozen specimens subjected to up to 2 freeze/thaw cycles have been evaluated.

### Specimen Shipping

Package and label specimens in compliance with applicable state, federal, and international regulations covering the transport of clinical specimens and infectious substances.

Do not exceed the storage limitations listed above.

## PROCEDURE

### Materials Provided

6R87 AdviseDx SARS-CoV-2 IgM Reagent Kit

### Materials Required but not Provided

- AdviseDx SARS-CoV-2 IgM assay file found on [www.corelaboratory.abbott](http://www.corelaboratory.abbott)
- 6R87-01 AdviseDx SARS-CoV-2 IgM Calibrator Kit
- 6R87-10 AdviseDx SARS-CoV-2 IgM Control Kit
- ARCHITECT Pre-Trigger Solution
- ARCHITECT Trigger Solution
- ARCHITECT Wash Buffer
- ARCHITECT Septum

For information on materials required for operation of the instrument, refer to the ARCHITECT System Operations Manual, Section 1.

For information on materials required for maintenance procedures, refer to the ARCHITECT System Operations Manual, Section 9.

### Assay Procedure

For a detailed description of how to run an assay, refer to the ARCHITECT System Operations Manual, Section 5.

- If using primary or aliquot tubes, refer to the ARCHITECT System Operations Manual, Section 5 to ensure sufficient specimen is present.
- Minimum sample cup volume is calculated by the system and printed on the Order List report. To minimize the effects of evaporation, verify adequate sample cup volume is present prior to running the test.
- Before loading the reagent kit on the system for the first time, the microparticle bottle requires mixing to resuspend microparticles that may have settled during shipment. After the first time the microparticles have been loaded, no further mixing is required.
  - **Invert the microparticle bottle 30 times.**
  - Visually inspect the bottle to ensure microparticles are resuspended. If microparticles are still adhered to the bottle, continue to invert the bottle until the microparticles have been completely resuspended.
  - **If the microparticles do not resuspend, DO NOT USE. Contact your local Abbott representative.**
  - Once the microparticles have been resuspended, place a septum on the bottle. For instructions about placing septums on bottles, refer to the Reagent Handling section of this package insert.
- Maximum number of replicates sampled from the same sample cup: 10
  - Priority:
    - Sample volume for first test: 75 µL
    - Sample volume for each additional test from same sample cup: 25 µL
  - ≤ 3 hours on the reagent and sample manager:
    - Sample volume for first test: 150 µL
    - Sample volume for each additional test from same sample cup: 25 µL
  - > 3 hours on the reagent and sample manager:
    - Replace with a fresh aliquot of sample.
- Refer to the AdviseDx SARS-CoV-2 IgM calibrator package insert [\[REF\]](#) 6R87-01 and/or AdviseDx SARS-CoV-2 IgM control package insert [\[REF\]](#) 6R87-10 for preparation and usage.
- For general operating procedures, refer to the ARCHITECT System Operations Manual, Section 5.
- For optimal performance, it is important to perform routine maintenance as described in the ARCHITECT System Operations Manual, Section 9. Perform maintenance more frequently when required by laboratory procedures.

## Sample Dilution Procedures

Dilution of samples for the AdviseDx SARS-CoV-2 IgM assay has not been verified.

## Calibration

For instructions on performing a calibration, refer to the ARCHITECT System Operations Manual, Section 6.

Calibrator is tested in triplicate.

A single sample of each control level must be tested to evaluate the assay using the ratio of the sample RLU to the cutoff RLU (S/C) for assay calibration.

Ensure that assay control values are within the S/C ranges specified in the control package insert.

Each assay control must be tested to evaluate the assay calibration. Once a calibration is accepted and stored, it may be used for 30 days. During this time, all subsequent samples may be tested without further calibration unless:

- A reagent kit with a new lot number is used.
- Daily quality control results are outside of quality control limits used to monitor and control system performance.

This assay may require recalibration after maintenance to critical parts or subsystems or after service procedures have been performed.

## Quality Control Procedures

The recommended control requirement for the AdviseDx SARS-CoV-2 IgM assay is that a single sample of each control level be tested once every 24 hours each day of use.

Additional controls may be tested in accordance with local, state, and/or federal regulations or accreditation requirements and your laboratory's quality control policy.

To establish statistically-based control limits, each laboratory should establish its own concentration target and ranges for new control lots at each clinically relevant control level. This can be accomplished by assaying a minimum of 20 replicates over several (3-5) days and using the reported results to establish the expected average (target) and variability about this average (range) for the laboratory. Sources of variation that should be included in this study in order to be representative of future system performance include:

- Multiple stored calibrations
- Multiple reagent lots
- Multiple calibrator lots
- Multiple processing modules (if applicable)
- Data points collected at different times of the day

Refer to published guidelines for information or general control recommendation, for example Clinical and Laboratory Standards Institute (CLSI) Guideline C24, 4th ed., or other published guidelines, for general quality control recommendations.<sup>14</sup>

- If more frequent control monitoring is required, follow the established quality control procedures for your laboratory.
- If quality control results do not meet the acceptance criteria defined by your laboratory, sample results may be suspect. Follow the established quality control procedures for your laboratory. Recalibration may be necessary. For troubleshooting information, refer to the ARCHITECT System Operations Manual, Section 10.
- Review quality control results and acceptance criteria following a change of reagent or calibrator lot.

Controls should be used according to the guidelines and recommendations of the control manufacturer. Concentration ranges provided in the control package insert should be used only for guidance.

For any control material in use, the laboratory should ensure that the matrix of the control material is suitable for use in the assay per the assay package insert.

## Quality Control Guidance

Refer to "Basic QC Practices" by James O Westgard, Ph.D. for guidance on laboratory quality control practices.<sup>15</sup>

## Verification of Assay Claims

To verify package insert claims, follow CLIA recommendations or internal laboratory procedures.

For protocols to verify package insert claims, refer to the ARCHITECT System Operations Manual, Appendix B.

## RESULTS

### Calculation

The ARCHITECT i System calculates the calibrator mean chemiluminescent signal from 3 calibrator replicates and stores the result. Results are reported by dividing the sample result by the stored calibrator result. The default result unit for the AdviseDx SARS-CoV-2 IgM assay is Index (S/C).

### Interpretation of Results

The cutoff is 1.00 Index (S/C).

As with all analyte determinations, the result should be used in conjunction with information available from clinical evaluation and other diagnostic procedures.

Index (S/C)	Interpretation
< 1.00	Negative
≥ 1.00	Positive

### Flags

Some results may contain information in the Flags field. For a description of the flags that may appear in this field, refer to the ARCHITECT System Operations Manual, Section 5.

## LIMITATIONS OF THE PROCEDURE

- **For use under an Emergency Use Authorization only.**
- This assay is for *in vitro* diagnostic use under FDA Emergency Use Authorization only.
- This assay is for clinical laboratory use only. It is not for home use.
- Results should be used in conjunction with other data; e.g., symptoms, results of other tests, and clinical impressions.
- This assay must not be used for testing samples from individuals who are immunocompromised. Immunocompromised individuals who have COVID-19 may have a delayed immune response and produce levels of antibody that may not be detected as positive by this assay.
- This assay should be used for testing samples collected 15 to 30 days after symptom onset.
- **It is not known at this time if the presence of antibodies to SARS-CoV-2 confers immunity.**
- A positive result may not indicate previous SARS-CoV-2 infection. Consider other information including clinical history and local disease prevalence, in assessing the need for a second but different serology test to confirm an immune response.
- Negative results do not rule out SARS-CoV-2 infection, particularly in those who have been in contact with the virus. Testing with a molecular diagnostic should be considered to evaluate for active infection in symptomatic individuals.
- Results from antibody testing should not be used to diagnose or exclude acute SARS-CoV-2 infection or to inform infection status.
- Pedigreed specimens with direct evidence of antibodies to non-SARS-CoV-2 coronavirus (common cold) strains such as HKU1, NL63, OC43, or 229E have not been evaluated with this assay.
- Not to be used to screen units of blood for SARS-CoV-2 infection.
- Potentially interfering disease states and other cross reactants have been evaluated and are represented in the SPECIFIC PERFORMANCE CHARACTERISTICS section of this package insert.

- Specimens from patients who have received preparations of mouse monoclonal antibodies for diagnosis or therapy may contain human anti-mouse antibodies (HAMA). Such specimens may show either falsely elevated or depressed values when tested with assay kits such as AdviseDx SARS-CoV-2 IgM that employ mouse monoclonal antibodies.<sup>16, 17</sup>
- Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with *in vitro* immunoassays. Patients routinely exposed to animals or to animal serum products can be prone to this interference, and anomalous values may be observed.<sup>18</sup>
- Rheumatoid factor (RF) in human serum can react with reagent immunoglobulins, interfering with *in vitro* immunoassays.<sup>18</sup>
- Specimens from patients who underwent hemodialysis may have autoantibodies in circulation that potentially interfere with *in vitro* immunoassays.<sup>19</sup>
- The performance of this device has not been established in individuals that have received a COVID-19 vaccine. The clinical significance of a positive or negative antibody result following COVID-19 vaccination has not been established, and the result from this assay should not be interpreted as an indication or degree of protection from infection after vaccination.
- The performance of this test was established based on the evaluation of a limited number of clinical specimens collected between March 2020 and May 2020, in the United States and Spain. The clinical performance has not been established in all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV-2 and their prevalence, which change over time.

## CONDITIONS OF AUTHORIZATIONS FOR THE LABORATORIES

The AdviseDx SARS-CoV-2 IgM Letter of Authorization, along with the authorized Fact Sheet for Healthcare Providers, the authorized Fact Sheet for Recipients, and authorized labeling are available on the FDA website: <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas> or at <https://www.corelaboratory.abbott/us/en/offerings/segments/infectious-disease/sars-cov-2>.

However, to assist clinical laboratories using the AdviseDx SARS-CoV-2 IgM assay (“your product” in the conditions below), the relevant Conditions of Authorization are listed below:

## SPECIFIC PERFORMANCE CHARACTERISTICS

Representative performance data are provided in this section. Results obtained in individual laboratories may vary.

### Precision

#### Within-Laboratory Precision

A study was performed based on guidance from CLSI EP05-A3.<sup>20</sup> Testing was conducted using 1 lot of the AdviseDx SARS-CoV-2 IgM Reagent Kit, 1 lot of the AdviseDx SARS-CoV-2 IgM Calibrator Kit, and 1 lot of the AdviseDx SARS-CoV-2 IgM Control Kit and 1 ARCHITECT i2000SR instrument. Two controls and 2 human plasma panels were assayed in replicates of 3, at 2 separate times per day, on 20 different days, for a total of 120 replicates for each test sample.

Sample	n	Mean (Index [S/C])	Repeatability (Within-Run)		Between-Run		Between-Day		Within-Laboratory <sup>a</sup>		
			SD	%CV	SD	%CV	SD	%CV	SD	SD 95% CI	%CV
Negative Panel	120	0.73	0.023	N/A	0.000	N/A	0.000	N/A	0.023	(0.020, 0.026)	N/A
Positive Panel	120	1.96	0.047	2.4	0.025	1.3	0.000	0.0	0.054	(0.047, 0.062)	2.7
Negative Control	120	0.02	0.001	N/A	0.000	N/A	0.000	N/A	0.001	(0.001, 0.001)	N/A
Positive Control	120	2.77	0.068	2.5	0.036	1.3	0.014	0.5	0.078	(0.069, 0.090)	2.8

<sup>a</sup> Includes repeatability (within-run), between-run, and between-day variability.

N/A = not applicable

- A. Authorized laboratories\* using your product **must** include with result reports, all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- B. Authorized laboratories **must use** your product as outlined in the **authorized labeling**. Deviations from the authorized procedures, including the authorized instruments, authorized clinical specimen types, authorized control materials, authorized other ancillary reagents and authorized materials required to use your product are not permitted.
- C. Authorized laboratories that receive your product **must** notify the relevant public health authorities of their intent to run your product prior to initiating testing.
- D. Authorized laboratories using your product **must** have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- E. Authorized laboratories **must** collect information on the performance of your product and report to DMD/OHT7-OIR/OPEQ/CDRH (via email: CDRH-EUA-Reporting@fda.hhs.gov) and Abbott Laboratories at <https://www.corelaboratory.abbott/us/en/offerings/segments/infectious-disease/sars-cov-2> any suspected occurrence of false **positive** or false **negative** results and significant deviations from the established performance characteristics of your product of which they become aware.
- F. All laboratory personnel using your product must be appropriately trained in automated immunoassay techniques and use appropriate laboratory and personal protective equipment when handling this kit, and use your product in accordance with the authorized labeling. All laboratory personnel using the assay must also be trained in and be familiar with the interpretation of results of the product.
- G. Abbott Laboratories, authorized distributors, and authorized laboratories using your product **must** ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.

\* The letter of authorization refers to, “Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, to perform moderate and high complexity tests” as “authorized laboratories.”

## Analytical Specificity

### Potentially Cross-Reacting Antibodies

The AdviseDx SARS-CoV-2 IgM assay was evaluated for potentially cross-reacting antibodies. A total of 143 specimens from 30 different categories were tested. One hundred forty-two (142) specimens were negative and 1 specimen was positive by the AdviseDx SARS-CoV-2 IgM assay. The data are summarized in the following table.

Category <sup>a</sup>	n	Positive	Negative
Anti-Hepatitis A Virus (HAV)	5	0	5
Anti-Hepatitis C Virus (HCV)	5	0	5
Anti-Hepatitis D Virus (HDV)	5	0	5
Anti-Herpes Simplex Virus (HSV)	5	0	5
Anti-Human T-Lymphotropic Virus (HTLV) Type 1	5	0	5
Anti-HTLV Type 2	5	0	5
Anti-Respiratory Syncytial Virus (RSV)	5	0	5
Anti-Varicella Zoster Virus	5	0	5
Antinuclear Antibody (ANA)	5	0	5
Cytomegalovirus (CMV) Immunoglobulin Class G	5	0	5
CMV IgM	5	0	5
Double-Stranded Deoxyribonucleic Acid (dsDNA) Antibody	5	0	5
Epstein-Barr Virus (EBV) IgG	5	0	5
EBV IgM	5	0	5
<i>Escherichia coli</i> ( <i>E. coli</i> ) Antibody	5	0	5
Enterovirus IgG	5	0	5
Enterovirus IgM	5	0	5
HAMA	5	0	5
Hepatitis B Core (HBc) IgM	4	0	4
Heterophilic Antibody Positive	5	0	5
Monoclonal Hyper IgG	5	0	5
Mycoplasma IgM	5	0	5
Parainfluenza IgG	5	0	5
Parainfluenza IgM	2	0	2
Polyclonal Hyper IgG	3	0	3
Rheumatoid Factor	5	1 <sup>b</sup>	4
Rubella IgG	5	0	5
Rubella IgM	5	0	5
Toxoplasmosis IgG	5	0	5
Toxoplasmosis IgM	4	0	4
Total	143	1	142

<sup>a</sup> Samples containing antibodies against the four coronavirus (common cold) strains 229E, NL63, OC43, HKU1, were not evaluated in the study.

<sup>b</sup> 1 out of 5 samples with antibodies to Rheumatoid Factor had a positive result with the AdviseDx SARS-CoV-2 IgM assay.

### Potentially Interfering Medical Conditions

The AdviseDx SARS-CoV-2 IgM assay was evaluated for potential cross-reactivity from individuals with potentially interfering medical conditions. A total of 65 specimens from 13 different categories were tested. Sixty-four (64) specimens were negative and 1 specimen was positive by the AdviseDx SARS-CoV-2 IgM assay. The data are summarized in the following table.

Category	n	Positive	Negative
Adenovirus	5	0	5
Autoimmune Hepatitis	5	0	5
Hemodialysis	5	1 <sup>a</sup>	4
Hepatitis B Virus (HBV)	5	0	5
Human Immunodeficiency Virus (HIV)	5	0	5
Influenza A	5	0	5
Influenza A/B	5	0	5
Influenza B	5	0	5
Influenza Vaccine	5	0	5
Lupus	5	0	5

Category	n	Positive	Negative
Picornavirus	5	0	5
Pregnant Females	5	0	5
Pregnant Females, Multiparous	5	0	5
Total	65	1	64

<sup>a</sup> 1 out of 5 samples from hemodialysis patients had a positive result with the AdviseDx SARS-CoV-2 IgM assay.

## Interference

### Potentially Interfering Endogenous Substances

A study was performed based on guidance from CLSI EP07, 3rd ed.<sup>21</sup> Each substance was tested at 2 levels of the analyte (approximately 0.80 Index [S/C] and 2.20 Index [S/C]). The observed interference was within  $\pm 0.10$  Index (S/C) or  $\pm 10.0\%$  at the following concentrations; therefore, the study showed no interference from these endogenous substances.

Potentially Interfering Endogenous Substance	Interferent Level
Unconjugated Bilirubin	40 mg/dL
Conjugated Bilirubin	40 mg/dL
Hemoglobin	1000 mg/dL
Triglyceride (Intralipid)	2000 mg/dL
Total Protein	15 g/dL

### Potentially Interfering Substances

A study was performed based on guidance from CLSI EP07, 3rd ed.<sup>21</sup> Each substance was tested at 2 levels of the analyte (approximately 0.80 Index [S/C] and 2.20 Index [S/C]). The observed interference was within  $\pm 0.10$  Index (S/C) or  $\pm 10.0\%$  at the following concentrations; therefore, the study showed no interference from these substances.

Potentially Interfering Substance	Interferent Level
Acetaminophen	15.6 mg/dL
Alprazolam	0.0258 mg/dL
Ascorbic Acid	5.25 mg/dL
Azithromycin	1.11 mg/dL
Biotin	4250 ng/mL
Captopril	0.264 mg/dL
Fluoxetine	14.2 mg/dL
Guaifenesin	0.450 mg/dL
Hydroxychloroquine	388.8 ng/mL
Ibuprofen	21.9 mg/dL
Remdesivir	27 $\mu$ mol/L

## Clinical Performance

A study was performed to determine the clinical performance of the AdviseDx SARS-CoV-2 IgM assay.

All specimens tested were assayed in replicates of one using one lot of the AdviseDx SARS-CoV-2 IgM Reagent Kit, one lot of the AdviseDx SARS-CoV-2 IgM Calibrator Kit, and one lot of the AdviseDx SARS-CoV-2 IgM Control Kit on one ARCHITECT i2000SR instrument.

To estimate the positive percent agreement (PPA) between the AdviseDx SARS-CoV-2 IgM assay and the polymerase chain reaction (PCR) comparator, 354 retrospective frozen serum and plasma specimens, collected at different times, were purchased from medical institutions, from a total of 111 subjects whose respiratory samples tested positive for SARS-CoV-2 by a US FDA authorized PCR method and who also presented with COVID-19 symptoms. Specimens from a total of 8 immunocompromised subjects were not included in the data analysis. Specimens from 2 subjects that were collected greater than 30 days post-symptom onset were not included in the data analysis. Specimens from the remaining 101 immunocompetent study subjects were included in the data analysis. The PPA and the 95% confidence interval (CI) were calculated using the initial sample collected in each of the 3 designated time frames after symptom onset (i.e.  $\leq 7$  days, 8-14 days, and 15-30 days), per subject. The performance summary data is illustrated in the table below.

### Positive Percent Agreement by Days Post-Symptom Onset

Days Post-Symptom Onset	n	Positive	Negative	PPA (95% CI)
≤ 7	54	23	31	42.59% (30.33, 55.84)
8 - 14	76	60	16	78.95% (68.50, 86.60)
15 - 30	40	38	2	95.00% (83.50, 98.62)

To estimate the negative percent agreement (NPA), frozen serum and plasma specimens from 2965 unique study subjects were tested using the AdviseDx SARS-CoV-2 IgM assay. All specimens were collected prior to September 2019 (pre-COVID-19 outbreak) and were therefore assumed to be negative. The NPA and the 95% CI were calculated. The performance summary data are illustrated in the table below.

### Negative Percent Agreement

AdviseDx SARS-CoV-2 IgM Results			
n	Positive	Negative	NPA (95% CI)
2965	13	2952	99.56% (99.25, 99.74)

### Longitudinal Study

From the positive agreement study above, a subset of 72 subjects with 2 or more blood draws post-symptom onset were assessed longitudinally. Out of the 72 subjects, 42 presented positive results in all bleeds, while 30 subjects showed SARS-CoV-2 IgM seroconversion. Representative AdviseDx SARS-CoV-2 IgM seroconversion results are provided below. Seroconversion was detected by the AdviseDx SARS-CoV-2 IgM assay at 10 days and 7 days post-symptom onset for subjects A and B, respectively.

Subject	Draw	Days Post-Symptom	Result	Interpretation
		Onset	(Index)	
A	1	1	0.03	Negative
	2	4	0.04	Negative
	3	7	0.83	Negative
	4	10	17.70	Positive
	5	15	27.90	Positive
	6	19	27.53	Positive
B	1	0	0.04	Negative
	2	4	0.12	Negative
	3	7	3.94	Positive
	4	14	24.72	Positive
	5	20	19.11	Positive

### Class Specificity

The anti-human IgM antibody used in the AdviseDx SARS-CoV-2 IgM assay demonstrates class-specific reactivity only to human SARS-CoV-2 IgM. No binding interactions were observed to human SARS-CoV-2 IgG.

A Class Specificity study was conducted to determine the impact of dithiothreitol (DTT) treatment on the detection of IgM and/or IgG positive samples by the Abbott AdviseDx SARS-CoV-2 IgM assay. DTT dissolves IgM antibody disulfide bonds and eliminates activity of the antibody. Upon treatment with DTT, five SARS-CoV-2 patient samples (initially positive for both IgG and IgM) were negative for IgM when tested with the Abbott AdviseDx SARS-CoV-2 IgM assay and remained positive for IgG when tested with the Abbott SARS-CoV-2 IgG assay. This establishes the specificity of the Abbott AdviseDx SARS-CoV-2 IgM kit to the IgM class of antibodies.

### BIBLIOGRAPHY

- van Dorp L, Acman M, Richard D, et al. Emergence of genomic diversity and recurrent mutations in SARS-CoV-2. *Infect Genet Evol* 2020;83:104351. doi:10.1016/j.meegid.2020.104351
- World Health Organization. *WHO Director-General's opening remarks at the media briefing on COVID-19 - 11 March 2020*. Accessed June 1, 2020. <https://www.who.int/dg/speeches/detail/who-director-general-s-opening-remarks-at-the-media-briefing-on-covid-19---11-march-2020>
- Holshue ML, DeBolt C, Lindquist S, et al. First case of 2019 novel coronavirus in the United States. *N Engl J Med* 2020;382:929-936. doi:10.1056/NEJMoa2001191
- World Health Organization. *Coronavirus disease 2019 (COVID-19) situation report-85, Data as received by WHO from national authorities by 10:00 CET, 14 April 2020*. Accessed June 1, 2020. <https://www.who.int/docs/default-source/coronaviruse/situation-reports/20200414-sitrep-85-covid-19>
- Zhao J, Yuan Q, Wang H, et al. Antibody responses to SARS-CoV-2 in patients of novel coronavirus disease 2019. *Clin Inf Dis* Accepted manuscript. Published online 28 March 2020. doi:10.1093/cid/ciaa344
- Xiao AT, Gao C, Zhang S. Profile of specific antibodies to SARS-CoV-2: the first report. *J Infect* 2020. doi:10.1016/j.jinf.2020.03.012
- Shen C, Wang Z, Zhao F, et al. Treatment of 5 critically ill patients with COVID-19 with convalescent plasma. *JAMA* Published online March 27, 2020. doi:10.1001/jama.2020.4783
- World Health Organization. *Laboratory testing strategy recommendations for COVID-19: interim guidance, 21 March 2020*. World Health Organization; 2020.
- US Department of Labor, Occupational Safety and Health Administration, 29 CFR Part 1910.1030, Bloodborne pathogens.
- US Department of Health and Human Services. *Biosafety in Microbiological and Biomedical Laboratories*. 6th ed. Washington, DC: US Government Printing Office; June 2020.
- World Health Organization. *Laboratory Biosafety Manual*. 3rd ed. Geneva: World Health Organization; 2004.
- Clinical and Laboratory Standards Institute (CLSI). *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. CLSI Document M29-A4. Wayne, PA: CLSI; 2014.
- Clinical and Laboratory Standards Institute (CLSI). *Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition*. CLSI Document GP44-A4. Wayne, PA: CLSI; 2010.
- Clinical and Laboratory Standards Institute (CLSI). *Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions*. 4th ed. CLSI Guideline C24. Wayne, PA: CLSI; 2016.
- Westgard JO. *Basic QC Practices*. 3rd ed. Madison, WI: Westgard Quality Corporation; 2010.
- Primus FJ, Kelley EA, Hansen HJ, et al. "Sandwich"-type immunoassay of carcinoembryonic antigen in patients receiving murine monoclonal antibodies for diagnosis and therapy. *Clin Chem* 1988;34(2):261-264.
- Schroff RW, Foon KA, Beatty SM, et al. Human anti-murine immunoglobulin responses in patients receiving monoclonal antibody therapy. *Cancer Res* 1985;45(2):879-885.
- Boscato LM, Stuart MC. Heterophilic antibodies: a problem for all immunoassays. *Clin Chem* 1988;34(1):27-33.
- Nolph KD, Husted FC, Sharp GC, et al. Antibodies to nuclear antigens in patients undergoing long-term hemodialysis. *Am J Med* 1976;60(5):673-676.
- Clinical and Laboratory Standards Institute (CLSI). *Evaluation of Precision of Quantitative Measurement Procedures: Approved Guideline—Third Edition*. CLSI Document EP05-A3. Wayne, PA: CLSI; 2014.
- Clinical and Laboratory Standards Institute (CLSI). *Interference Testing in Clinical Chemistry*. 3rd ed. CLSI Guideline EP07. Wayne, PA: CLSI; 2018.

## Key to Symbols

### ISO 15223 Symbols

	Consult instructions for use
	Manufacturer
	Sufficient for
	Temperature limitation
	Use by/Expiration date
<b>IVD</b>	<i>In Vitro</i> Diagnostic Medical Device
<b>LOT</b>	Lot Number
<b>REF</b>	List Number
<b>SN</b>	Serial number

### Other Symbols

<b>ASSAY I</b>	Assay Diluent
<b>CONJUGATE</b>	Conjugate
<b>CONT . I</b>	Contains Sodium Azide. Contact with acids liberates very toxic gas.
<b>CONTROL N .</b>	Control Number
<b>FOR USE I</b>	Identifies products to be used together
<b>G I</b>	Global Trade Item Number
<b>I</b>	Information needed for United States of America only
<b>M I</b>	Microparticles
<b>PRODUCT O I</b>	Product of Ireland
<b>PROTECT FROM I</b>	Protect from light
<b>REAGENT LOT</b>	Reagent Lot
<b>Rx ONLY</b>	For use by or on the order of a physician only (applicable to USA classification only).
<b>WAR . I</b>	Warning: May cause an allergic reaction.

Note for number formatting:

- A space is used as thousands separator (example: 10 000 specimens).
- A period is used to separate the integer part from the fractional part of a number written in decimal form (example: 3.12%).

ARCHITECT and related brand marks are trademarks of Abbott. Other trademarks are the property of their respective owners.



Abbott Ireland  
Diagnostics Division  
Finisklin Business Park  
Sligo  
Ireland  
+353-71-9171712

**Customer Service: Contact your local representative or find country-specific contact information on [www.corelaboratory.abbott](http://www.corelaboratory.abbott)**

Revised September 2021.

©2020, 2021 Abbott Laboratories



CoV-2 IgM

REF 6R87-10

H14979R02

# AdviseDx SARS-CoV-2 IgM Control Kit

FOR USE I

ARCHITECT

Read Highlighted Changes: Revised September 2021.

For use under an Emergency Use Authorization (EUA) Only  
Prescription Use only.

For *In Vitro* Diagnostic Use Only.

Package insert instructions must be carefully followed. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

For laboratory professional use only.

## NAME

AdviseDx SARS-CoV-2 IgM Control Kit (also referred to as CoV-2 IgM Ctrls or SARS-CoV-2 IgM Ctrls)

## INTENDED USE

The AdviseDx SARS-CoV-2 IgM Control Kit is for the estimation of test precision and the detection of systematic analytical deviations of the ARCHITECT i System when used for the qualitative detection of IgM antibodies to SARS-CoV-2 in human serum, serum separator tube, and plasma (dipotassium EDTA, tripotassium EDTA, lithium heparin, lithium heparin separator tube, sodium heparin).

For additional information, refer to the AdviseDx SARS-CoV-2 IgM reagent package insert and the ARCHITECT System Operations Manual.

The AdviseDx SARS-CoV-2 IgM assay is only for use under the Food and Drug Administration's Emergency Use Authorization.

## CONTENTS

The **CONTROL**  $\bar{-}$  contains human plasma.

The **CONTROL**  $\bar{+}$  contains inactivated, cell-free, human blood-derived material, reactive for anti-SARS-CoV-2 IgM.

Preservatives: sodium azide and antimicrobial agents.

The controls are at the following ranges. The ranges may be used for individual replicate control specifications on the ARCHITECT i System.

Control	Quantity	Anti-SARS-CoV-2 IgM
		RANGE
		(Index [S/C])
<b>CONTROL</b> $\bar{-}$	1 x 4.0 mL	$\leq 0.52$
<b>CONTROL</b> $\bar{+}$	1 x 4.0 mL	1.39 - 6.67

NOTE: The insert ranges for the controls are not lot specific and represent the total range of values which may be generated throughout the life of the product. It is recommended that each laboratory establish its own means and acceptable ranges which should fall within the package insert ranges. Sources of variation that can be expected include:

- Calibration
- Control lot
- Reagent lot
- Calibrator lot
- Instrument

## PRECAUTIONS

For Use Under An Emergency Use Authorization Only.

This assay is only for *in vitro* diagnostic use under the FDA Emergency Use Authorization.

- **IVD**
- For *In Vitro* Diagnostic Use
- **Rx ONLY**

- This product has not been FDA cleared or approved but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories.
- This product is for use with a test that has been authorized only for detecting the presence of IgM antibodies to SARS-CoV-2, not for any other viruses or pathogens.
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

## Safety Precautions

-  **CAUTION:** This product contains human-sourced and/or potentially infectious components. Refer to the CONTENTS section of this package insert. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced materials should be considered potentially infectious. It is recommended that this product, human specimens, and all consumables contaminated with potentially infectious materials be handled in accordance with the OSHA Standard on Bloodborne Pathogens. Biosafety Level 2 or other appropriate regional, national, and institutional biosafety practices should be used for materials that contain, are suspected of containing, or are contaminated with infectious agents.<sup>1-4</sup>
- The human-sourced materials used in the **CONTROL**  $\bar{+}$  have been tested and found to be reactive for anti-SARS-CoV-2 IgM and nonreactive for HBsAg, HIV-1 RNA or HIV-1 Ag, anti-HIV-1/HIV-2, and anti-HCV.
- The human-sourced material used in the **CONTROL**  $\bar{-}$  has been tested and found to be nonreactive for anti-SARS-CoV-2 IgM, HBsAg, HIV-1 RNA or HIV-1 Ag, anti-HIV-1/HIV-2, and anti-HCV.

The following warnings and precautions apply to: <b>CONTROL</b> $\bar{+}$ and <b>CONTROL</b> $\bar{-}$	
Contains sodium azide.	
EUH032	Contact with acids liberates very toxic gas.
P501	Dispose of contents / container in accordance with local regulations.

Follow local chemical disposal regulations based on your location along with recommendations and content in the Safety Data Sheet to determine the safe disposal of this product.

For the most current hazard information, see the product Safety Data Sheet.

Safety Data Sheets are available at [www.corelaboratory.abbott](http://www.corelaboratory.abbott) or contact your local representative.

For a detailed discussion of safety precautions during system operation, refer to the ARCHITECT System Operations Manual, Section 8.

## PREPARATION FOR USE

- Thaw completely at room temperature (15 to 30°C).
- Prior to each use, mix by gentle inversion.

## STORAGE

- This product is shipped on dry ice.
- Protect from light.
- Do not use past expiration date.

	Storage Temperature	Maximum Storage Time	Additional Storage Instructions
<b>Unopened</b>	-20°C or colder	Until expiration date	
<b>Opened</b>	2 to 8°C	Up to 45 days after thaw, not to exceed expiration date printed on the bottle	Return to original carton to protect from light. Store tightly capped. Store in upright position.

## INSTRUMENT PROCEDURE

- To obtain the recommended volume requirements for the controls, hold the bottle vertically, and dispense 4 drops of the negative control and 4 drops of the positive control into each sample cup in the assigned position.
- For information on configuring control data, refer to the ARCHITECT System Operations Manual, Section 2.
- For instructions on ordering and loading controls on the instrument, refer to the ARCHITECT System Operations Manual, Section 5.

## QUALITY CONTROL PROCEDURES

The recommended control requirement for the AdviseDx SARS-CoV-2 IgM assay is that a single sample of each control level be tested once every 24 hours each day of use.

## INDICATIONS OF INSTABILITY OR DETERIORATION

Instability or deterioration should be suspected if there are precipitates, visible signs of leakage, or if controls do not meet the appropriate package insert and/or ARCHITECT System Operations Manual criteria.

NOTE: The AdviseDx SARS-CoV-2 IgM controls may present a cloudy or turbid appearance following thaw. This is not necessarily a sign of deterioration if the controls meet the appropriate criteria.

## LIMITATIONS

Control values have not been established for assays other than the AdviseDx SARS-CoV-2 IgM assay. If the user wishes to use this control material with other assays, it is their responsibility to establish the appropriate ranges.

The controls are not calibrators and should not be used for assay calibration.

## BIBLIOGRAPHY

1. US Department of Labor, Occupational Safety and Health Administration, 29 CFR Part 1910.1030, Bloodborne pathogens.
2. US Department of Health and Human Services. *Biosafety in Microbiological and Biomedical Laboratories*. 6th ed. Washington, DC: US Government Printing Office; June 2020.
3. World Health Organization. *Laboratory Biosafety Manual*. 3rd ed. Geneva: World Health Organization; 2004.
4. Clinical and Laboratory Standards Institute (CLSI). *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. CLSI Document M29-A4. Wayne, PA: CLSI; 2014.

## Key to Symbols

ISO 15223 Symbols	
	Caution
	Consult instructions for use
	Manufacturer
	Temperature limitation
	Upper limit of temperature
	Use by/Expiration date
<b>CONTROL</b>   -	Negative Control
<b>CONTROL</b> +	Positive Control
<b>IVD</b>	<i>In Vitro</i> Diagnostic Medical Device
<b>LOT</b>	Lot Number
<b>REF</b>	List Number
Other Symbols	
<b>AFTER THAW</b>	After thaw store at
<b>CONT</b> .	Contains Sodium Azide. Contact with acids liberates very toxic gas.
<b>FOR USE</b>	Identifies products to be used together
<b>G</b>	Global Trade Item Number
	Information needed for United States of America only
<b>PRODUCT</b> O	Product of Ireland
<b>PROTECT FROM</b>	Protect from light
<b>RANGE</b>	Range
<b>Rx ONLY</b>	For use by or on the order of a physician only (applicable to USA classification only).
<b>UN</b>	Until first use store at

Note for number formatting:

- A space is used as thousands separator (example: 10 000 specimens).
- A period is used to separate the integer part from the fractional part of a number written in decimal form (example: 3.12%).

ARCHITECT and related brand marks are trademarks of Abbott. Other trademarks are the property of their respective owners.

Abbott Ireland  
Diagnostics Division  
Finisklin Business Park  
Sligo  
Ireland  
+353-71-9171712

**Customer Service: Contact your local representative or find country-specific contact information on [www.corelaboratory.abbott](http://www.corelaboratory.abbott)**

Revised September 2021.

©2020, 2021 Abbott Laboratories



CoV-2 IgM

REF 6R87-01

H14978R02

# AdviseDx SARS-CoV-2 IgM Calibrator Kit

FOR USE I

ARCHITECT

Read Highlighted Changes: Revised September 2021.

For use under an Emergency Use Authorization (EUA) Only  
Prescription Use only.

For *In Vitro* Diagnostic Use Only.

Package insert instructions must be carefully followed. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

For laboratory professional use only.

## NAME

AdviseDx SARS-CoV-2 IgM Calibrator Kit (also referred to as CoV-2 IgM Cal or SARS-CoV-2 IgM Cal)

## INTENDED USE

The AdviseDx SARS-CoV-2 IgM Calibrator Kit is for the calibration of the ARCHITECT i System when used for the qualitative detection of IgM antibodies to SARS-CoV-2 in human serum, serum separator tube, and plasma (dipotassium EDTA, tripotassium EDTA, lithium heparin, lithium heparin separator tube, sodium heparin).

For additional information, refer to the AdviseDx SARS-CoV-2 IgM reagent package insert and the ARCHITECT System Operations Manual.

The AdviseDx SARS-CoV-2 IgM assay is only for use under the Food and Drug Administration's Emergency Use Authorization.

## CONTENTS

The **CA I** contains inactivated, cell-free, human blood-derived material, reactive for anti-SARS-CoV-2 IgM.

Preservatives: sodium azide and antimicrobial agents.

Calibrator	Quantity
<b>CA I</b>	1 x 2.0 mL

## STANDARDIZATION

The SARS-CoV-2 IgM Calibrator is manufactured gravimetrically and is referenced to an internal reference standard.

## PRECAUTIONS

**For Use Under An Emergency Use Authorization Only.**

This assay is only for in vitro diagnostic use under the FDA Emergency Use Authorization.

- **IVD**
- For *In Vitro* Diagnostic Use
- **Rx ONLY**
- This product has not been FDA cleared or approved but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories.
- This product is for use with a test that has been authorized only for detecting the presence of IgM antibodies to SARS-CoV-2, not for any other viruses or pathogens.
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

## Safety Precautions



- **CAUTION:** This product contains human-sourced and/or potentially infectious components. Refer to the CONTENTS section of this package insert. No known test method can offer complete assurance that products derived from human sources or inactivated microorganisms will not transmit infection. Therefore, all human-sourced materials should be considered potentially infectious. It is recommended that this product, human specimens, and all consumables contaminated with potentially infectious materials be handled in accordance with the OSHA Standard on Bloodborne Pathogens. Biosafety Level 2 or other appropriate regional, national, and institutional biosafety practices should be used for materials that contain, are suspected of containing, or are contaminated with infectious agents.<sup>1-4</sup>
- The human-sourced materials used in the **CA I** have been tested and found to be reactive for anti-SARS-CoV-2 IgM and nonreactive for HBsAg, HIV-1 RNA or HIV-1 Ag, anti-HIV-1/HIV-2, and anti-HCV.

The following warnings and precautions apply to: <b>CA I</b>	
Contains sodium azide.	
EUH032	Contact with acids liberates very toxic gas.
P501	Dispose of contents / container in accordance with local regulations.

Follow local chemical disposal regulations based on your location along with recommendations and content in the Safety Data Sheet to determine the safe disposal of this product.

For the most current hazard information, see the product Safety Data Sheet.

Safety Data Sheets are available at [www.corelaboratory.abbott](http://www.corelaboratory.abbott) or contact your local representative.

For a detailed discussion of safety precautions during system operation, refer to the ARCHITECT System Operations Manual, Section 8.

## PREPARATION FOR USE

- Thaw completely at room temperature (15 to 30°C).
- Prior to each use, mix by gentle inversion.

## STORAGE

- This product is shipped on dry ice.
- Protect from light.
- Do not use past expiration date.

	Storage Temperature	Maximum Storage Time	Additional Storage Instructions
<b>Unopened</b>	-20°C or colder	Until expiration date	
<b>Opened</b>	2 to 8°C	Up to 45 days after thaw, not to exceed expiration date printed on the bottle	Return to original carton to protect from light. Store tightly capped. Store in upright position.

## INSTRUMENT PROCEDURE

- Test the calibrator in replicates of 3. The calibrator should be priority loaded.
- To obtain the recommended volume requirements for the calibrator, hold the bottle vertically, and dispense 4 drops of the calibrator into the sample cup in the assigned position.
- For information on ordering calibrations, refer to the ARCHITECT System Operations Manual, Section 6.

## QUALITY CONTROL PROCEDURES

A single sample of each control level must be tested to evaluate the assay calibration. Ensure that assay control values are within the ranges specified in the respective control package insert.

For information on ordering controls, refer to the ARCHITECT System Operations Manual, Section 5.

Once a calibration is accepted and stored, it may be used for 30 days. During this time, all subsequent samples may be tested without further calibration unless:

- A reagent kit with a new lot number is used.
- Daily quality control results are outside of quality control limits used to monitor and control system performance.

This assay may require recalibration after maintenance to critical parts or subsystems or after service procedures have been performed.

For additional information, refer to the AdviseDx SARS-CoV-2 IgM reagent package insert and the ARCHITECT System Operations Manual.

## INDICATIONS OF INSTABILITY OR DETERIORATION


Instability or deterioration should be suspected if there are precipitates, visible signs of leakage, if calibration does not meet the appropriate package insert and/or ARCHITECT System Operations Manual criteria, or if controls do not meet the appropriate criteria.

NOTE: The AdviseDx SARS-CoV-2 IgM calibrator may present a cloudy or turbid appearance following thaw. This is not necessarily a sign of deterioration if the controls meet the appropriate criteria.

## BIBLIOGRAPHY

1. US Department of Labor, Occupational Safety and Health Administration, 29 CFR Part 1910.1030, Bloodborne pathogens.
2. US Department of Health and Human Services. *Biosafety in Microbiological and Biomedical Laboratories*. 6th ed. Washington, DC: US Government Printing Office; June 2020.
3. World Health Organization. *Laboratory Biosafety Manual*. 3rd ed. Geneva: World Health Organization; 2004.
4. Clinical and Laboratory Standards Institute (CLSI). *Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Fourth Edition*. CLSI Document M29-A4. Wayne, PA: CLSI; 2014.

## Key to Symbols

ISO 15223 Symbols	
	Caution
	Consult instructions for use
	Manufacturer
	Temperature limitation
	Upper limit of temperature
	Use by/Expiration date
	<i>In Vitro</i> Diagnostic Medical Device
	Lot Number
	List Number
Other Symbols	
	After thaw store at
	Calibrator Or Whole Blood Calibrator
	Contains Sodium Azide. Contact with acids liberates very toxic gas.
	Identifies products to be used together
	Global Trade Item Number
	Information needed for United States of America only
	Product of Ireland
	Protect from light
	For use by or on the order of a physician only (applicable to USA classification only).
	Until first use store at

Note for number formatting:

- A space is used as thousands separator (example: 10 000 specimens).
- A period is used to separate the integer part from the fractional part of a number written in decimal form (example: 3.12%).

ARCHITECT and related brand marks are trademarks of Abbott.

Other trademarks are the property of their respective owners.



Abbott Ireland  
Diagnostics Division  
Finisklin Business Park  
Sligo  
Ireland  
+353-71-9171712

**Customer Service: Contact your local representative or find country-specific contact information on [www.corelaboratory.abbott](http://www.corelaboratory.abbott)**

Revised September 2021.

©2020, 2021 Abbott Laboratories

- REF** 6R87-20
- REF** 6R87-30
- REF** 6R87-01
- REF** 6R87-10

**For use under the Emergency Use Authorization (EUA) only.**

This assay is only for *in vitro* diagnostic use under the FDA Emergency Use Authorization.

- **IVD**
- For *In Vitro* Diagnostic Use
- **Rx ONLY**
- This is not the full instructions for use. To print the package insert(s), refer to [www.corelaboratory.abbott/IFU](http://www.corelaboratory.abbott/IFU). For complimentary printed package inserts, contact your local representative at 1-877-4ABBOTT.
- This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories;
- This product has been authorized only for detecting the presence of IgM antibodies to SARS-CoV-2, not for any other viruses or pathogens; and
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3 (b)(1), unless the declaration is terminated or authorization is revoked sooner.

**Key to Symbols**

**ISO 15223 Symbols**

	Manufacturer
<b>IVD</b>	<i>In Vitro</i> Diagnostic Medical Device
<b>REF</b>	List Number

**Other Symbols**

<b>FOR USE WITH</b>	Identifies products to be used together
<b>Rx ONLY</b>	For use by or on the order of a physician only (applicable to USA classification only).

ARCHITECT and related brand marks are trademarks of Abbott. Other trademarks are the property of their respective owners.

 Abbott Ireland  
Diagnostics Division  
Finisklin Business Park  
Sligo  
Ireland  
+353-71-9171712

Created August 2021.  
©2021 Abbott Laboratories