

BioCode® CoV-2 Flu Plus Assay

Instructions for Use

For Prescription Use Only

For *In Vitro* Diagnostic Use

For use under the Emergency Use Authorization (EUA) only

Catalog # 64-C0305

Contact Info

Customer Service:

Telephone: 1-833-Bio-Code
(1-833-246-2633 x882)

Email: Orders@ApBioCode.com

Technical Services:

Telephone: 1-833-BMB-Tech
(1-833-262-8324)

Email: TechSupport@ApBioCode.com

Website:

www.ApBioCode.com

Fax Number:

1-562-777-9884

Mailing Address:

12130 Mora Drive, Unit 2
Santa Fe Springs, CA 90670, USA



Applied BioCode, Inc.
12130 Mora Drive, Unit 2
Santa Fe Springs, CA 90670, USA



For *In Vitro* Diagnostic Use.

The information contained in this document is subject to change without notice. No part of this document may be reproduced or transmitted in any form or by any means, electronic or mechanical, for any purpose, without express written permission of Applied BioCode, Inc.

BioCode is a registered trademark of Applied BioCode, Inc.

All other names of products and brands appearing in this manual are trademarks or registered trademarks of their respective owners.

Table of Contents

NAME AND INTENDED USE.....	5
SUMMARY AND EXPLANATION OF THE TEST.....	5
PRINCIPLE OF PROCEDURE	8
Nucleic Acid Extraction.....	8
Overview of a BioCode® MDx-3000 Run	8
MATERIALS REQUIRED.....	9
Materials Required (Provided).....	9
Materials Required (But Not Provided).....	9
WARNINGS AND PRECAUTIONS	11
General Precautions.....	11
Precaution Related to Public Health Reporting.....	11
Laboratory Precautions.....	11
Safety Precautions	12
REAGENT STORAGE, HANDLING AND STABILITY.....	12
SAMPLE REQUIREMENTS.....	12
PROCEDURE.....	13
Extraction Methods.....	13
BioCode® CoV-2 Flu Plus Assay Set Up.....	14
INTERPRETATION OF RESULTS.....	15
External Negative Controls.....	15
External Positive Controls	15
Internal Control.....	16
Target Pathogen Interpretation	16
BioCode® CoV-2 Flu Plus Assay Test Report.....	18
LIMITATIONS OF THE PROCEDURE.....	18
CONDITIONS OF AUTHORIZATION FOR THE LABORATORY.....	20
PERFORMANCE CHARACTERISTICS	21
Clinical Performance	21
Limit of Detection	24

Analytical Reactivity (Inclusivity)	37
Analytical Specificity (Cross-Reactivity)	37
Competitive Inhibition BioCode® CoV-2 Flu Plus Assay	39
Cross-Contamination and Carryover	40
REFERENCES	41
TABLE OF SYMBOLS	43

NAME AND INTENDED USE

BioCode® CoV-2 Flu Plus Assay

The BioCode® CoV-2 Flu Plus Assay is a molecular assay based on reverse transcription polymerase chain reaction (RT-PCR) and end-point detection of amplified DNA sequences with analyte-specific probes that are coupled to barcoded magnetic beads (BMB) and is intended for the simultaneous qualitative detection and differentiation of RNA from SARS-CoV-2, influenza A (with H1 pdm09, H1 seasonal, H3 subtypes), influenza B and/or respiratory syncytial virus (RSV) in nasopharyngeal swab specimens collected from individuals suspected of respiratory viral infection, consistent with COVID-19, by their healthcare provider. Clinical signs and symptoms of respiratory viral infection due to SARS-CoV-2, influenza, and RSV can be similar. Testing is limited to laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform high complexity tests.

Results are for the identification and differentiation of SARS-CoV-2, influenza A (with H1 pdm09, H1 seasonal, H3 subtypes), influenza B and/or RSV. The BioCode CoV-2 Flu Plus Assay is not intended to detect the influenza C virus. RNA from influenza A, influenza B, RSV and SARS-CoV-2 are generally detectable in nasopharyngeal swab specimens during the acute phase of infection. Positive results are indicative of active infection but do not rule out bacterial infection or co-infection with other pathogens not detected by the test; clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. The agent detected may not be the definite cause of disease. Laboratories within the United States and its territories are required to report all SARS-CoV-2 results to the appropriate public health authorities.

Negative results do not preclude SARS-CoV-2, influenza A, influenza B, and/or RSV, infection and should not be used as the sole basis for patient management decisions. Negative results must be combined with clinical observations, patient history, and epidemiological information.

The BioCode® CoV-2 Flu Plus Assay is intended for use by qualified clinical laboratory personnel specifically instructed and trained in the operation of the BioCode MDx-3000 and in vitro diagnostic procedures. The BioCode® CoV-2 Flu Plus Assay is only for use under the Food and Drug Administration's Emergency Use Authorization.

SUMMARY AND EXPLANATION OF THE TEST

Acute respiratory infections (ARIs) are viral or bacterial infections that arise in the upper or lower respiratory systems. COVID-19 is caused by a novel coronavirus, SARS-CoV-2, which was initially known as 2019-nCoV. The original case cluster was observed in Wuhan, China and reported to World Health Organization (WHO) on December 31, 2019.¹ Since then the outbreak has spread to other regions of China, to neighboring countries, and subsequently to the rest of the world. WHO declared COVID-19 as a pandemic on March 11, 2020 with cases in numerous countries on 6 continents.² As of December 2021, over 267 million confirmed cases have been reported worldwide, and more than 5.2 million people, especially elderly and those with underlying medical conditions, have succumbed to COVID-19.³

The BioCode® CoV-2 Flu Plus Assay (CoV2P) is a multiplexed nucleic acid test intended for the qualitative detection and differentiation of SARS-CoV-2, influenza A (with subtypes), influenza B and RSV in nasopharyngeal swabs. The assay is designed to be used with the BioCode® MDx-3000 automated system, and test results from the BioCode® CoV-2P are available in about 5 hours.

Coronaviruses

Coronaviruses belong to the *Coronaviridae* virus family and contain positive-sense single-strand RNA. Human Coronaviruses OC43, HKU1, NL63, and 229E can all cause mild to moderate upper respiratory illness such as the common cold while infants, the elderly, and patients with weakened immune systems or cardiopulmonary disease are at a greater risk to develop lower respiratory infections including pneumonia and bronchitis.⁴ Two other important coronaviruses, MERS-CoV and SARS-CoV, are responsible for more severe respiratory diseases and have caused major outbreaks in recent years.⁴ As with other respiratory pathogens, coronaviruses are transmitted through droplets when an infected person coughs or sneezes.⁴ In the US coronavirus infections can occur year-round but are most common during the fall and winter months.⁴ The CDC tracks coronavirus cases with the National Respiratory and Enteric Virus Surveillance System (NREVSS) while the ECDC mainly tracks MERS-CoV and SARS-CoV.^{4,5,6} SARS-CoV-2 emerged as a respiratory virus outbreak in Wuhan, China in late 2019 and causes the disease COVID-19. During the initial outbreak in China, infection leads to respiratory distress or pneumonia in approximately 20% of confirmed COVID-19 cases, including critical illness in about 5% of cases requiring intubation or ventilators.⁷ The infectious agent for COVID-19 is a β -coronavirus similar to SARS-CoV and is easily transmitted through aerosol droplets and contaminated surfaces.⁸ In addition to transmission from symptomatic individuals, pre-symptomatic⁹ and asymptomatic transmissions¹⁰ have been reported. The CDC tracks COVID-19 cases within the United States using COVIDView.¹¹

The BioCode® CoV-2 Flu Plus Assay includes an assay for detection of a conserved region of SARS-CoV-2 N gene. The BioCode® CoV-2 Flu Plus Assay does not detect common human coronaviruses (OC43, HKU1, NL63, and 229E), MERS-CoV, or SARS-CoV.

Influenza A & B

Influenza viruses are negative-sense single-strand RNA orthomyxoviruses. Influenza A and B are the two major subtypes that cause annual flu epidemics in humans. Common flu symptoms include fever, chills, cough, sore throat, nasal congestion and runny nose, body aches, headaches, and fatigue, with young children sometimes experiencing vomiting and diarrhea. Influenza A subtypes are identified based on antigenic differences in two glycoproteins: hemagglutinin and neuraminidase. Genetic reassortment is responsible for the emergence of novel influenza A strains; notably, a new strain of H1N1 was responsible for the 2009 “swine flu” pandemic. Influenza B has only one serotype and two known lineages – B/Yamagata and B/Victoria – with significant influenza B epidemics occurring about every 2-3 years.¹² The BioCode® CoV-2 Flu Plus Assay detects and differentiates flu A subtype H1, flu A subtype H1 2009 pandemic (pdm), flu A subtype H3, and flu B.

The flu is spread by droplets when an infected individual coughs, sneezes, or talks.¹³ The virus may be present in the body for up to two days before a patient notices any symptoms, and most adults are able to infect others from the day before symptoms arise up to about a week after symptoms develop.¹³ Importantly, some individuals infected with a flu virus may be asymptomatic; these people are still capable of spreading the virus to others.¹³ A handful of antiviral drugs are available to treat the flu.¹³ While the flu circulates year-round in the United States, cases spike in the fall and winter months, with flu season usually beginning in October and lasting as late as May.¹³ The CDC closely tracks the number of respiratory specimens that test positive for influenza each flu season and publishes their findings online via FluView and FluView Interactive. Similarly, the European Centre for Disease Prevention and Control tracks flu cases in Europe and publishes weekly reports through the European Influenza Surveillance Network (EISN).^{14, 15}

The WHO has tracked global data on flu cases and the predominant subtypes effecting different countries since the 1950s and uses their data to issue recommendations on flu vaccine compositions.¹⁶ The WHO estimates that each year flu epidemics are responsible for between 3 and 5 million serious illnesses and up to 650,000 deaths worldwide, with most deaths in industrialized nations reported in adults over the age of 65 and the vast majority of deaths in children under the age of 5 occurring in developing countries.¹⁷ The WHO recommends annual influenza vaccinations for young children, the elderly, pregnant women, healthcare workers, and individuals with chronic medical conditions.¹⁷

Respiratory Syncytial Virus

Human respiratory syncytial virus (RSV), also known as human orthopneumovirus, is a negative-sense single-strand RNA virus divided into subgroups A and B. The BioCode® CoV-2 Flu Plus Assay has one assay that detects for both RSV A and RSV B. Like other respiratory viruses, RSV is spread through droplets expelled when an infected person coughs or sneezes.¹⁸ The virus generally causes mild cold-like symptoms such as runny nose, sneezing, and coughing, but may cause serious illness in young children and the elderly.¹⁸ RSV is the leading cause of severe respiratory illness in infants and young children, with 2.1 million outpatient visits and over 50,000 hospitalizations in children under five in the US annually.^{18,19} Worldwide it is estimated that between 66,000 and 199,000 children under the age of five die each year from RSV.²⁰ Among adults older than 65 in the US, RSV is responsible for an average of 177,000 hospitalizations and 14,000 deaths per year.²¹ In the US RSV cases spike during the fall, winter, and spring, however different countries experience RSV seasons at varying times and durations based on climate and latitude.^{18,19} There is no RSV vaccine currently available, although the monoclonal antibody palivizumab can be administered in monthly doses to high-risk individuals such as infants born prematurely to prevent RSV infection during the peak season.^{18,19} The CDC monitors RSV cases in the US via the National Respiratory and Enteric Virus Surveillance System (NREVSS) while the European Influenza Surveillance Network (EISN) tracks data on RSV and other influenza-like illnesses in Europe.

PRINCIPLE OF PROCEDURE

The BioCode® MDx-3000 is an automated system that integrates PCR amplification, target capture, signal generation and optical detection from nasopharyngeal swab (NPS) specimens, in either VTM or UTM. Nucleic acids from NPS are extracted with the NucliSENS® easyMAG® (bioMérieux) or MagNA Pure 96 (Roche) automated systems. Once the PCR plate is set up and sealed, the rest of the operations are automated on the MDx-3000.

Nucleic Acid Extraction

Nucleic acids (both RNA and DNA) are captured by coated magnetic beads and eluted on either the NucliSENS® easyMAG® or MagNA Pure 96 automated systems according to the manufacturer provided protocol.

Overview of a BioCode® MDx-3000 Run

Reverse Transcription and Multiplex PCR – Since the target of the BioCode® CoV-2 Flu Plus Assay is an RNA virus, a reverse transcription (RT) step is performed to convert the viral RNA into cDNA prior to amplification. The purified nucleic acid solution is combined with a freshly prepared reaction mix for one-step RT-PCR to amplify the target nucleic acids present in the sample. One of the target-specific primers for each assay is biotinylated at the 5'-end to generate labeled PCR product for subsequent detection.

Dispensing BMB-Probe Mix – Towards the end of PCR amplification, the robotic head dispenses BMB-Probe mix into the designated reaction wells of the capture plate using disposable pipette tips.

PCR Product Transfer – After PCR amplification is completed, the robotic head pierces the foil seal with disposable pipette tips and transfers PCR products into corresponding wells of the capture plate.

Target Capture – Amplified PCR products labeled with biotin are captured at a defined temperature by target-specific probes that are covalently coupled to designated Barcoded Magnetic Beads (BMBs). During this step, BMBs are kept in suspension by gentle agitation. Differentiation of captured targets is achieved by assigning a unique barcode pattern (BMB) for each assay and the internal control.

Signal Generation – After washing off unbound PCR products and unused primers, a streptavidin-phycoerythrin (SA-PE) conjugate is automatically added to the reaction by the robot. High affinity binding between biotin and streptavidin ensures that captured PCR products with the biotin moiety are labeled with phycoerythrin in close proximity to the BMBs.

Optical Detection – Optical detection is performed for each reaction well of the capture plate, an optically clear, flat-bottom microtiter plate. After washing off unbound SA-PE, excitation of the fluorophore at the designated wavelength emits fluorescence signal from BMBs tagged with SA-PE conjugates. Each reaction well is imaged at the defined emission wavelength for fluorescent signal and under bright field for identifying the barcode patterns (decoding).

Software - The BioCode® MDx-3000 Software controls the operation of the instrument, collects and analyzes data, and automatically generates interpretation for test reports at the end of the run. Fluorescent signals from BMBs with the same barcode are sorted and calculated to generate median fluorescence index (MFI) for each analyte. The presence or absence of a pathogen is determined by MFI relative to the validated assay cutoff. The software also analyzes the results of external and internal controls to validate the run, and individual specimen results for reporting.

MATERIALS REQUIRED

Materials Required (Provided)

Table 1 Reagents provided in the BioCode® CoV-2 Flu Plus Assay kit (64-C0305)

Component Name	Part No.	Contents	Storage
BioCode® Master Mix A	24M0011	500 µL x 2	Store at -20°C, after thaw store at 4°C for up to 30 days
BioCode® CoV-2 Flu Plus Primer Mix	44-P0505	500 µL x 2	Store at -20°C, after thaw store at 4°C for up to 30 days
BioCode® RT Mix	24R0006	60 µL x 1	Store at -20°C
BioCode® RNA-IC2	43-R0001	500 µL x 2	Store at -20°C, after thaw store at 4°C for up to 30 days
BioCode® CoV-2 Flu Plus BMB-Probe Mix	44-B0342	6000 µL x 1	Store at -20°C, after thaw vortex for 30 sec, store at 4°C up to 90 days

Materials Required (But Not Provided)

Table 2. General Reagents required for the BioCode MDx-3000

Component Name	Part No.	Contents	Storage
BioCode® SA-PE Mix	24-S0008	500 µL x 1	Single use; protect from light; store at 4°C. Do Not Freeze
BioCode® Buffer A	44-B0003	2 L x 1	Store at room temp

Table 3. BioCode MDx-3000 Consumables

Reagent	Source/Part No.	Quantity
Reagent Reservoirs	INTEGRA 4332	50 each x 4
Waste Bin and Lid	Applied BioCode 01-W0105	20 each
20 µL pipette tips	Beckman 717256	10 x 96 tips
250 µL pipette tips	Beckman 717252	10 x 96 tips
Bio-Rad 96-well hard shell plate 0.1 mL	Bio-Rad HSL9601	5 plates x 5 bags
PCR Adhesive Foil	Thermo Fisher Scientific AB-0626 or Eppendorf 0030127790	100 foils
Microtiter plate	Greiner bio-one 655101	10 plates x 10 bags
Microtiter plate lid	Nunc 5500	100 lids

Table 4. Reagents and Consumables for Validated Extraction Systems

Extraction System	Part Name (Part No.)	Quantity
Roche MagNA Pure 96	DNA and Viral Small Volume Kit (06 374 913 001)	576 extractions
	MagNA Pure 96 System Fluid (06640729001)	5500 mL x 1
	MagNA Pure 96 Processing Cartridge (06241603001)	36 cartridges
	MagNA Pure 96 Output Plate (06241611001)	60 plates
	MagNA Pure Tip 1000 µl (06241620001)	40 x 96 tips
	MagNA Pure Sealing foil (06241638001)	100 foils
bioMérieux NucliSENS® easyMAG®	EasyMAG® Magnetic Silica (280133)	24 x 1.2 mL
	EasyMAG® Lysis Buffer (280134)	4x 1000 mL
	EasyMAG® Buffer 1 (280130)	4x 1000 mL
	EasyMAG® Buffer 2 (280131)	4x 1000 mL
	EasyMAG® Buffer 3 (280132)	4x 1000 mL
	EasyMAG® Disposables (280135)	16 x 3 vessels and aspirator disposables
	Biohit Pipette Tips (280146)	10 x 96 tips

Equipment and Additional Consumables Required (But Not Provided)

- BioCode® MDx-3000
- NucliSENS® easyMAG® (bioMérieux) or MagNA Pure 96 (Roche) Extraction System
- Vortex
- Centrifuge
- Pipettes – single, multi-channel and/or repeater with accuracy range between 1-10 µL, 10-200 µL, and 100-1000 µL
- Sterile, RNase/DNase-free disposable aerosol-barrier micro pipettor tips
- 1.5 mL polypropylene micro centrifuge tubes and racks (RNase/DNase free recommended)
- Cooler racks for 1.5 mL tubes and 0.1 mL 96 well plate
- Biosafety cabinet (laminar flow hood) for extractions
- Freezer (manual defrost) at -10 to -30°C
- Freezer (manual defrost) at -60 to -90°C
- Refrigerator at 2 to 8°C

WARNINGS AND PRECAUTIONS

General Precautions

1. For Emergency Use Authorization (EUA) only.
2. For Prescription Use Only.
3. For *in vitro* diagnostic use.
4. This product has not been FDA cleared or approved but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories.
5. This product has been authorized only for the detection and differentiation of nucleic acid from SARS-CoV-2, influenza A, influenza B, and/or respiratory syncytial virus (RSV), not for any other viruses or pathogens.
6. The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated, or authorization is revoked sooner.
7. Results should be interpreted in combination with the patient's signs and symptoms and results from other diagnostic tests by a trained healthcare professional.
8. The BioCode® CoV-2 Flu Plus Assay is to be used with the BioCode® MDx-3000 with MDx software, and easyMAG (bioMerieux) or MagNA Pure 96 (Roche) automated extraction instruments.

Precaution Related to Public Health Reporting

Local, state, and federal rules and regulations for notification of reportable diseases are continually updated and include a number of organisms that are important for surveillance and outbreak investigations. Laboratories are responsible for following their state and/or local rules pertaining to reportable pathogens and should consult their local and/or state public health laboratories for isolate and/or clinical sample submission guidelines.

Laboratory Precautions

1. Perform the protocol as described in this package insert. Deviations from this protocol may produce erroneous results.
2. The BioCode® CoV-2 Flu Plus Assay should be performed in clearly defined work areas moving in one direction from pre-amplification areas to the amplification/detection area to reduce potential for contamination.
 - a. Begin with specimen preparation and reagent preparation before moving to amplification/detection.
 - b. Use dedicated equipment and supplies for each area (including personal protective equipment, such as lab coats and disposable gloves).
 - c. Clean work areas with 10% bleach or similar disinfectant followed by water before and after assay preparation.
3. A negative control must be tested for each run. If multiple lots are assayed at the same time, a negative control must be assayed for each lot.
4. Do not use reagents past the expiration date. Do not mix reagents or interchange kit components from different kit lots. Kit configurations are identified on the Kit outer carton and Kit Card.
5. Assay setup should be performed at room temperature. Keep Reaction Mix cold using a cooling block during formulation and loading of amplification plate.

Safety Precautions

1. Follow universal safety procedures. All patient specimens should be considered potentially infectious and handled accordingly.
2. Dispose of unused kit reagents and specimens according to local, state and federal regulations.
3. Wear appropriate personal protective equipment including, but not limited to, lab coats, gloves, and protective eyewear. Change gloves often.
4. Do not pipette by mouth.
5. BioCode® RT Mix is classified as an irritant. See SDS for details.

REAGENT STORAGE, HANDLING AND STABILITY

1. Store the kit components frozen (-20°C) prior to use.
2. Store RT Mix frozen (-20°C) except during use.
3. Once thawed, store Master Mix, Primer Mix, and RNA IC2 refrigerated (2-8°C) for up to 30 days.
4. Once thawed, store BMB-Probe Mix refrigerated (2-8°C) for up to 90 days.
5. SA-PE mix is for single use only. Store refrigerated (2-8°C). Protect from light. **DO NOT FREEZE.**
6. Store the Buffer A at room temperature (15-25°C).
7. Avoid storage of any materials near heating or cooling vents or in direct sunlight.
8. Always check the expiration date and do not use reagents beyond the expiration date printed.
9. Once RT-PCR reaction mix is prepared, the test run should be started as soon as possible (within 60 minutes).
10. Remove BMB-Probe Mix from MDx-3000 once the run is completed and store refrigerated (2-8°C).

SAMPLE REQUIREMENTS

This section describes the requirements for specimen collection, preparation, and handling that will help ensure accurate test results.

Nasopharyngeal Swab (NPS) collected according to the standard procedure. Immediately place in 1 – 3 mL of transport medium (VTM or UTM). Samples should be tested as soon as possible. They may be stored at the following conditions:

- 2 - 8°C for 72 hours after collection

Minimum Sample Volume - 200 µL of sample is required for testing.

PROCEDURE

Refer to the BioCode® MDx-3000 Operator's Manual for more detail and pictorial representations of the BioCode® MDx-3000 set up instructions.

Gloves and other Personal Protective Equipment (PPE) should be used when handling specimens and reagents. Once PCR reagents are prepared and sample is added to PCR plate, it should be promptly transferred to the instrument to start the run. After the run is complete, the PCR plate and capture plate should be sealed and discarded.

Extraction Methods

Note: It is strongly recommended that sample preparation be performed in a biosafety cabinet with gloves and appropriate personal protective equipment (PPE).

easyMAG Extraction

1. Pipet 10 µL RNA-IC2 into each well of the easyMAG cartridge
2. Transfer 200 µL of specimen or Control into easyMAG cartridge and load into easyMAG
3. Perform Protocol: **Generic 2.0.1, volume 0.200 mL, Eluate: 50.0 µL, Sample Type: Primary, Matrix: Other**
 - 3.1 Perform 10 min on-board incubation
 - 3.2 When prompted add magnetic silica
 - 3.2.1 Combine 550 µL nuclease-free water and 550 µL magnetic silica mix in one 1.5 mL tube per easyMAG cartridge
 - 3.2.2 Mix thoroughly and dispense 125 µL into each well of an 8-well ELISA strip per easyMAG cartridge.
 - 3.2.3 Add 100 µL to each easyMAG cartridge well and mix thoroughly
 - 3.3 Start remainder of run

MagNA Pure 96 Extraction

1. Pipet 10 µL RNA IC2 into each well of the MagNA Pure 96 processing cartridge (Be careful to pipet directly to the bottom of each well in the cartridge and not produce bubbles. Liquid on the side of the well and bubbles will lead to incorrect volume sensing and the extraction will be aborted.)
2. Transfer 200 µL of specimen or Control into the MagNA Pure 96 processing cartridge (Be careful to pipet directly to the bottom of each well in the cartridge and not produce bubbles. Liquid on the side of the well and bubbles will lead to incorrect volume sensing and the extraction will be aborted.)
3. Perform Protocol: **Pathogen Universal 200 3.1 for MagNA Pure Kit: DNA/Viral NA SV 2.0. Volume: 200 µL, Eluate: 50 µL.**

Nucleic Acid Storage Conditions

Transfer sample extracts from the cartridge into PCR grade micro-tubes, strips or plates and store samples in a 2-8°C refrigerator if testing within 12 hours. Store at -60°C or below if testing cannot be completed within 12 hours of extraction. Extracted nucleic acids may be stored at -60°C or below for up to 90 days.

BioCode® CoV-2 Flu Plus Assay Set Up

Note: Prepare the PCR Plate in a dedicated reaction mix prep area.

1. Thaw Primer Mix, Master Mix and BMB-Probe Mix at room temperature. Perform a quick vortex (2-3 seconds) and centrifuge to collect reagents at the bottom of the tube.
2. Prepare the reaction mix in a polypropylene microcentrifuge tube as described below:

Table 5. Reaction Mix Formulation

Component	Reaction Mix Volume (µL) per reaction	Reaction Mix Volume (µL) per 10 reactions
BioCode® Master Mix A	10.0 µL	100 µL
BioCode® CoV-2 Flu Plus Primer Mix	9.5 µL	95 µL
BioCode® RT Mix	0.5 µL	5 µL
Reaction Mix Volume (µL)	20 µL	200 µL

3. Mix reaction mix by pipetting up and down 8 to 10 times and centrifuge to collect contents at the bottom of the tube. Store at 2-8°C or on a cooling block until ready to set up PCR (not to exceed one hour). Do NOT vortex reaction mix.
4. Pipette 20 µL of reaction mix into appropriate wells of a 96-well plate.
5. Pipette 5 µL of each extracted sample into the wells.
6. Pipette 5 µL extracted negative control into the NC well.
7. Seal plate with pierceable foil. Store at 2-8°C or on a cooling block until ready to load onto the BioCode® MDx-3000 (not to exceed one hour from the time the reaction mix is prepared).
8. Briefly centrifuge plate to collect samples at the bottom of the plate.
9. Load plate onto BioCode® MDx-3000.
10. Vortex thawed room temperature BMB-Probe Mix for 30 seconds at high speed and load onto the BioCode® MDx-3000. **(Note: Precipitates may appear at cold temperatures. If precipitates are present, allow the BMB-Probe Mix to warm to room temperature and vortex additional 30 seconds.)**
11. Load reagents and consumables as prompted by graphic user interface and run BioCode® CoV2P Protocol.

INTERPRETATION OF RESULTS

The BioCode® MDx-3000 software analyzes the data based on plate validity, sample validity and Median Fluorescent Intensity (MFI) compared to an MFI threshold. The software will suppress results if Internal or Negative controls are invalid. The software will indicate if external positive controls are valid or invalid, but will not suppress results if the positive control is not valid.

External Negative Controls

External negative controls can be transport media or well characterized negative specimens. The negative control should go through all processing steps (extractions, amplification, and detection). At least one negative control is required for each plate/kit lot. The BioCode® MDx-3000 software will suppress results for all samples if the Negative Control(s) are not valid (see table below).

Table 6. Criteria for Valid Negative Control

Control	Targets	RNA IC	Description
Negative Control	Not Detected	Detected	Plate Status: Valid. Samples can be interpreted.
Negative Control	Detected	N/A	Plate Status: Invalid. Samples results cannot be interpreted. Results suppressed by software.
Negative Control	N/A	Not Detected	Plate Status: Invalid. Samples results cannot be interpreted. Results suppressed by software.

External Positive Controls

Each laboratory should establish its own Quality Control ranges and frequency of QC testing based on applicable local laws, regulations and good laboratory practices.

External positive controls can be well characterized clinical samples, inactivated virus or synthetic constructs. The positive controls should go through all processing steps (extractions, amplification, and detection). It is recommended that at least one positive control be included for each assay run. Wells identified as Positive Controls will be trended by the BioCode® MDx-3000 software and the report will indicate a valid or invalid result on the report header (see table below). The software will not suppress results based on positive control results. If a positive control does not perform as expected, the user should review all samples in that batch to determine if results should be reported.

During validation, Applied BioCode used the AccuPlex SARS-CoV-2 Reference material (LGC SeraCare; 0505-0126) and the Influenza B, Influenza A H1, Influenza A H3, Respiratory Syncytial Virus A, and Influenza A H1N1pdm (Zeptomatrix; NATRRPP-ABC) each at a 1:4 dilution, individually or combined in pairs extracted per IFU.

Table 7. Criteria for Valid Positive Control

Control	Targets	RNA IC	Recommendations
Positive Control	Expected Target Detected	N/A	Report will indicate positive control is Valid. No user intervention required.
Positive Control	Expected Target Not Detected	N/A	Report will indicate positive control is Invalid. User should review results prior to release.

Internal Control

An RNA Internal Control (RNA IC2: bacteriophage MS2) is added to each sample during extraction. The internal control monitors the efficiency of the extraction, reverse transcription, amplification and detection stages of the assay. Positive results may be reported in the absence of RNA IC detection. However, the BioCode® MDx-3000 software will suppress negative results for any wells with invalid RNA IC results (see table).

Table 8. Criteria for RNA Internal Control (RNA IC)

Targets	RNA IC	Recommendations
Detected	Detected	Well status: Valid. Report all results.
Detected	Not Detected	Well status: Invalid. Detected results may be reported. Repeat/reflex testing.
Not Detected	Not Detected	Well status: Invalid. Not Detected results suppressed by software. Repeat/reflex testing.

Lack of RNA IC signal may indicate sample-associated inhibition or reagent/instrumentation issues. Samples suspected of being inhibitory should be repeated from extraction. If reagent or instrument issues are suspected specimens may be repeated from stored nucleic acid extracts.

Target Pathogen Interpretation

Fluorescent signals from BMBs with the same barcode are sorted and the median fluorescence intensity (MFI) is calculated for each analyte. The BMB assay is considered “Detected” by comparing the MFI to a validated assay cutoff. The SARS-CoV-2 and the Respiratory Syncytial Virus each have a single BMB assay in the multiplex. The software will report these as detected if the single BMB assay produces MFI above the threshold.

Table 9. Single BMB Assay names with corresponding results.

Assay Name	Assay Result	Report Result
RSV	Detected	Respiratory Syncytial Virus Detected
CoV-2	Detected	SARS-CoV-2 Detected

In contrast, the test results for several targets rely on the combination of multiple BMB assays. These include the Influenza A, A subtyping and Influenza B assays. The tables below outline possible assay results for these targets and the corresponding test reports.

Table 10. Possible Assay results and corresponding report results for Influenza B (2 BMB assays).

Assay Name/Result	Assay Name/Result	Report Result
FluB1 / Not Detected	FluB2 / Not Detected	Influenza B Not Detected
FluB1 / Detected	FluB2 / Any Result	Influenza B Detected
FluB1 / Any Result	FluB2 / Detected	Influenza B Detected

Table 11. Possible Assay results and corresponding report results for Influenza A and HA subtyping assays

Target Assay BioCode® CoV2P Result	FluA	FluA/H1	FluA/H1 pdm09	FluA/H3	Action
Influenza A Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Report result
Influenza A and A/H1 Detected	Detected	Detected	Not Detected	Not Detected	Report result
Influenza A/H1pdm09 Detected	Detected	Not Detected	Detected	Not Detected	Report result
Influenza A/H3 Detected	Detected	Not Detected	Not Detected	Detected	Report result
Influenza A/H1, and A/H1pdm09 Detected	Detected	Detected	Detected	Not Detected	Multiple infections are possible but rare, retest to confirm result ^a
Influenza A/H1, and A/H3 Detected	Detected	Detected	Not Detected	Detected	Multiple infections are possible but rare, retest to confirm result ^a
Influenza A/H1pdm09, and A/H3 Detected	Detected	Not Detected	Detected	Detected	Multiple infections are possible but rare, retest to confirm result ^a
Influenza A/H1, A/H1pdm09, and A/H3 Detected	Detected	Detected	Detected	Detected	Multiple infections are possible but rare, retest to confirm result ^a
Influenza A (no subtype detected)	Detected	Not Detected	Not Detected	Not Detected	Retest (see the section on Influenza A, no subtype detected below)
Influenza A Indeterminate	Not Detected	Detected	Detected or Not Detected	Detected or Not Detected	Retest ^b
	Not Detected	Detected or Not Detected	Detected	Detected or Not Detected	
	Not Detected	Detected or Not Detected	Detected or Not Detected	Detected	

^a Repeated multiple positive should be further confirmed by an FDA-cleared influenza subtyping assay.

^b If the retest result confirms the original result, it is recommended that the sample be further investigated using an FDA-cleared influenza A subtyping assay and/or sending the residual sample to local public health laboratory for further testing.

Influenza A (no subtype detected):

If the FluA assay is positive, but none of the hemagglutinin (HA) subtyping assays are positive, then the interpretation is Influenza A (no subtype detected). This result could occur when the titer of the virus in the specimen is low and not detected by the subtyping assays. This result could also indicate the presence of a novel Influenza A strain or a seasonal Influenza A/H3 or A/H1pdm09 strain with critical sequence mismatches to the primers and/or probes of the BioCode CoV2P influenza A HA subtyping assays. In both

cases, the sample in question should be retested. If the retest provides a different result, test the sample a third time to ensure the accuracy of the result. If the retest provides the same result, then the function of the BioCode CoV2P should be verified by testing with appropriate external control materials (known positive samples for Influenza A/H1, Influenza A/H3 and Influenza A/H1pdm09), and a negative control should also be run to test for PCR-product contamination. If the BioCode CoV2P accurately identifies the external positive and negative controls, contact the appropriate public health authorities for confirmatory testing.

BioCode® CoV-2 Flu Plus Assay Test Report

The analyzed BioCode® MDx-3000 results are displayed in two report formats: Run Report for the entire run including multiple specimens, or Sample Report for individual specimens. Both reports can be exported as a PDF or CSV file. Each report includes fully analyzed and interpreted results for specimens and/or controls but is formatted differently. Refer to operator manual for more details and examples of the BioCode® MDx-3000 reports.

The Run Report displays analyzed results in a tabular format for all wells (specimens/controls) in a run from a specific Kit lot. If more than one lot is run together, separate Run Reports will be generated by the software for each lot. Possible results by target are: Detected, Not detected, Indeterminate (for Influenza A only), Invalid or N/A (if not ordered and result is masked).

Note: Masking of SARS-CoV-2 results is not appropriate for an assay under EUA. Results for SARS-CoV-2 must be reported.

The Sample report displays results for a single well (specimen/control). In addition to results for each target, the Sample Reports include a results summary section which allows positive results to be reviewed at a glance. The Sample Report results summary will also indicate well validity based on BMB counts, background MFI, and external and internal controls. Sample reports also include any samples specific comments entered during setup.

Both report headers provide traceability information for: Run name, Run start and finish time, User ID, Software version, Instrument ID, Kit Name, and Reagent lots and expiration dates. The headers also include sections for Run Status and External Controls status. The Run Status section will specify if the run is Incomplete, Valid or Invalid based on the Negative Control results for the specific run/kit lot. The External Controls section indicates the results for the negative controls (Valid or Invalid) and Positive Controls (Valid, Invalid, or N/A if not assayed). The Run Status and Controls sections should be reviewed prior to review of target results. In addition to these summaries, the software will also mask results in the detailed tabular sections based on plate and well validity requirements (see interpretation of results for details).

Completed reports can be electronically reviewed. Reviewer comments will be added to the report footer for traceability under the review section. In addition, MFI (Median Florescence Intensity) reports are available for information only for administrator level users.

LIMITATIONS OF THE PROCEDURE

- The use of this assay as an *in vitro* diagnostic under FDA Emergency Use Authorization (EUA) is limited to laboratories that are certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. § 263a, and meet requirements to perform high complexity tests.

- The BioCode® CoV-2 Flu Plus Assay is to be used with the BioCode® MDx-3000 with MDx software, and easyMAG (bioMerieux) or MagNA Pure 96 (Roche) automated extraction instruments. Results of this test should be interpreted by a trained clinician in conjunction with clinical history, epidemiological data and any other laboratory data.
- This assay is qualitative and does not provide a quantitative value for the pathogen(s) present in the sample.
- The performance of the BioCode® CoV-2 Flu Plus Assay has been validated with nasopharyngeal swab (NPS) specimens in VTM or UTM. It has not been validated for other specimen types.
- The performance of this test has not been established for patients without signs of symptoms of respiratory infection.
- The performance of the BioCode® CoV-2 Flu Plus Assay is dependent upon proper sample collection, handling, transportation, storage, and preparation. Failure to observe proper procedures in any one of these steps can lead to incorrect results. There is a risk of false positive and false negative results caused by improperly collected, transported, or handled specimens. The internal control (RNA IC2) will not indicate whether or not nucleic acid has been lost due to inadequate collection, transport or storage of specimens. Particular attention should be given to the Laboratory Precautions noted under the Warnings and Precautions section.
- The performance of this test was established based on the evaluation of a limited number of clinical specimens. The clinical performance has not been established with all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV2 and their prevalence, which change over time.
- Negative results do not exclude the possibility of infection. Negative test results may occur from sequence variants in the region targeted by the assay, the presence of inhibitors, technical error, sample mix-up, or an infection caused by a pathogen not detected by the panel. Test results may also be affected by concurrent antiviral therapy or viral load in the sample that are below the limit of detection for the test. Negative results should not be used as the sole basis for diagnosis, treatment, or other management decisions. Negative results in the setting of a respiratory illness may be due to infection with pathogens that are not detected by this test or lower respiratory tract infection that is not detected by a nasopharyngeal swab specimen.
- Nucleic acids may persist independently of transmissibility or it may be asymptotically carried. Therefore, a positive result does not necessarily indicate the presence of transmissible virus level.
- There is a risk of false positive results due to cross-contamination with other samples, viral nucleic acids or amplified product. Particular attention should be given to the Laboratory Precautions noted under the *Warnings and Precautions* section above.
- There is a risk of false positive results due to non-specific amplification. Erroneous results due to cross-reactivity with organisms found in respiratory tract that were not evaluated or new variant sequences that emerge are possible.
- The performance of this test has not been established for monitoring treatment of COVID-19 infection.
- The performance of this device has not been assessed in a population vaccinated against COVID-19.
- The performance of this test has not been evaluated for immunocompromised individuals.
- The effect of antibiotic treatment on test performance has not been evaluated.

- The performance of this test has not been established for screening of blood or blood products.

CONDITIONS OF AUTHORIZATION FOR THE LABORATORY

The BioCode® CoV-2 Flu Plus Assay Letter of Authorization, along with the authorized Fact Sheet for Healthcare Providers, the authorized Fact Sheet for Patients and authorized labeling are available on the FDA website:

<https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas>

To assist clinical laboratories using the BioCode® CoV-2 Flu Plus Assay the relevant Conditions of Authorization are listed verbatim below, and are required to be met by laboratories performing the EUA test.

- a) Authorized laboratories using your product¹ must include with test result reports all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- b) Authorized laboratories using your product must use your product as outlined in the authorized labeling. Deviation from the authorized procedures, including the authorized instruments, authorized extraction methods, authorized clinical specimen types, authorized control materials, authorized other ancillary reagents and authorized materials required are not permitted.
- c) Authorized laboratories that receive your product must notify the relevant public health authorities of their intent to run your product prior to initiating testing.
- d) Authorized laboratories using your product must have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- e) Authorized laboratories must collect information on the performance of your product and report to DMD/OHT7-OIR/OPEQ/CDRH (via email: CDRH_EUA-Reporting@fda.hhs.gov) and you (TechSupport@ApBioCode.com) any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of your product of which they become aware.
- f) All laboratory personnel using your product must be appropriately trained in PCR techniques and use appropriate laboratory and personal protective equipment when handling this kit and use this product in accordance with the authorized labeling.
- g) Applied BioCode, authorized distributors, and authorized laboratories using your product must ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.

¹ “Your product” refers to the BioCode CoV-2 Flu Plus Assay. The letter of authorization refers to, “Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform high complexity tests” as “authorized laboratories”.

PERFORMANCE CHARACTERISTICS

The BioCode® CoV-2 Flu Plus Assay was developed by combining a set of primers and probes from the BioCode® SARS-CoV-2 EUA assay with a subset of primers and probes from the FDA-cleared BioCode® Respiratory Pathogen Panel (RPP). For all assays shared between the two panels, the sequences of primers and probes were unchanged. In addition, the BioCode® CoV-2 Flu Plus Assay also includes an additional reverse primer with a single base variation for influenza A H3 to broaden sequence variation coverage. Studies were performed to demonstrate that the performance characteristics of the assays were unaffected by the panel modification. The original studies of the BioCode® Respiratory Pathogen Panel for the included targets and the BioCode® SARS-CoV-2 assay are presented below as well as these additional studies.

Clinical Performance

Clinical comparison study of BioCode® CoV-2 Flu Plus Assay– Retrospective samples

The BioCode® CoV-2 Flu Plus Assay was evaluated with 115 frozen de-identified nasopharyngeal swab samples in UTM/VTM. The samples were collected by qualified personnel and stored in compliance with conditions stated in the IFU. As influenza H1 positive samples were unavailable, an additional 10 unique negative NPS samples in VTM were spiked with influenza A H1. This data with contrived samples is presented separately. Randomized samples were tested in a blinded fashion and compared to the FDA-cleared molecular assay (for Flu A including subtypes, Flu B and RSV) and FDA authorized SARS-CoV-2 molecular test (for SARS-CoV-2) after extraction per instructions for use. The BioCode CoV-2 Flu Plus Assay was extracted and assayed per instructions for use with both the easyMAG and MagNA Pure 96 system. Separate extractions for each sample were performed prior to testing with the subject assay or the comparison assays. A summary of the demographic information of the tested samples is provided in **Table 12**.

Table 12. Demographics of retrospective clinical specimens

Retrospective Clinical Specimens	
Total Specimens	115
Gender	n/N (%)
Male	54 (47%)
Female	61 (53%)
Age Category	n/N (%)
0-5 years	8/115 (7%)
6-21 years	14/115 (12%)
22-64 years	76/115 (66%)
65+ years	17/115 (15%)

There were two samples that were invalid with the BioCode CoV2P assay after initial extraction with the easyMAG. Both were valid after repeat extraction. There were no invalid samples for the BioCode CoV2P MP96 extractions or either reference assay. One sample was indeterminate for Flu A and one was Flu A no subtype by the reference method. One sample was inconclusive for SARS-CoV-2 by the

comparator assay on initial testing and was detected upon repeat testing.

The Retrospective study results stratified by target and BioCode CoV2P using both the MagNA Pure and easyMAG extraction systems are presented in the tables below.

Table 13. Summary of Clinical Study results compared to the BioCode CoV2P with MagNA Pure 96 extraction system: Retrospective specimens stratified by target

Target	(n)	Positive Agreement		Negative Agreement	
		PPA(%)	95% CI	NPA (%)	95% CI
Influenza A	115	23/23 (100%)	85.7%,100.0%	91/91 ^d (100%)	95.9%,100.0%
Influenza A H1 ^a	115	N/A	N/A	114/114 ^d (100%)	96.7%,100.0%
Influenza A H1 2009pdm	115	10/10 (100%)	72.2%,100.0%	104/104 ^d (100%)	96.4%,100.0%
Influenza A H3	115	12/12 (100%)	75.8%,100.0%	101/102 ^{bd} (99.0%)	94.7%,99.8%
Influenza B	115	10/10 (100%)	72.2%,100.0%	105/105 (100%)	96.5%,100.0%
Respiratory Syncytial Virus	115	10/10 (100%)	72.2%,100.0%	103/105 ^c (98.1%)	93.3%,99.5%
SARS-CoV-2	115	30/30 (100%)	88.6%,100.0%	85/85 (100%)	95.7%,100.0%

- a. Archived specimens were not available, therefore spiked negative NPS was used. See Table 15 for breakdown of organisms and concentrations.
- b. One specimen was detected as Influenza A/H3 by BioCode CoV2P and only as Flu A (no subtype) by the comparator assay. Bi-directional sequencing of this sample confirmed as H3 with sequence similarity to the A/Kansas/14/2017 (H3N2) strain.
- c. Two specimens were detected as RSV by BioCode CoV2P and negative by the comparator assay. Repeat of these specimens were negative by both methods. Placement on the plate suggests a pipetting error for the initial false positive.
- d. One specimen was indeterminate for Flu A with the reference method. These were not included in the Flu A agreement calculations.

Table 14. Summary of Clinical Study results compared to the BioCode CoV2P with easyMAG extraction system: Retrospective specimens stratified by target

Target	(n)	Positive Agreement		Negative Agreement	
		PPA (%)	95% CI	NPA (%)	95% CI
Influenza A	115	23/23 (100%)	85.7%,100.0%	90/91 (98.9%) ^{c, d}	94.0%,99.8%
Influenza A H1 ^a	115	N/A	N/A	114/114 (100%) ^d	96.7%,100.0%
Influenza A H1 2009pdm	115	10/10 (100%)	72.2%,100.0%	103/104 (99.0%) ^{c, d}	94.7%,99.8%
Influenza A H3 ^b	115	12/12 (100%)	75.8%,100.0%	101/102 (99.0%) ^{b, d}	94.7%,99.8%
Influenza B	115	10/10 (100%)	72.2%,100.0%	104/105 (99.0%) ^c	94.8%,99.8%
Respiratory Syncytial Virus	115	10/10 (100%)	72.2%,100.0%	105/105 (100%)	96.5%,100.0%

Target	(n)	Positive Agreement		Negative Agreement	
		PPA (%)	95% CI	NPA (%)	95% CI
SARS-CoV-2 ^e	115	29/30 (96.7%)	83.3%,99.4%	84/85 (98.8%) ^c	93.6%,99.8%

- Archived specimens were not available, therefore spiked negative NPS was used. See Table 15 for breakdown of organisms and concentrations.
- One specimen was detected as Influenza A/H3 by BioCode CoV2P and only as Flu A (no subtype) by the comparator assay. Bi-directional sequencing of this sample confirmed as H3 with sequence similarity to the A/Kansas/14/2017 (H3N2) strain.
- Three specimens (one each for influenza A/H1N1pdm, influenza B and SARS-CoV-2) were positive by BioCode CoV2P and negative by the comparator assay. Two of the three were negative on repeat by both methods.
- One specimen was indeterminate for Flu A with the comparator assay. It was not included in the Flu A agreement calculations.
- One specimen was detected as SARS-CoV-2 with the reference method and not detected by the BioCode CoV2P. The sample had undergone 2 freeze thaws before extraction with the EasyMAG. It was detected by BioCode CoV2P on the initial testing prior to freeze thaw (see MP96 result).

Clinical comparison study of BioCode® CoV-2 Flu Plus Assay– Contrived Influenza A H1 specimens

Due to the rare occurrence of influenza A H1, archived positive specimens could not be obtained. To compare performance with the comparator assay, contrived clinical specimens were prepared using 10 unique prescreened natural NPS specimens UTM. Contrived specimens were spiked at concentrations of 1.3X LOD or greater using 2 different strains of influenza A H1. The spiked samples were blinded and compared to the FDA-cleared molecular assay and after extraction with the MagNA Pure 96 system in parallel with the retrospective samples. The results and breakdown of the strains and concentrations tested are presented in Table 15.

Table 15. Results from Influenza A H1 spiked NPS specimens

Target	Source	Strain/Isolate	Fold LoD ^a	Concentration	PPA (%)	95% CI
Influenza A H1N1	Zeptomatrix 0810036CF	A/New Caledonia/20/99	1.3	20 TCID ₅₀ / mL	2/2 (100%)	34.2%, 100%
			4	60 TCID ₅₀ / mL	2/2 (100%)	34.2%, 100%
			12	180 TCID ₅₀ / mL	2/2 (100%)	34.2%, 100%
	Zeptomatrix 0810244CF	A/Brisbane/59/07	1.3	20 TCID ₅₀ / mL	2/2 (100%)	34.2%, 100%
			4	60 TCID ₅₀ / mL	1/1 (100%)	20.7%, 100%
			12	180 TCID ₅₀ / mL	1/1 (100%)	20.7%, 100%
	Combined					10/10 (100%)

a. LoD as presented in the comparator assay Package Insert

Limit of Detection

Comparison Testing Near Limit of Detection for BioCode® CoV-2 Flu Plus Assay

A study was performed to compare the performance of the BioCode® CoV-2 Flu Plus Assay on the BioCode® MDX-3000 near the Limit of Detection compared to the BioCode® RPP or BioCode® SARS-CoV-2. Quantified stocks were serially diluted and extracted in quadruplicate with both the EasyMAG and MagNA Pure 96 extraction systems and assayed in parallel with BioCode® CoV2P and BioCode® RPP or BioCode SARS-CoV-2 assays to confirm equivalent LoDs. All targets were detected equivalently (within 3-fold) or better with BioCode® CoV2P (Table 16). Additionally, the Influenza A/Kansas/14/2017(H3N2) was detected as Flu A (no subtype) by BioCode® RPP and Influenza A H3 by BioCode® CoV2P which contains the additional primer for improved detection.

Table 16. Results from LoD for BioCode® CoV-2 Flu Plus Assay and BioCode® RPP or BioCode® SARS-CoV-2 Assays with easyMAG extraction.

Target	Species/Strain/Isolate	Source	Concentration (EasyMAG)		Concentration (MagNA Pure 96)	
			RPP or SARS-CoV-2	CoV2P	RPP or SARS-CoV-2	CoV2P
Influenza A H1N1	A/New Caledonia/20/99	Zeptomatrix 0810036CF	5.0x10 ⁰ TCID ₅₀ /mL	1.7x10 ⁰ TCID ₅₀ /mL	5.6x10 ⁻¹ TCID ₅₀ /mL	5.6x10 ⁻¹ TCID ₅₀ /mL
	A/Brisbane/59/2007	Zeptomatrix 0810244CF	4.5x10 ¹ TCID ₅₀ /mL	5.0x10 ⁰ TCID ₅₀ /mL	4.5x10 ¹ TCID ₅₀ /mL	1.5x10 ¹ TCID ₅₀ /mL
Influenza A H1N1 pdm09	A/California/07/09	Zeptomatrix 0810165CF	1.3x10 ⁻¹ TCID ₅₀ /mL	1.3x10 ⁻¹ TCID ₅₀ /mL	1.3x10 ⁻¹ TCID ₅₀ /mL	4.0x10 ⁻¹ TCID ₅₀ /mL
	A/Mexico/4108/09	Zeptomatrix 0810166CF	7.3x10 ⁻¹ TCID ₅₀ /mL	2.4x10 ⁻¹ TCID ₅₀ /mL	2.2x10 ⁰ TCID ₅₀ /mL	2.2x10 ⁰ TCID ₅₀ /mL
Influenza A H3N2	A/Wisconsin/67/05	Zeptomatrix 0810252CF	1.3x10 ⁰ TCID ₅₀ /mL	3.9x10 ⁰ TCID ₅₀ /mL	1.3x10 ⁰ TCID ₅₀ /mL	1.3x10 ⁰ TCID ₅₀ /mL
	A/Kansas/14/2017	CDC	>4.5x10 ³ EID ₅₀ /mL ^a	1.7x10 ² EID ₅₀ /mL	>4.5x10 ³ EID ₅₀ /mL ^a	1.7x10 ² EID ₅₀ /mL
Influenza B	B/Wisconsin/1/10 (Yamagata)	Zeptomatrix 0810241CF	2.7x10 ⁻¹ TCID ₅₀ /mL	2.7x10 ⁻¹ TCID ₅₀ /mL	9.0x10 ⁻² TCID ₅₀ /mL	9.0x10 ⁻² TCID ₅₀ /mL
	B/Malaysia/2560/2004 (Victoria)	Zeptomatrix 0810258CF	1.1x10 ⁻² TCID ₅₀ /mL	3.3x10 ⁻² TCID ₅₀ /mL	3.3x10 ⁻² TCID ₅₀ /mL	3.3x10 ⁻² TCID ₅₀ /mL
Respiratory Syncytial Virus	Type A	Zeptomatrix 0810040ACF	1.1x10 ⁻¹ TCID ₅₀ /mL	3.3x10 ⁻¹ TCID ₅₀ /mL	1.1x10 ⁻¹ TCID ₅₀ /mL	1.1x10 ⁻¹ TCID ₅₀ /mL

Target	Species/Strain/Isolate	Source	Concentration (EasyMAG)		Concentration (MagNA Pure 96)	
			RPP or SARS-CoV-2	CoV2P	RPP or SARS-CoV-2	CoV2P
	B WV/14617/85	ATCC VR-1400	6.0x10 ⁻² TCID ₅₀ /mL	6.0x10 ⁻² TCID ₅₀ /mL	2.0x10 ⁻² TCID ₅₀ /mL	2.0x10 ⁻² TCID ₅₀ /mL
SARS-CoV-2	USA-WA1/2020	Zeptomatrix 0810587CFHI	1.7x10 ⁻² TCID ₅₀ /mL	1.7x10 ⁻² TCID ₅₀ /mL	5.7x10 ⁻³ TCID ₅₀ /mL	5.7x10 ⁻³ TCID ₅₀ /mL

a. Influenza A H3N2 (A/Kansas/14/2017) was detected as Flu A (No Subtype) with BioCode RPP at 1.7x10² EID₅₀/mL.

Table 17. Comparison at or near LoD for BioCode® CoV-2 Flu Plus Assay and BioCode® RPP or BioCode® SARS-CoV-2 Assays with easyMAG extraction.

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
Influenza A H1N1 (A/New Caledonia/20/99)	Zeptomatrix 0810036CF	5.0 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.7 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	3/4	FluA (no subtype)	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.6 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	0/4	Indeterminate	FluA (no subtype)	Not Detected	Not Detected
			BioCode CoV2P	0/4	Not Detected	FluA (no subtype)	Not Detected	FluA (no subtype)
Influenza A H1N1 (A/Brisbane/59/2007)	Zeptomatrix 0810244CF	4.5 x 10 ¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		1.5 x 10 ¹ TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Detected	FluA (no subtype)
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.0 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	1/4	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.7 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	1/4	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)	Detected
			BioCode CoV2P	2/4	Detected	FluA (no subtype)	Detected	FluA (no subtype)
Influenza A H1N1 pdm09 (A(H1N1)/California/07/09)	Zeptomatrix 0810165CF	4.0 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		4.4 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	1/4	Indeterminate	FluA (no subtype)	Detected	FluA (no subtype)
			BioCode CoV2P	0/4	Not Detected	Indeterminate	Indeterminate	Indeterminate
Influenza A H1N1 pdm09 (A/Mexico/4108/09)	Zeptomatrix 0810166CF	7.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
			BioCode RPP	3/4	Detected	Detected	Detected	Indeterminate
		2.4 x 10 ⁻¹ TCID ₅₀ /mL	BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		8.2 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	0/4	Indeterminate	Indeterminate	Indeterminate	FluA (no subtype)
			BioCode CoV2P	1/4	Not Detected	FluA (no subtype)	Detected	Not Detected
Influenza A H3N2 (A/Wisconsin/67/05)	Zeptomatrix 0810252CF	3.9 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.3 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	3/4	Detected	Detected	FluA (no subtype)	Detected
		4.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	3/4	FluA (no subtype)	Detected	Detected	Detected
			BioCode CoV2P	3/4	Detected	Detected	Detected	Indeterminate
Influenza H3N2 (A/Kansas/14/2017)	CDC VP2019	5.0 x 10 ² EID ₅₀ /mL	BioCode RPP	0/4	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		1.7 x 10 ² EID ₅₀ /mL	BioCode RPP	0/4	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.6 x 10 ¹ EID ₅₀ /mL	BioCode RPP	0/4	Not Detected	FluA (no subtype)	FluA (no subtype)	FluA (no subtype)
			BioCode CoV2P	0/4	FluA (no subtype)	Not Detected	FluA (no subtype)	FluA (no subtype)
Influenza B (Wisconsin/1/10) Yamagata	Zeptomatrix 0810241CF	2.7 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		9.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Detected	Not Detected
			BioCode CoV2P	3/4	Not Detected	Detected	Detected	Detected
		3.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	1/4	Not Detected	Not Detected	Detected	Not Detected
			BioCode CoV2P	2/4	Detected	Not Detected	Detected	Not Detected
Influenza B (Malaysia/2560/2004) Victoria	Zeptomatrix 0810258CF	3.3 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.1 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
			BioCode CoV2P	2/4	Not Detected	Detected	Detected	Not Detected
		3.7 x 10 ⁻³ TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Detected	Not Detected
			BioCode CoV2P	0/4	Not Detected	Not Detected	Not Detected	Not Detected
Respiratory Syncytial Virus (Type A)	Zeptomatrix 0810040ACF	3.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.1 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	3/4	Detected	Detected	Not Detected	Detected
		3.7 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	2/4	Detected	Detected	Not Detected	Not Detected
			BioCode CoV2P	3/4	Detected	Detected	Detected	Not Detected
Respiratory Syncytial Virus (Type B WV/14617/85)	ATCC VR-1400	1.8 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		6.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		2.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Not Detected	Detected
			BioCode CoV2P	2/4	Detected	Not Detected	Not Detected	Detected
SARS-CoV-2 (USA-WA1/2020)	Zeptomatrix 0810587CFHI	5.1 x 10 ⁻² TCID ₅₀ /mL	BioCode CoV2	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.7 x 10 ⁻² TCID ₅₀ /mL	BioCode CoV2	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.7 x 10 ⁻³ TCID ₅₀ /mL	BioCode CoV2	2/4	Detected	Not Detected	Detected	Not Detected
			BioCode CoV2P	2/4	Detected	Detected	Not Detected	Not Detected

Table 18. Comparison at or near LoD for BioCode® CoV-2 Flu Plus Assay and BioCode® RPP or BioCode® SARS-CoV-2 Assays with MagNA Pure 96 extraction.

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
Influenza A H1N1 (A/New Caledonia/20/99)	Zeptomatrix 0810036CF	1.7 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.6 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.9 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	0/4	Flu A (No subtype)	Flu A (No subtype)	Not Detected	Indeterminate
			BioCode CoV2P	2/4	Flu A (No subtype)	Not Detected	Detected	Detected
Influenza A H1N1 (A/Brisbane/59/2007)	Zeptomatrix 0810244CF	4.5 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.5 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	3/4	Flu A (No subtype)	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.0 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	3/4	Flu A (No subtype)	Detected	Detected	Detected
			BioCode CoV2P	3/4	Detected	Detected	Flu A (No subtype)	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
Influenza A H1N1 pdm09 (A(H1N1)/California/07/09)	Zeptomatrix 0810165CF	4.0 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	2/4	Detected	Indeterminate	Detected	Indeterminate
		4.4 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	3/4	Not Detected	Detected	Detected	Detected
			BioCode CoV2P	1/4	Flu A (No subtype)	Indeterminate	Detected	Not Detected
Influenza A H1N1 pdm09 (A/Mexico/4108/09)	Zeptomatrix 0810166CF	6.6 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		2.2 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		7.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Detected	Indeterminate
			BioCode CoV2P	3/4	Detected	Detected	Indeterminate	Detected
Influenza A H3N2 (A/Wisconsin/67/05)	Zeptomatrix 0810252CF	3.9 x 10 ⁰ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		1.3 x 10 ⁰ TCID ₅₀ /mL	BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
			BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		4.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	3/4	Indeterminate	Detected	Detected	Detected
			BioCode CoV2P	3/4	Detected	Flu A (No subtype)	Detected	Detected
Influenza A H3N2 (A/Kansas/14/2017)	CDC VP2019	5.0 x 10 ² EID ₅₀ /mL	BioCode RPP	1/4	Detected	Flu A (No subtype)	Flu A (No subtype)	Flu A (No subtype)
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.7 x 10 ² EID ₅₀ /mL	BioCode RPP	0/4	Flu A (No subtype)			
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.6 x 10 ¹ TCID ₅₀ /mL	BioCode RPP	0/4	Flu A (No subtype)			
			BioCode CoV2P	2/4	Flu A (No subtype)	Flu A (No subtype)	Detected	Detected
Influenza B (Wisconsin/1/10) Yamagata	Zeptomatrix 0810241CF	2.7 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		9.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		3.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	1/4	Detected	Not Detected	Not Detected	Not Detected
			BioCode CoV2P	1/4	Detected	Not Detected	Not Detected	Not Detected
Influenza B (Malaysia/2560/2004) Victoria	Zeptomatrix 0810258CF	1.0 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		3.3 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.1 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	3/4	Detected	Detected	Detected	Not Detected
			BioCode CoV2P	3/4	Detected	Detected	Not Detected	Detected
Respiratory Syncytial Virus (Type A)	Zeptomatrix 0810040ACF	3.3 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		1.1 x 10 ⁻¹ TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		3.7 x 10 ⁻² TCID ₅₀ /mL	BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
			BioCode RPP	2/4	Detected	Detected	Not Detected	Not Detected
			BioCode CoV2P	1/4	Detected	Not Detected	Not Detected	Not Detected
Respiratory Syncytial Virus (Type B WV/14617/85)	ATCC VR-1400	6.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		2.0 x 10 ⁻² TCID ₅₀ /mL	BioCode RPP	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		6.7 x 10 ⁻³ TCID ₅₀ /mL	BioCode RPP	2/4	Not Detected	Not Detected	Detected	Detected
			BioCode CoV2P	1/4	Not Detected	Not Detected	Not Detected	Detected
SARS-CoV-2 (USA-WA1/2020)	Zeptomatrix 0810587CFHI	1.7 x 10 ⁻² TCID ₅₀ /mL	BioCode CoV2	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected
		5.7 x 10 ⁻³ TCID ₅₀ /mL	BioCode CoV2	4/4	Detected	Detected	Detected	Detected
			BioCode CoV2P	4/4	Detected	Detected	Detected	Detected

Organism	Source	Concentration	Target Probe	Number Detected (n of n)	Replicate 1	Replicate 2	Replicate 3	Replicate 4
		1.9 x 10 ⁻³ TCID ₅₀ /mL	BioCode CoV2	3/4	Detected	Not Detected	Detected	Detected
			BioCode CoV2P	0/4	Not Detected	Not Detected	Not Detected	Not Detected

Analytical Reactivity (Inclusivity)

Analytical Reactivity (Inclusivity) – BioCode® CoV-2 Flu Plus Assay

In silico Inclusivity analysis for the SARS-CoV-2 assay.

In silico analysis for the SARS-CoV-2 assay was performed against US sequences in the GISAID database as of September 20th, 2021. The analysis included 1,033,881 sequences. Sequences were evaluated for percent homology and predicted binding efficiency for identified mismatches. Oligo binding is predicted for >99.9% of all sequences evaluated.

In silico Inclusivity analysis for Flu A, Flu B and RSV Assays

In silico inclusivity analysis was performed to evaluate the primers and probes for each assay based on percent sequence homology and review of predicted binding efficiency for identified mismatches. For the Flu A and Flu B primers and probe sequences for the BioCode CoV-2 Flu Plus Assay was performed against the US sequences in the GISAID database as of March 11th, 2021. The Flu A general assay analysis included 27,922 sequences including sequences for Flu A/H1, A/H1pdm09 and A/H3N2, oligo binding is predicted for all evaluated sequences. For the Flu A/H1pdm09 assay, 12,048 sequences were evaluated. Oligo binding is predicted for 99.97% of the evaluated sequences. For the Flu A/H1 seasonal assay 1392 sequences were evaluated. Of these, oligo binding is predicted for 99.0%. For the Flu A/H3 assay, 16502 sequences since January 2015 were evaluated. Oligo binding is predicted for 99.6% of the evaluated sequences. For the FluB1 and FluB2 assays, 11090 and 9927 sequences were evaluated (respectively). Oligo binding is predicted for all sequences for the FluB1 assay and 99.1% for the FluB2 assay. The RSV assay was evaluated with 2560 sequences from the NCBI database, oligo binding is predicted for 99.3% of the evaluated sequences.

Analytical Specificity (Cross-Reactivity)

In Silico Analysis – BioCode® CoV-2 Flu Plus Assay

In silico analysis was performed for each primer and probe using NCBI BLAST. In summary, no significant homology ($\leq 80\%$) was found between primers and probes and analyzed organisms with the exception of SARS-CoV and *Pseudomonas aeruginosa*. For SARS-CoV, 92% homology was observed for the reverse primer and probe for the N target. Detection of SARS-CoV is not expected because of limited homology in each reaction for at least one primer/probe needed for the detection and amplification of an analyte. For *Pseudomonas aeruginosa* 85% homology was observed with the forward primer and 91% homology with the probe. Detection of *Pseudomonas aeruginosa* is not expected because of limited homology with the reverse primer, which is required for the detection and amplification of an analyte. Additionally, no evidence of cross-reactivity with *Pseudomonas aeruginosa* was observed in wet testing.

Table 19. Results from *in silico* cross reactivity analysis for influenza A assays in the BioCode CoV2P.

Organism	Influenza A Matrix			Influenza A H1 (seasonal)				Influenza A H1N1 pdm09			Influenza A H3			
	INFA 1 Pb	FLUA F	FLUA R Bio	FluA H1 P1	FLUA H1 F	FLUA H1 R(A) Bio	FLUA H1 R(B) Bio	CDC H1N1 probe	2009 H1N1 F	2009 H1N1 R bio	FLUH3 P	FLUH3 F	FLUA H3 R Bio	FLUA H3 Rc Bio
Human coronavirus 229E	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human coronavirus HKU1	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human coronavirus NL63	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human coronavirus OC43	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
MERS-coronavirus	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
SARS-coronavirus	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%

Organism	Influenza A Matrix			Influenza A H1 (seasonal)				Influenza A H1N1 pdm09			Influenza A H3			
	INFA 1 Pb	FLUA F	FLUA R Bio	FluA H1 P1	FLUA H1 F	FLUA H1 R(A) Bio	FLUA H1 R(B) Bio	CDC H1N1 probe	2009 H1N1 F	2009 H1N1 R bio	FLUH3 P	FLUH3 F	FLUA H3 R Bio	FLUA H3 Rc Bio
Adenovirus Species B Serotype 7A	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Adenovirus Species C Serotype 2	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Adenovirus Species E Serotype 4	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Enterovirus D68	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human Metapneumovirus	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human Rhinovirus Type A1	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Human Influenza A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Influenza B	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Parainfluenza Virus 1	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Parainfluenza Virus 2	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Parainfluenza Virus 3	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Parainfluenza Virus 4a	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
Respiratory Syncytial Virus (Type A)	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Bordetella pertussis</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Candida albicans</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Chlamydia pneumoniae</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Haemophilus influenzae</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Legionella pneumophila</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Mycobacterium tuberculosis</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Mycoplasma pneumoniae</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Pneumocystis jirovecii (PJP)</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Pseudomonas aeruginosa</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Staphylococcus epidermidis</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Streptococcus salivarius</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Streptococcus pneumoniae</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%
<i>Streptococcus pyogenes</i>	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%	<80%

Table 20: Results from *in silico* cross reactivity analysis for the influenza B, RSV, SARS-CoV-2 and Internal Control assays in the BioCode CoV2P.

Organism	Influenza B						Respiratory Syncytial Virus			SARS-CoV-2 N gene (Assay A)			RNA Internal Control (MS2)		
	FLUB F (A)	FLUB R Bio (A)	FluB1 P	FLUB F (B)	FLUB R Bio (B)	FluB 2P	RSVA F	RSVA R	RSVP	Na F	Na R Bio	Na P	MS2 F	MS2 R Bio	MS2 P4
Human coronavirus 229E	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Human coronavirus HKU1	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Human coronavirus NL63	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Human coronavirus OC43	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
MERS-coronavirus	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
SARS-coronavirus	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	92%	92%	≤80%	≤80%	≤80%
Adenovirus Species B Serotype 7A	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Adenovirus Species C Serotype 2	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Adenovirus Species E Serotype 4	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Enterovirus D68	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Human Metapneumovirus	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Human Rhinovirus	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%

Organism	Influenza B						Respiratory Syncytial Virus			SARS-CoV-2 N gene (Assay A)			RNA Internal Control (MS2)		
	FLUB F (A)	FLUB R Bio (A)	FluB1 P	FLUB F (B)	FLUB R Bio (B)	FluB 2P	RSVA F	RSVA R	RSVP	Na F	Na R Bio	Na P	MS2 F	MS2 R Bio	MS2 P4
Type A1															
Human Influenza A	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Influenza B	N/A	N/A	N/A	N/A	N/A	N/A	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Parainfluenza Virus 1	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Parainfluenza Virus 2	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Parainfluenza Virus 3	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Parainfluenza Virus 4a	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
Respiratory Syncytial Virus (Type A)	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	N/A	N/A	N/A	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Bordetella pertussis</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Candida albicans</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Chlamydia pneumoniae</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Haemophilus influenzae</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Legionella pneumophila</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Mycobacterium tuberculosis</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Mycoplasma pneumoniae</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Pneumocystis jirovecii (PJP)</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Pseudomonas aeruginosa</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	85%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Staphylococcus epidermidis</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Streptococcus salivarius</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Streptococcus pneumoniae</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%
<i>Streptococcus pyogenes</i>	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%	≤80%

Competitive Inhibition BioCode® CoV-2 Flu Plus Assay

A study was performed to evaluate the potential for inhibition samples with mixed infections with the BioCode® CoV-2 Flu Plus Assay on the MDx-3000. Samples were prepared by spiking one target at high concentration ($\geq 10^5$ units/mL when possible) and one target at medium concentration ($\leq 3x$ LoD) in simulated NPS matrix (sNPS). All targets were detected.

Table 21. Results from Competitive inhibition study for CoV-2 Flu Plus Assay

Panel Designation	Viral Strain	Level	Titer Tested	Result (n of 3 Detected)
Competitive Inhibition Sample 1	Influenza A H3N2 A/Wisconsin/67/05a	High	1.0x10 ⁵ TCID ₅₀ /mL	3/3
	Respiratory Syncytial Virus	Medium	9.9x10 ⁻¹ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 2	Respiratory Syncytial Virus	High	1.0x10 ⁵ TCID ₅₀ /mL	3/3
	Influenza A H3N2 A/Wisconsin/67/05a	Medium	1.2x10 ¹ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 3	Influenza A H1N1 A/New Caledonia/20/99	High	1.0x10 ⁵ TCID ₅₀ /mL	3/3
	Influenza B Wisconsin/1/10	Medium	8.1x10 ⁻¹ TCID ₅₀ /mL	3/3

Panel Designation	Viral Strain	Level	Titer Tested	Result (n of 3 Detected)
Competitive Inhibition Sample 4	Influenza B Wisconsin/1/10	High	1.0x10 ⁴ TCID ₅₀ /mL	3/3
	Influenza A H1N1 A/New Caledonia/20/99	Medium	5.1x10 ⁰ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 5	Influenza A H1N1 pdm California/07/09	High	1.0x10 ⁵ TCID ₅₀ /mL	3/3
	SARS-CoV-2 USA-WA1/2020	Medium	5.1x10 ⁻² TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 6	SARS-CoV-2 USA-WA1/2020	High	1.0x10 ⁴ TCID ₅₀ /mL	3/3
	Influenza A H1N1 pdm California/07/09	Medium	3.9x10 ⁻¹ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 8	SARS-CoV-2 USA-WA1/2020	High	1.0x10 ⁴ TCID ₅₀ /mL	3/3
	Influenza B Wisconsin/1/10	Medium	8.1x10 ⁻¹ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 9	Influenza B Wisconsin/1/10	High	1.0x10 ⁴ TCID ₅₀ /mL	3/3
	SARS-CoV-2 USA-WA1/2020	Medium	5.1x10 ⁻² TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 10	SARS-CoV-2 USA-WA1/2020	High	1.0x10 ⁴ TCID ₅₀ /mL	3/3
	Respiratory Syncytial Virus	Medium	9.9x10 ⁻¹ TCID ₅₀ /mL	3/3
Competitive Inhibition Sample 11	Respiratory Syncytial Virus	High	1.0x10 ⁵ TCID ₅₀ /mL	3/3
	SARS-CoV-2 USA-WA1/2020	Medium	5.1x10 ⁻² TCID ₅₀ /mL	3/3

Cross-Contamination and Carryover

Carry-over contamination studies have been performed for the BioCode® MDx-3000 system in conjunction with the easyMAG® (K180041) and MagNA Pure 96 systems (K190585). Since this study is not assay-specific, no additional testing was performed for BioCode® CoV2P.


REFERENCES

1. Patel A., Jernigan D.B. Initial Public Health Response and Interim Clinical Guidance for the 2019 Novel Coronavirus Outbreak — United States, December 31, 2019–February 4, 2020. *MMWR Morb Mortal Wkly Rep* 2020. **69**, 140–146 (2020). Retrieved from [http://dx.doi.org/10.15585/mmwr.mm6905e1external icon](http://dx.doi.org/10.15585/mmwr.mm6905e1external%20icon)
2. World Health Organization. WHO Director-General’s opening remarks at the media briefing on COVID-19 – 11 March 2020. Retrieved from <https://www.who.int/dg/speeches/detail/who-director-general-s-opening-remarks-at-the-media-briefing-on-covid-19---11-march-2020>
3. World Health Organization. Coronavirus disease (COVID-19) Pandemic. Retrieved from <https://www.who.int/coronavirus>
4. European Center for Disease Prevention and Control. Middle East respiratory syndrome coronavirus (MERS-CoV). Retrieved from <https://ecdc.europa.eu/en/middle-east-respiratory-syndrome-coronavirus>
5. European Centre for Disease Prevention and Control. Severe acute respiratory syndrome (SARS). Retrieved from <https://ecdc.europa.eu/en/severe-acute-respiratory-syndrome>
6. Wu Z., McGoogan J.M. Characteristics of and Important Lessons From the Coronavirus Disease 2019 (COVID-19) Outbreak in China. Summary of a Report of 72 314 Cases From the Chinese Center for Disease Control and Prevention. *JAMA*. **323**(13), 1239-1242 (2020).
7. Centers for Disease Control and Prevention. Coronavirus (COVID-19). Retrieved from <https://www.cdc.gov/coronavirus/2019-ncov/>
8. Wei W.E., Li Z., Chiew C.J., Yong S.E., Toh M.P., Lee V.J. Presymptomatic Transmission of SARS-CoV-2 — Singapore, January 23–March 16, 2020. *MMWR Morb Mortal Wkly Rep*. ePub: 1 April 2020. Retrieved from <http://dx.doi.org/10.15585/mmwr.mm6914e1>
9. Clinical Course and Molecular Viral Shedding Among Asymptomatic and Symptomatic Patients With SARS-CoV-2 Infection in a Community Treatment Center in the Republic of Korea Lee et al. *JAMA Intern Med*. doi:10.1001/jamainternmed.2020.3862
10. Centers for Disease Control and Prevention. COVIDView. Retrieved from https://www.cdc.gov/coronavirus/2019-ncov/covid-data/covidview/index.html?CDC_AA_refVal=https%3A%2F%2Fwww.cdc.gov%2Fcoronavirus%2F2019-ncov%2F2019-ncov-data%2F2019-ncov-covidview.html
11. Hay, A.J., Gregory, V., Douglas, A. R., Lin, Y. P. The evolution of human influenza viruses. *Philos Trans R Soc Lond B Biol Sci*. **356**(1416), 1861-1870 (2001).
12. Centers for Disease Control and Prevention. Influenza (Flu). Retrieved from <https://www.cdc.gov/flu/index.htm>
13. European Centre for Disease Prevention and Control. European Influenza Surveillance Network (EISN). Retrieved at <https://ecdc.europa.eu/en/about-us/partnerships-and-networks/disease-and-laboratory-networks/eisn>
14. European Centre for Disease Prevention and Control. Seasonal influenza. In: ECDC. Annual epidemiological report for 2017. Stockholm, ECDC (2018).
15. World Health Organization. Influenza. Global Influenza Surveillance and Response System (GISRS). Retrieved from https://www.who.int/influenza/gisrs_laboratory/en/
16. World Health Organization. Influenza (Seasonal). Retrieved from [https://www.who.int/news-room/fact-sheets/detail/influenza-\(seasonal\)](https://www.who.int/news-room/fact-sheets/detail/influenza-(seasonal))
17. Centers for Disease Control and Prevention. Respiratory Syncytial Virus Infection (RSV). Retrieved at <https://www.cdc.gov/rsv/index.html>

18. Broberg, E. K., Waris, M., Johansen, K., Snacken, R., Penttinen, P. Seasonality and geographical spread of respiratory syncytial virus epidemics in 15 European countries, 2010 to 2016. *Euro Surveill.* **23**(5), (2018).
19. Nair, H., et al. Global burden of acute lower respiratory infections due to respiratory syncytial virus in young children: a systematic review and meta-analysis. *Lancet.* **375**(9725), 1545-1555 (2010).
20. Falsey, A. R. , Hennessey, P. A., Formica M .A., Cox, C., Walsh, E. E. Respiratory syncytial virus infection in elderly and high-risk adults. *New Engl J Med.* **352**(17), 1749-1759 (2005).

TABLE OF SYMBOLS

The following symbols are used on the BioCode® CoV-2 Flu Plus Assay kit components and/or in this package insert.

	Batch code		Keep away from sunlight		Temperature limitations
	Catalog number		Contains sufficient for <n> tests		Consult instructions for use
	Use by YYYY-MM-DD		Do Not Reuse		Manufacturer
	Emergency Use Authorization		For Prescription Use Only		<i>In vitro</i> diagnostic device

- 1. **PRODUCT NAME:** BioCode CoV-2 Flu Plus Assay
- 2. **PRODUCT STOCK CODE:** 64-C0305
- 3. **DESCRIPTION:** Printed card for BioCode CoV-2 Flu Plus Assay kit
- 4. **SPECIFICATIONS:** 3.75"X2.5"

Front

BioCode[®] CoV-2 Flu Plus Assay kit

For use under the Emergency Use Authorization (EUA) Only

- This product has not been FDA cleared or approved, but been authorized for emergency use by FDA under an EUA for use by authorized laboratories;
- This product has been authorized only for the detection and differentiation of nucleic acid from SARS-CoV-2, influenza A, influenza B, and respiratory syncytial virus (RSV), not for any other viruses or pathogens; and
- The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.
- Full labeling can be accessed at <https://apbiocode.com/literatures.htm> or printed version may be requested from techsupport@apbiocode.com at no additional cost.



Applied BioCode, Inc., 12130 Mora Dr., Unit 2
Santa Fe Springs, CA 90670, USA
www.ApBioCode.com

Not actual size

Back

BioCode[®] CoV-2 Flu Plus Assay kit

REF **64-C0305**

LOT **AX-MMXXX**

⌚ **YYYY-MM-DD**

🧊 **-30°C to -10°C**

Name	Color Code	EC Symbol	Part No.	Lot No.	Exp. Date	Contents
BioCode [®] Master Mix A	Green	REAG/A	24M0011	0000XXXXXX	YYYY-MM-DD	500 µL x 2
BioCode [®] RT Mix	Yellow	REAG/B	24R0006	0000XXXXXX	YYYY-MM-DD	60 µL x 1
BioCode [®] CoV-2 Flu Plus Primer Mix	White	REAG/C	44-P0506	PX-MMXXX	YYYY-MM-DD	500 µL x 2
BioCode [®] CoV-2 Flu Plus BMB-Probe Mix	Blue	REAG/D	44-B0342	PX-MMXXX	YYYY-MM-DD	6 mL x 1
BioCode [®] RNA-IC2	Clear	Control/IC	43-R0001	PX-MMXXX	YYYY-MM-DD	500 µL x 2

For use under the Emergency Use Authorization (EUA) Only



Applied BioCode, Inc.,
12130 Mora Dr., Unit 2
Santa Fe Springs, CA 90670, USA
www.ApBioCode.com







851034007097
64-C0305
AX-MMDD
YYYY-MM-DD
CoV2P
1
PK-MMXXX

Not actual size

The colors are defined as follows:

Green color:	Red: 0	Green: 176	Blue: 80
Yellow color:	Red: 255	Green: 255	Blue: 0
White color:	Red: 255	Green: 255	Blue: 255
Blue color:	Red: 0	Green: 176	Blue: 240

5. Scanned 2-D barcode script and Human readable:

851034007097,64-C0305,AX-MMXXX,YYYY-MM-DD,CoV2P,1,PX-MMXXX,

Where:

851034007097	is UDI code of BioCode CoV-2 Flu Plus kit
64-C0305	is the part number of BioCode CoV-2 Flu Plus kit
AX-MMXXX	is the lot number of BioCode CoV-2 Flu Plus kit
YYYY-MM-DD	is the kit expiration date in YYYY-MM-DD format
CoV2P	is the BioCode CoV-2 Flu Plus Assay
1	is the BioCode CoV-2 Flu Plus Protocol Ver. 1
PX-MMXXX	is the Biocode CoV-2 Flu Plus BMB-Probe lot number

Human readable: contains above information

6. BLANK LABEL PRODUCT CODE: GRAYTEX paper, Vendor PN: 211-8511-100 8^{1/2}"X11",
ABC PN: 05-L0072

7. INSPECTION METHOD

Use QF-0024: Label Inspection Sheet

8. RELEASING ACTIVITY

QF-0692 BioCode CoV-2 Flu Plus Kit Assembly