



April 21, 2020

Bright Spine  
% Bahram Parvinian  
Founder  
Lighthouse Regulatory Consulting Group  
3 Harrowgate CT  
Potomac, Maryland 20854

Re: K192145

Trade/Device Name: Galileo Vertebral Body Replacement Device  
Regulation Number: 21 CFR 888.3060  
Regulation Name: Spinal Intervertebral Body Fixation Orthosis  
Regulatory Class: Class II  
Product Code: PLR  
Dated: March 20, 2020  
Received: March 24, 2020

Dear Bahram Parvinian:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent L. Showalter, Ph.D.  
Assistant Director (Acting)  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K192145

Device Name

Galileo Vertebral Body Replacement Device

Indications for Use (Describe)

When used as a single-level partial vertebral body replacement device, the Galileo™ devices are indicated for use in the cervical spine (C2-T1) of skeletally mature patients for the partial replacement (i.e., partial vertebrectomy) of a vertebral body resected or excised for the treatment of tumors, or trauma/fracture, or osteomyelitis, or to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The device is intended for use with autograft or allogenic bone graft comprising cancellous and/or corticocancellous bone graft. When the Galileo™ device is used as a single-level, partial vertebral body replacement device, supplemental fixation may be used.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**510(k) Summary: Galileo Vertebral Body Replacement Device**

In accordance with 21 CFR 807.92 of the Federal Code of Regulations

<b>Date Prepared</b>	April 16, 2020
<b>Submitted By</b>	Bright Spine LLC Robert Simonson t 561-289-9378 r@brightspine.com
<b>Primary Contact</b>	Bahram Parvinian M.S. Principal Consultant Lighthouse Regulatory Consulting Group LLC (301)938-7669 Bahram@lighthouse regulatory.com
<b>Trade Name</b>	Galileo Vertebral Body Replacement Device
<b>Common Name</b>	Vertebral body replacement device
<b>Classification Name</b>	Spinal intervertebral body fixation orthosis
<b>Class</b>	II
<b>Product Code</b>	PLR
<b>CFR Section</b>	21 CFR section 888.3060
<b>Device Panel</b>	Orthopedic
<b>Primary Predicate Device</b>	Choice Spine Hawkeye VBR system (K183588)
<b>Additional Predicate Devices</b>	Next Spine Matrixx Intervertebral fusion device (K171140) Bengal System (K103488)
<b>Reference Predicate Devices</b>	Bright Spine Intervertebral Fusion Device (K102449)
<b>Device Description</b>	The Galileo Vertebral Body Replacement device is a surgical grade titanium (Ti-6Al-4V) device and is available in various widths and heights. It has openings of various sizes to allow for the placement of bone graft and for the free flow of cells between the bone graft and the bone of the patient. It is intended for partial vertebral body replacement in a single vertebra and to hold bone graft material.
<b>Materials</b>	Titanium (Ti-6Al-4V)

<b>Intended Use</b>	Partial cervical vertebral body replacement
<b>Substantial Equivalence Claimed to Predicate Devices</b>	Galileo Vertebral Body Replacement (VBR) is substantially equivalent to the predicate devices in terms of intended use, design, materials used, mechanical safety and performances.
<b>Indications for Use</b>	When used as a single-level vertebral body replacement device, the Galileo™ devices are indicated for use in the cervical spine (C2-T1) of skeletally mature patients for the partial replacement (i.e., partial vertebrectomy) of a vertebral body resected or excised for the treatment of tumors, or trauma/fracture, or osteomyelitis, or to achieve decompression of the spinal cord and neural tissues in cervical degenerative disorders. The device is intended for use with autograft or allogenic bone graft comprising cancellous and/or corticocancellous bone graft. When the Galileo™ device is used as a single-level, partial vertebral body replacement device, supplemental fixation may be used.
<b>Summary of the technological characteristics compared to predicate</b>	<p>Galileo VBR has the same intended use and similar indication as Choice Spine Hawkeye VBR system (K183588) which is for cervical spine (C2-T1) and is indicated for use with FDA-cleared supplemental spinal fixation systems. Galileo VBR has the same intended use as predicate Next Spine Matrixx device (K171140) which includes partial vertebrectomy and to hold autograft or allograft in place. Galileo VBR uses an identical surgical procedure to K103488 (Bengal System).</p> <p>Galileo VBR is indicated for same patient population and same disease condition as K183588. Similar to predicates, Galileo VBR is composed of titanium (Ti-6Al-4V), and is made of comparable sizes and shapes. The method of sterilization for Galileo VBR is the same as predicate Hawkeye VBR system (K183588).</p>
<b>Non-clinical Test Summary</b>	Mechanical performance tests were conducted following the ASTM F2077 (Test methods for intervertebral body fusion devices) and ASTM F2267 (Standard test method for measuring load induced subsidence of intervertebral body fusion device under static axial compression), and expulsion testing. All tests were passed.
<b>Clinical Test Summary</b>	Clinical literature review demonstrates the clinical utility of the device sizes and operative procedure for single level anterior cervical partial vertebrectomy as outlined in the labeling.
<b>SE conclusion</b>	The intended use and technological characteristics of Galileo VBR is the same as the identified predicates. There are differences in shape and design of Galileo VBR compared to the predicates, however these differences do not result in new questions of safety and effectiveness. Where necessary, performance testing has been provided to demonstrate Galileo VBR is substantially equivalent to the identified predicates.