



GE Healthcare (Tianjin) Company Limited
% Glen Sabin
Regulatory Affairs Director
GE Healthcare (GE Medical Systems, LLC)
3200 N Grandview Blvd.
WAUKESHA WI 53188

January 16, 2022

Re: K211980
Trade/Device Name: SIGNA Prime
Regulation Number: 21 CFR 892.1000
Regulation Name: Magnetic resonance diagnostic device
Regulatory Class: Class II
Product Code: LNH
Dated: December 14, 2021
Received: December 15, 2021

Dear Glen Sabin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

CAPT Patrick Hintz, MSIH, CIH, USPHS
Associate Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211980

Device Name
SIGNA Prime

Indications for Use (Describe)

The SIGNA Prime is a whole body magnetic resonance scanner designed to support high resolution, high signal-to-noise ratio, and short scan times. It is indicated for use as a diagnostic imaging device to produce axial, sagittal, coronal, and oblique images, spectroscopic images, parametric maps, and/or spectra, dynamic images of the structures and/or functions of the entire body, including, but not limited to, head, neck, TMJ, spine, breast, heart, abdomen, pelvis, joints, prostate, blood vessels, and musculoskeletal regions of the body.

Depending on the region of interest being imaged, contrast agents may be used. The images produced by SIGNA Prime reflect the spatial distribution or molecular environment of nuclei exhibiting magnetic resonance.

These images and/or spectra when interpreted by a trained physician yield information that may assist in diagnosis.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”



510(k) Summary

K211980

In accordance with 21 CFR 807.92 the following summary of information is provided:

Date:	January 10, 2022
Submitter:	GE Healthcare (Tianjin) Company Limited No. 266 Jingsan Road, Tianjin Airport Economic Area Tianjin, P.R. China 300308
Distributor	GE Medical Systems, LLC 3200 N Grandview BLVD. Waukesha, WI USA 53188
Primary Contact Person:	Glen Sabin Director, Regulatory Affairs GE Healthcare Phone: 262- 5216848 E-mail: glen.sabin@ge.com
Secondary Contact Person:	Huande Li Regulatory Affairs Manager GE Healthcare Phone: 86-010-57083027 E-mail: huande.li@ge.com
Device Trade Name:	SIGNA™ Prime
Common/Usual Name:	Magnetic Resonance Diagnostic Device
Classification Names:	Magnetic Resonance Diagnostic Device per 21 CFR 892.1000
Product Code:	LNH
Predicate Device(s):	SIGNA™ Creator (K143251)
Reference Device:	SIGNA™ Premier (K193282)



<p>Device Description:</p>	<p>SIGNA™ Prime is a whole body magnetic resonance scanner designed to support high resolution, high signal to-noise ratio, and short scan times. The systems use a combination of time-varying magnet fields (Gradients) and RF transmissions to obtain information regarding the density and position of elements exhibiting magnetic resonance. The system can image in the sagittal, coronal, axial, oblique, and oblique planes, using various pulse sequences, imaging techniques and reconstruction algorithms. The system features a 1.5T superconducting magnet with 60cm bore size. The system is designed to conform to NEMA DICOM standards (Digital Imaging and Communications in Medicine).</p>
<p>Indications for Use</p>	<p>The SIGNA Prime is a whole body magnetic resonance scanner designed to support high resolution, high signal-to-noise ratio, and short scan times. It is indicated for use as a diagnostic imaging device to produce axial, sagittal, coronal, and oblique images, spectroscopic images, parametric maps, and/or spectra, dynamic images of the structures and/or functions of the entire body, including, but not limited to, head, neck, TMJ, spine, breast, heart, abdomen, pelvis, joints, prostate, blood vessels, and musculoskeletal regions of the body.</p> <p>Depending on the region of interest being imaged, contrast agents may be used. The images produced by SIGNA Prime reflect the spatial distribution or molecular environment of nuclei exhibiting magnetic resonance.</p> <p>These images and/or spectra when interpreted by a trained physician yield information that may assist in diagnosis.</p>
<p>Technology:</p>	<p>The SIGNA™ Prime employs the same fundamental scientific technology as its predicate devices.</p> <p>SIGNA™ Prime builds on the 1.5T IPM magnet, newly designed Gradient Driver, new designed RF transmit architecture, new</p>



	designed RF receiving chain and new software platform application suite.
Comparison of Indications for Use	<p>The changes in technology do not impact the indications for use. The indications for use have not changed, other than to reflect the SIGNA™ Prime product name.</p> <p>Therefore, the intended use is the same as the predicate device in accordance with the FDA's guidance document "The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]", dated 28 July 2014.</p>
Comparison of Technological Characteristics	<p>Overall, the SIGNA™ Prime employs the same fundamental scientific technology as the predicate device.</p> <p>System Design: There are five notable technological differences between the SIGNA™ Prime and the predicate device: the 1.5T IPM magnet, newly designed Gradient Driver, new designed RF transmit architecture, new designed RF receiving chain and new software platform application suite.</p> <p>Operating Principles: The SIGNA™ Prime functions using the same operating principles as the predicate device.</p> <p>Materials: The SIGNA™ Prime and the predicate device both use flame retardant materials.</p> <p>Safety and Performance Testing: Both the SIGNA™ Prime and the predicate device comply with the same safety and performance testing (see Determination of Substantial Equivalence, below). These technological differences do not raise any different questions regarding safety and effectiveness. Both devices must address questions of whether they provide an adequate level of image quality appropriate for diagnostic use. The performance data described in this submission include results of both bench testing and clinical testing that show the image quality performance of SIGNA™ Prime compared to the predicate device.</p>



<p>Determination of Substantial Equivalence:</p>	<p><u>Summary of Non-Clinical Tests:</u></p> <p>The SIGNA™ Prime and the predicate device were subject to similar risk management testing to demonstrate substantial equivalence of safety and performance.</p> <p>Testing to the following voluntary standards included:</p> <ul style="list-style-type: none">• ANSI/AAMI ES60601-1• IEC 60601-1-2• IEC 60601-2-33• IEC 62304• IEC 60601-1-6• IEC 62366-1• ISO 10993-1 <p>In addition, the SIGNA™ Prime complies with applicable NEMA MS standards for MRI and NEMA PS3 standard for DICOM, as does the predicate device.</p> <p>Both the SIGNA™ Prime and the predicate device have a successful biocompatibility track record, as demonstrated by ISO 10993 testing and by their history of use in previously cleared devices.</p> <p>The following quality assurance measures were applied to the development of the subject device, as they were for the predicate device:</p> <ul style="list-style-type: none">• Risk Analysis• Requirements Reviews• Design Reviews• Testing on unit level (Module verification)• Integration testing (System verification)• Performance testing (Verification)• Simulated use testing (Validation) <p><u>Summary of Clinical Tests:</u></p> <p>The subject of this premarket submission, the SIGNA™ Prime, did not require clinical studies to support substantial equivalence. Sample clinical images have been included in this submission.</p>
--	---



	<p>The sample clinical images demonstrate acceptable diagnostic image performance of the SIGNA™ Prime in accordance with the FDA Guidance “Submission of Premarket Notifications for Magnetic Resonance Diagnostic Devices” issued on November 18, 2016. The image quality of the SIGNA™ Prime is substantially equivalent to that of the predicate device.</p> <p><u>Substantial Equivalence Conclusion:</u></p> <p>The indications for use of the proposed device are comparable to the claimed predicate device. The SIGNA™ Prime employs equivalent technology to the claimed predicate device. Additionally, the results from the above non-clinical tests demonstrate that the device performs as intended. Therefore, the SIGNA™ Prime is substantially equivalent to the predicate device to which it has been compared.</p>
Conclusion:	<p>In conclusion, GE Healthcare considers the SIGNA™ Prime to be as safe, as effective, with performance that is substantially equivalent to the predicate device.</p>