



December 16, 2020

Aesthetic Technology Ltd  
% Richard Hamer  
US Agent/Consultant  
Richard Hamer Associates LLC  
705 Spring Lakes Blvd  
Bradenton, Florida 34210

Re: K202028  
Trade/Device Name: Dermalux Flex MD  
Regulation Number: 21 CFR 890.5500  
Regulation Name: Infrared Lamp  
Regulatory Class: Class II  
Product Code: ILY, OLP, OHS  
Dated: November 19, 2020  
Received: November 23, 2020

Dear Richard Hamer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Purva Pandya  
Acting Assistant Director  
DHT4A: Division of General Surgery Devices  
OHT4: Office of Surgical and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K202028

Device Name

Dermalux Flex MD

Indications for Use (Describe)

Intended to deliver heat in the IR spectrum to provide topical heating for the purpose of elevating tissue temperature; for the temporary relief of minor muscle and joint pain, arthritis and muscle spasm; relieving stiffness; promoting the relaxation of muscle tissue; and to temporarily increase local blood circulation.

The blue spectrum light is intended to reduce mild to moderate inflammatory acne vulgaris.

Intended to emit energy in the red and infrared spectrum for the use in dermatology for the treatment of periorbital and full face wrinkles.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## **510(k) SUMMARY**

### **I. ADMINISTRATIVE**

**Submitter:**

Aesthetic Technology Ltd.  
211 Europa Blvd.  
Warrington, Cheshire WA5 7TN  
United Kingdom  
+44 (0)845 689 1789

**Contact Person:** Dale Needham

**Date of Preparation:** December 11, 2020

### **II. DEVICE NAME**

**Proprietary Name:** Dermalux® Flex MD

**Common Name:** Infrared Lamp; Powered Light Based Non-Laser Surgical Instrument

**Classification Name:** Infrared Lamp; Laser surgical instrument for use in general and plastic surgery and in dermatology

**Regulation Number:** 21 CFR §878.4810 and §890.5500

**Regulatory Class:** Class II

**Product Code:** ILY, OLP, OHS

### **III. PREDICATE DEVICES**

Biophotas Celluma 3 Phototherapy System; K152280 and K171323, Biophotas Inc.; Rejuvalite MD; K133896, Trophy Skin Inc. and Opera Lebody; K201107, GTG Wellness Co. Ltd.

### **IV. DEVICE DESCRIPTION**

The Dermalux® Flex MD is a Class II medical device for use on a treatment bed. The device emits specific wavelengths of low level, narrow band light for the treatment of certain

dermatological and musculoskeletal indications. The wavelengths used in the Dermalux® Flex MD system are Blue 415nm, Red 633nm and Near Infrared 830nm.

The Dermalux® Flex MD enables treatment of the face and the body via a flexible LED Array. The system is operated by a capacitive touch Controller, which allows one of 3 treatments to be selected, through selecting various wavelength combinations. The light is generated by Light Emitting Diodes (LED's) that are contained within the flexible LED Array. The LED Array can be placed into the Base, holding in the appropriate position to administer a facial treatment. For use on the body, the LED Array is removed from the Base unit and placed over the appropriate body part.

## V. INDICATIONS FOR USE

The Dermalux Flex MD is intended to deliver heat in the IR spectrum to provide topical heating for the purpose of elevating tissue temperature; for the temporary relief of minor muscle and joint pain, arthritis and muscle spasm; relieving stiffness; promoting the relaxation of muscle tissue; and to temporarily increase local blood circulation. The blue spectrum light is intended to reduce mild to moderate inflammatory acne vulgaris. The Dermalux Flex MD is intended to emit energy in the red and infrared spectrum for use in dermatology for the treatment of periorbital and full face wrinkles.

## VI. COMPARISON TO PREDICATE DEVICES

Parameter	Subject Device	Predicate Devices (K152280 and K171323)	Secondary Predicate K133896	Secondary Predicate K201107
Product name	Dermalux® Flex MD Phototherapy System	Biophotas Celluma3	Rejuvalite MD	Opera Lebody
Product code	ILY, OHS, OLP	ILY, OHS, OLP	OHS	OHS
Regulation No.	21 CFR 878.4810 and 21 CFR 890.5500	21 CFR 878.4810 and 21 CFR 890.5500	21 CFR 878.4810	21 CFR 878.4810
Class	Class II	Class II	Class II	Class II
Intended use	Use of the red, blue and near-infrared regions of the spectrum to emit energy to treat dermatological and musculoskeletal conditions.	Use of the red, blue and near-infrared regions of the spectrum to emit energy to treat dermatological and musculoskeletal conditions	Use of the red and near-infrared regions of the spectrum to emit energy to treat full face wrinkles	Use of the red and near-infrared regions of the spectrum to emit energy to treat full face wrinkles
Indications for use	Intended to deliver heat in the IR spectrum to provide topical heating for the purpose of elevating tissue temperature; for the temporary relief of minor muscle and joint pain, arthritis and muscle spasm; relieving stiffness; promoting the relaxation	The BioPhotas Celluma3 is intended to deliver heat in the IR spectrum to provide topical heating for the purpose of elevating tissue temperature; for the temporary relief of minor muscle and joint pain, arthritis and muscle spasm; relieving stiffness;	Intended for use in the treatment of full face wrinkles	Intended for use in the treatment of full face wrinkles

	of muscle tissue; and to temporarily increase local blood circulation. The blue spectrum light is intended to reduce mild to moderate inflammatory acne vulgaris. Intended to emit energy in the red and infrared spectrum for the use in dermatology for the treatment of periorbital and full face wrinkles.	promoting relaxation of muscle tissue; and to temporarily increase local blood circulation. The blue spectrum is intended to reduce mild to moderate inflammatory acne vulgaris. The Celluma3 is intended to emit energy in the red and infrared region of the spectrum for use in dermatology for the treatment of periorbital (K152280) and full face (K171323) wrinkles		
Parameter	Subject Device	Predicate Devices (K152280 and K171323)	Secondary Predicate K133896	Secondary Predicate K201107
Power supply	100-240Vac, 50/60Hz, 4.6 -1.85A, 460-430W.	110 – 120V	110-120V	Not stated
Wavelength	Red light: 633nm±5nm Blue light: 415nm±5nm NIR light: 830nm±5nm	Red light: 640nm +/-25nm Blue light: 465nm NIR light: 880nm +/-50nm	Red Light: 600, 622, 600nm NIR Light: 860nm	Red Light: 630 nm NIR Light: 830 nm
Panels Type	1 Panel	1 Panel	1 Panel	1 Panel (Mask)
Light frequency	N/A as DC Power	N/A as DC Power	N/A as DC Power	N/A as DC Power
Output Power	Red - 633nm (11.5mW/cm <sup>2</sup> ) Blue - 415nm (5.52mW/cm <sup>2</sup> ) NIR - 830nm (5.5mW/cm <sup>2</sup> )	6.5 mW/cm <sup>2</sup>	62 mW/cm <sup>2</sup> (at 4 “ distance)	50 mW/cm <sup>2</sup>
Standard dose in Joules	Red: 20J/cm <sup>2</sup> Blue: 10J/cm <sup>2</sup> NIR: 10J/cm <sup>2</sup>	11.7 J/cm <sup>2</sup>	11.2 J/cm <sup>2</sup>	30J/cm <sup>2</sup>
Treatment area	792 cm <sup>2</sup>	773 cm <sup>2</sup>	90 cm <sup>2</sup>	Not stated
Treatment time	Up to 30 minutes	Up to 30 minutes	3 minutes	10 minutes
Operation interface	Device uses a pre-set timer and software to control treatment duration.	Device uses a timer and software to control treatment duration.	Device uses a timer to control treatment duration.	Device uses a timer to control treatment duration.
Software Use	Yes Rx	Yes OTC	No OTC	No OTC

Substantial equivalence is readily evidenced by comparison of product attributes summarized in above. Apparent differences in output power and standard dose do not impact safety and effectiveness because the amount of energy delivered to the skin is similar due to differences in treatment duration and distance from the LED source.

## VII. PERFORMANCE TESTING

**Bench:** Performance testing of the Dermalux® Flex MD was conducted to verify that the device met all design specifications. The test results demonstrate that the Dermalux® Flex MD complies with all requirements, including international and FDA-recognized consensus standards:

EN/IEC 60601-1 (FDA #19-4) Medical electrical equipment - Part 1: General requirements for basic safety and essential performance: 2005

EN/IEC 60601-1-2 (FDA #19-8) Medical electrical equipment -Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic disturbances - Requirements and tests: 2018

EN/IEC 60601-1-11 (FDA #19-14) Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment: 2015

EN/IEC 62304 (FDA #13-79) Medical device software - Software life cycle processes: 2006

EN/IEC 62471 (FDA #12-249) Photobiological safety of lamps and lamp systems: 2008

IEC 60601-2-57 (FDA #12-242) Medical electrical equipment – Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use: 2011

IEC 62366-1 (FDA #5-113) Medical devices — Part 1: Application of usability engineering to medical devices: 2015

EN/IEC 60601-1-6 (FDA #5-89) Medical electrical equipment -Part 1-6: General requirements for basic safety and essential performance – Collateral Standard: Usability: 2013

Photometric Testing of LED Light Panel

ISO 10993-1 Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process

**Animal:** No studies were performed.

**Clinical:** No studies were performed.

## VIII. CONCLUSION

Based on design considerations and testing of product attributes, we conclude that the Dermalux® Flex MD performs at least as well as the above-mentioned predicate devices. The Dermalux® Flex MD is therefore considered to be as safe, as effective, and to perform as well as the legally marketed predicate devices.