



January 5, 2023

Treace Medical Concepts Inc.  
Melissa Coloroso  
Principal Regulatory Affairs Specialist  
203 Fort Wade Road, Suite 150  
Ponte Vedra, Florida 32081

Re: K221181

Trade/Device Name: Treace Medical Concepts (TMC) Plating System  
Regulation Number: 21 CFR 888.3030  
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories  
Regulatory Class: Class II  
Product Code: HRS, HWC  
Dated: April 22, 2022  
Received: April 25, 2022

Dear Melissa Coloroso:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
Shumaya Ali-S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair  
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K221181

Device Name  
Treace Medical Concepts (TMC) Plating System

### Indications for Use (Describe)

The TMC Plating System is intended for use in stabilization of fresh fractures, revision procedures, joint fusion and reconstruction of the feet. The system can be used in both adult and pediatric patients aged >12 years. In the foot, the system can be used for the following specific examples:

- Osteotomy of the 1st metatarsal for the treatment of deformity (e.g., hallux valgus) such as:
  - Opening base wedge osteotomy
  - Closing base wedge osteotomy
  - Crescentic osteotomy
  - Proximal Chevron osteotomy
  - Distal Chevron osteotomy (Austin)
  - Transverse osteotomy
- Arthrodesis of the tarsometatarsal (TMT) joints or the 1st metatarsophalangeal (MTP) joint for the treatment of deformity (e.g., hallux valgus, hallux rigidus, metatarsus adductus) and/or arthritis
- First metatarsal fracture fixation
- Flatfoot Osteotomies
  - Lateral Column Lengthening (Evans Osteotomy)
  - Plantar Flexion Opening Wedge Osteotomy of the Medial Cuneiform (Cotton Osteotomy)
- Mid / Flatfoot Fusions
  - LisFranc Arthrodesis and/or Stabilization
  - Intercuneiform Fusions
  - Navicular-Cuneiform (NC) Fusion
  - Talo-Navicular (TN) Fusion
  - Calcaneo-Cuboid (CC) Fusion
- Medial Column Fusion

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

Treace Medical Concepts (TMC) Plating System

April 21, 2022

**Company:** Treace Medical Concepts, Inc.  
100 Palmetto Park Place  
Ponte Vedra, FL 32081

**Establishment  
Registration:** 3011623994

**Primary Contact:** Melissa Coloroso, Principal Regulatory Affairs Specialist  
Phone: 303-641-8592  
Fax: 904-834-7169  
Email: mcoloroso@treace.net

**Secondary Contact:** Kristina Hall, Director, Regulatory Affairs  
Phone: 904-373-5904 ext. 1321  
Fax: 904-834-7169  
Email: khall@treace.net

**Trade Name:** Treace Medical Concepts (TMC) Plating System

**Common Name:** Plate, fixation, bone  
Screw, fixation, bone

**Classification:** Class II

**Regulation Number:** 21 CFR 888.3030 (Single/multiple component metallic bone fixation appliance and accessories and accessories) (primary)  
21 CFR 888.3040 (Smooth or threaded metallic bone fixation fastener)

**Panel:** 87- Orthopedic

**Product Code(s):** HRS, HWC

### **Predicate Device(s):**

Primary Predicate:

- K220136 Treace Medical Concepts (TMC) Plating System

Additional Predicates:

- K181872 CrossRoads Extremity Systems, LLC. MIS Bunion System
- K101962, OrthoHelix Surgical Designs Inc., Mini MaxLock Extreme Plating System

**Device Description:**

The Treace Medical Concepts (TMC) Plating System includes straight, L-shaped, H- shaped, anatomically curved plates, and intramedullary plates. The system includes 2.5mm, 2.7mm, and 3.0mm diameter cannulated and non-cannulated, locking, and non-locking screws ranging in lengths from 10-32mm. The plates and screws are intended for use in stabilization and fixation of fractures, revision procedures, fusions, and reconstructions (osteotomy) of the foot.

The purpose of this traditional 510(k) submission is to introduce additional plate and screw options within the Treace Medical Concepts (TMC) Plating System.

All implantable components are manufactured from implant grade titanium alloy (Ti-6Al-4V-ELI) per ASTM F136 and are provided sterile by gamma irradiation.

**Indications for Use:**

The TMC Plating System is intended for use in stabilization of fresh fractures, revision procedures, joint fusion and reconstruction of the feet. The system can be used in both adult and pediatric patients aged >12 years. In the foot, the system can be used for the following specific examples:

- Osteotomy of the 1<sup>st</sup> metatarsal for the treatment of deformity (e.g., hallux valgus) such as:
  - Opening base wedge osteotomy
  - Closing base wedge osteotomy
  - Crescentic osteotomy
  - Proximal Chevron osteotomy
  - Distal Chevron osteotomy (Austin)
  - Transverse osteotomy
- Arthrodesis of the tarsometatarsal (TMT) joints or the 1st metatarsophalangeal (MTP) joint for the treatment of deformity (e.g., hallux valgus, hallux rigidus, metatarsus adductus) and/or arthritis
- First metatarsal fracture fixation
- Flatfoot Osteotomies
  - Lateral Column Lengthening (Evans Osteotomy)
  - Plantar Flexion Opening Wedge Osteotomy of the Medial Cuneiform (Cotton Osteotomy)
- Mid / Flatfoot Fusions
  - LisFranc Arthrodesis and/or Stabilization
  - Intercuneiform Fusions
  - Navicular-Cuneiform (NC) Fusion
  - Talo-Navicular (TN) Fusion
  - Calcaneo-Cuboid (CC) Fusion
- Medial Column Fusion

**Substantial Equivalence:**

The subject TMC Plating System is substantially equivalent to the following predicate devices:

Primary Predicate:

- Treace Medical Concepts (TMC) Plating System, K220136

Additional Predicates:

- CrossRoads Extremity Systems LLC., MIS Bunion System, K181872
- OrthoHelix Surgical Designs Inc., Mini MaxLock Extreme Plating System, K101962

The subject TMC Plating System is manufactured from implant grade titanium alloy (Ti-6Al-4V-ELI) per ASTM F136 and is intended to be used in stabilization of fresh fractures, revision procedures, joint fusion and reconstruction of the feet, identical to the primary predicate device. The subject TMC Plating System shares similar materials, geometry, construction, and overall design with the predicate devices. Thus, it can be concluded that the subject devices do not raise new questions about safety and effectiveness and are substantially equivalent to the predicate devices.

**Performance Testing:**

Mechanical testing including static and dynamic 4-point bend testing and static torsion, driving torque, removal torque, and axial pullout based on ASTM F382 and ASTM F543 were performed. Further, engineering analysis of the worst-case cross-sections and geometries on the subject and predicate devices were evaluated. The mechanical testing, cross-sectional, and geometrical analysis demonstrated the subject devices to be substantially equivalent to the predicate devices.

**Conclusion:**

The Treace Medical Concepts (TMC) Plating System has similar intended use, overall design, materials, and mechanical properties to that of the predicate devices. Therefore, it can be concluded that the subject device is at least as safe and effective and are substantially equivalent to the predicate devices.