



April 27, 2023

Inari Medical, Inc.  
Kaitlyn Weinkauf  
Senior Regulatory Affairs Specialist  
6001 Oak Canyon, Suite 100  
Irvine, California 92618

Re: K223210

Trade/Device Name: ClotTrievers XL Catheter  
Regulation Number: 21 CFR 870.5150  
Regulation Name: Embolectomy Catheter  
Regulatory Class: Class II  
Product Code: QEW, KRA  
Dated: October 14, 2022  
Received: October 17, 2022

Dear Kaitlyn Weinkauf:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Gregory W. O'Connell -S  
Digitally signed by Gregory W. O'Connell  
Date: 2023.04.27 12:05:46 -04'00'

Gregory O'Connell  
Assistant Director  
DHT2C: Division of Coronary  
and Peripheral Intervention Devices  
OHT2: Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K223210

Device Name  
ClotTrier XL Catheter

### Indications for Use (Describe)

The ClotTrier Thrombectomy System is indicated for:

- The non-surgical removal of thrombi and emboli from blood vessels.
- Injection, infusion, and/or aspiration of contrast media and other fluids into or from a blood vessel.

The ClotTrier Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

**510(k) SUMMARY**

Date prepared	April 27, 2023
Name	Inari Medical, Inc. 6001 Oak Canyon, Suite 100 Irvine, CA 92618 949-600-8433
Contact person	Kaitlyn Weinkauf Sr. Regulatory Affairs Specialist
Name of Device	ClotTriever <sup>®</sup> Thrombectomy System
Device Trade Name	ClotTriever <sup>®</sup> XL Catheter
Common name	Embolectomy catheter
Regulation name	Embolectomy catheter
Classification number	21 CFR 870.5150
Primary product code	QEW
Secondary product code	KRA
Regulatory class	II
Predicate device	ClotTriever <sup>®</sup> Thrombectomy System (K212632)
Description	<p>The ClotTriever Thrombectomy System is a single-use, sterile medical device designed to remove thrombi and emboli from the peripheral vasculature. The ClotTriever Thrombectomy System consists of ClotTriever 13 Fr and 16 Fr Sheaths, Protrieve Sheath, ClotTriever Catheter, ClotTriever BOLD Catheter, and the ClotTriever XL Catheter, each packaged separately.</p> <p>The ClotTriever Catheter, ClotTriever BOLD Catheter, and the ClotTriever XL Catheter are comprised of reinforced polymeric coaxial shafts terminating in an expandable coring element and thrombus collection bag. Three ports terminating in two stopcocks and a luer lock connection are provided for de-airing the catheter shafts. To aid in fluoroscopic visualization, the ClotTriever Sheaths, Protrieve Sheath, and ClotTriever Catheters have radiopaque distal tips. The ClotTriever XL Catheter is intended for vessels 10 to 28 mm in the peripheral vasculature.</p>
Indications for Use	<p>The introduction of the ClotTriever XL Catheter variant does not change the indications for use of the ClotTriever Thrombectomy System.</p> <p>The ClotTriever Thrombectomy System is indicated for:</p> <ul style="list-style-type: none"> <li>• The non-surgical removal of thrombi and emboli from blood vessels.</li> </ul>

	<ul style="list-style-type: none"> <li>• Injection, infusion, and/or aspiration of contrast media and other fluids into or from a blood vessel.</li> </ul> <p>The ClotTriever Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).</p>		
Device modifications	<p>The purpose of this submission is to introduce a ClotTriever BOLD Catheter variant, the ClotTriever XL Catheter, that is larger in diameter and longer in length.</p> <p>There have been no changes to the ClotTriever 13 Fr and 16 Fr Sheaths, the Protrieve Sheath, the ClotTriever Catheter, or the ClotTriever BOLD Catheter.</p>		
Comparison of Technological Characteristics with the Predicate Device	<p>The proposed device and predicate device have a similar design and materials of construction. With the exception of the modifications to the catheter OD, catheter length, bag length, coring element length and diameter, and a minor material change, the predicate and proposed devices are the same. These modifications do not change the basic design or the principles of operation from the predicate device. There are no new or different questions of safety or efficacy.</p> <p>There is no change of intended use or fundamental scientific technology between the proposed device and predicate device. The changes made for the ClotTriever XL Catheter do not change the technological characteristics of the ClotTriever Thrombectomy System.</p>		
	Device	ClotTriever XL Catheter (Proposed)	ClotTriever Thrombectomy System ClotTriever BOLD Catheter Predicate (K212632)
	Manufacturer	Inari Medical	Inari Medical
	Product Code	QEW, KRA	QEW
	Intended Use	The ClotTriever Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).	The ClotTriever Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).
	Indications for Use	<p>The ClotTriever Thrombectomy System is indicated for:</p> <ul style="list-style-type: none"> <li>• The non-surgical removal of thrombi and emboli from blood vessels.</li> <li>• Injection, infusion, and/or aspiration of contrast media and other fluids into or from a blood vessel.</li> </ul> <p>The ClotTriever Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).</p>	<p>The ClotTriever Thrombectomy System is indicated for:</p> <ul style="list-style-type: none"> <li>• The non-surgical removal of thrombi and emboli from blood vessels.</li> <li>• Injection, infusion, and/or aspiration of contrast media and other fluids into or from a blood vessel.</li> </ul> <p>The ClotTriever Thrombectomy System is intended for use in the peripheral vasculature including deep vein thrombosis (DVT).</p>
	Device Description	The ClotTriever Thrombectomy System consists of the ClotTriever Sheaths, the Protrieve Sheath, the ClotTriever	The ClotTriever Thrombectomy System consists of the ClotTriever Sheaths, the Protrieve Sheath, the

	<p>Catheter, the ClotTrieve BOLD Catheter, and the ClotTrieve XL Catheter. The ClotTrieve/Protrieve Sheaths are comprised of a reinforced polymeric sheath equipped with a self-expanding distal mesh funnel, a flush/aspiration port, and a proximal hemostasis valve. The ClotTrieve Catheter, ClotTrieve BOLD Catheter, and ClotTrieve XL Catheter are comprised of reinforced polymeric coaxial shafts terminating in an expandable coring element and thrombus collection bag. Other accessories provided include a pre-dilator, the funnel loading tool, and a large bore 60 cc syringe.</p>	<p>ClotTrieve Catheter, and the ClotTrieve BOLD Catheter. The ClotTrieve/Protrieve Sheaths are comprised of a reinforced polymeric sheath equipped with a self-expanding distal mesh funnel, a flush/aspiration port, and a proximal hemostasis valve. The ClotTrieve and ClotTrieve BOLD Catheters are comprised of reinforced polymeric coaxial shafts terminating in an expandable coring element and thrombus collection bag. Other accessories provided include a pre-dilator, the funnel loading tool, and a large bore 60 cc syringe.</p>
Principles of Operation	<p>The ClotTrieve XL Catheter is advanced into the vessel and beyond the clot. The self-expanding braided nitinol wire net is deployed. The expanded net cores, separates, and entraps thrombus from the vessel as it is being drawn to the funnel opening of the ClotTrieve/Protrieve Sheath. The net is collapsed and pulled into and through the ClotTrieve/Protrieve Sheath with the entrapped clot. A 60 cc syringe is provided for the aspiration of clot in the sheath and the infusion of contrast media and other fluids.</p>	<p>The ClotTrieve BOLD Catheter is advanced into the vessel and beyond the clot. The self-expanding braided nitinol wire net is deployed. The expanded net cores, separates, and entraps thrombus from the vessel as it is being drawn to the funnel opening of the ClotTrieve/Protrieve Sheath. The net is collapsed and pulled into and through the ClotTrieve/Protrieve Sheath with the entrapped clot. A 60 cc syringe is provided for the aspiration of clot in the sheath and the infusion of contrast media and other fluids.</p>
Target Vessel	Peripheral vessels 10-28 mm	Peripheral vessels 6-16 mm
Catheter OD/ID	OD: 16 Fr ID: 0.178"	OD: 12 Fr ID: 0.124"
Catheter Length	105 cm	80 cm
Collection Bag	Length: 10 cm Wire Diameter: 0.0040"	Length: 18 cm Wire Diameter: 0.0045"
Catheter Deployed Length	126 cm	111 cm
Coring Element OD	30 mm	16 mm
Delivery Catheter Materials	<u>Liner:</u> Pebax 72D, Propell	<u>Liner:</u> PTFE

		<u>Proximal segment:</u> Pebax 63D, Cool Grey 4C  <u>Intermediate segment:</u> Pebax 63D, Violet C, fluoro-safe marker  <u>Distal segment:</u> Pebax 63D, Cool Grey 4C Pebax 55D, Violet C (distal tip)	<u>Proximal segment:</u> Pebax 63D, Cool Grey 4C  <u>Intermediate segment:</u> Pebax 63D, Violet C, fluoro-safe marker  <u>Distal segment:</u> Pebax 55D, Violet C
	Guidewire compatibility	0.035"	0.035"
	Shelf-Life	6 months	2 years
	Sterile	SAL 10 <sup>-6</sup> , EO	SAL 10 <sup>-6</sup> , EO
	How provided	Sterile, single use	Sterile, single use
Summary of substantial equivalence	<p>The proposed device, the ClotTrievers XL Catheter, and the predicate device, the ClotTrievers BOLD Catheter, have the same indications for use, intended use, principles of operation, and fundamental scientific technology.</p> <p><u>Non-Clinical Testing</u></p> <p>In accordance with the design failure modes and effects analysis, verification and validation tests were identified to support the substantial equivalence of the ClotTrievers XL Catheter to the predicate device. This testing demonstrated compliance with relevant product specifications.</p> <p>The following tests were performed on the proposed device to establish substantial equivalence:</p> <ul style="list-style-type: none"> <li>• Pouch Seal Visual Inspection and Dye Penetration</li> <li>• Packaging Usability</li> <li>• Visual and Dimensional Inspection</li> <li>• Guidewire Compatibility</li> <li>• Protrieve and ClotTrievers Sheath Compatibility</li> <li>• Radial Force</li> <li>• Retraction Force into ClotTrievers Sheath</li> <li>• Deployment/Retraction Force</li> <li>• Retraction force to expand the collection bag</li> <li>• Kink Radius</li> <li>• Fluid Leakage - Protrieve, ClotTrievers Sheath, Catheter</li> <li>• Vacuum Testing - Protrieve, ClotTrievers Sheath</li> <li>• Simulated Use, Track and Rotation (Torque)</li> <li>• Simulated Use, Track and Tensile</li> <li>• Particulate Matter</li> <li>• Conical Fitting with 6% Luer Taper</li> <li>• Pre-clinical/GLP Animal Safety Evaluation (including In Vivo Functional Testing/Radiopacity Verification)</li> </ul>		

The following testing was leveraged from the predicate ClotTrier BOLD Catheter (K212632):

- Pouch peel and seal strength
- Corrosion Resistance
- Handle Torque
- Sterilization Validation

The following biocompatibility tests were performed on the proposed device as suggested by ISO 10993-1:

- Cytotoxicity
- Sensitization
- Intracutaneous Reactivity
- Acute Systemic Toxicity (Systemic Injection Test and Rabbit Pyrogen Test)
- Hemocompatibility - Hemolysis
- Hemocompatibility - Complement Activation
- Hemocompatibility – Thrombogenicity (*in vivo*)
- Hemocompatibility – Thrombogenicity (*in vitro*) (Platelet and Leukocyte)
- Hemocompatibility – Thrombogenicity (*in vitro*) (Partial Thromboplastin Time Test)

Clinical testing was not required to support substantial equivalence.

Test results demonstrated that all acceptance criteria were met; therefore, the device conforms to established product specifications.

#### Conclusion

The proposed device modifications to the ClotTrier Thrombectomy System do not change its intended use nor do they change the principles of operation. The verification and validation results demonstrate that the proposed ClotTrier XL Catheter is substantially equivalent to the predicate device.