



March 29, 2023

Anika Therapeutics, Inc.
% Mehdi Kazemzadeh-Narbat, PhD, PMP, CQA
Associate Director, Regulatory Affairs
Mcra LLC.
803 7th Street NW
Floor 3
Washington, District of Columbia 20001

Re: K223915

Trade/Device Name: Tactoset
Regulation Number: 21 CFR 888.3045
Regulation Name: Resorbable Calcium Salt Bone Void Filler Device
Regulatory Class: Class II
Product Code: MQV
Dated: December 29, 2022
Received: December 29, 2022

Dear Dr. Kazemzadeh-Narbat:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Sara S. Thompson -S

For

Laurence D. Coyne, Ph.D.

Director

DHT6C: Division of Restorative, Repair,
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K223915

Device Name
Tactoset

Indications for Use (Describe)

Tactoset® Injectable Bone Substitute is a synthetic, biocompatible bone graft substitute material that hardens and converts to a poorly crystalline hydroxyapatite at body temperature. It is indicated for filling bone voids or defects of the skeletal system (i.e. extremities and pelvis) that are not intrinsic to the stability of bony structure. These defects may be surgically created osseous defects or defects created from traumatic injury to the bone. The device provides an injectable, self-setting, osteoconductive bone graft substitute that resorbs and is replaced by the growth of new bone during the healing process and may be combined with autogenous bone marrow. Tactoset® Injectable Bone Substitute can augment hardware and support bone fragments during the surgical procedure. The cured paste acts only as a temporary support media and is not intended to provide structural support during the healing process.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) SUMMARY

Device Trade Name: Tactoset

Manufacturer: Anika Therapeutics, Inc.
32 Wiggins Avenue
Bedford, MA 01730

Contact: Wei Zhao
Executive Director, Regulatory Affairs
Mobile: (978)888-5948
E-Mail: wzhao@anika.com

Prepared by: Mehdi Kazemzadeh-Narbat, PhD, PMP, CQA
Director, Regulatory Affairs,
MCRA, LLC
803 7th Street NW, Floor 3
Washington, DC 20001
Office: 202.552.6011
mkazemzadeh@mcra.com

Date Prepared: Mar 23, 2023

Classifications: 21 CFR §880.3045, Resorbable calcium salt bone void filler device

Class: II

Product Codes: MQV

Primary Predicate: Tactoset (K212083)

Additional Predicates: Tactoset (K173008)
MasterGraft Putty (K051386)
MagnetOs Granules (K213111)

Indications For Use:

Tactoset® Injectable Bone Substitute is a synthetic, biocompatible bone graft substitute material that hardens and converts to a poorly crystalline hydroxyapatite at body temperature. It is indicated for filling bone voids or defects of the skeletal system (i.e. extremities and pelvis) that are not intrinsic to the stability of bony structure. These defects may be surgically created osseous defects or defects created from traumatic injury to the bone. The device provides an injectable, self-setting, osteoconductive bone graft substitute that resorbs and is replaced by the growth of new bone during the healing process and may be combined with autogenous bone marrow. Tactoset® Injectable Bone Substitute can augment hardware and support bone fragments during the surgical procedure. The cured paste acts only as a temporary support media and is not intended to provide structural support during the healing process.

Device Description:

Tactoset is an injectable, settable osteoconductive calcium phosphate bone graft substitute material. It is provided to the end-user as two components (a dry powder and an aqueous solution in separate pre-loaded syringes) that must be mixed intra-operatively prior to implantation using the supplied mixing system to form a cohesive paste. The dry powder component is composed of the alpha phase of tricalcium phosphate $[Ca_3(PO_4)_2]$, calcium carbonate $[CaCO_3]$, and monocalcium phosphate $[Ca(H_2PO_4)_2]$. The liquid component is composed of sodium phosphate dibasic $[Na_2HPO_4]$, citric acid $[C_6H_8O_7]$, hyaluronic acid (HA), and water for injection. At a specified volume, the aqueous setting solution can be combined with autogenous bone marrow, i.e., bone marrow aspirate. Tactoset is provided sterile for single use.

Predicate Device:

Anika Therapeutics, Inc. submits the following information in this Premarket Notification to demonstrate that, for the purposes of FDA's regulation of medical devices, Tactoset is substantially equivalent in indications, design principles, and performance to the following predicate devices, which have been determined by FDA to be substantially equivalent to pre-amendment devices:

Primary Predicate:	Tactoset (K212083)
Additional Predicates:	Tactoset (K173008) MasterGraft Putty (K051386) MagnetOs Granules (K213111)

Performance Testing Summary:

Bench testing data including analytical characterization, chemical composition, physical properties, mechanical testing (Pull-Out, Compressive Strength, Bone Alignment) and animal functional study were conducted for the subject Tactoset and the predicate devices. The results of the study demonstrate substantial equivalence to the predicate devices.

Substantial Equivalence:

The subject device is substantially equivalent to the Tactoset (K212083) in terms of material composition, manufacturing process, sterilization, and packaging and both are manufactured in the same facility.

The subject device and the predicate devices (primary and additional) have the same intended uses, the same product classification and product code (MQV) and have similar "Indications for Use" statements in the extremities and pelvis. The subject device and the predicate devices are bone void fillers that are intended for bony voids or gaps that are not intrinsic to the stability of the bony structure. The subject device and the primary predicate incorporate the same basic design, incorporates similar materials, have identical manufacturing process, identical packaging and are both provided sterile for single-patient, single-use in the same device volume.

The performance of Tactoset + BMA was compared to Tactoset in a rabbit critical sized femoral defect model. At the 12-week timepoint, animal study data demonstrated new bone formation averages of 14.90% in the Tactoset + BMA group, 14.65% in the Tactoset predicate group, and 5% in the empty defect negative control group. Animal study data demonstrated that

approximately 54.3% of implant material remained in the Tactoset + BMA group and 60.9% remained in the Tactoset group 12 weeks following implantation. Clinical performance has not been evaluated.

No clinical data were required to support the substantial equivalence.

Table 1. Substantial Equivalence of the Subject and Predicate Devices.

Information	Subject Device (Tactoset with Bone Marrow Aspirate)	Primary Predicate
Manufacturer	Anika Therapeutics, Inc.	Anika Therapeutics, Inc.
Trade Name	Tactoset	Tactoset
510(k) Number	TBD	K212083
Product Code	MQV	MQV
Classification	Resorbable calcium salt bone void filler device	Resorbable calcium salt bone void filler device
Materials	Powder: 4g - α -Tricalcium phosphate - Calcium carbonate - Monocalcium phosphate Liquid: 4mL setting solution, as supplied - Sodium hyaluronate - Citric acid - Sodium phosphate dibasic 3mL of the supplied setting solution may be combined with 1 mL of autogenous bone marrow	Powder: 4g - α -Tricalcium phosphate - Calcium carbonate - Monocalcium phosphate Liquid: 4mL setting solution - Sodium hyaluronate - Citric acid - Sodium phosphate dibasic
Physical Form	Injectable Paste/Putty	Injectable Paste/Putty
Dimension (CC)	4	4
Mixing Time (Min)	1	1
Min Working Time (Min)	7-18	7-18
Setting Time (Min)	10	10
Sterility Method	Single Use, Sterile, SAL 10^{-6} Setting Solution: Autoclave and VHP Powder Component: Gamma Irradiation	Single Use, Sterile, SAL 10^{-6} Setting Solution: Autoclave and VHP Powder Component: Gamma Irradiation
Pyrogenicity	LAL < 20 EU/device	LAL < 20 EU/device
Packaging	The liquid syringe is double packaged in Tyvek pouches. The cement powder is provided in a mixing syringe packaged in a foil pouch. The foil pouch is then packaged in a plastic tray with a Tyvek lid.	The liquid syringe is double packaged in heat-sealed Tyvek pouches. The cement powder is provided in a mixing syringe packaged in a foil pouch. The foil pouch is then packaged in a plastic tray with a Tyvek lid.

Conclusion:

The subject device and the predicate devices have the same intended use, have similar technological characteristics, and are made of similar materials. The subject and predicate devices are packaged in similar materials and are sterilized using similar methods. The data included in this submission demonstrate substantial equivalence to the predicate devices listed above. Tactoset is as safe, as effective, and performs as well as, or better, than the predicate devices.