



November 22, 2022

Carlsmed, Inc.
Karen Liu
VP Quality and Regulatory
1800 Aston Ave., Ste. 100
Carlsbad, California 92008

Re: K222009

Trade/Device Name: aprevo® anterior lumbar interbody fusion device with interfixation
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: OVD
Dated: October 24, 2022
Received: October 25, 2022

Dear Ms. Liu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Katherine D. Kavlock -S

for

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K222009

Device Name

aprevo® anterior lumbar interbody fusion device with interfixation

Indications for Use (Describe)

The aprevo® anterior lumbar interbody fusion device with interfixation (ALIF-X) is intended for interbody fusion in skeletally mature patients and is to be used with supplemental fixation instrumentation cleared for use in the lumbar spine. The aprevo® anterior lumbar interbody fusion device with interfixation is indicated for use as an adjunct to fusion at one or more levels of the lumbar spine in patients having an ODI >40 and diagnosed with severe symptomatic adult spinal deformity (ASD) conditions. These patients should have had six months of non-operative treatment. The devices are intended to be used with autograft and/or allogenic bone graft comprised of cancellous bone and/or cortico-cancellous bone. These implants may be implanted via a variety of open or minimally invasive approaches. The aprevo® anterior lumbar interbody fusion device with interfixation is intended to be used with the screws that accompany each implant and with supplemental fixation.

The aprevo® anterior lumbar interbody fusion device with interfixation (ALIF-X) is indicated for use at one or more levels of the lumbosacral spine as an adjunct to fusion in patients with the following indications: degenerative disc disease (DDD), disc herniation (with myelopathy and/or radiculopathy), spondylolisthesis, deformity (degenerative scoliosis or kyphosis), spinal stenosis, and failed previous fusion (pseudarthrosis). DDD is defined as discogenic back pain with degeneration of the disc as confirmed by history and radiographic studies. These patients should be skeletally mature and have had at least six (6) months of non-operative treatment. aprevo® lumbar interbody fusion devices are to be filled with autograft bone and/or allogenic bone graft composed of cancellous and/or corticocancellous bone. These devices are intended to be used with supplemental fixation systems that have been cleared for use in the thoracolumbosacral spine (e.g., posterior pedicle screw and rod systems). These implants may be implanted via a variety of open or minimally invasive approaches. The aprevo® anterior lumbar interbody fusion device with interfixation is intended to be used with the screws that accompany each implant and with supplemental fixation.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Submitter's Name:	Carlsmed, Inc.
Submitter's Address:	1800 Aston Ave., Ste 100 Carlsbad, CA 92008
Submitter's Telephone:	760-766-1926
Contact Person:	Karen Liu, VP Quality and Regulatory Carlsmed, Inc. 760-766-1926 regulatory@carlsmed.com
Date Summary was Prepared:	06-July-2022
Trade or Proprietary Name:	aprevo® anterior lumbar interbody fusion device with interfixation
Common or Usual Name:	Intervertebral Fusion Device with Bone Graft, Lumbar
Classification:	Class II per 21 CFR §888.3080
Product Code:	OVD
Classification Panel:	Orthopedic Devices

DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

The aprevo® anterior lumbar interbody fusion device with interfixation, (ALIF-X), is designed to stabilize the lumbar spinal column and facilitate fusion. The personalized aprevo® devices incorporate patient specific features to allow the surgeon to tailor the deformity correction to the individual needs of the patient. The individualized surgical correction plan and device configurations are developed using patient CT scans. The aprevo® devices are manufactured from Titanium Alloy (Ti-6Al-4V) per ASTM F3001 and ASTM F136. The aprevo® ALIF-X device has a cavity intended for the packing of bone graft.

INDICATIONS FOR USE

The aprevo® anterior lumbar interbody fusion device with interfixation (ALIF-X) is intended for interbody fusion in skeletally mature patients and is to be used with supplemental fixation instrumentation cleared for use in the lumbar spine. The aprevo® anterior lumbar interbody fusion device with interfixation is indicated for use as an adjunct to fusion at one or more levels of the lumbar spine in patients having an ODI >40 and diagnosed with severe symptomatic adult spinal deformity (ASD) conditions. These patients should have had six months of non-operative treatment. The devices are intended to be used with autograft and/or allogenic bone graft comprised of cancellous bone and/or cortico-cancellous bone. These implants may be implanted via a variety of open or minimally invasive approaches. The aprevo® anterior lumbar interbody fusion device with interfixation is intended to be used with the screws that accompany each implant and with supplemental fixation.

The aprevo® anterior lumbar interbody fusion device with interfixation (ALIF-X) is indicated for use at one or more levels of the lumbosacral spine as an adjunct to fusion in patients with the following indications: degenerative disc disease (DDD), disc herniation (with myelopathy and/or radiculopathy), spondylolisthesis, deformity (degenerative scoliosis or kyphosis), spinal stenosis, and failed previous fusion (pseudarthrosis). DDD is defined as discogenic back pain with degeneration of the disc as confirmed by history and radiographic studies. These patients should be skeletally mature and have had at least six (6) months of non-operative treatment. aprevo® lumbar interbody fusion devices are to be filled with autograft bone and/or allogenic bone graft composed of cancellous and/or corticocancellous bone. These devices are intended to be used with supplemental fixation systems that have been cleared for use in the thoracolumbosacral spine (e.g., posterior pedicle screw and rod systems). These implants may be implanted via a variety of open or minimally invasive approaches. The aprevo® anterior lumbar interbody fusion device with interfixation is intended to be used with the screws that accompany each implant and with supplemental fixation.

TECHNOLOGICAL CHARACTERISTICS

The aprevo® anterior lumbar interbody fusion device with interfixation is made from titanium alloy that conforms to ASTM F3001 and ASTM F136. The subject and predicate devices have nearly identical technological characteristics and the minor differences do not raise any new issues of safety and effectiveness. Specifically, the following characteristics are identical between the subject and predicates:

- Indications for use
- Technological characteristics
- Sizes
- Materials of manufacture
- Patient specific adaptable features
- Mechanical functionality

Table 5-1 Predicate Devices

510k Number	Trade or Proprietary or Model Name	Manufacturer	Predicate Type
K191391	HEDRON™ IA Integrated Lumbar Spacers	Globus Medical Inc.	Primary
K202034	aprevo® Intervertebral Body Fusion Device	Carlsmed, Inc.	Additional

PERFORMANCE DATA

The aprevo® anterior lumbar interbody fusion device with interfixation has been tested in the following test modes:

- Static axial compression per ASTM F2077
- Static compression shear per ASTM F2077
- Dynamic axial compression per ASTM F2077
- Dynamic compression shear per ASTM F2077
- Subsidence per ASTM F2267
- Static Screw Pushout testing
- Clinical Evaluation of Interfixated Implant Usability, Fit and Accuracy

The results of this testing show that the mechanical performance of the aprevo® anterior lumbar interbody fusion device with interfixation is sufficient for its intended use and is substantially equivalent to legally marketed predicate devices.

CONCLUSION

The overall indications for use, technology characteristics, and mechanical performance data lead to the conclusion that the aprevo® anterior lumbar interbody fusion device with interfixation is substantially equivalent to the predicate device.