

November 17, 2022

Canadian Pioneer Medical Technology Corporation (CPMT LASER)
Rashid Sayah
Managing Director
210 Drumlin Circle, Concord
Vaughan, Ontario L4K 3E3
Canada

Re: K222673

Trade/Device Name: Alexandrite and Nd:YAG Laser models CANLASE and TORLASE

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In

Dermatology

Regulatory Class: Class II Product Code: GEX Dated: September 6, 2022 Received: September 6, 2022

Dear Rashid Sayah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to https://www.fda.gov/medical-device-problems.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance) and CDRH Learn (https://www.fda.gov/training-and-continuing-education/cdrh-learn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice">https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jianting Wang -S

Jianting Wang
Acting Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

Indications for Use

intensity of black and/or blue-black tattoos) and plaques.

Form Approved: OMB No. 0910-0120 Expiration Date: 06/30/2020

Expiration Date: 06/30/2020 See PRA Statement below.

indications for OSE	See I NA Statement below.
510(k) Number (if known)	
K222673	
Device Name	
Alexandrite and Nd:YAG laser models CANLASE, TORLASE	
Indications for Use (Describe)	
755nm:	
Temporary hair reduction. Stable long-term or permanent reduction through selective	e targeting of melanin in hair follicles.
Permanent hair reduction is defined as long-term stable reduction in the number of h	airs regrowing after a treatment
regime. Permanent hair reduction is defined as the long-term, stable reduction in the	number of hairs regrowing when
measured at 6, 9, and 12 months after the completion of a treatment regime. On all si	kin types (Fitzpatrick I- VI) including
tanned skin.	
Treatment of benign pigmented lesions.	
Treatment of wrinkles.	
The photocoagulation of dermatological vascular lesions (such as port-wine stains, h	emangiomas, telangiectasias).
1064nm:	
Removal of unwanted hair, for stable long term or permanent hair reduction and for	treatment of PFB. Permanent hair
reduction is defined as the long-term, stable reduction in the number of hairs regrow	ing when measured at 6, 9, and 12
months after the completion of a treatment regime. The lasers are indicated on all sk	in types Fitzpatrick I-VI including
tanned skin. Photocoagulation and hemostasis of pigmented and vascular lesions suc	th as but not limited to port wine
stains, hemangioma, warts, telangiectasia, rosacea, venus lake, leg veins and spider v	veins. Coagulation and hemostasis of
soft tissue. Benign pigmented lesions such as, but not limited to, lentigos (age spots)	, solar lentigos (sun spots), cafe au
lait macules sehorrheic keratosis nevi chloasma verrucae skin tags keratosis tatto	oos (significant reduction in the

The laser is indicated for pigmented lesions to reduce lesion size, for patients with lesions that would potentially benefit from aggressive treatment, and for patients with lesions that have not responded to other laser treatments.

Reduction of red pigmentation in hypertrophic and keloid scars where vascularity is an integral part of the scar. Treatment of wrinkles.

Temporary increase of clear nail in patients with onychomycosis (e.g., dermatophytes, Trichophyton rubrum and T. mentagrophytes, and/or yeast Candida Albicans, etc.)

CONTINUE ON A SEPARATE PAGE IF NEEDED.			
Prescription Use (Part 21 CFR 801 Subpart D)	Over-The-Counter Use (21 CFR 801 Subpart C)		
Type of Use (Select one or both, as applicable)			
mentagrophytes, and/or yeast Candida Albicans, etc.)			

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K222673

Section 5 510(k) summary

I Submitter

Canadian Pioneer Medical Technology Corporation (CPMT LASER) 210 Drumlin Circle #2, Vaughan, Ontario, L4K 3E3, Canada

Contact person:

Dr. Rashid Reza Mir Sayah

Managing Director Phone: 4377727788

Email:

Canadianpioneer@yahoo.com

Canadianpioneermedical@gmail.com Date of preparation: Sep 06, 2022

II Subject Device

Trade/Device Name: Alexandrite Nd:YAG laser, Models: CANLASE, TORLASE

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And

In Dermatology
Regulatory Class: II
Product code: GEX

Manufacturer: Canadian Pioneer Medical Technology Corporation

III Predicate Devices

Trade/Device Name: GentleMax Pro Plus Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And

In Dermatology Regulatory Class: II Product code: GEX

510(k) number: k201111 Manufacturer: Candela Corporation

IV Device description

Alexandrite Nd:YAG laser Models CPMT CANLASE, CPMT TORLASE

contain two separate laser heads (Alexandrite and Nd:YAG), which produce laser light outputs of 755 nm and 1064 nm, respectively. The output of each laser head is optically combined on the laser rail, so that their beam paths are identical as they exit the laser system.

This allows the use of a single delivery system which can output either 755 nm or 1064 nm wavelengths. The laser system creates a beam of high intensity light that penetrates deep into the skin tissue where it delivers a controlled amount of therapeutic heat.

V Indications for use

755nm:

Temporary hair reduction. Stable long-term or permanent reduction through selective targeting of melanin in hair follicles. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowing after a treatment regime. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime. On all skin types (Fitzpatrick I- VI) including tanned skin.

Treatment of benign pigmented lesions. Treatment of wrinkles.

The photocoagulation of dermatological vascular lesions (such as port-wine stains, hemangiomas, telangiectasias).

1064nm:

Removal of unwanted hair, for stable long term or permanent hair reduction and for treatment of PFB. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime. The lasers are indicated on all skin types Fitzpatrick I-VI including tanned skin. Photocoagulation and hemostasis of pigmented and vascular lesions such as but not limited to port wine stains, hemangioma, warts, telangiectasia, rosacea, venus lake, leg veins and spider veins. Coagulation and hemostasis of soft tissue. Benign pigmented lesions such as, but not limited to, lentigos (age spots), solar lentigos (sun spots), cafe au lait macules, seborrheic keratosis, nevi, chloasma, verrucae, skin tags, keratosis, tattoos (significant reduction in the intensity of black and/or blue-black tattoos) and plaques.

The laser is indicated for pigmented lesions to reduce lesion size, for patients with lesions that would potentially benefit from aggressive treatment, and for patients with lesions that have not responded to other laser treatments.

Reduction of red pigmentation in hypertrophic and keloid scars where vascularity is an integral part of the scar. Treatment of wrinkles.

Temporary increase of clear nail in patients with onychomycosis (e.g., dermatophytes, Trichophyton rubrum and T. mentagrophytes, and/or yeast Candida Albicans, etc.)

VI Comparison of technological characteristics with the predicate devices

The indication of proposed device is covered by the predicated devices.

Device feature		
	Alexandrite and Nd:YAG Laser	Laser treatment system

	(subject device)	GentleMax Family of Laser Systems (K201111)
Product code	GEX	GEX
Manufacturer	Canadian Pioneer Medical Technology corporation (CPMT LASER)	Candela Corp.
Regulation	21 CFR 878.4810	21 CFR 878.4810
number		
Indications for use	Temporary hair reduction. Stable long-term or permanent reduction through selective targeting of melanin in hair follicles. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowing after a treatment regime. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime. On all skin types (Fitzpatrick I- VI) including tanned skin. Treatment of benign pigmented lesions. Treatment of wrinkles. The photocoagulation of dermatological vascular lesions (such as port-wine stains, hemangiomas, telangiectasias). 1064nm: Removal of unwanted hair, for stable long term or permanent hair reduction and for treatment of PFB. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9,	Temporary hair reduction. Stable long-term or permanent reduction through selective targeting of melanin in hair follicles. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowing after a treatment regime. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime. On all skin types (Fitzpatrick I- VI) including tanned skin. Treatment of benign pigmented lesions. Treatment of wrinkles. The photocoagulation of dermatological vascular lesions (such as port-wine stains, hemangiomas, telangiectasias). 1064nm: Removal of unwanted hair, for stable long term or permanent hair reduction and for

12 and months after the completion of a treatment regime. The lasers are indicated on all skin types Fitzpatrick I-VI including tanned skin. Photocoagulation and hemostasis of pigmented and vascular lesions such as but not limited to port wine stains, hemangioma, warts, telangiectasia, rosacea, venus lake, leg veins and spider veins. Coagulation and hemostasis of soft tissue. Benign pigmented lesions such as, but not limited to, lentigos (age spots), solar lentigos (sun spots), cafe au lait macules, seborrheic keratosis, nevi, chloasma, verrucae, skin tags, keratosis, tattoos (significant reduction in the intensity of black and/or blue-black tattoos) and plaques.

The laser is indicated for pigmented lesions to reduce lesion size, for patients with lesions that would potentially benefit from aggressive treatment, and for patients with lesions that have not other responded to laser treatments.

Reduction of red pigmentation in hypertrophic and keloid scars where vascularity is an integral part of the scar. Treatment of wrinkles.

Temporary increase of clear nail in patients with onychomycosis (e.g., dermatophytes, Trichophyton rubrum and T. mentagrophytes, and/or yeast Candida Albicans, etc.)

treatment of PFB. Permanent hair reduction is defined as the long-term, stable reduction in the number of hairs regrowing when measured at 6, 9, and 12 months after the completion of a treatment regime. The lasers are indicated on all skin types Fitzpatrick I-VI including tanned skin. Photocoagulation and hemostasis of pigmented and vascular lesions such as but not limited to port wine stains, hemangioma, warts, telangiectasia, rosacea, venus lake, leg veins and spider veins. Coagulation hemostasis of soft tissue. Benign pigmented **lesions** such as, but not limited to, lentigos (age spots), lentigos (sun spots), cafe au lait macules, seborrheic nevi, chloasma, keratosis, verrucae, skin tags, keratosis, tattoos (significant reduction in the intensity of black and/or blue-black tattoos) and plaques.

The laser is indicated for pigmented lesions to reduce lesion size, for patients with lesions that would potentially benefit from aggressive treatment, and for patients with lesions that have not responded to other laser treatments.

Reduction of red pigmentation in hypertrophic and keloid

		scars where vascularity is an integral part of the scar. Treatment of wrinkles. Temporary increase of clear nail in patients with onychomycosis (e.g., dermatophytes, Trichophyton rubrum and T.
		mentagrophytes, and/or yeast Candida Albicans, etc.)
Laser Type	Alexandrite and Nd:YAG laser	Alexandrite and Nd:YAG laser
Laser classification	Class IV	Class IV
Wavelength	755 & 1064	755 & 1064
Spot Size	1.5 mm, 3 mm, 3x10 mm, 5 mm, 6 mm, 8 mm, 10 mm, 12 mm, 15 mm, 18 mm, 20 mm, 22 mm, 24 mm, 26 mm	1.5 mm, 3 mm, 3x10 mm, 5 mm, 6 mm, 8 mm, 10 mm, 12 mm, 15 mm, 18 mm, 20 mm, 22 mm, 24 mm, 26 mm
Frequency	0.5 - 10Hz	1-10 Hz
Maximum Energy (J)	60 Joules (J) ALEX(optional 70 J) ; 110 J Nd:YAG	68 Joules (J) ALEX; 90 J Nd:YAG
Accuracy of Output	‡ 20%	‡ 20%
Pulse Duration	0.25 – 100 ms	0.25 – 100 ms
System Cooling	Ambient Air	Ambient Air
Electrical	Comply with IEC60601-1,	Comply with IEC60601-1,
Safety	IEC60601-2-22	IEC60601-2-22
EMC	Comply with IEC60601-1-2,	Comply with IEC60601-1-2,
Laser safety	Comply with IEC60825-1, IEC60601-2-22	Comply with IEC60825-1, IEC60601-2-22

VII Performance data

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility testing

Biocompatibility of the Laser Treatment System was evaluated in accordance with ISO 10993 for the body contact category of "Surface –intact skin" The following tests were performed, as recommended: Cytotoxicity, Irritation and Sensitization. All evaluation acceptance criteria were met

Electrical safety and electromagnetic compatibility (EMC)

Electrical safety and EMC testing were conducted on the Laser Treatment System. The system has been tested to comply with the following standards:

- IEC 60601-1 Medical Electrical Equipment Part 1: General Requirements For Basic Safety And Essential Performance;
- IEC 60601-2-22, Medical Electrical Equipment Part 2-22: Particular Requirements For Basic Safety And Essential Performance Of Surgical, Cosmetic, Therapeutic And Diagnostic Laser Equipment;
- IEC 60825-1, Safety of laser products Part 1: Equipment classification and requirements.
- IEC 60601-1-2 Medical electrical equipment- Part 1-2: General requirements for basic safety and essential performance- Collateral standard: Electromagnetic compatibility- Requirements and tests.

VIII Conclusion

The Laser Treatment System is substantially equivalent to its predicate devices. The non-clinical testing demonstrates that the device is as safe, as effective and performs as well as the legally marketed device.