



January 30, 2020

Osstem Implant Co., Ltd.
% Peter Lee
QA/RA Manager
Hiossen Inc.
85 Ben Fairless Dr.
Fairless Hills, Pennsylvania 19030

Re: K192941
Trade/Device Name: HySil Super Fast Impression Materials
Regulation Number: 21 CFR 872.3660
Regulation Name: Impression Material
Regulatory Class: Class II
Product Code: ELW
Dated: November 4, 2019
Received: November 5, 2019

Dear Peter Lee:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Srinivas "Nandu" Nandkumar
Director
DHT1B: Division of Dental Devices
OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT and Dental Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

510(k) Number: K192941

Device Name: HySil Super Fast Impression Materials

Indications for Use:

HySil Heavy Super Fast is to be used as heavy-bodied materials for:

- One-step impression technique (simultaneous technique) using single or dual viscosities
- Two-step impression technique using dual viscosities
- Functional impressions

HySil Mono Super Fast is to be used as a medium-bodied tray or syringeable impression material for:

- Taking impressions over fixed/removable restorations and implants (i.e., transferring impressions on posts and bridge components)
- Functional impressions
- Fabricating crown and bridgework or inlays
- Fabricating full or partial dentures
- Reline impressions
- Use in the simultaneous mixing technique as well as the putty-wash and triple tray techniques
- Transferring root posts when fabricating posts and cores indirectly

HySil Light Super Fast is to be used as syringeable impression materials for:

- Two-step putty-wash impression technique
- One-step putty-wash impression technique
- Two-step impression technique using dual viscosities
- Reline impressions
- Fabricating full or partial dentures

Prescription Use (21 CFR 801 Subpart D)

Over-The Counter Use (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com

510(k) Summary

Date: January 29, 2020

1. Company and Correspondent making the submission

- Submitter's Name : Osstem Implant Co., Ltd.
- Address : 66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, 48002, Republic of Korea
- Contact : Ms. Jungmin Yoo
- Phone : +82-51-850-2500

- Correspondent's Name : Hiossen Inc.
- Address : 85 Ben Fairless Dr. Fairless Hills, PA 19030
- Contact : Mr. Peter Lee
- Phone : +1-267-759-7031

2. Proposed Device

- Trade or (Proprietary) Name : HySil Super Fast Impression Materials
- Classification Name : Impression Material
- Regulation Number : 21CFR872.3660
- Device Classification : Class II
- Classification Product Code : ELW

3. Predicated Device

- Primary Predicate
HySil Impression Materials, Osstem Implant Co., Ltd. (K170736)

4. Description

HySil Super Fast Impression Materials are addition-curing, elastomeric materials with hydrophilic property. They are intended to be placed on an impression tray (or injected directly into the mouth, depending on the technique and device) and used to reproduce the structure of a patient's teeth and gums.

HySil Super Fast Impression Materials meet ISO 4823 standard and consist of three different viscosities (heavy-bodied, medium-bodied, and light-bodied) in delivery system as standard 1:1 50 ml cartridges.



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com

5. Substantial Equivalence Matrix

	Proposed Devices	Predicated Devices	Remark
Device Name	HySil Super Fast Impression Materials Heavy, Mono, Light	HySil Impression Materials Putty, Heavy, Mono, Light, Extra Light, Bite	Different
510(k) No.	K192941	K170736	Different
Manufacturer	Osstem Implant Co., Ltd.	Osstem Implant Co., Ltd.	Same
Indications for Use	<p>HySil Heavy Super Fast is to be used as heavy-bodied materials for:</p> <ul style="list-style-type: none"> - One-step impression technique (simultaneous technique) using single or dual viscosities - Two-step impression technique using dual viscosities - Functional impressions <p>HySil Mono Super Fast is to be used as a medium-bodied tray or syringeable impression material for:</p> <ul style="list-style-type: none"> - Taking impressions over fixed/removable restorations and implants (i.e., transferring impression posts and bridge components) - Functional impressions - Fabricating crown and bridgework or inlays - Fabricating full or partial dentures - Reline impressions - Use in the simultaneous mixing technique as well as the putty-wash and triple tray techniques - Transferring root posts when fabricating posts and cores indirectly <p>HySil Light Super Fast is to be used as syringeable impression materials for:</p> <ul style="list-style-type: none"> - Two-step putty-wash impression technique 	<p>HySil Putty is to be used as preliminary materials for:</p> <ul style="list-style-type: none"> - Two-step Putty-wash impression technique - One-step Putty-wash impression technique <p>HySil Heavy is to be used as heavy-bodied materials for:</p> <ul style="list-style-type: none"> - One-step impression technique (simultaneous technique) using single or dual viscosities - Two-step impression technique using dual viscosities - Functional impressions <p>HySil Mono is to be used as a medium-bodied tray or syringeable impression material for:</p> <ul style="list-style-type: none"> - Taking impressions over fixed/removable restorations and implants (i.e., transferring impression posts and bridge components) - Functional impressions - Fabricating crown and bridgework or inlays - Fabricating full or partial dentures - Reline impressions - Use in the simultaneous mixing technique as well as the putty-wash and triple tray techniques - Transferring root posts when fabricating posts and cores indirectly 	<p>Different since the predicated devices have 6 types of devices, Putty, Heavy, Mono, Light, Extra Light, and Bite while the proposed devices have 3 types of devices including Heavy, Mono, and Light.</p> <p>Although there are differences in number of types, the Indications for Use Statements of Heavy, Mono and Light type for both proposed and predicated devices are same.</p>



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
 Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com

	<ul style="list-style-type: none"> - One-step putty-wash impression technique - Two-step impression technique using dual viscosities - Reline impressions - Fabricating full or partial dentures 	<p>HySil Light is to be used as syringeable impression materials for:</p> <ul style="list-style-type: none"> - Two-step putty-wash impression technique - One-step putty-wash impression technique - Two-step impression technique using dual viscosities - Reline impressions - Fabricating full or partial dentures <p>HySil Extra Light is to be used as syringeable impression materials for:</p> <ul style="list-style-type: none"> - Two-step putty-wash impression technique - One-step putty-wash impression technique - Two-step impression technique using dual viscosities - Reline impressions - Fabricating full or partial dentures <p>HySil Bite is used for impression as below.</p> <ul style="list-style-type: none"> - Taking occlusal surfaces - Confirming occlusal surfaces - Recording after putting the articulator 	
Principle of Operation	<ul style="list-style-type: none"> - Be placed on an impression tray (or injected directly into the mouth, depending on the technique and device) and used to reproduce the structure of a patient's teeth and gums - Provide models for study and for production of restorative prosthetic devices 	<ul style="list-style-type: none"> - Be placed on an impression tray (or injected directly into the mouth, depending on the technique and device) and used to reproduce the structure of a patient's teeth and gums - Provide models for study and for production of restorative prosthetic devices 	Same
Description of Material	Vinylpolysiloxane	Vinylpolysiloxane	Compose with same affiliated material, but ratios of each



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com

			component in use are different
Standard Conformed	ISO 4823	ISO 4823	Same
Working Time	Over 1 min. 15 sec.	Over 2 ~ 2 min. 30 sec.	Proposed devices have shorter working time
Shelf-life	2 years	2 years	Same
S.E.	<p>Similarities The proposed devices and the predicated devices are made with same affiliated material called Vinylpolysiloxane conformed to ISO 4823 standard. Both are delivered in 50 ml cartridge with base and catalyst ratio of 1 to 1; have same indications for use; have same principle of operation; and have same shelf-life.</p> <p>Differences Compared to the predicated devices, the proposed devices have different composition ratios which results in shorter working time. However, based on the results of the performance and the biocompatibility testing, the proposed and the predicated devices both passed the requirements. Also, there are differences in Indications for Use Statement since the predicated devices have 6 types of devices, Putty, Heavy, Mono, Light, Extra Light, and Bite while the proposed devices have 3 types of devices including Heavy, Mono, and Light. Although there are differences in number of types, the Indications for Use Statements of Heavy, Mono and Light type for both proposed and predicated devices are same. Thus, the differences in Indications for Use Statement do not affect the substantial equivalence of proposed devices.</p> <p>∴ Therefore, we stated that proposed devices (HySil Heavy, Mono, Light Super Fast Impression Materials) are substantially equivalent to the predicated devices (HySil Heavy, Mono, Light Impression Materials) cleared in K170736.</p>		

6. Indications for Use

HySil Heavy Super Fast is to be used as heavy-bodied materials for:

- One-step impression technique (simultaneous technique) using single or dual viscosities
- Two-step impression technique using dual viscosities
- Functional impressions

HySil Mono Super Fast is to be used as a medium-bodied tray or syringeable impression material for:

- Taking impressions over fixed/removable restorations and implants (i.e., transferring impression posts and bridge components)
- Functional impressions
- Fabricating crown and bridgework or inlays
- Fabricating full or partial dentures
- Reline impressions



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com

- Use in the simultaneous mixing technique as well as the putty-wash and triple tray techniques
- Transferring root posts when fabricating posts and cores indirectly

HySil Light Super Fast is to be used as syringeable impression materials for:

- Two-step putty-wash impression technique
- One-step putty-wash impression technique
- Two-step impression technique using dual viscosities
- Reline impressions
- Fabricating full or partial dentures

7. Summary of Non-clinical Performance Testing

Non-clinical testing data are submitted to demonstrate substantial equivalence.

Biocompatibility Evaluation

Biocompatibility testing was considered followed the FDA Guidance Document *Use of International Standard ISO 10993-1, "Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process,"* and the ISO 10993 suite of standards.

Sterilization Validation and Shelf-life

Proposed devices are delivered in non-sterile. Therefore, sterilization validation was not considered. The Shelf-life testing was considered followed the FDA Guidance Document *Shelf Life of Medical Device*, ISO 4823 and ISO 3417; and determined shelf-life of proposed devices are 2 years.

Mechanical Properties

Mechanical testing was evaluated following ISO 4823, and it was found that the proposed devices are substantially equivalent to the predicate devices from the result of bench test.

8. Summary of Clinical Testing

No clinical studies are submitted.

9. Conclusion

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807, and based on the information provided in this premarket notification, Osstem Implant Co., Ltd. concludes that the HySil Super Fast Impression Materials are substantially equivalent to the predicated devices as herein.



Osstem Implant Co., Ltd.

66-16, Bansong-ro 513beon-gil, Haeundae-gu, Busan, Republic of Korea
Tel: +82 51 850 2500 Fax: +82 51 861 4693 www.osstem.com
