



September 12, 2022

OsteoCentric Technologies  
% Mr. Ken Riordan  
Regulatory Project Manager  
Telos Partners, LLC  
2850 Frontier Drive  
Warsaw, Indiana 46582

Re: K222512

Trade/Device Name: Integrity-SI Fusion System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: OUR  
Dated: August 17, 2022  
Received: August 19, 2022

Dear Mr. Riordan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.  
Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K222512

Device Name

Integrity-SI Fusion System

Indications for Use (Describe)

The Integrity-SI Fusion System is intended for sacroiliac joint fusion for the following conditions:

1. Sacroiliac joint dysfunction that is a direct result of sacroiliac joint disruption and degenerative sacroiliitis. This includes conditions whose symptoms began during pregnancy or in the peripartum period and have persisted postpartum for more than 6 months.
2. To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.
3. Acute, non-acute, and non-traumatic fractures involving the sacroiliac joint

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### I. Submitter

OsteoCentric Technologies  
75 West 300 N, Suite 150  
Logan, UT 84321  
Phone: 1-800-969-0639

Submitter Contact Person: Todd Evans, Senior Director of Quality and Regulatory Affairs  
Submission Contact Person: Ken Riordan, Regulatory Project Manager, Telos Partners, LLC

Date Prepared: August 11, 2022

### II. Device

Device Proprietary Name:	Integrity-SI Fusion System
Common or Usual Name:	Sacroiliac Joint Fixation
Classification Name:	Smooth or threaded metallic bone fixation fastener
Regulation Number:	21 CFR 888.3040
Product Code:	OUR
Device Classification	Class II

### III. Predicate Device

Substantial equivalence is claimed to the following devices:

- CoorsTek Integrity-SI Fusion System, K161755 (Primary Predicate)
- SI-BONE iFuse Implant System, K193524

### IV. Device Description

The Integrity-SI Fusion System consists of partially and fully threaded, self-tapping cannulated titanium implants designed to be inserted across sacroiliac joint providing stability for joint arthrodesis when used in conjunction with allograft or autograft. The surgical implants are available in various sizes to accommodate patient anatomy. The 10mm and 12mm diameter screws are offered in partially and fully threaded version in lengths from 40-110mm, in 5mm increments. All fastener sizes are available in non-coated or hydroxyapatite-coated (HA) versions. The 10mm and 12mm screws also include a pre-assembled washer for improved joint compression. The fully threaded 6.5mm diameter, optional secondary fasteners are offered for additional rotational stability in lengths of 30 – 70 mm, in 5mm increments and are only intended for use in conjunction with a primary 10mm or 12mm screw.

**V. Indications for Use**

The Integrity-SI Fusion System is intended for sacroiliac joint fusion for the following conditions:

- Sacroiliac joint dysfunction that is a direct result of sacroiliac joint disruption and degenerative sacroiliitis. This includes conditions whose symptoms began during pregnancy or in the peripartum period and have persisted postpartum for more than 6 months.
- To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.
- Acute, non-acute, and non-traumatic fractures involving the sacroiliac joint.

**VI. Comparison of Technological Characteristics to the Predicate**

There are no changes to the technological characteristics of the OsteoCentric Integrity-SI Fusion System. The device is identical to the predicate device with respect to the design, materials, sterilization, and manufacturing methods. The change to the Indications for Use does not change the intended use of the device, does not introduce new indications, and does not change the therapeutic effect or use of the implants.

**VII. Performance Data**

There were no changes in the design, materials, sterilization, or manufacturing methods of the device. Therefore, no new performance data was required to demonstrate the substantial equivalence of the subject device to the predicate device.

**VIII. Conclusion**

The technological characteristics and intended use have not changed compared to the predicate device. Therefore, the OsteoCentric Integrity-SI Fusion System is substantially equivalent to the predicate device.