



April 20, 2022

Boston Scientific  
Daniel Root  
Sr. Regulatory Affairs Specialist  
Two Scimed Place  
Maple Grove, Minnesota 55311

Re: K220866

Trade/Device Name: EKOS+ Endovascular Device  
Regulation Number: 21 CFR 870.5150  
Regulation Name: Embolectomy Catheter  
Regulatory Class: Class II  
Product Code: QEY, KRA  
Dated: March 23, 2022  
Received: March 25, 2022

Dear Daniel Root:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Gregory O'Connell  
Assistant Director  
DHT2C: Division of Coronary  
and Peripheral Intervention Devices  
OHT2: Office of Cardiovascular Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K220866

Device Name  
EKOS+ Endovascular Device

### Indications for Use (Describe)

The EKOS+ Endovascular System is indicated for the:

- Ultrasound facilitated, controlled and selective intravascular infusion of physician-specified fluids, including thrombolytics, for the treatment of pulmonary embolism and/or deep vein thrombosis.
- Controlled and selective infusion of physician-specified fluids, including thrombolytics, into the pulmonary arteries and/or peripheral vasculature.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



Boston Scientific Corporation  
 300 Boston Scientific Way  
 Marlborough, MA 01752  
 (508) 683-4000  
 www.bostonscientific.com

## 510(k) Summary

<b>Sponsor</b>	Boston Scientific Corporation 300 Boston Scientific Way Marlborough, Massachusetts 01752
<b>Contact Name and Information</b>	Daniel Root Three Scimed Place Maple Grove, MN 55311-1566 Phone: 425-395-5820 Email: Daniel.Root@bsci.com
<b>Date Prepared</b>	April 18, 2022
<b>Proprietary Name</b>	EKOS+ Endovascular Device
<b>Common Name</b>	Continuous Flush Catheter
<b>Product Code (Primary/Secondary)</b>	QEY, KRA
<b>Classification (Primary/Secondary)</b>	Mechanical Thrombolysis Catheter (21 CFR §870.5150) Catheter, Continuous Flush (21 CFR §870.1210)
<b>Predicate Device</b>	The EKOS+ Endovascular Device is substantially equivalent to another legally marketed device. This predicate device is the EKOS+ Endovascular Device (K213422).
<b>Device Description</b>	The EKOS+ Endovascular System consists of an EKOS+ Endovascular Device and EKOS Control Unit (Control Unit 4.0 and Connector Interface Cables). The EkoSonic Endovascular Device consists of a single-use, disposable infusion catheter with removable ultrasound core. The infusion catheter contains multiple side holes distributed over the length of the treatment zone. The ultrasound core contains up to 20 ultrasound elements, evenly spaced over the treatment zone. Thermal sensors in the treatment zone monitor catheter temperature. The Control System generates and controls the delivery of radiofrequency energy to the ultrasound core while monitoring and controlling the temperature of the treatment zone.
<b>Indications for Use/ Intended Use</b>	The EKOS+ Endovascular System is indicated for the: <ul style="list-style-type: none"> <li>• Ultrasound facilitated, controlled and selective intravascular infusion of physician-specified fluids, including thrombolytics, for the treatment of pulmonary embolism and/or deep vein thrombosis</li> <li>• Controlled and selective infusion of physician-specified fluids, including thrombolytics, into the pulmonary arteries and/or peripheral vasculature.</li> </ul>
<b>Device Technology Characteristics and Comparison to Predicate Device</b>	The EKOS+ Endovascular Device incorporates the following changes from the predicate EKOS+ Endovascular Device (K213422): an updated luer design that is ISO 80369-7 compliant; luer material change to a Cyrolite® Polymer; and an updated distal catheter tip design and material.

<b>Characteristic</b>	<b>EKOS+ Endovascular Device (Subject Device)</b>	<b>EKOS+ Endovascular Device (Predicate Device)</b>
<b>510(k) Number</b>	K220866	K213422
<b>Product Code</b>	QEY, KRA	QEY, KRA
<b>Indications for Use</b>	<p>The EKOS+ Endovascular System is indicated for the:</p> <ul style="list-style-type: none"> <li>• Ultrasound facilitated, controlled and selective intravascular infusion of physician-specified fluids, including thrombolytics, for the treatment of pulmonary embolism and/or deep vein thrombosis</li> <li>• Controlled and selective infusion of physician-specified fluids, including thrombolytics, into the pulmonary arteries and/or peripheral vasculature.</li> </ul>	<p>The EKOS+ Endovascular System is indicated for the:</p> <ul style="list-style-type: none"> <li>• Ultrasound facilitated, controlled and selective intravascular infusion of physician-specified fluids, including thrombolytics, for the treatment of pulmonary embolism and/or deep vein thrombosis</li> <li>• Controlled and selective infusion of physician-specified fluids, including thrombolytics, into the pulmonary arteries and/or peripheral vasculature.</li> </ul>
<b>Principle of Operation</b>	The EKOS+ Endovascular System [EKOS+ Endovascular Device with Control Unit] employs ultrasound to facilitate the delivery of thrombolytic agents into vascular blood clots.	The EKOS+ Endovascular System [EKOS+ Endovascular Device with Control Unit] employs ultrasound to facilitate the delivery of thrombolytic agents into vascular blood clots.
<b>Infusion Hole Pattern</b>	Multiple side-holes	Multiple side-holes
<b>Catheter Working Length</b>	106 cm or 135 cm	106 cm or 135 cm
<b>Treatment Zone Length</b>	8 cm — 20 cm	8 cm — 20 cm
<b>Compatible Guide Wire</b>	0.035"	0.035"
<b>Labeled Outer Diameter</b>	7.8 Fr	7.7 Fr
<b>Sterilization Method</b>	Ethylene Oxide Sterilization	Ethylene Oxide Sterilization
<b>SAL</b>	10 <sup>-6</sup>	10 <sup>-6</sup>
<b>Infusion Hole Pattern</b>	Multiple side-holes	Multiple side-holes

Characteristic	EKOS+ Endovascular Device (Subject Device)	EKOS+ Endovascular Device (Predicate Device)
<b>Radiopaque Markers</b>	One tungsten-loaded Pebax tip and one platinum iridium marker band on the IC. USC ultrasound elements are radiopaque	One tungsten-loaded Pebax tip and one platinum iridium marker band on the IC. USC ultrasound elements are radiopaque
<b>Single-Use</b>	Yes	Yes
<b>Biocompatibility</b>	Meets all the requirements in accordance with ISO 10993-1	Meets all the requirements in accordance with ISO 10993-1
<b>Infusion Catheter Tip Material</b>	Pebax Tungsten Blend	Pebax Tungsten Blend
<b>Luer and Manifold Design</b>	ISO 594 Compliant	ISO 80369-7 Compliant
<b>Luer Material</b>	Polycarbonate	Cyrolite® Polymer (acrylic copolymer)

---

**Non-Clinical Performance Data**

Determination of substantial equivalence is based on an assessment of non-clinical performance bench testing, including bench-top performance evaluations and biological safety.

**Bench Testing:**

Bench testing was performed to evaluate physical integrity, functionality, and performance of the catheter. Performance criteria includes: guidewire/sheath compatibility, tip pressure, luer lock/connector performance.

**Biological Safety Testing:**

Biocompatibility testing in accordance with ISO 10993-1, microbial assessments including bioburden and endotoxin, and pyrogenicity and sterility assurance testing show the device has acceptable biological safety for its intended use.

---

**Clinical Testing**

Performance testing from clinical studies is not required to demonstrate substantial equivalence of EKOS+ Endovascular Device.

---

**Conclusion**

Based on the indications for use, technological characteristics, and performance testing, EKOS+ Endovascular Device has been shown to be appropriate for its intended use and is considered to be substantially equivalent to EKOS+ Endovascular Device, K213422.